


# MODEL 184 5 MHz SWEEP GENERATOR



# WAVETEK



**MODEL 184  
5 MHz  
SWEEP GENERATOR**

**WAVETEK®**

9045 Balboa Ave., San Diego, Calif. 92123  
P. O. Box 651, San Diego, California 92112  
Tel 714/279-2200 TWX 910-335-2007

*Manual Revision 2/80  
Instrument Release C-5/79*

# CONTENTS

<b>SECTION 1</b>	<b>GENERAL DESCRIPTION</b>	
1.1	THE MODEL 184 . . . . .	1-1
1.2	SPECIFICATIONS . . . . .	1-1
<b>SECTION 2</b>	<b>INITIAL PREPARATION</b>	
2.1	UNPACKING INSPECTION . . . . .	2-1
2.2	PREPARATION FOR USE . . . . .	2-1
2.3	ELECTRICAL ACCEPTANCE CHECK . . . . .	2-1
<b>SECTION 3</b>	<b>OPERATION</b>	
3.1	CONTROLS AND CONNECTORS . . . . .	3-1
3.2	OPERATION . . . . .	3-4
3.2.1	Signal Termination . . . . .	3-4
3.2.2	Continuous Operation . . . . .	3-5
3.2.3	Trigger Mode . . . . .	3-5
3.2.4	Triggering . . . . .	3-6
3.2.5	GATED (or Tone Burst) Mode . . . . .	3-6
3.2.6	SWEEP Mode . . . . .	3-6
3.2.7	Voltage Control – VCG . . . . .	3-6
<b>SECTION 4</b>	<b>CIRCUIT DESCRIPTION</b>	
4.1	VOLTAGE CONTROLLED GENERATOR . . . . .	4-1
4.2	SWEEP GENERATOR . . . . .	4-2
<b>SECTION 5</b>	<b>CALIBRATION</b>	
5.1	FACTORY REPAIR . . . . .	5-1
5.2	REQUIRED TEST EQUIPMENT . . . . .	5-1
5.3	REMOVING GENERATOR COVER . . . . .	5-1
5.4	CALIBRATION . . . . .	5-1
<b>SECTION 6</b>	<b>TROUBLESHOOTING</b>	
6.1	INTRODUCTION . . . . .	6-1
6.2	CIRCUIT BOARD ACCESS . . . . .	6-1
6.3	BASIC TECHNIQUES . . . . .	6-1
6.4	TROUBLESHOOTING INDIVIDUAL COMPONENTS . . . . .	6-1
6.5	TROUBLESHOOTING GUIDE . . . . .	6-2
<b>SECTION 7</b>	<b>PARTS AND SCHEMATICS</b>	
7.1	DRAWINGS . . . . .	7-1
7.2	ORDERING PARTS . . . . .	7-1
7.3	ADDENDA . . . . .	7-1

# SECTION 1

## GENERAL DESCRIPTION

### 1.1 THE MODEL 184

Wavetek Model 184, the 5 MHz Sweep Generator, is a precision source of sine, triangle, square, positive pulse and negative pulse waveforms plus dc voltage. Frequency of the waveforms is manually and remotely variable from 100  $\mu$ Hz to 5 MHz.

The generator can repetitively sweep between two individually set frequencies and at a particular sweep rate. The amplitude of waveforms is variable from 20V p-p, open circuit maximum, to -80 dB.

DC reference of the waveforms can be offset positively and negatively.

The symmetry of the waveforms is continuously adjustable from approximately 1:19 to 19:1. Varying symmetry provides variable duty cycle pulses, sawtooth and asymmetrical sine waveforms.

A voltage representing generator frequency, a fixed-amplitude pulse train of that frequency, and a voltage ramp representing frequency sweep rate are provided as front panel outputs.

### 1.2 SPECIFICATIONS

Specifications (waveform, frequency, and amplitude selection), operating modes, precision (accuracy), and waveform purity (quality) are listed in the following paragraphs.

#### 1.2.1 Versatility

##### Waveforms

Five selectable waveforms, sine , triangle , square , positive pulse , negative pulse , plus variable DC output. Symmetry of all waveform outputs is continuously adjustable from approximately 1:19 to 19:1. Varying symmetry provides variable duty cycle pulses, sawtooth, or asymmetrical sine waveforms. Separate sync output is included.

##### Control

Frequency can be controlled manually, with external voltage (VCG) or with internally generated ramp voltage.

##### Operating Frequency Range

Frequency selectable from 0.0001 Hz to 5 MHz in the following linear ranges:

X 0.001	0.0001 Hz to 0.005 Hz
X 0.01	0.001 Hz to 0.05 Hz
X 0.1	0.005 Hz to 0.5 Hz
X 1	0.05 Hz to 5 Hz
X 10	0.5 Hz to 50 Hz
X 100	5 Hz to 500 Hz
X 1K	50 Hz to 500 Hz
X 10K	500 Hz to 500 kHz
X 100K	5 kHz to 500 kHz
X 1M	50 kHz to 5 MHz

##### NOTE

*When SYMMETRY control is used, the output frequency is different from the dial indicated frequency. The maximum symmetry ratio obtainable is also dependent on the frequency dial setting.*

##### Main Output

, , ; variable to 20V p-p into open circuit and 10V p-p into 50 $\Omega$  load. DC offset of waveform (or DC if selected) is adjustable to  $\pm$ 10 volts open circuit and  $\pm$ 5 volts into 50 $\Omega$  load.

, ; DC: 0 to +10 or -10 volts into open circuit and 0 to +5 or -5 volts into 50 $\Omega$  load.

Output dc voltage is limited to approximately  $\pm$ 10 volt open circuit and output current is limited to approximately 130 mA.

Output can be attenuated from 0 dB to -80 dB: -60 dB in 20 dB steps, plus a 20 dB vernier for continuous variation (20 dB vernier does not affect offset or DC).

##### Pulse Output

Output voltage is TTL compatible. Rise and fall times are typically 15 ns. Sync is normally a symmetrical square waveform; with SYMMETRY control ON, it is rectangular.

### Sweep Output

SWEEP OUT connector provides a nominal 0 to +5 volt (open circuit) ramp from a 600 $\Omega$  source impedance.

### DC Offset

DC offset of  $\sim$ ,  $\wedge$ , or  $\square$  waveform, or DC if selected, is adjustable to  $\pm 10$  volts open circuit and  $\pm 5$  volts into a 50 $\Omega$  load. Output current is limited to approximately 130 mA. Waveform + offset is limited to  $\pm 10$ V into an open circuit.

### GCV Output

A dc voltage proportional to the instantaneous frequency of the generator output. 0 to +5V, open circuit, 1 k $\Omega$  source impedance.

## 1.2.2 Operating Modes

### Continuous

Operating as a standard VCG (voltage controlled generator), frequency output is determined by front panel control settings in conjunction with external control voltage at VCG IN.

### Triggered

Only one complete cycle of output appears at 50 $\Omega$  OUT connector for each pulse applied to TRIG IN connector (or press of MAN TRIG switch).

### Gated

Same as triggered mode except that output oscillations continue for duration of gating signal applied to TRIG IN connector (or as the MAN TRIG switch is depressed).

### Sweep

The internal ramp generator can sweep the main generator up in frequency with a linear ramp. The end frequency points may be individually set and the time for each ramp can be varied from 100s to 1 ms.

## 1.2.3 Voltage Controlled Generator

VCG Control Range: Up to 1000:1 frequency change with external voltage input.

Input Impedance: 5 k $\Omega$ .

VCG Voltage: 0 to 5V.

VCG Slew Rate: 2% of range per  $\mu$ s.

VCG Linearity 0.0001 Hz to 50 kHz:  $\pm 0.5\%$ .

## 1.2.4 Triggered Generator

Trigger pulse is 1V p-p to  $\pm 10$ V; input impedance is 10 k $\Omega$ ,

33 pF; minimum pulse width is 50 ns; maximum repetition rate is 5 MHz.

## 1.2.5 Horizontal Precision

### Dial Accuracy (Symmetrical Waveform)

$\pm 2\%$  of full scale for 0.005 Hz to 5 MHz.

$\pm 4\%$  of reading and  $\pm 2\%$  of full scale for 0.0005 Hz to 0.005 Hz.

### Frequency Vernier

Approximately 1% of range. Vernier affects calibration of frequency dial.

### Time Symmetry

$\pm 1\%$  for 0.005 to 500 kHz.

## 1.2.6 Vertical Precision

### Amplitude Change With Frequency (Sine)

Less than 0.1 dB to 100 kHz.

Less than 0.2 dB to 1 MHz.

Less than 1 dB to 5 MHz.

### Step Attenuator Accuracy

$\pm 0.3$  dB per 20 dB step.

### Stability

Short Term:  $\pm 0.05\%$  for 10 minutes.

Long Term:  $\pm 0.25\%$  for 24 hours.

Percentages apply to amplitude, dc offset and main generator frequency in the linear mode.

### Amplitude Symmetry

$\pm 1\%$  of amplitude range to 1 MHz for all symmetrical waveforms.

## 1.2.7 Purity

### Sine Distortion

Less than 0.5% for 10 Hz to 100 kHz.

Less than 1% for 0.005 to 1 MHz.

All harmonics at least 30 dB down for 1 to 5 MHz.

### Triangle Linearity

Greater than 99% for 0.0005 Hz to 100 kHz.

### Square Wave Rise and Fall Time

Less than 25 ns terminated into 50 $\Omega$  load.

### Square Wave Total Aberrations

Less than 5%.

## 1.2.8 Environmental

All specifications listed are for 25 $^{\circ}$ C  $\pm 5^{\circ}$ C. For operation from 0 $^{\circ}$ C to 55 $^{\circ}$ C, specifications including horizontal pre-

cision, amplitude symmetry, and sine wave distortion are derated by a factor of 2.

#### 1.2.9 Mechanical

##### Dimensions

11 $\frac{1}{4}$  in./28.6 cm wide; 5 in./14.5 cm high; 10 $\frac{1}{2}$  in./27.3 cm deep.

##### Weight

8.5 lb/3.8 kg net; 12 lb/5.5 kg shipping.

#### 1.2.10 Power

90V to 110V, 105V to 125V, 180V to 220V or 210V to 250V; 50 Hz to 400 Hz; less than 25 watts.

#### NOTE

*Specifications apply from 10 to 100% of a selected frequency range with SYMMETRY control OFF.*

0 1 2 3 4 5 6 7 8 9 10 11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32 33 34 35 36 37 38 39 40 41 42 43 44 45 46 47 48 49 50 51 52 53 54 55 56 57 58 59 60 61 62 63 64 65 66 67 68 69 70 71 72 73 74 75 76 77 78 79 80 81 82 83 84 85 86 87 88 89 90 91 92 93 94 95 96 97 98 99 100 101 102 103 104 105 106 107 108 109 110 111 112 113 114 115 116 117 118 119 120 121 122 123 124 125 126 127 128 129 130 131 132 133 134 135 136 137 138 139 140 141 142 143 144 145 146 147 148 149 150 151 152 153 154 155 156 157 158 159 160 161 162 163 164 165 166 167 168 169 170 171 172 173 174 175 176 177 178 179 180 181 182 183 184 185 186 187 188 189 190 191 192 193 194 195 196 197 198 199 200 201 202 203 204 205 206 207 208 209 210 211 212 213 214 215 216 217 218 219 220 221 222 223 224 225 226 227 228 229 230 231 232 233 234 235 236 237 238 239 240 241 242 243 244 245 246 247 248 249 250 251 252 253 254 255 256 257 258 259 260 261 262 263 264 265 266 267 268 269 270 271 272 273 274 275 276 277 278 279 280 281 282 283 284 285 286 287 288 289 290 291 292 293 294 295 296 297 298 299 300 301 302 303 304 305 306 307 308 309 310 311 312 313 314 315 316 317 318 319 320 321 322 323 324 325 326 327 328 329 330 331 332 333 334 335 336 337 338 339 340 341 342 343 344 345 346 347 348 349 350 351 352 353 354 355 356 357 358 359 360 361 362 363 364 365 366 367 368 369 370 371 372 373 374 375 376 377 378 379 380 381 382 383 384 385 386 387 388 389 390 391 392 393 394 395 396 397 398 399 400 401 402 403 404 405 406 407 408 409 410 411 412 413 414 415 416 417 418 419 420 421 422 423 424 425 426 427 428 429 430 431 432 433 434 435 436 437 438 439 440 441 442 443 444 445 446 447 448 449 450 451 452 453 454 455 456 457 458 459 460 461 462 463 464 465 466 467 468 469 470 471 472 473 474 475 476 477 478 479 480 481 482 483 484 485 486 487 488 489 490 491 492 493 494 495 496 497 498 499 500 501 502 503 504 505 506 507 508 509 510 511 512 513 514 515 516 517 518 519 520 521 522 523 524 525 526 527 528 529 530 531 532 533 534 535 536 537 538 539 540 541 542 543 544 545 546 547 548 549 550 551 552 553 554 555 556 557 558 559 560 561 562 563 564 565 566 567 568 569 570 571 572 573 574 575 576 577 578 579 580 581 582 583 584 585 586 587 588 589 590 591 592 593 594 595 596 597 598 599 600 601 602 603 604 605 606 607 608 609 610 611 612 613 614 615 616 617 618 619 620 621 622 623 624 625 626 627 628 629 630 631 632 633 634 635 636 637 638 639 640 641 642 643 644 645 646 647 648 649 650 651 652 653 654 655 656 657 658 659 660 661 662 663 664 665 666 667 668 669 670 671 672 673 674 675 676 677 678 679 680 681 682 683 684 685 686 687 688 689 690 691 692 693 694 695 696 697 698 699 700 701 702 703 704 705 706 707 708 709 710 711 712 713 714 715 716 717 718 719 720 721 722 723 724 725 726 727 728 729 730 731 732 733 734 735 736 737 738 739 740 741 742 743 744 745 746 747 748 749 750 751 752 753 754 755 756 757 758 759 760 761 762 763 764 765 766 767 768 769 770 771 772 773 774 775 776 777 778 779 780 781 782 783 784 785 786 787 788 789 790 791 792 793 794 795 796 797 798 799 800 801 802 803 804 805 806 807 808 809 810 811 812 813 814 815 816 817 818 819 820 821 822 823 824 825 826 827 828 829 830 831 832 833 834 835 836 837 838 839 840 841 842 843 844 845 846 847 848 849 850 851 852 853 854 855 856 857 858 859 860 861 862 863 864 865 866 867 868 869 870 871 872 873 874 875 876 877 878 879 880 881 882 883 884 885 886 887 888 889 890 891 892 893 894 895 896 897 898 899 900 901 902 903 904 905 906 907 908 909 910 911 912 913 914 915 916 917 918 919 920 921 922 923 924 925 926 927 928 929 930 931 932 933 934 935 936 937 938 939 940 941 942 943 944 945 946 947 948 949 950 951 952 953 954 955 956 957 958 959 960 961 962 963 964 965 966 967 968 969 970 971 972 973 974 975 976 977 978 979 980 981 982 983 984 985 986 987 988 989 990 991 992 993 994 995 996 997 998 999 1000

# SECTION 2

## INITIAL PREPARATION

### 2.1 UNPACKING INSPECTION

After carefully unpacking the instrument, inspect the external parts for damage to knobs, dials, indicators, surface areas, etc. If there is damage, file a claim with the carrier who transported the instrument. Retain the shipping container and packing material for use in case reshipment is required.

### 2.2 PREPARATION FOR USE

Before connecting the instrument to line power, be sure the rear panel 115/230V and HI/LO switches are set to the value nearest the line voltage and that the fuse is correct for the switch setting. Be sure that the plug on the power cord is the proper mate for the line receptacle.

AC Line Voltage	Switch A	Switch B	Fuse (SB)
90 - 110	115	LO	1/4 amp
105 - 125	115	HI	1/4 amp
180 - 220	230	LO	1/8 amp
210 - 250	230	HI	1/8 amp

### 2.3 ELECTRICAL ACCEPTANCE CHECK

This checkout procedure verifies the generator operation. If a malfunction is found, refer to the Warranty in the front of this manual. An oscilloscope, 50Ω coax cable and 50Ω feedthru are needed for this procedure (figure 2-1).

Preset the generator front panel controls as follows:

Control	Position
GEN MODE . . . . .	CONT
WAVEFORM . . . . .	~
SYMMETRY . . . . .	NORMAL
FREQ Range . . . . .	X 1K
FREQ VERNIER . . . . .	CAL
Frequency Dial . . . . .	1
AMPLITUDE . . . . .	0 dB
AMPLITUDE VARIABLE . . . . .	Full clockwise
DC OFFSET . . . . .	12 o'clock
TRIG LEVEL . . . . .	Full counterclockwise
SWEEP MODE . . . . .	CONT SWEEP
SWEEP WIDTH . . . . .	MAX
SWEEP TIME . . . . .	12 o'clock

Perform the steps in table 2-1. Only approximate values are required to verify operation.




Figure 2-1. Acceptance Test Setup

Table 2-1. Performance Checkout

Step	Control	Position/Operation	Observe at 50Ω OUT
	<b>Function</b>		
1	POWER	Push on	Sine wave, 1 kHz, 10V p-p
2	WAVEFORM	~	Triangle wave
3	WAVEFORM	⌊	Square wave
4	WAVEFORM	⌋	Positive pulse
5	WAVEFORM	⌋	Negative pulse

Performance Checkout (Continued)

Step	Control	Position/Operation	Observe at 50Ω OUT
6	SYMMETRY	cw	Frequency ÷ 10, decreasing negative pulse width
7	SYMMETRY	ccw	Decreasing positive pulse width
8	SYMMETRY	NORMAL	---
9	WAVEFORM	∩ (OFFSET)	Sine wave, 1 kHz
<b>Frequency</b>			
10	FREQ Range	X 1M	Frequency = 1 MHz
11	FREQ VERNIER	ccw	Frequency decreases by at least 50 kHz
12	FREQ VERNIER	CAL	---
13	FREQ Range	X .001	Detect generator running
14	FREQ Range	X 1K	---
15	Frequency Dial	5	Frequency = 5 kHz
16	Frequency Dial	.05	Frequency = 50 Hz
17	Frequency Dial	1	---
<b>Amplitude</b>			
18	AMPLITUDE Range	-60 dB	10 mV p-p
19	AMPLITUDE VARIABLE	ccw	1 mV p-p
20	AMPLITUDE Range	0	1V p-p
21	DC OFFSET	cw	Positive slew; about +5V positive peak
22	DC OFFSET	ccw	Negative slew; about -5V negative peak
23	WAVEFORM	∩ (No OFFSET)	Triangle wave
<b>Trigger &amp; Gate</b>			
24	GEN MODE	TRIG	0 Vdc
25	MAN TRIG	Push	Generate one cycle
26	GEN MODE	GATED	0 Vdc
27	MAN TRIG	Push and hold	Continuous ∩ waveforms

Performance Checkout (Continued)

Step	Control	Position/Operation	Observe at 50Ω OUT
28	MAN TRIG	Release	0 Vdc
29	Sweep GEN MODE	SWEEP	Frequency sweeps
30	SWEEP TIME	cw	Sweep rate increases
31	SWEEP TIME	ccw	Sweep rate decreases

Disconnect the cable at 50Ω OUT and connect it to SWP OUT: observe a 5 V ramp waveform. Connect it to GCV

OUT: observe a ramp plus dc. Connect it to PULSE OUT: observe a 2.4 V positive pulse.



# SECTION 3 OPERATION

## 3.1 CONTROLS AND CONNECTORS

The generator front panel controls and connectors are shown in figure 3-1 and keyed to the following descriptions:

### ① POWER Switch

Power is turned on and off with the POWER push-button. The frequency dial index ①A lights when power is turned on.

### ② Frequency Dial

Frequency settings of the dial multiplied by frequency range ①B determine output frequency. In frequency sweep operation, this dial determines the frequency from which sweep is started.

### ③ SWEEP WIDTH Control

The SWEEP WIDTH control determines the amount of frequency change that will occur in sweep mode. After the sweep start frequency is set by the frequency dial ②, SWEEP WIDTH determines the sweep stop frequency (see figure 3-2).

### ④ SWEEP MODE Selector

The SWEEP MODE selector is enabled by the GEN MODE selector ④ set to SWEEP. After the start frequency of the generator is determined by the frequency dial ②, the stop frequency is determined by the SWEEP WIDTH control ③. The SWEEP START and SWEEP STOP settings will hold the output signal at the start and stop




Figure 3-1. Front Panel Controls and Connectors




Figure 3-2. Effect of Sweep Time and Width on Output Frequency

frequencies, respectively, while the frequency dial (2) and SWEEP WIDTH control (3) are adjusted. CONT SWEEP allows frequency sweeping between the two set limits.

(5) SWEEP TIME Control

Frequency of the internal sweep ramp, and thus, the sweep repetition rate, is governed by the SWEEP TIME control (see figure 3-2).

(6) Main Output Connector

Maximum output of 10V p-p signals into a 50Ω load (20V p-p open circuit) is provided at the 50Ω OUT connector; all generator mode signals are delivered at this connector. See (7) for amplitude of output.

(7) AMPLITUDE Control

The AMPLITUDE switch affects waveforms, dc output and waveform dc offset. The VARIABLE control affects waveforms only. Maximum waveform amplitude is with the 0 dB setting of the AMPLITUDE control and with the VARIABLE control fully cw (see table 3-1). Amplitude is decreased 20 dB with VARIABLE control fully ccw.

Table 3-1. Maximum Voltage at 0 dB

Function	50Ω	
	Open Circuit	Termination
~ , ^ , □	20V p-p	10V p-p
⌌	0 to +10V	0 to +5V
⌍	0 to -10V	0 to -5V
DC	±10V	±5V

(8) Synchronizing Pulse Output Connector

A fixed amplitude (0 to about 5V) TTL pulse of the generator frequency is provided at the PULSE OUT connector. This output can be used as a synchronizing reference for the main output (6). Phase of the waveforms relative to the sync output is shown in figure 3-3.




Figure 3-3. Pulse/Waveform Phase Relationship and Waveform Reference Lines

(9) DC OFFSET Control

Offset of waveforms and dc voltage are controlled by the DC OFFSET control. The WAVEFORM switch (10) must be in one of the four right-hand settings. Center of the waveform reference (figure 3-3) is skewed positive with clockwise rotation, negative with counterclockwise rotation. Offset and dc voltage maximums are ±5V (50Ω terminated).

(10) WAVEFORM Selector

Sine ~ , triangle ^ , and square □ waveforms, and positive and negative square pulse trains ⌌ , ⌍ are selected for output by the WAVEFORM selector, with or without dc offset. When set for dc offset, the inner knob (9) controls the ±5V (50Ω terminated) offset. DC is selected for dc output with voltage controlled by the inner knob (9).

(11) Manual Trigger Control

In TRIG mode (14), the MAN TRIG pushbutton is used to trigger a single cycle of waveform output

and, in the GATED mode, to gate the output of waveforms until released.

**NOTE**

The TRIG LEVEL control (13) must be fully CCW.

**(12) Remote Trigger Input Connector**

The TRIG IN connector accepts voltage level inputs that trigger and gate the generator in TRIG and GATED modes (14), respectively. The trigger level control (13) determines the level at which the TRIG IN input is accepted for triggering or gating. A positive-going excursion through a voltage level, which can be set in the range of -7.5V to +7.5V by the TRIG LEVEL control triggers or gates the generator operation.

A negative-going dc excursion through the trigger level ends gated operation. Figure 3-4 shows triggering and gating of the generator waveforms at time t1. Once triggered or gated, a full cycle of the selected waveform is output to its completion: when gating is removed at time t2, for example, the last full cycle of waveform completes itself at time t3.




Figure 3-4. Generator Trigger and Gated Control

**(13) Trigger Level Control**

The TRIG LEVEL control determines the level at which the input at the TRIG IN connector (12) is accepted as a trigger in the generator trigger and gated modes (14). The trigger level can be varied from fully clockwise, where a positive-going excursion thru -7.5V is a trigger, to fully counterclockwise, where a positive-going excursion thru +7.5V is a trigger.

**(14) Generator Mode Selector**

Generator modes are:

1. Continuous – An uninterrupted output of the selected waveform at the selected frequency and amplitude.
2. Triggered – One cycle of the selected waveform at the selected frequency and amplitude when the trigger signal is detected at TRIG IN (12) or when manually triggered at (11).
3. Gated – A burst of the selected waveform at the selected frequency and amplitude, which starts when the gate signal is detected at TRIG IN (12) and lasts through the completion of the last cycle started before the removal of the gate signal, or starts and stops when manually gated at (11).
4. Automatically Swept Frequency – Sweep as determined by the SWEEP MODE control (4).

**(15) Sweep Ramp Output Connector**

The internal sweep generator ramp is available at the SWP OUT connector. Ramp frequency is varied by the SWEEP TIME control. Output is a 0 to +5V ramp, 600Ω source impedance.

**(16) Waveform SYMMETRY Control**

Normal symmetrical output results when SYMMETRY is set to NORMAL; an asymmetrical, or unbalanced, waveform results when SYMMETRY is set between  $\nabla$  and  $\nabla$ . (Asymmetric operation reduces generator frequency to approximately 1/10th the normal output.) Figure 3-5 shows the effect of SYMMETRY control on the waveforms.

**NOTE**

When **SYMMETRY** control is used, the output frequency is different from the dial indicated frequency. The maximum symmetry ratio obtainable also depends on the frequency dial setting. A typical example is shown in table 3-2.




Figure 3-5. Effect of **SYMMETRY** Control

**17** GCV Output Connector

GCV OUT provides dc excursions of 0 to about +5V which represent the output frequency in the selected range. Source impedance is 1 k $\Omega$ .

**18** Frequency Range Control

The selected range settings of the **FREQ** selector, multiplied with the frequency dial **2** setting determine output frequency.

**19** Frequency **VERNIER** Control

The frequency is as labeled on **18** and **2**, when the **VERNIER** control is set fully clockwise to **CAL** (calibrated). Rotating the **VERNIER** control counterclockwise decreases output frequency. The range is approximately 1% of the selected frequency range.

**20** VCG Input Connector

DC voltage excursions of 0 to  $\pm 5$  volts at the VCG IN connector control frequency within the selected range. Positive inputs increase frequencies set by the frequency dial **2** and range control **18**, and negative inputs decrease the frequencies. Input impedance is 5 k $\Omega$ . Frequency excursions of 1000:1 are possible.

**3.2 OPERATION**

Operation is discussed in terms of the four generator modes: continuous, triggered, gated and sweep plus VCG.

**3.2.1 Signal Termination**

Proper signal termination, or loading, of the generator connectors is necessary for its specified operation. For example, the proper termination of the main output is shown in figure 3-6. Placing the 50 $\Omega$  terminator, or 50 $\Omega$  resistance, in parallel with a higher impedance matches the receiving




Figure 3-6. Signal Termination

Table 3-2. Dial Setting

Frequency Range	X 100K					
	5	4	3	2	1	0.5
Indicated Frequency	500 kHz	400 kHz	300 kHz	200 kHz	100 kHz	50 kHz
Output Frequency	54 kHz	44 kHz	33 kHz	23 kHz	12 kHz	6.5 kHz
Maximum Symmetry Ratio	18:1	18:1	18:1	17:1	16:1	15:1

instrument input impedance to the generator output impedance, thereby minimizing signal reflection or power loss on the line due to phase angle mismatch.

The input and output impedance of the generator connectors are listed below:

Connector	Impedance
50Ω OUT	50Ω
TRIG IN	10 kΩ
PULSE OUT (TTL)	*
SWP OUT	600Ω
VCG IN	5 kΩ
GCV OUT	1 kΩ

\*The PULSE OUT connector can drive up to 20 Transistor-Transistor Logic (TTL) loads (low level between 0V and 0.4V, and high level between 2.4V and 5V).

### 3.2.2 Continuous Operation

Basic, or continuous, operation of the generator involves turning on power, selecting a continuous output mode, selecting a waveform, and setting the output signal frequency and amplitude. When operation is critical, allow a one-half hour warm-up period. The following steps demonstrate use as a basic function generator:

Step	Control/Connector	Setting
1	50Ω OUT	Connect circuit (refer to paragraph 3.2.1).
2	PULSE OUT	Use for external synchronization, if required.
3	GEN MODE	CONT
4	WAVEFORM	Choose one of the left-hand set of waveforms. If dc or dc offset is desired, use right-hand set.
5	SYMMETRY	NORMAL or desired asymmetry. (Affects frequency calibration.)
6	FREQ	As desired for frequency range.
7	Frequency Dial	As desired for exact frequency.
8	FREQ VERNIER	CAL, unless extreme frequency accuracy is re-

Step	Control/Connector	Setting
		quired, in which case, monitor with a frequency counter.
9	AMPLITUDE	As desired.
10	AMPLITUDE VARIABLE	As desired.
11	DC OFFSET	As desired (step 4, right-hand set of waveforms must be chosen). CW positive offset, ccw negative offset. See figure 3-7 for restrictions.




Figure 3-7. DC OFFSET Control

### 3.2.3 Trigger Mode

Operation as a triggered one cycle generator is as for a basic function generator, only the operating mode is triggered (TRIG) instead of continuous (CONT), and a manual or remote trigger (MAN TRIG, TRIG IN) is used to start the single cycle of waveform. Perform the steps given in paragraph 3.2.2, only set the GEN MODE control in step 3 to TRIG. Refer to paragraph 3.2.4 for triggering.

#### NOTE

The signal at SWP OUT can be used as source of repetitive trigger inputs.

### 3.2.4 Triggering

Manual trigger as follows:

Step	Control/Connector	Setting
1	TRIG LEVEL	Full ccw.
2	MAN TRIG	Press for each cycle desired.

Remote trigger as follows:

Step	Control/Connector	Setting
1	TRIG LEVEL	Rotate the TRIG LEVEL control cw to set negative thresholds as low as $-7.5V$ through which a positive-going TRIG IN connector input can pass to provide triggering. CCW sets positive thresholds of up to $+7.5V$ through which a positive-going TRIG IN level can pass to provide triggering.
2	TRIG IN	Apply a positive-going voltage (through the threshold set in the preceding step) to the TRIG IN connector to provide remote triggering.

#### CAUTION

Avoid voltages greater than  $\pm 50V$  at TRIG IN to prevent damage to the generator.

### 3.2.5 GATED (or Tone Burst) Mode

Operation as a gated or tone burst generator is as for a triggered generator, only the operating mode is GATED, and releasing the MAN TRIG or removing the remote trigger voltage ends the burst of output waveform. Perform the steps of paragraph 3.2.2, only set the GEN MODE control to GATED. Refer to paragraph 3.2.4 for triggering.

### 3.2.6 SWEEP Mode

The generator can be set for a repetitive sweep of output frequencies within a given range. Operation is like continuous mode, only a separately controlled, internal ramp generator provides an additional VCG input to control frequency. (This internally-generated ramp is also available

at the SWP OUT connector.) Perform the steps given in paragraph 3.2.2, only set the GEN MODE control in step 3 to SWEEP and include the following steps:

Step	Control/Connector	Setting
1	SWEEP MODE	SWEEP START.
2	Frequency Dial	Desired start sweep frequency.
3	SWEEP MODE	SWEEP STOP.
4	SWEEP WIDTH	Desired stop sweep frequency.
5	SWEEP MODE	CONT SWEEP
6	SWEEP TIME	As desired.

### 3.2.7 Voltage Control — VCG

Operation with voltage control can be done in any mode but is usually done in continuous mode; the frequency




Figure 3-8. VCG Voltage-to-Frequency Nomograph

within a particular range is additionally controlled with dc levels within  $\pm 5V$ , injected at the VCG IN connector. Perform the steps given in paragraph 3.2.2, only set the frequency dial to determine a reference from which the frequency is to be voltage controlled:

1. For frequency control with positive dc inputs at VCG IN, set the dial for a lower limit from which frequency is to be increased.
2. For frequency control with negative dc inputs at VCG IN, set the dial for an upper limit from which frequency is to be decreased.
3. For modulation with an ac input at VCG IN, set the dial at the desired center frequency. Do not

exceed the maximum dynamic range of the selected frequency range.

Figure 3-8 is a nomograph with examples of the frequency dial effect as a reference for VCG IN voltages. Example 1 shows that with 0V VCG input (2nd column), frequency (3rd column) is as determined by the frequency dial setting of 2 (1st column). Example 2 shows that with a positive VCG input, output frequency is increased. Example 3 shows that with a negative VCG input, output frequency is decreased. (Note that the Factor of  $50\Omega$  OUT Frequency column must be multiplied by the frequency range in order to give the actual  $50\Omega$  OUT frequency.) For full 1000:1 linear mode VCG sweep of the generator frequencies, set the FREQ VERNIER full ccw.






Figure 4-1. Simplified Block Diagram



# 4

## SECTION 4

### CIRCUIT DESCRIPTION

#### 4.1 VOLTAGE CONTROLLED GENERATOR

The heart of the generator consists of the positive and negative current sources, the current switch, timing capacitors, triangle amplifier, and hysteresis switch (figure 4-1).

The positive and negative current sources generate equal but opposite polarity currents which charge and discharge the timing capacitor selected by the range selector. The current switch, which is controlled by the hysteresis switch, selects either the positive or the negative current as the input to the capacitor. Since the capacitor is being charged by a current source which changes polarity periodically, the voltage across the capacitor forms a triangle waveform. This waveform is fed through the triangle buffer amplifier to the hysteresis switch. The hysteresis switch determines when the triangle waveform reaches predetermined positive and negative peak values. When this occurs, the output of the hysteresis changes state and causes the current switch to select the opposite polarity current. The output of the hysteresis switch is a square wave whose edges correspond to the triangle peak values.

The magnitude of the current produced by the current sources is dependent upon the output of the VCG amplifier. By varying the output of the VCG amplifier, the frequency of the triangle and square waveforms may be controlled.

In order to generate sine waves, the triangle waveform is sine shaped in the sine converter circuit with nonlinear elements. The waveform switch selects the waveform of interest and a portion of the signal is selected by the amplitude potentiometer and applied to the output amplifier. The output amplifier is capable of driving a  $50\Omega$  load and may be dc offset. The amplifier output is routed to a  $50\Omega$  attenuator which can provide 60 dB of attenuation in 20 dB steps. An additional 20 dB of attenuation can be obtained from the amplitude control.

The square wave from the hysteresis switch is also applied to the TTL sync pulse generator, whose square wave output is TTL compatible.

To change frequency ranges, different timing capacitors may be selected by the frequency range switch. On the very slow frequencies the capacitance multiplier becomes active.

This circuit senses the capacitor charging current and then subtracts a certain percentage of it from the capacitor. As a result, the capacitor does not charge as fast, and the frequency, as a result, is lower.

Several things can affect the frequency of the generator by varying the output of the VCG amplifier. One is the frequency dial of the function generator which feeds a voltage to the VCG amplifier. In addition to the frequency dial, the frequency vernier feeds in a voltage to the VCG amplifier. The range of the vernier is approximately 1% of the full scale frequency. Finally, an external voltage applied to the VCG input can control the frequency of the generator loop. The VCG input allows frequency modulation of the generator by an external signal.

Under normal conditions the generator loop runs with the positive and negative current sources balanced. This results in symmetrical sine, triangle and square waveforms, or in the case of the square waveform, a duty cycle of 50%. By varying the symmetry control, the current sources may be unbalanced which results in the generation of asymmetrical waveforms. This allows the generation of pulses, ramps, and other waveshapes.

In the trigger mode, the generator is stopped by the trigger amplifier. This amplifier compares the output of the triangle amplifier to ground. Its output draws just the right amount of current away from the capacitor to keep it at zero volts. This level is known as the trigger baseline. When an external signal is applied to the trigger input, it is shaped into a fast rise time pulse by the squaring circuit and is applied to the trigger logic circuit. This circuit in turn shuts off the trigger amplifier for one cycle of the output waveform. Trigger input may also be made manually by the manual trigger switch.

The trigger logic circuit also allows the generator to run in the gated mode. In this mode the generator will run as long as the trigger input signal is positive. When it goes negative, the generator will continue to run until the last cycle is complete and then remain at the trigger baseline level.

The GCV output is an analog output voltage proportional to the instantaneous output frequency of the generator. This is from the GCV amplifier which senses the positive

current source output and generates a voltage proportional to the current.

#### 4.2 SWEEP GENERATOR

The internal sweep generator can frequency modulate the main generator. It generates a positive-going linear ramp signal which can be controlled in both period and amplitude. This signal is added to that of the main frequency dial and is applied to the VCG amplifier. This results in a linearly increasing FM sweep.

The linear sweep ramp is obtained by charging the capacitor with a constant current from the log current source. When the positive peak of the ramp reaches a predetermined value, the comparator turns on the discharge current source

and quickly discharges the capacitor to zero volts. When this point is reached, the comparator resets and turns off the discharge current source. A new ramp cycle now begins.

The buffer amplifier provides a high input impedance to the capacitor, thus assuring a highly linear ramp. It also provides the power required to drive the main generator VCG circuit.

The log current source has a built-in log converter circuit which allows a single sweep rate control to provide sweep rates of 1 ms to 100s: a  $10^5$  control range.

A sweep output signal can drive X-Y recorders or other devices.

# 5

## SECTION 5 CALIBRATION

### 5.1 FACTORY REPAIR

Wavetek maintains a factory repair department for those customers not possessing the necessary personnel or test equipment to maintain the instrument. If an instrument is returned to the factory for calibration or repair, a detailed description of the specific problem should be attached to minimize turnaround time.

### 5.2 REQUIRED TEST EQUIPMENT

Spectrum Analyzer . . . . . 600 kHz to 5 MHz  
 Voltmeter . . . . . Microvolt dc measurement (0.01% accuracy)  
 Oscilloscope, Dual Channel . . . . . 150 MHz bandwidth  
 Distortion Analyzer . . . . . To 600 kHz  
 Counter . . . . . To 10 MHz (0.1% accuracy)  
 50Ω Load . . . . . ±0.1% accuracy

### 5.3 REMOVING GENERATOR COVER

For main circuit board access, invert the instrument, remove the four screws in the cover, and lift off the bottom cover.

### 5.4 CALIBRATION

After referring to the following preliminary data, perform calibration, as necessary, per table 5-1. If performing partial calibration, check previous settings and adjustments for applicability.

1. Unless otherwise noted, all measurements made at the 50Ω OUT connector should be terminated into a 50Ω (±0.1%) load.
2. Test Points (TPs) and adjustments are on the main board unless noted otherwise.
3. Before connecting the unit to an ac source, check the ac line circuit to make sure the 115/230 and HI/LO switches are set at the correct position (see paragraph 2.2).

4. Start the calibration by setting the front panel switches as follows:

FREQ Range . . . . . X 10K  
 Frequency Dial . . . . . 5  
 FREQ VERNIER . . . . . CAL  
 SYMMETRY . . . . . NORMAL  
 GEN MODE . . . . . CONT  
 WAVEFORM . . . . . □ (No Offset)  
 AMPLITUDE . . . . . 0  
 AMPLITUDE VARIABLE . . . . . Max cw  
 SWEEP MODE . . . . . SWEEP START  
 SWEEP TIME . . . . . 100s

5. Allow the unit to warm up at least 30 minutes for final calibration.

Table 5-1. Calibration Chart

Step	Check	Tester	Cal Points	Control Settings	Adjust	Desired Results	Remarks
1	Power Supply Regulators	Voltmeter	C84 (+)	—	R206	+15 Vdc ±50 mV	Ground is C84 (—).
2			C88 (—)			—15 Vdc ±150 mV	
3			C80 (+)			+5V ±250 mV	
Cover the instrument and allow a 30 minute warm-up. Keep covered as much as possible during calibration.							
4	Amplifier Offset	Voltmeter	Q19 emitter	GEN MODE: TRIG WAVEFORM: ~	R192	0V ±5 mV	~ amplifier output.

Table 5-1. Calibration Chart (Continued)

Step	Check	Tester	Cal Points	Control Settings	Adjust	Desired Results	Remarks	
5	Amplifier Offset	Voltmeter	50Ω OUT		R 124	0V ± 10mV		
6				AMPLITUDE VARIABLE: max ccw	R 156		Repeat steps 5 and 6	
7	Time Symmetry	Dual channel scope	50Ω OUT	GEN MODE: CONT WAVEFORM:  FREQ: X 1K Dial: 5 Scope time base: 20 μs/div	R 32	Time symmetry < 0.1%	Follow procedure in figure 5-1.	
8				FREQ: X 100K Dial: .05	R 35		Follow procedure in figure 5-1.	
9							Repeat steps 6 and 7.	
10	VCG Zero			Same as for step 7	R 13	Minimum frequency shift while shorting and opening VCG IN BNC to ground		
11	Sine Distortion	Distortion analyzer (with 50Ω termination)	50Ω OUT	FREQ: X 1K VERNIER: CAL Dial: 5 WAVEFORM: 	R 68, R 71	Distortion < 0.16%	If minimum distortion cannot be met, refer to table 6-1.	
12				Dial: 1			Distortion < 0.2%	If adjustment was necessary, repeat step 10.
13				FREQ: X 10K				
14	High Freq Sine Distortion			FREQ: X 1M Dial: 1 WAVEFORM: 	C 64	Minimum rise time with minimum overshoot		
15		Spectrum analyzer		WAVEFORM: 	None	All harmonics below -32 dB from 1 to 5 MHz	If not, refer to table 6-1.	
16	Frequency	Counter	50Ω OUT	WAVEFORM:  FREQ: X 10K Dial: 5	R 21	50 kHz ± 100 Hz		
17				FREQ: X 1M	C 22	5 MHz ± 20 kHz	Repeat steps 15 and 16.	
18				FREQ: X 100K	-C 17	500 ± 1 kHz	Change C 16 if necessary	

Table 5-1. Calibration Chart (Continued)

Step	Check	Tester	Cal Points	Control Settings	Adjust	Desired Results	Remarks
19	Frequency	Counter	50Ω OUT	FREQ: X 100	None	500 ±10 Hz	
20				FREQ: X 1K		5 kHz ±100 Hz	
21				FREQ: X 10K		50 ±1 kHz	
22	Time Symmetry	Dual channel scope		FREQ: X 10 Scope time base: 0.1s/div	R92	Time symmetry < 0.1%	Follow procedure in figure 5-1.
23	Frequency	Counter	50Ω OUT	Dial: 5	R88	50 ±0.1 Hz or 20 ms ±40 μs	Change R87 if necessary.
24				FREQ: X .001	None	5 MHz ± 0.3 mHz 189s to 217s.	
25				FREQ: X .01		.05 Hz ± 1 mHz or 20s ± 400ms	
26				FREQ: X .1		0.5 Hz ± 10mHz or 2s ± 40ms	
27				FREQ: X 1		5 Hz ± 100mHz or 0.2s ± 4ms	
28				FREQ: X 1K Dial: 5, 4, 3, 2, 1, .5		Dialed Freq ±100 Hz	
29				FREQ: X 1M Dial: .5, 1, 2, 3, 4, 5		Dialed Freq ± 100kHz	

Remove the four screws attaching the main board to the long standoffs. Put the bottom cover on, but do not insert the screws. Place the instrument on its feet and remove the top cover for sweep board access.

30	Sweep	Oscilloscope	SWP OUT	SWEEP MODE: CONT SWEEP			With SWEEP TIME, set SWP OUT to near 100 Hz. Alternate SWEEP MODE between CONT SWEEP and SWEEP STOP while performing step 31.
----	-------	--------------	---------	---------------------------	--	--	---

Table 5-1 Calibration Chart (Continued)

Step	Check	Tester	Cal Points	Control Settings	Adjust	Desired Results	Remarks
31	Sweep	Oscilloscope	SWP OUT	No change	Sweep board R12	CONT SWEEP amplitude = SWEEP STOP amplitude $\pm 10$ mV	



1. ADJUST OSCILLOSCOPE.  
 TRIGGER: INTERNAL AND ALTERNATE  
 CHANNEL A: NORMAL  
 CHANNEL B: INVERTED
2. ADJUST FREQUENCY DIAL/VERNIER FOR ONE CYCLE ON SCOPE.
3. SWITCH X 10 SWEEP MAGNIFIER ON.




Figure 5-1. Time Symmetry Measurement

# SECTION 6

## TROUBLESHOOTING

### 6.1 INTRODUCTION

This section is organized as follows:

Circuit Board Access  
Basic Techniques  
Troubleshooting Individual Components  
Troubleshooting Guide

(Refer to paragraph 5.2 for required test equipment.)

#### NOTE

*Wavetek maintains a factory repair department for those customers not possessing the necessary personnel or test equipment to maintain the instrument. If an instrument is returned to the factory for calibration or repair, a detailed description of the specific problem should be attached to minimize turnaround time.*

### 6.2 CIRCUIT BOARD ACCESS

Turn the instrument over, remove the four screws in the bottom cover and remove the bottom cover. For sweep board access, remove the four screws holding the main board to its long standoffs, place the instrument right side up and remove the top cover.

### 6.3 BASIC TECHNIQUES

Troubleshooting requires no special technique. Listed below are a few reminders of basic electronic fault isolation.

1. Check control settings carefully. Many times a seemingly malfunction is an incorrect control setting, or a knob that has loosened on its shaft.
2. Check associated equipment connections. Make sure that all connections are securely connected to the correct connector.
3. Perform the calibration procedure. Many out-of-specification indications can be corrected by performing specific calibration procedures.

4. Visually check the interior of the instrument. Look for such indications as broken wires, charred components, and loose leads.

### 6.4 TROUBLESHOOTING INDIVIDUAL COMPONENTS

#### 6.4.1 Transistor

A transistor is defective if more than one volt is measured across its base emitter junction in the forward direction.

A transistor when used as a switch may have a few volts reverse bias voltage.

If the collector and emitter voltages are the same, but the base emitter voltage is less than 500 mV forward voltage (or reversed bias), the transistor is defective.

A transistor is defective if its base current is larger than 10% of its emitter current (calculate currents from voltage across the base and emitter series resistors).

#### 6.4.2 Diode

A diode is defective if there is greater than 1 volt (typically 0.7 volt) forward voltage across it (except Zener and LED).

#### 6.4.3 Operational Amplifier

The "+" and "-" inputs of an operational amplifier will have less than 15 mV voltage difference when operating under normal conditions.

If the output voltage stays at maximum positive, its "+" input voltage should be more positive than its "-" input voltage, or vice versa; otherwise, the operational amplifier is defective.

#### 6.4.4 FET Transistor

No gate current should be drawn by the gate of an FET transistor. If so, the transistor is defective.

The gate-to-source voltage is always reverse biased under a normal operating condition; e.g., the source voltage is more positive than the gate voltage for 2N5485, and the source

voltage is more negative than gate voltage for a 2N5462. Otherwise, the FET is defective.

#### 6.4.5 Capacitor

Shorted capacitors have zero volts across their terminals.

Opened capacitor can be located (but not always) by using a good capacitor connected in parallel with the capacitor under test and observing the resulting effect.

### 6.5 TROUBLESHOOTING GUIDE

Table 6-1 provides a list of possible malfunction symptoms, their probable causes, and the prescribed remedies. Localize the fault to a specific stage by checking the parameters given for the test points. Then check the dc operating voltages at the pins of solid-state devices. Check associated passive elements with a high input impedance ohmmeter (power off) before replacing a suspected semiconductor element.

Table 6-1. Troubleshooting Guide

Symptom	Corrective Procedures
<b>POWER SUPPLY PROBLEM</b>	
Blown fuse	<ol style="list-style-type: none"> <li>1. Check that the HI/LO and 115/230 switches at the rear panel are set correctly. (Refer to paragraph 2.2.)</li> <li>2. Replace fuse; if fuse blows again, refer to the following steps.</li> <li>3. Examine circuit boards and wiring for source of short circuit.</li> <li>4. Use an ohmmeter to detect possible short circuits between power supply and ground and between individual power supplies.</li> <li>5. Isolate each part of the circuit by unplugging the sweep board and unsoldering the jumpers along the power supply path. Plug in the sweep board and replace the jumpers one-by-one to isolate the overloading circuit. Frequently, a shorted capacitor is the problem.</li> </ol>
±15V supply voltage below normal	Isolate the power supply from most of the generator circuits by unsoldering the two jumpers near the "+" end of C81 on the main circuit board. If supply voltage returns to normal, there is an extra loading current from a generator circuit; otherwise, troubleshoot the power supply circuitry.
±15V supply voltage above normal	Power supply circuit malfunction.
+5V regulator voltage abnormal	Isolate the regulator from generator circuits by unsoldering any leads at E15, E16 and E17. If regulator voltage returns to normal, there is an extra loading current from a generator circuit; otherwise, the trouble is in the regulator. Replace IC10.
Index (lighted indicator) on front panel abnormally bright or dim	HI/LO switch at the rear panel is not set correctly. (Refer to paragraph 2.2.)
<b>OUTPUT WAVEFORM PROBLEM</b>	
No output waveform at 50Ω OUT and PULSE OUT (GEN MODE at CONT)	<ol style="list-style-type: none"> <li>1. Ensure power supply voltages are normal.</li> <li>2. Temporarily remove Q44 on main board. If generator runs, problem is in the trigger and gate logic circuit. Otherwise, trouble is in the generator loop.</li> </ol>

Table 6-1. Troubleshooting Guide (Continued)

Symptom	Corrective Procedures
No output waveform at 50Ω OUT, but PULSE OUT normal, or all waveforms greatly distorted	Set the AMPLITUDE VARIABLE full ccw and set WAVEFORM to DC. If the output voltage at 50Ω OUT can be adjusted to ±10V into open circuit with the DC OFFSET control, the problem is in the preamplifier; otherwise, check the output amplifier.
Both waveform amplitude and frequency jittering	<ol style="list-style-type: none"> <li>1. Power supply out of regulation due to ac line voltage being too low. Check line voltage. Make sure the HI/LO switch setting on rear panel is correct. (Refer to paragraph 2.2.)</li> <li>2. Power supply malfunction. (Refer to Power Supply Problem.)</li> </ol>
Distorted sine and square waveforms, but triangle waveform normal	Sine converter and square shaper malfunction. Check for defective diode.  <i>NOTE</i>  <i>If a diode is bad, the entire set of eight diodes should be replaced with a new matched set, or select a diode that gives minimum sine distortion at 1 kHz.</i>
Half of sine and square waveforms missing	<ol style="list-style-type: none"> <li>1. Defective diodes CR17 or CR21.</li> <li>2. Defective switch wafer or loose contact of SW3-A and SW3-B.</li> </ol>
Distorted triangle and sine waveforms at one particular frequency range	<ol style="list-style-type: none"> <li>1. Check for defective timing capacitor of the range (C15 thru C23).</li> <li>2. Check C8 thru C10, C13, C25 and C94, if distortion shown at X 1 MHz range.</li> </ol>
Distorted waveform or generator not running when X .001 Hz thru X 10 Hz selected	Capacitance multiplier malfunction.
Sine distortion out of specification at frequency below 500 kHz	<ol style="list-style-type: none"> <li>1. Square wave time symmetry is not calibrated correctly.</li> <li>2. Defective component in sine converter and square shaper.</li> </ol> <i>NOTE</i>  <i>If a diode is bad, the entire set of eight diodes should be replaced with a new matched set, or select a diode that gives minimum sine distortion at 1 kHz.</i> <ol style="list-style-type: none"> <li>3. Resistor R109, R111, R112 or R114 is out of tolerance. Connect 10 kΩ trim potentiometers in locations marked R111 and R112. Adjust the two trim potentiometers and also R68 and R71 to obtain less than 0.16% distortion. Remove the potentiometers, measure the resistance and replace with standard 1/8W resistors. If 0.16% distortion still cannot be achieved, remove both R110 and R113 and connect a 500Ω trim potentiometer in each location. Adjust the two trim potentiometers R68 and R71 for less than 0.16% distortion. Replace potentiometers with standard 1/8W resistors.</li> <li>4. If sine distortion is OK at 1 kHz, but out of specification at 10 kHz, check for defective C31, C32, C38, Q6 and Q14.</li> </ol>

Table 6-1. Troubleshooting Guide (Continued)

Symptom	Corrective Procedures
Sine distortion out of specification at frequency greater than 500 kHz	<ol style="list-style-type: none"> <li>1. Check square wave for slow rise/fall time. If so, check for defective capacitor in the pre-amplifier and output amplifier.</li> <li>2. Frequency dial accuracy and sine distortion problems at X 1M range are due to the excess peaking or roll-off of the triangle waveform. Capacitors C28 and C35, also C29 and C34, need to be selected for maximum flatness of the triangle peak voltages at emitter of Q19. To check the flatness of the triangle peak voltage, a high frequency oscilloscope and a X 10 scope probe (&gt; 150 MHz bandwidth) should be used. The oscilloscope probe should be correctly compensated and its ground lead length should keep to minimum.</li> <li>3. If triangle is distorted, check for defective capacitors C8 thru C10, C13, C25 and C94.</li> <li>4. Check for defective diodes CR10 or CR11.</li> </ol>

TIME SYMMETRY PROBLEM

Positive slope of triangle remains constant when frequency dial varied	<ol style="list-style-type: none"> <li>1. Defective Q5, Q6, C9, IC3 and associated circuitry.</li> <li>2. Defective Q9 thru Q12 and CR6 thru CR9.</li> </ol>
Negative slope of triangle remains constant when frequency dial varied	<ol style="list-style-type: none"> <li>1. Defective IC3 and associated circuitry.</li> <li>2. Defective Q9 thru Q12 and CR6 thru CR9.</li> </ol>
Symmetry cannot be adjusted to specification	Defective Q6, Q14, R33, R34, R40 and R41.
Symmetry worse at low frequency end of dial	Check for high leakage components Q6, Q9 thru Q12 and Q14.
Symmetry out of specification at X 10 frequency range or below	Defective IC6 and associated circuitry.

FREQUENCY ACCURACY PROBLEM

Frequency accuracy out of specification at X 1 kHz range	<ol style="list-style-type: none"> <li>1. Mismatched dial and potentiometer, if frequency is out of specification at the same portion of the dial in every range. Ensure that the number on the back of the dial matches the number on the potentiometer.</li> <li>2. Defective dial potentiometer.</li> <li>3. VCG amplifier (IC2) or current source (IC3) is saturated when frequency dial is set to the top (5.0). Check for defective Q1, Q6, IC2 and associated circuitry.</li> </ol>
Frequency accuracy out of specification at X 10K and X 100 kHz ranges	Check for defective C30 thru C33, C38 and R61 thru R66.
Frequency accuracy out of specification at X 1 MHz range	Check for defective C25, C28, C29, C34, C35, R60, R67, CR10 and CR11.

Table 6-1. Troubleshooting Guide (Continued)

Symptom	Corrective Procedures
Frequency accuracy out of specification at X .001 to X 10 Hz ranges	<ol style="list-style-type: none"> <li>1. R90 and R94 thru R96 are mismatched. Defective R97.</li> <li>2. Defective IC5, IC6 and associated circuitry.</li> <li>3. If triangle is distorted when dial is set to the top (5.0), defective regulator Q22 and Q23.</li> </ol>
<b>MODE OF OPERATION PROBLEM</b>	
Output not in agreement with GEN MODE switch setting	Trigger and gate logic circuit or IC8 malfunction.
Generator running in trigger or gated mode	If voltage at pin 11 of IC8 is 0 to 0.4V (logic zero) when TRIG mode is selected, the problem is in the trigger amplifier (Q42 thru Q45). Otherwise, troubleshoot IC8, IC9 and associated circuitry.
Generator can be triggered by operating MAN TRIG switch, but not by external signal	Squaring circuit malfunction.
<b>FREQUENCY SWEEP PROBLEM (All components on sweep board)</b>	
Generator frequency not sweeping; no ramp signal at SWP OUT (GEN MODE at SWEEP, SWEEP MODE at CONT SWEEP)	<ol style="list-style-type: none"> <li>1. If IC1 pins 2 and 3 do not have same voltage level, the buffer amplifier (IC1 and Q6) is bad.</li> <li>2. Comparator flip-flop (Q7 thru Q10) is bad.</li> <li>3. Log current source (Q3 thru Q5) is bad.</li> </ol>



# SECTION 7

## PARTS AND SCHEMATICS

### 7.1 DRAWINGS

The following assembly drawings (with parts lists) and schematics are in the arrangement shown below.

### 7.2 ORDERING PARTS

When ordering spare parts, please specify part number, circuit reference, board, serial number of unit, and if applicable, the function performed.

### 7.3 ADDENDA


Under Wavetek's product improvement program, the latest electronic designs and circuits are incorporated into each Wavetek instrument as quickly as development and testing permit. Because of the time needed to compose and print instruction manuals, it is not always possible to include the most recent changes in the initial printing. Whenever this occurs, addendum pages are prepared to summarize the changes made and are inserted immediately inside the rear cover. If no such pages exist, the manual is correct as printed.

	Drawing No.
<b>CHASSIS</b>	
Schematic	0004-00-0054
Assembly Drawing	0102-00-0313
Parts List	1101-00-0056
<b>MAIN BOARD</b>	
Schematic	0103-00-0126
Assembly Drawing	0101-00-0126
Parts List	1100-00-0128
<b>SWEEP BOARD</b>	
Schematic	0103-00-0123
Assembly Drawing	0101-00-0123
Parts List	1100-00-0123









THIS DOCUMENT CONTAINS PROPRIETARY INFORMATION AND IS THE PROPERTY OF WAVE-TEK CORPORATION. IT IS TO BE KEPT CONFIDENTIAL AND NOT BE REPRODUCED FOR ANY PURPOSES WITHOUT WRITTEN AUTHORIZATION.

ALL CONNECTIONS TO BE MECHANICALLY SECURE TO LUGS PRIOR TO SOLDERING GROUND WIRES

NOTE: UNLESS OTHERWISE SPECIFIED

DATE	DATE	BY	DATE	APP
8/1/79	8/1/79	RD	5/2/78	
<p>REMOVE ALL BURRS AND BREAK SHARP EDGES MATERIAL</p> <p>FINISH: WAVE-TEK PROCESS</p> <p>DO NOT SCALE DIMS</p>				
<p>WAVE-TEK SAN DIEGO - CALIFORNIA</p> <p>CHASSIS ASSEMBLY</p> <p>MODEL NO. 184 183</p> <p>FIG. NO. C102-00-0313</p> <p>23338</p> <p>SHEET 1 OF 1</p>				

THIS DOCUMENT CONTAINS PROPRIETARY INFORMATION AND DESIGN RIGHTS BELONGING TO WAVETEK CORPORATION. IT IS TO BE USED FOR MAINTENANCE ONLY. REPRODUCTION, OPERATION AND MAINTENANCE WITHOUT WRITTEN AUTHORIZATION.

REV	BY	DATE	APP
1			

REFERENCE DESIGNATIONS	PART DESCRIPTION	UM10-MFG-PART-NO	MFG	WAVETEK NO.	QTY/PI
NONE	ASSY DMWG CHASSIS	0102-00-0315	WYK	0102-00-0315	1
NONE	OVAL ASSY	183-025	WYK	1201-00-0025	1
NONE	TRANSFORMER	5800-00-0022	WYK	1604-00-00H3	1
NONE	END BELL	110-333	WYK	1800-00-0178	1
NONE	PLATE/NAME	189-305	WYK	1800-00-0180	1
NONE	INDICATION OVAL	180-503	WYK	1800-00-0070	1
NONE	COVER, TOP	180-300-1	WYK	1800-00-3000	1
NONE	EXPANDER	180-581	WYK	1800-00-5016	2
NONE	POST	180-502	WYK	1800-00-5020	4
NONE	COVER, BOTTOM	180-500-2	WYK	1800-00-5050	1
NONE	PANEL, REAR	182-301	WYK	1800-00-5113	1
NONE	COND. MODULE	1800-00-5600	WYK	1800-00-5600	1
NONE	PANEL, FT	1800-00-6730	WYK	1800-00-6730	1
NONE	LABEL, I.D.	1800-00-5950	WYK	1800-00-5950	1
NONE	BIC CONN	KL-7946	KLN	2100-01-5002	6
NONE	SOLDER LUS	1897	SMIM	2100-04-0012	6
NONE	SOLDER LUS	1885-6	SMIM	2100-04-0025	4
NONE	EVAN	08595-6	AMP	2100-00-0017	2

WAVETEK PARTS LIST  
 TITLE: 810 CHASSIS  
 ASSEMBLY NO: 1101-00-0056  
 PAGE: 1

REFERENCE DESIGNATIONS	PART DESCRIPTION	UM10-MFG-PART-NO	MFG	WAVETEK NO.	QTY/PI
NONE	STD. NUB	HB-87-1-88-M	MUGN	2800-01-0008	4
NONE	EUAX NUB SET	HB-87-1-88-M-9	MUGN	2800-01-0009	4
NONE	LAMP	CH-7576	DMEN	2800-02-0013	1
NONE	FUSE, 125V, 1/4A, 5F	313-250	LIFU	2800-05-0008	1
NONE	NUB HOLD	345001	LIFU	2800-05-0012	1
NONE	BURNING MILIBER	4L2FF	TRGM	2800-01-0002	6
NONE	STANDOFF	1875-M03-F05-032	UNLCP	2800-02-0016	4
NONE	STANDOFF	1895-M03-F05-040	UNLCP	2800-02-0011	4
NONE	PAIL ASSY W/FT	180-500	WTA	2800-08-0016	1
NONE	SPEEDONT, SELF METAL	1795-032-4	11W	2800-04-0003	6
NONE	WASHER, SHROUDER	2858	SMIM	2800-27-0009	12
NONE	STRAIN RELIEF BUSH	W550-1	MEYCO	2800-37-0005	1
NONE	SW ASSY SLIDE #1-LU	S482-F1-HK	ULI	5105-00-0001	1
NONE	SWITCH ASSY SLIDE	W250-LF	SGLT	5105-00-0002	1
NONE	SOLDER BOARD	08255-LF-SG	SGLT	5105-04-0001	2
NONE	PHR CARD	0-778V-008-07	PALMG	0801-88-0008	1

WAVETEK PARTS LIST  
 TITLE: 810 CHASSIS  
 ASSEMBLY NO: 1101-00-0056  
 PAGE: 2

REMOVE ALL DIMENSIONS FROM THIS DRAWING UNLESS OTHERWISE SPECIFIED. ALL DIMENSIONS ARE IN INCHES UNLESS OTHERWISE SPECIFIED. DO NOT SCALE DIMENSIONS.

DATE: \_\_\_\_\_  
 TITLE: PARTS LIST CHASSIS  
 WAVETEK PROCESS  
 SCALE: \_\_\_\_\_  
 SHEET NO: 184  
 OF: 23338  
 REV: A

REV	BY	DATE	APP
1			

REFERENCE DESIGNATIONS	PART DESCRIPTION	UM10-MFG-PART-NO	MFG	WAVETEK NO.	QTY/PI
NONE	ASSY DMWG CHASSIS	0102-00-0315	WYK	0102-00-0315	1
NONE	OVAL ASSY	183-025	WYK	1201-00-0025	1
NONE	TRANSFORMER	5800-00-0022	WYK	1604-00-00H3	1
NONE	END BELL	110-333	WYK	1800-00-0178	1
NONE	PLATE/NAME	189-305	WYK	1800-00-0180	1
NONE	INDICATION OVAL	180-503	WYK	1800-00-0070	1
NONE	COVER, TOP	180-300-1	WYK	1800-00-3000	1
NONE	EXPANDER	180-581	WYK	1800-00-5016	2
NONE	POST	180-502	WYK	1800-00-5020	4
NONE	COVER, BOTTOM	180-500-2	WYK	1800-00-5050	1
NONE	PANEL, REAR	182-301	WYK	1800-00-5113	1
NONE	COND. MODULE	1800-00-5600	WYK	1800-00-5600	1
NONE	PANEL, FT	1800-00-6730	WYK	1800-00-6730	1
NONE	LABEL, I.D.	1800-00-5950	WYK	1800-00-5950	1
NONE	BIC CONN	KL-7946	KLN	2100-01-5002	6
NONE	SOLDER LUS	1897	SMIM	2100-04-0012	6
NONE	SOLDER LUS	1885-6	SMIM	2100-04-0025	4
NONE	EVAN	08595-6	AMP	2100-00-0017	2

WAVETEK PARTS LIST  
 TITLE: 810 CHASSIS  
 ASSEMBLY NO: 1101-00-0056  
 PAGE: 1

REFERENCE DESIGNATIONS	PART DESCRIPTION	UM10-MFG-PART-NO	MFG	WAVETEK NO.	QTY/PI
NONE	STD. NUB	HB-87-1-88-M	MUGN	2800-01-0008	4
NONE	EUAX NUB SET	HB-87-1-88-M-9	MUGN	2800-01-0009	4
NONE	LAMP	CH-7576	DMEN	2800-02-0013	1
NONE	FUSE, 125V, 1/4A, 5F	313-250	LIFU	2800-05-0008	1
NONE	NUB HOLD	345001	LIFU	2800-05-0012	1
NONE	BURNING MILIBER	4L2FF	TRGM	2800-01-0002	6
NONE	STANDOFF	1875-M03-F05-032	UNLCP	2800-02-0016	4
NONE	STANDOFF	1895-M03-F05-040	UNLCP	2800-02-0011	4
NONE	PAIL ASSY W/FT	180-500	WTA	2800-08-0016	1
NONE	SPEEDONT, SELF METAL	1795-032-4	11W	2800-04-0003	6
NONE	WASHER, SHROUDER	2858	SMIM	2800-27-0009	12
NONE	STRAIN RELIEF BUSH	W550-1	MEYCO	2800-37-0005	1
NONE	SW ASSY SLIDE #1-LU	S482-F1-HK	ULI	5105-00-0001	1
NONE	SWITCH ASSY SLIDE	W250-LF	SGLT	5105-00-0002	1
NONE	SOLDER BOARD	08255-LF-SG	SGLT	5105-04-0001	2
NONE	PHR CARD	0-778V-008-07	PALMG	0801-88-0008	1


WAVETEK PARTS LIST  
 TITLE: 810 CHASSIS  
 ASSEMBLY NO: 1101-00-0056  
 PAGE: 2

REMOVE ALL DIMENSIONS FROM THIS DRAWING UNLESS OTHERWISE SPECIFIED. ALL DIMENSIONS ARE IN INCHES UNLESS OTHERWISE SPECIFIED. DO NOT SCALE DIMENSIONS.

DATE: \_\_\_\_\_  
 TITLE: PARTS LIST CHASSIS  
 WAVETEK PROCESS  
 SCALE: \_\_\_\_\_  
 SHEET NO: 184  
 OF: 23338  
 REV: A







THIS DOCUMENT CONTAINS PROPRIETARY INFORMATION OF WAVEJETEK CORPORATION. IT IS THE PROPERTY OF WAVEJETEK CORPORATION AND IS LOANED TO YOU. IT IS NOT TO BE REPRODUCED OR TRANSMITTED IN ANY FORM OR BY ANY MEANS, ELECTRONIC OR MECHANICAL, INCLUDING PHOTOCOPYING, RECORDING, OR BY ANY INFORMATION STORAGE AND RETRIEVAL SYSTEM, WITHOUT WRITTEN AUTHORIZATION.

WAVEJETEK  
 Schematic  
 MAIN BOARD

DATE: 11/18/83  
 BY: J. J. JONES  
 CHECKED: J. J. JONES  
 APPROVED: J. J. JONES  
 SCALE: DO NOT SCALE DRAWING  
 MODEL NO: 183 184 185 0103-00-0126  
 SHEET 2 OF 3

Δ MATCHED PAIR  
 REMOVE ALL BUBBLES AND BRICK SHAPED EGGS  
 MATCHED PAIR

WAVEJETEK PROCESS

DO NOT SCALE DRAWING


SCALE

MODEL NO: 183 184 185 0103-00-0126

SHEET 2 OF 3

\* FD TTT MATCHED SET OF 8  
 NOTE: UNLESS OTHERWISE SPECIFIED

REV	ECN	BY	DATE	APP
A	REVISD PER ENG.	WA	8-4-85	
B	ECN 1527	RO	2-18-85	
C	ECN 1407/1408	RO	1-14-85	
D	ECN 1719	RA	5-28-85	



DATE	1985	WAVETEK
TITLE	SILKSCREEN ASSEMBLY PRINT (MAIN BOARD)	
SCALE	2/1	
MODEL NO.	183 184 185	010-00-0261D
REV	23338	SHEET 1 OF 2

ETZ1 WAS E7Z  
 E481 " E48  
 E521 " E52  
 E111 WAS E9  
 E10 ADDED TO  
 HOLE UNDER ETZ1

2. SOLDER R227 AT E7 LOCATION ON 183, 184, 185  
 ADD R226 AT E111 ON MODEL 185.  
 1. 9 PIN MOLEX TO CRYSTAL BD TO BE  
 WIRED BACKWARDS SO THAT BROWN  
 WIRE IS PIN 9 AND WHITE WIRE IS  
 PIN 1.

MACROVEL 8-3384E








THIS DOCUMENT CONTAINS PROPRIETARY INFORMATION AND DESIGN RIGHTS BELONGING TO NATIONAL INSTRUMENTS CORPORATION. REPRODUCTION, OPERATION, AND MAINTENANCE WITHOUT WRITTEN AUTHORIZATION IS PROHIBITED.

REV	DATE	BY	APP
A	ECM	RD	S-77





- J1 P1 BLANK
- 1 5V
- 2 BLANK
- 3 BLANK
- 4 VCG
- 5 15V
- 6 15V
- 7 15V

REMOVE ALL BURRS AND BREAK SHARP EDGES MATERIAL	DATE 4-19-53	DESIGNER J. J. WILSON	WAVETEK SAN DIEGO - CALIFORNIA
FINISH WAVEFORM PROCESS	PROJECT NAME 183-184	TOLERANCE UNLESS SPECIFIED: XX - 10% UNLESS NOTED	SCHEMATIC SWEEP BOARD
DO NOT SCALE DIMS	SCALE 183-184	DO NOT SCALE DIMS	0103-00-0123
			REV. 4
			23338
			SHEET 1 OF 1

LAST DES USED  
 R15  
 CR4  
 Q1  
 Q2  
 ICI  
 SW1

1. \* MISMATCHED PAIR.  
 NOTE: UNLESS OTHERWISE SPECIFIED



DRAWN	DATE	WAVETEK	DATE
BA	1975	JANIDRO/DAUF	
ENGR		TITLE	SILKSCREEN ASSEMBLY PRINT (SWEEP BD)
RELEASE APPROV		MODEL NO	183-184
TELEPHONE		DATE '85	010-00-0123
SCALE		REV	A
		23338	SHEET 1 OF 1





