
VisiTrigger Quick Start

*for HP 16715A, HP 16716A, and HP 16717A
logic analyzer modules*

Read this if you are familiar with the trigger user interface on HP 16500, HP 16600, or HP 16700 logic analysis systems.

This quick start guide covers:

- Using Trigger Functions
- Editing Trigger Sequence Levels
- Storage Qualification
- Inter-Module Triggering

These user interface changes will occur only with the HP 16715A, HP 16716A, and HP 16717A logic analysis modules, which require Version A.01.40.00 software. For the user interface changes in Version A.01.40.00, see the *Version A.01.40.00 User Interface Changes* booklet for more information.

Hewlett-Packard Company
Printed in the USA

Manual Part Number
16715-92000, April 1999



Using Trigger Functions


Functions are same as "Macros" in previous versions.

To see the matching graphic, single-click a function.

To replace the selected sequence level, double-click a function.

The advanced functions are at the bottom of the list. Use them for multi-way branching, timers and counters.

Click here and select "Replace label.." to change the label. Select "Insert label.." to create an AND.



These buttons use the currently selected function.

Editing Trigger Sequence Levels

General rule: Click on a button to get a list of options.

Click here and select "Insert Branch" to add another branch.

Click here to change the operator to equals, edge, range, >, etc.

Multiple actions are allowed in each branch. Click here and select "Insert action" to add another action.

To create an AND/OR, click here and select "Insert event". This replaces the Combo dialog.

Trigger Sequence

1	If	ADDR	=	1234	Hex
		occurs	1	time	
then		Timer	1	Start from reset	
		Flag	1	Set	
		Goto	Next		
Else if		ADDR	In range	0000	1325 Hex Or
		ADDR	>=	F345	Hex
then		Goto	2		

Multiple branches, timers, counters and flags are only available in advanced functions.


Storage Qualification

Default Storing means “unless a sequence level says otherwise, this is what to store”. This eliminates the need to specify storage for each sequence level.

There are actions to turn default storing On and Off.

This overrides the default storing. This function is the same as the “While Storing” macro in previous versions.

This sequence level doesn’t specify what is stored, so the Default Storing applies.




Intermodule Triggering

The Intermodule dialog still controls the arming order. It is easier if the arming order is set up before specifying the triggers for each module.

Specify which analyzer is to send the arm out signal. This only appears if both analyzers in a module are active.

There's a "Wait for Arm In" trigger function. Use it to specify which sequence level is to wait for the "arm in" signal.

Arm out signals are sent when this module triggers. This is set up in the Intermodule dialog.



Summary

- Trigger “macros” have become trigger “functions”.
- Click on a trigger function in the list; then, click on the “Replace”, “Insert Before”, or “Insert After” buttons. Double-clicking a function is a short cut for “Replace”.
- To create an “AND”, click on the label button and select “Insert”. This replaces the combo dialog.
- “While storing” becomes “Store pattern until pattern occurs”.
- Within each branch, multiple actions are allowed, one of which must be a Go To or Trigger action.
- Advanced functions are located at the bottom of the list. Use them for multiple branches, timers, flags, and global counters.
- The Arming Control dialog has been replaced by the trigger functions “Wait for Arm in” and “Wait for other analyzer to trigger”.