



OPERATING AND SERVICE MANUAL

37203A

HP-IB EXTENDER (Including Option 001)

SERIAL NUMBERS

This manual applies directly to instruments with serial number prefixed 2009U. For additional important information about serial numbers, see INSTRUMENTS COVERED BY MANUAL in Section I.

© Copyright HEWLETT-PACKARD LIMITED 1980
SOUTH QUEENSFERRY, WEST LOTHIAN, SCOTLAND

Manual Part Number 37203-90000
Microfiche Part Number 37203-90025

Printed: AUGUST 1980



Figure 1-1 Model 37203A HP-IB Extender

SECTION I GENERAL INFORMATION

1-1 INTRODUCTION

1-2 This Operating and Service Manual contains information required to install, operate, test, adjust and service the Hewlett-Packard Model 37203A HP-IB Extender. Figure 1-1 shows the 37203A.

1-3 This section of the manual describes the instrument and includes information on the identification, accessories, characteristics, safety and other basic information.

1-4 Listed on the title page of this manual is a microfiche part number. This number can be used to order 4 x 6 inch microfilm transparencies of the manual. Each microfiche contains up to 96 photo duplicates of the manual pages. The microfiche package also includes the latest Manual Changes supplement.

1-5 OPERATING CHARACTERISTICS

1-6 These Operating Characteristics summarise the features and nominal performance of the 37203A. They do not constitute warrantable specifications.

1-7 HP-IB Data-Byte Rate

1-8 The HP-IB data-byte rate differs from system to system, depending on the following factors:

Serial transmission rate
Serial link propagation delay
Handshake rate of devices connected to the HP-IB Extenders.

1-9 The recommended coaxial cable or fibre optic cable was used in achieving the maximum transfer rates given in Table 1-1 and the speed of the 37203A was not reduced by the effect of other devices connected to the bus.

1-10 HP-IB Interface

1-11 The 37203A may be connected directly to a maximum of 14 other HP-IB devices subject to the normal cabling restrictions imposed by the interface standard. It will provide extension for the full range of HP-IB functions.

1-12 HP-IB is Hewlett-Packard's implementation of IEEE Standard 488-1978. The 37203A is in general compliance with each of the following standards and supports their major capabilities:

IEEE Standard 488-1978
ANSI Standard MC1.1
IEC Standard 625-1

1-13 Total compatibility among independently designed products interconnected via the 37203A is beyond the control and scope of this product, see Section 6 of the above referenced standards.

Table 1-1 Maximum Transfer Rates (Nominal)

	Coaxial Cable (see Note 2)				Fibre Optic Cable (Opt 001)		
	Short (at normal speed) See Note 1	250m (max range at normal speed)	500m (max range at 1/4 speed)	1000m (max range at 1/16 speed)	Short See Note 1	250m	1000m
Max HP-IB byte transfer rate (k byte/s)	50	40	14.5	2.75	50	39	25
Max SRQ propagation delay (µs)	14	18	55	200	14	20	30
Max Parallel Poll response time (µs)	20	25	75	270	20	25	40

Note 1 For distances <250m, interpolate between Short and 250m columns.

Note 2 These results were achieved using the recommended coaxial cable (Belden cable type 9248)

1-26 One 37203A converts the bit parallel HP-IB protocol into a bit serial stream that is transmitted over coaxial cable or dual fibre optic cable for distances up to 1000 metres. The 37203A at the other end of the serial link performs the reverse function. The entire range of HP-IB functions may be extended to the remote sites but the timing of the parallel poll response will be slightly altered due to the transmission delay in an extended system.

1-27 A pair of HP-IB Extenders communicate with each other using 22 bit data frames. These data frames shuttle continuously back and forth between ends, continually updating each end with the current state of the devices and HP-IB Extender at the other.

1-28 Each transmitted data frame includes a cyclic redundancy check code which is rechecked when the frame is received. Any errors which are detected will cause the erroneous frame to be rejected and one 37203A will restart the data frame shuttle.

1-29 Two HP-IB Extenders can extend the full range of HP-IB functions. The HP-IB Extenders provide a transparent interface i.e. it is usually possible to insert two HP-IB Extenders at any point in the HP-IB without altering the controller-programming. Due to transmission delay in an extended system the parallel poll response time from remote devices will increase slightly. The parallel poll response time from local devices will not be affected.

1-30 The maximum number of HP-IB devices that can be connected to a single 37203A is 14. A complete system using one pair of Extenders therefore, could consist of up to 14 devices, plus a 37203A at the local end, and another 14 HP-IB device, plus a 37203A at the remote end.

1-31 OPTIONS

1-32 Option 001

1-33 The 37203A Option 001 adds a fibre optic transmitter and fibre optic receiver to the standard 37203A. This provides the capability of extending the HP-IB over either dual fibre optic cable or coaxial cable. The 37203A Option 001 transmits the bit serial stream over dual fibre optic cable for distances up to 1000 metres. The 37203A Option 001 enables a higher HP-IB data-byte transfer rate to be attained for transmission distances greater than 250 metres than is possible with the standard 37203A (see Table 1-1). The fibre optic cable removes the metallic connection between the Bus Extenders and therefore eliminates any electromagnetic pick-up on the serial link. Option 001 is recommended for use in severe electrical environments or where the use of electrical signalling is not acceptable.

1-34 Option 301

1-35 The 37203A Option 301 provides a kit for mounting a single 37203A into a full module width rack, see Figure 1-3.




Figure 1-3 Option 301

SECTION I

GENERAL INFORMATION

1-1 INTRODUCTION

1-2 This Operating and Service Manual contains information required to install, operate, test, adjust and service the Hewlett-Packard Model 37203A HP-IB Extender. Figure 1-1 shows the 37203A.

1-3 This section of the manual describes the instrument and includes information on the identification, accessories, characteristics, safety and other basic information.

1-4 Listed on the title page of this manual is a microfiche part number. This number can be used to order 4 x 6 inch microfilm transparencies of the manual. Each microfiche contains up to 96 photo duplicates of the manual pages. The microfiche package also includes the latest Manual Changes supplement.

1-5 OPERATING CHARACTERISTICS

1-6 These Operating Characteristics summarise the features and nominal performance of the 37203A. They do not constitute warrantable specifications.

1-7 HP-IB Data-Byte Rate

1-8 The HP-IB data-byte rate differs from system to system, depending on the following factors:

Serial transmission rate

Serial link propagation delay

Handshake rate of devices connected to the HP-IB Extenders.

1-9 The recommended coaxial cable or fibre optic cable was used in achieving the maximum transfer rates given in Table 1-1 and the speed of the 37203A was not reduced by the effect of other devices connected to the bus.

1-10 HP-IB Interface

1-11 The 37203A may be connected directly to a maximum of 14 other HP-IB devices subject to the normal cabling restrictions imposed by the interface standard. It will provide extension for the full range of HP-IB functions.

1-12 HP-IB is Hewlett-Packard's implementation of IEEE Standard 488-1978. The 37203A is in general compliance with each of the following standards and supports their major capabilities:

IEEE Standard 488-1978

ANSI Standard MC1.1

IEC Standard 625-1

1-13 Total compatibility among independently designed products interconnected via the 37203A is beyond the control and scope of this product, see Section 6 of the above referenced standards.

Table 1-1 Maximum Transfer Rates (Nominal)

	Coaxial Cable (see Note 2)				Fibre Optic Cable (Opt 001)		
	Short (at normal speed) See Note 1	250m (max range at normal speed)	500m (max range at 1/4 speed)	1000m (max range at 1/16 speed)	Short See Note 1	250m	1000m
Max HP-IB byte transfer rate (k byte/s)	50	40	14.5	2.75	50	39	25
Max SRQ propagation delay (μ s)	14	18	55	200	14	20	30
Max Parallel Poll response time (μ s)	20	25	75	270	20	25	40

Note 1 For distances <250m, interpolate between Short and 250m columns.

Note 2 These results were achieved using the recommended coaxial cable (Belden cable type 9248)

1-14 General

Power Requirements: 100/120/220/240 ac +5 -10%; 48 to 66Hz, 25VA max.

Operating Temperature: 0 to +55°C.

Dimensions: 89mm high, 213mm wide, 356mm deep.

Weight: 3.5kg (7.7lb), net. 5kg (11lb), shipping.

1-15 SAFETY CONSIDERATIONS

1-16 This Safety Class 1 instrument (provided with a protective earth terminal) has been designed and tested according to international safety standards. Information with regard to safety is presented at appropriate places throughout the manual.

1-17 INSTRUMENTS COVERED BY MANUAL

1-18 Attached to the instrument is a serial number plate. The serial number is in the form: 0000U00000. It is in two parts; the first four digits and the letter are the serial prefix and the last five digits are the suffix. The prefix is the same for all identical instruments; it changes only when a change is made to the instrument. The suffix, however, is assigned sequentially and is different for each instrument. The contents of this manual apply to instruments with the serial number prefix listed under SERIAL NUMBERS on the title page.

1-19 An instrument manufactured after the printing of this manual may have a serial number prefix that is not listed on the title page. This unlisted serial number prefix indicates the instrument is different from those described in this manual. The manual for this newer instrument is accompanied by a Manual Changes supplement. This supplement contains 'change information' that explains how to adapt the manual to the newer instrument.

1-20 In addition to change information, the supplement may contain information for correcting errors in the manual. To keep this manual as current and accurate as possible, Hewlett-Packard recommends that you periodically request the latest Manual Changes supplement. The supplement for this manual is identified with the manual print date and part number, both of which appear on the manual title page. Complementary copies of the supplement are available from Hewlett-Packard.

1-21 For information concerning a serial number prefix that is not listed on the title page or in the Manual Change supplement, contact your nearest Hewlett-Packard office.

1-22 DESCRIPTION

1-23 The 37203A HP-IB Extender enables the distance between groups of devices interfaced by the Hewlett-Packard Interface Bus (HP-IB) to be extended beyond the limits imposed by direct HP-IB cabling. Functional HP-IB operation of a programmed HP-IB system will usually be identical, with and without the 37203A's, except that due to the transmission delay in an extended system the parallel poll response time will increase slightly. There is no restriction in passing control between devices.

Note: HP-IB is Hewlett-Packard's implementation of IEEE Std. 488-1978 "Standard Digital Interface for Programmable Instrumentation".

1-24 The 37203A has a maximum HP-IB data-byte transfer rate of 50kbyte/s whether data is transmitted over coaxial cable or dual fibre optic cable. The actual HP-IB data-byte transfer rate will be governed principally by the speed of the slowest local or remote HP-IB device.

1-25 By using the 37203A HP-IB Extenders and a coaxial cable or dual fibre optic cable, transmission distances of up to 1000 metres are attainable (see Figure 1-2).




Figure 1-2 System Configuration - Point-to-Point

1-26 One 37203A converts the bit parallel HP-IB protocol into a bit serial stream that is transmitted over coaxial cable or dual fibre optic cable for distances up to 1000 metres. The 37203A at the other end of the serial link performs the reverse function. The entire range of HP-IB functions may be extended to the remote sites but the timing of the parallel poll response will be slightly altered due to the transmission delay in an extended system.

1-27 A pair of HP-IB Extenders communicate with each other using 22 bit data frames. These data frames shuttle continuously back and forth between ends, continually updating each end with the current state of the devices and HP-IB Extender at the other.

1-28 Each transmitted data frame includes a cyclic redundancy check code which is rechecked when the frame is received. Any errors which are detected will cause the erroneous frame to be rejected and one 37203A will restart the data frame shuttle.

1-29 Two HP-IB Extenders can extend the full range of HP-IB functions. The HP-IB Extenders provide a transparent interface i.e. it is usually possible to insert two HP-IB Extenders at any point in the HP-IB without altering the controller-programming. Due to transmission delay in an extended system the parallel poll response time from remote devices will increase slightly. The parallel poll response time from local devices will not be affected.

1-30 The maximum number of HP-IB devices that can be connected to a single 37203A is 14. A complete system using one pair of Extenders therefore, could consist of up to 14 devices, plus a 37203A at the local end, and another 14 HP-IB device, plus a 37203A at the remote end.

1-31 OPTIONS

1-32 Option 001

1-33 The 37203A Option 001 adds a fibre optic transmitter and fibre optic receiver to the standard 37203A. This provides the capability of extending the HP-IB over either dual fibre optic cable or coaxial cable. The 37203A Option 001 transmits the bit serial stream over dual fibre optic cable for distances up to 1000 metres. The 37203A Option 001 enables a higher HP-IB data-byte transfer rate to be attained for transmission distances greater than 250 metres than is possible with the standard 37203A (see Table 1-1). The fibre optic cable removes the metallic connection between the Bus Extenders and therefore eliminates any electromagnetic pick-up on the serial link. Option 001 is recommended for use in severe electrical environments or where the use of electrical signalling is not acceptable.

1-34 Option 301

1-35 The 37203A Option 301 provides a kit for mounting a single 37203A into a full module width rack, see Figure 1-3.




Figure 1-3 Option 301

1-36 Option 302

1-37 The 37203A Option 302 provides a kit for mounting a 37203A plus another unit of the same width and height or another 37203A into a full module width rack, see Figure 1-4.

1-38 ACCESSORIES SUPPLIED

1-39 Supplied with the 37203A is a power cord appropriate to the country of destination (see Section 2).




Figure 1-4 Option 302

1-40 ACCESSORIES AVAILABLE

1-41 Table 1-2 lists the accessories available from Hewlett-Packard which can be used in assembling an Extended HP-IB System.

Table 1-2 Accessories Available

Description		HP Part Number	
HP-IB Cables	1 metre	10833A	
	2 metres	10833B	
	4 metres	10833C	
	0.5 metre	10833D	
Fibre Optic Cable (with connectors)	≤1000 metres	Simplex Cable	Duplex Cable
		39200A	39200B
See Note 1			

Note 1: 39200A/B defines an optical cable for user specified length supplied with factory installed and tested connectors. Length must be specified in metres and can be any 1 metre increment from 1 to 1000 metres. Length information is shown as option 001 to the base product with quantity equal to the number of cable assemblies ordered. The length should include some excess to allow for limited repair.

*Examples: For two lengths of simplex cable 245 metres long specify
39200A Optic Cable Assy Quantity 2
Option 001 245 metres long Quantity 2*

*For seven lengths of duplex cable 1000 metres specify
39200B Optic Cable Assy Quantity 7
Option 001 1000 metres long Quantity 7*

39200A cable is simplex cable, therefore 2 lengths will be required per serial link.

1-42 RECOMMENDED TEST EQUIPMENT

1-43 Equipment required to maintain the 37203A is listed in Table 1-3. Other equipment may be substituted if it meets or exceeds the critical specifications listed in the table.

Table 1-3 Recommended Test Equipment

Instrument	Critical Specifications	Recommended Model	Use*
Oscilloscope and plug-ins	100MHz bw, 0.5µs/div, 4 Channels	HP 180C/1809A/1825A	A, T
Frequency Counter	Frequency range 10Hz to 25MHz	HP 5302A	A, T
Signature Analyser	Unique	HP 5004A	T
Digital Multimeter	±0.1 at ±5V	HP 3476A/B	T
Desktop Computer	Unique	HP 9825A	P, T
Interface Card (2 off)	Unique	HP 98034A	P, T
Performance Verification Tape	Unique	HP 37203- 12101	P, T
String & Advanced Programming ROM	Unique	HP 98210A	P, T
General I/O & Extended I/O ROM	Unique	HP 98213A	P, T
Logic Probe	TTL Compatible	HP 545A	T
Logic Pulser	TTL Compatible	HP 546A	T

* A = Adjustments, P = Performance Tests, T = Troubleshooting

CAUTION

Before connecting this instrument to a power outlet, ensure the voltage selector is correctly set for the voltage of the power source and a fuse of the correct rating is fitted.

2-11 Power Cable

2-12 In accordance with the international safety standards, this instrument is equipped with a three-wire power cable. When connected to an appropriate ac power receptacle, this cable grounds the instrument cabinet. The type of power cable supplied with the instrument depends on the country of destination. Figure 2-2 illustrates the standard power plugs commonly used. The number shown below each plug is the HP Part Number of the power cord equipped with that plug. If the appropriate power cord is not included with the instrument, notify the nearest Hewlett-Packard office and a replacement will be provided.

2-13 The colour codes used in each power cable are:

- Line - Brown
- Neutral - Blue
- Ground - Green/Yellow

2-14 Operating Environment

2-15 **Temperature.** The instrument may be operated in temperatures from 0°C to +55°C.

2-16 **Humidity.** The instrument should be protected from extreme temperature changes which may cause condensation within the instrument.

2-17 **Altitude.** The instrument may be operated at altitudes up to 4600 metres (15,000 feet).

2-18 Rack Mounting

2-19 Rack mounting kits are available and can be purchased through your nearest Hewlett-Packard office.

- Rack Mount Kit (for single unit) 5061-0072
- Rack Mount Kit (for two units) 5061-0094 and 5061-0074

WARNING

To avoid the possibility of injury or death, the following precautions must be followed before the instrument is switched on:

- (a) Note that the protection provided by grounding the instrument cabinet may be lost if any power cable other than the three-pronged type supplied is used to couple the ac line voltage to the instrument.
- (b) If this instrument is to be energized via an auto-transformer to reduce or increase the line voltage, make sure that the common terminal is connected to the neutral pole of the power source.
- (c) The power cable plug shall only be inserted into a socket outlet provided with a protective earth contact. The protective action must not be negated by the use of an extension cord without a protective conductor (grounding).




Figure 2-2 Power Receptacle

SECTION II

INSTALLATION

2-1 INTRODUCTION

2-2 This section contains information and instructions required to install the 37203A HP-IB Extender. This section also includes information about initial inspection, operating environment and storage and shipment.

2-3 INITIAL INSPECTION

2-4 Inspect the shipping container for damage. If the shipping container or cushioning material is damaged, it should be kept until the contents of the shipment have been checked for completeness and the instrument has been checked both mechanically and electrically. The contents of the shipment should be as shown in Figure 1-1. Procedures for performance verification are given in Section IV. If the contents of the shipment are incomplete, if there is mechanical damage or defect, or if the instrument does not pass the Performance Verification, notify the nearest Hewlett-Packard office. If the shipping container is damaged, or the cushioning material shows signs of stress, notify the carrier as well as the Hewlett-Packard office.

Keep the shipping material for carrier's inspection. The Hewlett-Packard office will arrange for repair or replacement at HP option without waiting for claim settlement.

2-5 PREPARATION FOR USE

2-6 Power Requirements

2-7 The 37203A HP-IB Extender requires a power source of 100, 120, 220, or 240V ac $\pm 5\%$ – 10% at 48 to 66Hz single phase. Power consumption is less than 25VA.

2-8 A timed fuse of 250mA, 250V normal blow (2110-0201) is required for 100/120V operation, and 125mA, 250V normal blow (2110-0318) is required for 220/240V.

2-9 Line Voltage Selection

2-10 Figure 2-1 provides instructions for line voltage and fuse selection.




Figure 2-1 Line Selector

CAUTION

Before connecting this instrument to a power outlet, ensure the voltage selector is correctly set for the voltage of the power source and a fuse of the correct rating is fitted.

2-11 Power Cable

2-12 In accordance with the international safety standards, this instrument is equipped with a three-wire power cable. When connected to an appropriate ac power receptacle, this cable grounds the instrument cabinet. The type of power cable supplied with the instrument depends on the country of destination. Figure 2-2 illustrates the standard power plugs commonly used. The number shown below each plug is the HP Part Number of the power cord equipped with that plug. If the appropriate power cord is not included with the instrument, notify the nearest Hewlett-Packard office and a replacement will be provided.

2-13 The colour codes used in each power cable are:

- Line — Brown
- Neutral — Blue
- Ground — Green/Yellow

2-14 Operating Environment

2-15 **Temperature.** The instrument may be operated in temperatures from 0°C to +55°C.

2-16 **Humidity.** The instrument should be protected from extreme temperature changes which may cause condensation within the instrument.

2-17 **Altitude.** The instrument may be operated at altitudes up to 4600 metres (15,000 feet).

2-18 Rack Mounting

2-19 Rack mounting kits are available and can be purchased through your nearest Hewlett-Packard office.

- Rack Mount Kit (for single unit) 5061-0072
- Rack Mount Kit (for two units) 5061-0094 and 5061-0074

WARNING

To avoid the possibility of injury or death, the following precautions must be followed before the instrument is switched on:

- (a) Note that the protection provided by grounding the instrument cabinet may be lost if any power cable other than the three-pronged type supplied is used to couple the ac line voltage to the instrument.
- (b) If this instrument is to be energized via an auto-transformer to reduce or increase the line voltage, make sure that the common terminal is connected to the neutral pole of the power source.
- (c) The power cable plug shall only be inserted into a socket outlet provided with a protective earth contact. The protective action must not be negated by the use of an extension cord without a protective conductor (grounding).




Figure 2-2 Power Receptacle

2-20 HP-IB

2-21 The instrument is connected to the HP-IB by means of an HP-IB interface cable from the connector on the rear panel of the instrument. Each end of the cable has both a male and female "piggyback" connector which simplifies interconnection of instruments and cables by allowing connectors to be stacked. Up to 15 HP-IB devices, including a 37203A, may be interconnected by HP-IB interface cables.

2-22 In order to maintain proper line voltages and timing relationships, restrictions are placed on the length of HP-IB cable used in direct HP-IB coupling. When connecting devices directly on to the HP-IB, at the local and remote ends, the following rules should be observed to ensure correct operation:

The total HP-IB cable length for the devices connected to a 37203A must be.

- (1) Less than or equal to 20 metres (65.6 feet).
- (2) Less than or equal to 2 metres (6.56 feet) times the total number of devices connected to the bus but the total length must not exceed 20 metres.

2-23 A list of HP-IB interconnecting cables available is given in Table 2-1.

Table 2-1 HP-IB Interconnecting Cables

Length	Accessory Number
1 metre	10833A
2 metres	10833B
4 metres	10833C
0.5 metre	10833D

2-24 HP-IB Connector

2-25 Figure 2-3 shows the HP-IB connector pin allocation and signal names. A description of the signals on this connector follows.

HP-IB Pins 1-4 (DIO1-DIO4) and
 HP-IB Pins 13-16 (DIO5-DIO8) are the Data Input/Output Lines of the HP-IB.

HP-IB Pin 5 (EOI). This line, End or Identity, is used to indicate the end of a multiple byte message, and is also used for parallel polling.

HP-IB Pin 6 (DAV)
 HP-IB Pin 7 (NRFD)
 HP-IB Pin 8 (NDAC). Data Valid, Not Ready for Data and No Data Accepted lines are the Handshake lines which control the transfer of data bytes between addressed devices.

HP-IB Pin 9 (IFC). This is the "Interface Clear" line. When the system controller sets IFC low true all HP-IB instruments are unaddressed.

HP-IB Pin 10 (SRQ). This is the "Service Request" line and is set low true by any instrument requiring service.

HP-IB Pin 11 (ATN). This is the "Attention" line which is pulled low true to set the HP-IB in the Command mode, and is also used for parallel polling.

HP-IB Pin 12 (SHIELD) is the ground to chassis pin at the HP-IB connector.

2-26 For further information on the HP-IB, refer to the "Condensed Description of the Hewlett-Packard Interface Bus" HP Part Number 59401-90030.




Figure 2-3 HP-IB Connector

2-27 COAXIAL CABLE

2-28 The use of Belden type 9248 cable is recommended. However any other cable which meets the specifications as listed in paragraph 3-20 may be used. A 75Ω BNC connector should be attached to each end of the cable. Trompeter type UPL 20-41 is suitable for use with the recommended Belden type 9248 cable.

2-29 TEST SWITCHES AND BOARD

2-30 The RUN/TEST switch A1S1 should be set to R (RUN) on both 37203A's (see Figure 2-4).

2-31 The test board A1TL1 should be in the factory preset RUN position (see Figure 2-4).

2-32 Switches 1 to 6 of switch A1S2 should be in the factory preset OFF position (see Figure 2-5).

2-33 37203A CONFIGURATIONS

2-34 The following paragraphs outline the steps required to configure the 37203A. The configuration of the 37203A will depend whether the serial link is coaxial cable or dual fibre optic cable. The configurations described are Point-to-Point using coaxial cable and dual fibre optic cable. For system configurations (Star and Tandem) see Paragraph 3-22.




Figure 2-4 Run/Test Switch and Test Board




Figure 2-5 Test Switch

2-35 Point-to-Point Coaxial Configuration

2-36 Figure 2-6 shows how two 37203A's can be connected to extend the operating distance between instruments on the HP-IB by using up to 1000 metres of coaxial cable. Up to 14 HP-IB devices can be connected to the 37203A at the local and remote sites.

2-37 The COAX/OPT switch A1S5 should be set to the COAX position on both 37203A's, see Figure 2-8.

2-38 The SERIAL DATA RATE switch on the rear panel of both 37203A's see Figure 3-4, should be set to the rate according to the distance between 37203A's, see Table 3-1.

2-39 The MASTER/SLAVE switch on the rear panel should be set to MASTER on the unit at one end of the serial link and SLAVE on the unit at the other end, see Figure 3-4. As the switch is purely concerned with communication start up, in the event of a serial data error occurring, it is unimportant which way round the MASTER and SLAVE are assigned.

WARNING

When the coaxial cable is used in an out of doors environment, do not remove or connect the coaxial cable from or to the HP-IB Extenders during an electrical storm. To do so could result in a lethal electric shock!

2-40 Point-to-Point Fibre Optic Configuration (Option 001)

2-41 Figure 2-7 shows how two 37203A's can be connected to extend the operating distance between instruments on the HP-IB by using up to 1000 metres of dual fibre optic cable. Up to 14 HP-IB devices can be connected to a 37203A at the local and remote sites.

2-42 The COAX/OPT switch A1S5 should be set to the OPT position on both 37203A's, see Figure 2-8.

2-43 The SERIAL DATA RATE switch on the rear panel of both 37203A's may be set to any rate, see Figure 3-4.




Figure 2-6 Point-to-Point Coaxial Configuration




Figure 2-7 Point-to-Point Fibre Optic Configuration




Figure 2-8 COAX/OPT Switch

2-44 The MASTER/SLAVE switch on the rear panel should be set to MASTER on the unit at one end of the serial link and SLAVE on the unit at the other end, see Figure 3-4. As the switch is purely concerned with communication start up, in the event of a serial data error occurring, it is unimportant which way round the MASTER and SLAVE are assigned.

2-45 FIBRE OPTIC CABLE INSTALLATION

2-46 For best results cable installers should observe the following cautions.

2-47 **Storage Temperature.** Cable should be stored at temperatures between -40°C and $+85^{\circ}\text{C}$. Storage at lower temperatures could cause temporary degradation of the optical (loss) properties of the cable. Storage above 85°C may cause softening of polyurethane outer cable jacket.

2-48 **Operating Temperature.** HP optical cable assemblies are guaranteed to meet all specified mechanical and performance parameters over the range of 0°C to 70°C operating. Cables operating outside this range but within the storage temperature range -40°C to $+85^{\circ}\text{C}$ will not be damaged. They will, however, increase in optical attenuation (loss) at the lower temperatures. This increase in optical loss will reduce the operating margin of the optical transmission system and may result in an increase in bit errors below 0°C .

2-49 **Pull Force on Cable.** Maximum pull force on cable should not exceed 30kg (66 lbs) per channel or 60kg (132 lbs) for duplex. Greater forces may cause stretching and breaking of the optical fibre inside the cable. The light weight of HP optical cable should allow easy installation in most applications without exceeding this limit.

2-50 **Pull Force on Connector.** Maximum pull force applied to the optical connector should not exceed 5kg (11 lbs). This pull strength is designed for reliable and repeated connections and disconnections of the connectors. The connector is not designed to support the pull force of drawing the cable through trays or conduits. Pulling and installation tools should be attached to the cable jacket several inches away from the optical connectors.

2-51 **Bending Radius of Cable.** The minimum safe bend diameter of the HP optical cable is 1.4cm (0.6"). Any tighter bending places a severe stress on the internal glass fibre and can result in breakage of the optical path.

2-52 Care should be taken that at no time during installation or use the cable is kinked, knotted, or bent into a loop smaller than 0.6". All installation pulleys, bends in conduits and corners to be turned should be checked for or built up to this diameter. HP optical cable has been designed and tested to withstand at least 50,000 repeated bends to the 1.4cm specified diameter, so observing these procedures should assure that no cable is damaged during installation.

2-53 **Crushing Force and Impact Resistance.** HP optical cable has been designed and tested to withstand crush loads and impacts equal to or greater than wire cables can tolerate. The cables can be stepped or even driven over without damage if laid on a flat surface. (This is not recommended as standard practice.) They can be installed along floorboards or under carpets if required. Standard precautions must be observed to avoid cutting the cable with a sharp object, snagging it with passing equipment, slamming it in a door or tripping over it. These problems could break the cable completely, or could exceed the minimum bending diameter and break the inner glass fibre.

2-54 Scuffing and Abrasion. The cable jacket is polyurethane, a tough, smooth plastic material with excellent abrasion resistance. Surface scuffs and abrasion to this jacket, should they occur, will not affect cable performance, since the optical fibre path lies in the centre of the cable, protected by buffering jackets and tough, aramid fibre strength members.

2-55 Suspending the Cable. Due to the light weight and high tensile strength of the HP optical cable, it can support almost 4000 metres (13,000 ft) of its own weight. It can thus be suspended along walls or over dropped ceilings if necessary.

2-56 Cable to Cable Interconnection. Only one length of cable can be used between 37203A units, i.e. no cable to cable interconnections are permitted.

2-57 Resistance to Solvents and Moisture. HP optical cables are specified for operation in environments with relative humidity up to 95%. Cable jacketing is polyurethane. The cable should perform as well as or better than standard copper cable similarly jacketed.

2-58 Outdoor Use. Although the cable jacket contains additives to retard damage by ultra-violet rays (sunlight) it is not specified or tested for outdoor use. If the cable must be used outdoors, make every effort to control the cable environment to stated specifications; otherwise, performance and useable life of the cable may be degraded.

2-59 Connecting Fibre Optic Cable

2-60 The HP optical connectors contain precision ferrules which maintain close tolerance optical fibre alignment to the optical ports of the 37203A. These connectors should be screwed in snugly by holding the connector in one hand and tightening the coupling nut "finger tight" with the other.

Note: Coupling nut "finger tight" torque is defined as 0.05 to 0.10 Newton-metres.

CAUTION

No wrenches, gloves, rags or tools should be used for added leverage on the connector nut. Excessive tightening of the knurled nut will result in misalignment and possible permanent damage to the connector or the optical ports.

2-61 STORAGE AND SHIPMENT

2-62 Environment

2-63 The instrument may be stored or shipped in environments within the following limits:

Temperature	-40°C to +75°C
Altitude	up to 15,300 metres (50,000 feet)

The instrument should also be protected from temperature extremes which could cause condensation within the instrument.

2-64 Packaging

2-65 Tagging for Service. If the instrument is being returned to Hewlett-Packard for service, please complete one of the blue repair tags at the end of this section and attach it to the instrument.

2-66 Original Packaging. Containers and materials identical to those used in factory packaging are available through Hewlett-Packard offices. If the instrument is being returned to Hewlett-Packard for servicing, attach a tag indicating the type of service required, return address, model number, and full serial number. Also, mark the container 'FRAGILE' to ensure careful handling. In any correspondence, refer to the instrument by model number and full serial number.

2-67 Other Packaging. The following general instructions should be used for re-packing with commercially available materials:

- Wrap instrument in heavy paper or plastic. (If shipping to Hewlett-Packard office or service centre, attach tag indicating type of service required, return address, model number, and full serial number.)
- Use strong shipping container. A double-wall carton made of 200-pound test material is adequate.
- Use a layer of shock absorbing material 50 to 75mm (2 to 3 inch) thick around all sides of the instrument to provide firm cushioning and prevent movement inside container. Protect the control panel with cardboard.
- Seal shipping container securely.
- Mark shipping container FRAGILE to ensure careful handling.
- In any correspondence, refer to instrument by model number and full serial number.

SECTION III

OPERATION

3-1 INTRODUCTION

3-2 This section of the manual explains the operating characteristics and the function of the controls and indicators of the Model 37203A HP-IB Extender.

3-3 OPERATING CHARACTERISTICS

3-4 A basic description of the 37203A is given in Section I Paragraph 1-22 to 1-30. The following paragraphs describe some aspects of the 37203A Operating Characteristics in more detail.

3-5 Rejection of Electrical Interference

3-6 The coaxial cable driver/receiver circuit is isolated by optical couplers which enables the circuit to float free from ground, limited to approximately $\pm 24V$ by a varistor. This provides protection from differing ground potentials at the two ends and from the effects of induced interference in the cable, by preventing earth currents. Data is transferred to and from the floating circuit through the optical couplers. Rejection of all but extremely severe electrical interference is assured when using coaxial cable with 100% screening.

3-7 Transmitted data frames are also checked by a 4-bit Cyclic Redundancy Check Code. This provides some additional protection from HP-IB errors.

3-8 Transparent HP-IB Extension

3-9 'Transparent', in the context used here, means that HP-IB functional operation is maintained when the 37203A's are introduced. Occasionally, however, program alterations may be necessary to accommodate slightly changed timing relationships.

3-10 The function of 'transparent' extension provided by each of the 37203A's is to accept data from, or source data onto, each bus on behalf of instruments at the opposite end. Bytes are passed between distant buses effectively without transit buffer storage, which means that at no time does unwanted data need to be flushed from the buffers, e.g. upon change of data direction, as in other extenders.

3-11 No special programming is required when using the 37203A's which are fully transparent. The timing relationships of operational events may, however, be slightly altered in the case of fast devices when the 37203A's are introduced.

3-12 Performance with the Parallel Poll Function

3-13 The 37203A supports the Parallel Poll Function but results in a small departure from the IEEE 488 Standard. Due to absolute transmission delay, no Extender, including the 37203A, can guarantee a correct and up-to-date Parallel Poll response from devices distant from the controller within the $2\mu s$ which the Standard requires. Instead a pair of 37203A's return the response from distant devices, as rapidly as possible, to the polling controller, and within a time shown in Table 3-1.

3-14 Between the poll being conducted by the controller and the response becoming available from the remote end, the DIO lines at the local Extender are held in the passive (false) condition to prevent the lines being interpreted erroneously.

3-15 Performance with Service Request Function

3-16 The pair of 37203A's communicate the Service Request message (SRQ) to the controller within the times shown in Table 3-1.

Table 3-1 Distance and Speed of Operation at the different Serial Data Rates

	Coaxial Cable (see Note 2)				Fibre Optic Cable (Opt 001)		
	Short (at normal speed) <i>see Note 1</i>	250m (max range at normal speed)	500m (max range at 1/4 speed)	1000m (max range at 1/16 speed)	Short <i>See Note 1</i>	≤250m	≤1000m
Max HP-IB byte transfer rate (k byte/s)	50	40	14.5	2.75	50	39	25
Max SRQ propagation delay (μs)	14	18	55	200	14	20	30
Max Parallel Poll response time (μs)	20	25	75	270	20	25	40

Note 1: For distances <250m, interpolate between Short and 250 columns.

Note 2: These results were achieved using the recommended coaxial cable (Belden cable type 9248)

3-17 Serial Data Rate Settings

3-18 When the serial link is dual fibre optic cable the serial data rate may be set to NORMAL regardless of the length of dual fibre optic cable.

3-19 When the serial link is coaxial cable the serial data rate must be reduced as the length of the coaxial cable increases, see Table 3-1. This is due to the following characteristics of coaxial cable:

As the length of the coaxial cable increases the attenuation increases and the bandwidth decreases. These characteristics result in the high frequency components of the Transmit waveform of the data frame being distorted. This degradation of the waveform ultimately results in serial data errors. Reducing the serial data rate minimises the degradation and so maintains the integrity of the serial link.

3-20 The recommended coaxial cable has:

- 75Ω impedance.
- loss not exceeding 6.9dB per 100 metres at 100MHz.

- dc resistance (inner plus outer conductors) not exceeding 4.1Ω per 100 metres.
- 100% shield coverage, see Paragraph 3-6.

3-21 Inferior cable may be used providing that the total loss and resistance does not exceed that of the recommended cable at its maximum distance. Coaxial cable with reduced shield coverage may be used if the environment is not electrically noisy.

3-22 System Configurations

3-23 The 37203A can only be used in pairs, so multiple drops from the cable are not possible. However, further pairs of 37203A's can be used to support additional clusters of remote devices in arrangements which avoid loops. In these system configurations the serial link may be either coaxial cable or dual fibre optic cable.

3-24 The 37203A pairs may be arranged in a star configuration where the pairs are connected in parallel, see Figure 3-1. In this configuration a total of 14 HP-IB devices can be connected to any remote bus. The total number of devices in a system is limited by the addressing capability of the HP-IB.

3-25 The 37203A pairs may be arranged in a tandem configuration where the pairs are connected in series, see Figure 3-2. In this configuration a total of 13 HP-IB devices

can be connected to any remote bus. The total number of devices in a system is limited by the addressing capability of the HP-IB.




Figure 3-1 Star Configuration





Figure 3-2 Tandem Configuration


3-26 FRONT AND REAR PANEL FEATURES

3-27 The front and rear panel features on the 37203A are described in Figures 3-3 and 3-4.



1. **LINE.** The LINE switch, switches the ac supply ON or OFF. A line-on indicator is illuminated to give a visual indication of when the instrument is in the 'line-on' condition.
2. **DATA LOSS.** The DATA LOSS indicator gives a visual indication of a loss of data frames.
3. **DATA ERRORS.** The DATA ERRORS indicator gives a visual indication that an error has been detected by the Cyclic Redundancy Check Code. The indicator flashes when an error has been detected but will stay on continuously if many errors occur.

Figure 3-3 37203A HP-IB Extender Front Panel Features



1. Connector for the HP-IB; used to connect the 37203A to the HP-IB.
2. Connector for the coaxial cable; used for interconnecting 37203A's for distances up to 1000 metres.
3. **MASTER/SLAVE.** The MASTER/SLAVE switch should be set to Master on the unit at one end of the serial link and Slave on the unit at the other end. It is unimportant which way round the Master and Slave are assigned.
4. **SERIAL DATA RATE.** The SERIAL DATA RATE switch sets the speed of operation of the 37203A. Reducing the serial data rate allows the maximum transmission distance by coaxial cable to increase up to a maximum of 1000 metres.
5. Power Module connection receptacle with fuse and line voltage selector. Figure 2-1 indicates the method of selecting the line voltage.
6. Connectors for the fibre optic cables (Option 001); used for interconnecting 37203A's for distances up to 1000 metres.

Figure 3-4 37203A HP-IB Extender Rear Panel Features

3-28 OPERATOR'S MAINTENANCE

3-29 Fibre Optics

3-30 No grease, dirt or other foreign material should be allowed to collect on the alignment ferrule of the optical connector. Dirt at this point can reduce the optical signal and may scratch the fibre end during connector insertion. If a dirty ferrule is observed or suspected, the ferrule tip can be cleaned with a swab moistened in alcohol.

3-31 Service and Repair

3-32 HP optical cables are not customer or field repairable. They contain optical waveguide and precision optical interfaces which must be repaired or replaced at the factory.

3-33 A damaged connector or a cable broken at the connector would normally be repaired by cutting off the damaged section and installing new connectors.

3-34 Consult with the local Hewlett-Packard Sales and Service Office, listed at the rear of this manual, for directions on what action to take.

SECTION IV

PERFORMANCE VERIFICATION

4-1 INTRODUCTION

4-2 The 37203A's are used in pairs and are designed to work with virtually all bus-compatible equipment, provided such equipment conforms to IEEE Standard 488-1978. The 37203A is transparent to all HP-IB functions. The performance verification routines check that the 37203A supports these HP-IB functions.

4-3 PERFORMANCE VERIFICATION ROUTINES

4-4 The Performance Verification routines are supplied on data cartridge HP 37203-12101. A list of routines is given in Table 4-1.

4-5 EQUIPMENT REQUIRED

4-6 Equipment required to run the Performance Verification is as given in table 4-2.

Table 4-1 Performance Verification Routines

Routine
REMOTE ENABLE
SERVICE REQUEST TEST
PARALLEL POLL TEST
CONTROLLER SENDS LISTEN ADDRESS
CONTROLLER SENDS UNLISTEN
CONTROLLER SENDS TALK ADDRESS
CONTROLLER SENDS UNTALK
CONTROLLER SOURCES 0-255
CONTROLLER ACCEPTS 0-255
CONTROL PASSES TO INTERFACE 8

Table 4-2 Equipment Required

Quantity	Item	Type
1	Desktop Computer	HP 9825A
1	String & Advanced Programming ROM	HP 98210A
1	General I/O & Extended I/O ROM	HP 98213A
2	HP-IB Interface Card	HP 98034A
2	HP-IB Bus Extender	HP 37203A
1	Coax Cable	—
1	Performance Verification Tape	HP 37203-12101

4-7 SELECT CODE AND INTERFACE ADDRESS

4-8 Before the Performance Verification can be run one of the HP 98034A HP-IB Interface Cards select code and interface address needs to be changed.

4-9 The select code switch is accessible through a small hole on top of the interface case. The switch is preset to select code 7 at the factory. The select code should be set to 8 by rotating the switch using a small screwdriver.

4-10 The interface address switch is situated internally and the interface covers should be removed, see Figure 4-1.

4-11 Follow these steps to change the interface address switches.

1. Remove the 8 screws as shown in Figure 4-1.

2. Remove edge connector.
3. Carefully separate the two printed circuit boards.
4. Set the switches as illustrated in Figure 4-1. Switches 1 to 5 select the address (octal 26, bit 5 is the most significant bit and setting each switch to the "ON" position corresponds to a "0"). Switch 6 the System Controller enable switch should be set to OFF because only one system controller is allowed in a system.
5. To reassemble ensure that the five pin connectors on one board are aligned with their sockets on the other board.
6. Reconnect the edge connector.
7. Position the cable wires so that they are not crimped as the interface case is pressed together.
8. Replace all the screws.




Figure 4-1 HP 98034A Interface Card

4-12 The other HP 98034A HP-IB Interface Card should be left set to the factory preset values; select code 7, interface address octal 25, system control enable on.

Note: Both interface cards Parallel Poll Sense and Parallel Poll Bit should be left set to their factory preset values, position 1 and Bit 1 respectively.

4-13 Table 4-3 summarises the HP 98034A HP-IB Interface Cards switch settings.

4-14 PERFORMANCE VERIFICATION SYSTEM CONFIGURATION

4-15 Connect the Desktop Computer, HP-IB Interface Cards and HP-IB Extenders as shown in Figure 4-2 and explained in paragraph 4-16.

4-16 These steps should be followed carefully to ensure correct configuration.

1. Connect the two 37203A's using a coax cable.
2. Ensure that both 37203A's have switch A1S5 set

to COAX; A1S1 set to RUN; A1S2 switches all set to OFF and A1TL1 in RUN mode.

3. Set the Serial Data Rate switch to Normal on both 37203A's.
4. Set the 'Local' 37203A to MASTER.
5. Set the 'Remote' 37203A to SLAVE.
6. Connect the HP-IB Interface Card with select code 7 to the Master 37203A HP-IB connector.
7. Connect the HP-IB Interface Card with select code 8 to the Slave 37203A HP-IB connector.
8. Insert both cards into separate I/O slots on the rear panel of the HP 9825A.

4-17 DESKTOP COMPUTER & HP-IB INTERFACE CARD VERIFICATION

4-18 Before the Performance Verification is run the desktop computer and HP-IB interface cards should be checked to ensure they have been configured correctly.

Table 4-3 HP 98034A Switch Settings

Function	Setting	
Select Code	7	8
Desktop Computer Address	Talk = U Listen = 5	Talk = V Listen = 6
System Controller	ON	OFF
Parallel Poll Sense	Negative — True Logic (Position 1)	Negative — True Logic (Position 1)
Parallel Poll Bit	Bit 1 (least significant bit)	Bit 1 (least significant bit)




Figure 4-2 Performance Verification System Configuration

4-19 Insert the Performance Verification Cartridge 37203-12101 into the HP 9825A tape transport.

4-20 Switch on all the equipment .

4-21 Ensure that both DATA LOSS and DATA ERROR leds are off. If either of the lamps are on, check that the Serial Data Rate is set to NORMAL on both 37203A's and that one 37203A is set to Master and the other Slave. If either lamp is still on refer to the General Service Sheet G1 in Section VIII.

4-22 Press Special Function Key f_0 .

4-23 Verify the display:

0.00 213.00 68.00 76.00

4-24 This verifies the HP-IB Interface Card with select code 7.

4-25 If the display is verified proceed to Paragraph 4-26. If the display is not verified switch the 37203A's OFF then ON and repeat Paragraph 4-22. If the display is still not verified check the HP-IB Interface Card switch settings (see Table 4-3); if this is correct the HP-IB Interface Card is defective.

4-26 Press Special Function Key f_1 .

4-27 Verify the display:

0.00 214.00 68.00 4.00

4-28 This verifies the HP-IB Interface Card with select code 8.

4-29 If the display is not verified switch the 37203A's OFF then ON and repeat Paragraph 4-26. If the display is still not verified check the HP-IB Interface Card switch settings (see Table 4-3); if this is correct the HP-IB Interface Card is defective.

4-30 This completes the Desktop Computer and HP-IB Interface Card verification. The system is ready to verify that the two 37203A's function correctly.

4-31 Before running the Performance Verification routines ensure that the DATA LOSS and DATA ERRORS leds are off. If either of the lamps are on check the Serial Data Rate switch is set to Normal on both 37203A's and that one 37203A is set to Master and the other Slave. If either lamp is still on refer to General Service Sheet G1 in Section VIII.

4-32 To run the Performance Test routines press f_2 .

4-33 A successful verification is indicated by:

PROGRAM SUCCESSFULLY COMPLETED

being displayed on the 9825.

4-34 If a Performance Verification routine fails an error message will be displayed. Power both 37203A's OFF then ON and repeat paragraph 4-32. If the routine fails again refer to General Service Sheet G1 in Section VIII.

SECTION V

ADJUSTMENTS

5-1 INTRODUCTION

5-2 This section describes the adjustments required to return the instrument to peak operating capabilities when repairs have been made. Included in this section are adjustment procedures, adjustment location diagrams and a table of adjustable components.

5-3 Adjustments should only be made after ascertaining that the instrument is out of calibration. To avoid any interaction between adjustments, the procedures in this section should be performed in the order given.

5-4 Table 5-1 is a list of the adjustable components with related information. The location of each of the adjustable components is shown in Figure 5-1.

5-5 EQUIPMENT REQUIRED

5-6 Equipment required for the adjustments is listed in

the table of Recommended Test Equipment in Section I. Any equipment that satisfies the critical specifications given in the table may be substituted for the recommended models.

5-7 SAFETY CONSIDERATIONS

5-8 This section contains warnings and cautions that must be followed for your protection and to avoid damage to the equipment.

WARNING

Maintenance described herein is performed with protective covers removed and power applied to the instrument. Maintenance should be performed only by service trained personnel who are aware of the hazards involved.

Table 5-1 Adjustable Components

Adjustment Name	Reference Designator	Adjustment Paragraph	Service Sheet	Description
24MHz OSCILLATOR	A1C25	5-11	A1	Adjusts 24MHz crystal oscillator
TRANSMIT DATA-SKEW	A1R17	5-12	A1	Adjusts input to optocoupler U62
RECEIVE DATA-SKEW	A1R30	5-13	A1	Adjusts input to optocoupler U63




Figure 5-1 Location of Adjustable Components

5-9 ADJUSTMENT PROCEDURE

5-10 Preliminary Procedure

1. Remove top cover.

5-11 24MHz OSCILLATOR ADJUSTMENT

DESCRIPTION

This adjustment tunes the oscillator to 24MHz.

EQUIPMENT

Frequency counter hp 5302A

PROCEDURE

1. Connect the counter to Test Point MCLCK on A1.
2. Adjust A1C25 to obtain a frequency of 24 ± 0.25 MHz.

5-12 TRANSMIT DATA – SKEW ADJUSTMENT

DESCRIPTION

This adjustment provides the correct mark-space ratio of the output signal to the coaxial cable, to compensate for the opto-coupler.

EQUIPMENT

Oscilloscope hp 180C, 1809A, 1825A
10:1 Probe hp 10004D

PROCEDURE

1. Set the SERIAL DATA RATE switch on the rear panel to NORMAL.
2. Set the MASTER/SLAVE switch on the rear panel to MASTER.
3. Set the COAX/OPT switch A1S5 to COAX.
4. Set the R/T switch A1S1 to R.
5. Set switch 6 of A1S2 to ON and switches 1-5 switch OFF.
6. Set the plug-in card A1TL1 located in J2 to the RUN position.
7. Connect the oscilloscope to Test Point OUT on A1 using Test Point FGND on A1 to ground the oscilloscope probe.
8. Switch the 37203A OFF for 5 seconds then ON.
9. Set the oscilloscope to $0.05\mu\text{s}/\text{cm}$, negative trigger and when using the 10:1 probe to $0.2\text{V}/\text{cm}$.
10. Adjust X-POSITION to centre transition between bits 1 and 2 on the vertical middle line of the graticule, see Figure 5-2.

11. Adjust Y-POSITION to put ground level in the middle of the trace.




Figure 5-2 Transmit Skew

12. Adjust A1R17 to make the duration of bits 1 and 2 equal at the mid-amplitude of the waveform. Use the X-POSITION to exactly centre the transition on the vertical middle line of the oscilloscope as the adjustment is being made.
13. Disconnect Oscilloscope.

5-13 RECEIVE DATA – SKEW ADJUSTMENT

DESCRIPTION

This adjustment provides the correct mark-space ratio of the incoming data signal to compensate for the optocoupler.

EQUIPMENT

Oscilloscope hp 180C, 1809A, 1825A
10:1 Probe hp 10004D

PROCEDURE

1. Set the SERIAL DATA RATE switch on the rear panel to NORMAL.
2. Set the MASTER/SLAVE switch on the rear panel to MASTER.
3. Set the COAX/OPT switch on A1S5 to coax.
4. Set the R/T switch A1S1 to R.
5. Set switch 6 of the A1S2 to ON and switches 1-5 switch OFF.
6. Set the plug-in card A1TL1 located in J2 to the RUN position.

7. Connect the oscilloscope to Test Point RX on A1.
8. Switch the 37203A OFF for 5 seconds then ON.
9. Set the oscilloscope to $0.05\mu\text{s}/\text{cm}$, positive trigger and when using the 10:1 probe $0.2\text{V}/\text{cm}$.
10. Adjust the X-POSITION to centre the transition between bits 1 and 2 on the vertical middle line of the graticule, see Figure 5-3.
11. Adjust the Y-POSITION to put ground on the 0V line.




Figure 5-3 Receive Skew

12. Adjust A1R30 to make the durations of bits 1 and 2 equal at the +1.4V level. Use the X-POSITION to exactly centre the transition on the +1.4V level as the adjustment is being made.
13. Disconnect oscilloscope.

SECTION VI

REPLACEABLE PARTS

6-1 INTRODUCTION

6-2 This section contains information for ordering parts. Table 6-1 lists abbreviations used in the parts list and throughout the manual. Table 6-2 lists all replaceable parts in reference designator order. Table 6-3 contains the names and addresses that correspond to the manufacturers code numbers.

6-3 ABBREVIATIONS

6-4 Table 6-1 lists all abbreviations used in the parts list, the schematics and throughout the manual. In some cases two forms of the abbreviation are used, one all in capital letters, and one partial or no capitals. This occurs because the abbreviations in the parts list are always all capitals. However, in the schematics and other parts of the manual, other abbreviation forms are used with both lower and upper case letters.

6-5 REPLACEABLE PARTS LIST

6-6 Table 6-2 is the list of replaceable parts and is organized as follows:

- (a) Electrical assemblies and their components in alpha-numeric order by reference designation.
- (b) Chassis-mounted parts in alpha-numeric order by reference designation.
- (c) Miscellaneous parts.

The information given for each part consists of the following:

- (a) The Hewlett-Packard part number.
- (b) Part number check digit (CD).
- (c) The total quantity (Qty) in the instrument.
- (d) The description of the part.

- (e) A typical manufacturer of the part in a five-digit code.
- (f) The manufacturers number for that part.

The total quantity for each part is given only once – at the first appearance of the part in the list.

6-7 ORDERING INFORMATION

6-8 To order a part listed in the replaceable parts table, quote the Hewlett-Packard part number (with the check digit), indicate the quantity required, and address the order to the nearest Hewlett-Packard office. The check digit will ensure accurate and timely processing of your order.

6-9 To order a part that is not listed in the replaceable parts table, include the instrument model number, instrument serial number, the description and function of the part, and the number of parts required. Address the order to the nearest Hewlett-Packard Office.

6-10 DIRECT MAIL ORDER SYSTEM

6-11 Within the USA, Hewlett-Packard can supply parts through a direct mail order system. Advantages of using the system are as follows:

- (a) Direct ordering and shipment from the HP Parts Centre in Mountain View, California.
- (b) No maximum or minimum on any mail order (there is a minimum order amount for parts ordered through a local HP office when the orders require billing and invoicing).
- (c) Prepaid transportation (there is a small handling charge for each order).
- (d) No invoices – to provide these advantages, a cheque or money order must accompany each order.

6-12 Mail Order forms and specific ordering information are available through your local HP office. Addresses and phone numbers are located at the back of this manual.

Table 6-1 Reference Designations and Abbreviations

REFERENCE DESIGNATIONS

A assembly	E miscellaneous electrical part	P electrical connector (movable portion); plug	U integrated circuit; microcircuit
AT attenuator; isolator; termination	F fuse	Q transistor: SCR; triode thyristor	V electron tube
B fan; motor	FL filter	R resistor	VR voltage regulator; breakdown diode
BT battery	H hardware	RT thermistor	W cable; transmission path; wire
C capacitor	HY circulator	S switch	X socket
CP coupler	J electrical connector (stationary portion); jack	T transformer	Y crystal unit (piezo-electric or quartz)
CR diode; diode thyristor; varactor	K relay	TB terminal board	Z tuned cavity; tuned circuit
DC directional coupler	L coil; inductor	TC thermocouple	
DL delay line	M meter	TP test point	
DS annunciator; signaling device (audible or visual); lamp; LED	MP miscellaneous mechanical part		

ABBREVIATIONS

A ampere	COMPL complete	FET field-effect transistor	LF low frequency
ac alternating current	CONN connector	F/F flip-flop	LG long
ACCESS accessory	CP cadmium plate	FH flat head	LH left hand
ADJ adjustment	CRT cathode-ray tube	FIL H fillister head	LIM limit
A/D analog-to-digital	CTL complementary transistor logic	FM frequency modulation	LIN linear taper (used in parts list)
AF audio frequency	CW continuous wave	FP front panel	lin linear
AFC automatic frequency control	cw clockwise	FREQ frequency	LK WASH lock washer
AGC automatic gain control	em centimeter	FXD fixed	LO low; local oscillator
AL aluminum	D/A digital-to-analog	g gram	LOG logarithmic taper (used in parts list)
ALC automatic level control	dB decibel	GE germanium	log logarithm(ic)
AM amplitude modulation	dBm decibel referred to 1 mW	GHz gigahertz	LPF low pass filter
AMPL amplifier	dc direct current	GL glass	LV low voltage
APC automatic phase control	deg degree (temperature interval or difference)	GRD ground(ed)	m meter (distance)
ASSY assembly	° degree (plane angle)	H henry	mA milliampere
AUX auxiliary	°C degree Celsius (centigrade)	h hour	MAX maximum
avg average	°F degree Fahrenheit	HET heterodyne	MΩ megohm
AWG American wire gauge	°K degree Kelvin	HEX hexagonal	MEG meg (10 ⁶) (used in parts list)
BAL balance	DEPC deposited carbon	HD head	MET FLM metal film
BCD binary coded decimal	DET detector	HDW hardware	MET OX metallic oxide
BD board	diam diameter	HF high frequency	MF medium frequency; microfarad (used in parts list)
BE CU beryllium copper	DIA diameter (used in parts list)	HG mercury	MFR manufacturer
BFO beat frequency oscillator	DIFF AMPL differential amplifier	HI high	mg milligram
BH binder head	div division	HP Hewlett-Packard	MHz megahertz
BKDN breakdown	DPDT double-pole, double-throw	HPF high pass filter	mH millihenry
BP bandpass	DR drive	HR hour (used in parts list)	mho mho
BPF bandpass filter	DSB double sideband	HV high voltage	MIN minimum
BRS brass	DTL diode transistor logic	Hz Hertz	min minute (time)
BWO backward-wave oscillator	DVM digital voltmeter	IC integrated circuit	... ' minute (plane angle)
CAL calibrate	ECL emitter coupled logic	ID inside diameter	MINAT miniature
ccw counter-clockwise	EMF electromotive force	IF intermediate frequency	mm millimeter
CER ceramic	EDP electronic data processing	IMPG impregnated	MOD modulator
CHAN channel	ELECT electrolytic	in inch	MOM momentary
cm centimeter	ENCAP encapsulated	INCD incandescent	MOS metal-oxide semiconductor
CMO cabinet mount only	EXT external	INCL include(s)	ms millisecond
COAX coaxial	F farad	INP input	MTG mounting
COEF coefficient		INS insulation	MTR meter (indicating device)
COM common		INT internal	mV millivolt
COMP composition		kg kilogram	mVac millivolt, ac
		kHz kilohertz	mVdc millivolt, dc
		kΩ kilohm	mVpk millivolt, peak
		kV kilovolt	
		lb pound	
		LC inductance-capacitance	
		LED light-emitting diode	

NOTE

All abbreviations in the parts list will be in upper-case.

Table 6-1 Reference Designations and Abbreviations (continued)

mVp-p . . . millivolt, peak-to-peak	P peak (used in parts list)	REF reference	TERM terminal
mVrms millivolt, rms	PAM pulse-amplitude modulation	REG regulated	TFT thin-film transistor
mW milliwatt	PC printed circuit	REPL replaceable	TGL toggle
MUX multiplex	PCM pulse-code modulation; pulse-count modulation	RF radio frequency	THD thread
MY mylar	PDM pulse-duration modulation	RFI radio frequency interference	THRU through
μA microampere	pF picofarad	RH round head; right hand	TI titanium
μF microfarad	PH BRZ . . . phosphor bronze	RLC resistance-inductance-capacitance	TOL tolerance
μH microhenry	PHL Phillips	RMO rack mount only	TRIM trimmer
μmho micromho	PIN positive-intrinsic-negative	rms root-mean-square	TSTR transistor
μs microsecond	PIV peak inverse voltage	RND round	TTL transistor-transistor logic
μV microvolt	pk peak	ROM read-only memory	TV television
μVac microvolt, ac	PL phase lock	R&P rack and panel	TVI television interference
μVdc microvolt, dc	PLO phase lock oscillator	RWV reverse working voltage	TWT traveling wave tube
μVpk microvolt, peak	PM phase modulation	S scattering parameter	U micro (10 ⁶) (used in parts list)
μVp-p microvolt, peak-to-peak	PNP positive-negative-positive	s second (time)	UF microfarad (used in parts list)
μVrms microvolt, rms	P/O part of	” second (plane angle)	UHF ultrahigh frequency
μW microwatt	POLY polystyrene	S-B slow-blow (fuse) (used in parts list)	UNREG unregulated
nA nanoampere	PORC porcelain	SCR silicon controlled rectifier; screw	V volt
NC no connection	POS positive; position(s) (used in parts list)	SE selenium	VA voltampere
N/C normally closed	POSN position	SECT sections	Vac volts, ac
NE neon	POT potentiometer	SEMICON semiconductor	VAR variable
NEG negative	P-P peak-to-peak	SHF superhigh frequency	VCO voltage-controlled oscillator
nF nanofarad	PP peak-to-peak (used in parts list)	SI silicon	Vdc volts, dc
NI PL nickel plate	PPM pulse-position modulation	SIL silver	VDCW volts, dc, working (used in parts list)
N/O normally open	PREAMPL preamplifier	SL slide	V(F) volts, filtered
NOM nominal	PRF pulse-repetition frequency	SNR signal-to-noise ratio	VFO variable-frequency oscillator
NORM normal	PRR pulse repetition rate	SPDT single-pole, double-throw	VHF very-high frequency
NPN negative-positive-negative	ps picosecond	SPG spring	Vpk volts, peak
NPO negative-positive zero (zero temperature coefficient)	PT point	SR split ring	Vp-p volts, peak-to-peak
NRFR not recommended for field replacement	PTM pulse-time modulation	SPST single-pole, single-throw	Vrms volts, rms
NSR not separately replaceable	PWM pulse-width modulation	SSB single sideband	VSWR voltage standing wave ratio
ns nanosecond	PWV peak working voltage	SST stainless steel	VTO voltage-tuned oscillator
nW nanowatt	RC resistance-capacitance	STL steel	VTVM vacuum-tube voltmeter
OBD order by description	RECT rectifier	SQ square	V(X) volts, switched
OD outside diameter		SWR standing-wave ratio	W watt
OH oval head		SYNC synchronize	W/ with
OP AMPL operational amplifier		T timed (slow-blow fuse)	WIV working inverse voltage
OPT option		TA tantalum	WW wirewound
OSC oscillator		TC temperature compensating	W/O without
OX oxide		TD time delay	YIG yttrium-iron-garnet
oz ounce			Z ₀ characteristic impedance
Ω ohm			

NOTE

All abbreviations in the parts list will be in upper-case.

MULTIPLIERS

Abbreviation	Prefix	Multiple
T	tera	10 ¹²
G	giga	10 ⁹
M	mega	10 ⁶
k	kilo	10 ³
da	deka	10
d	deci	10 ⁻¹
c	centi	10 ⁻²
m	milli	10 ⁻³
μ	micro	10 ⁻⁶
n	nano	10 ⁻⁹
p	pico	10 ⁻¹²
f	femto	10 ⁻¹⁵
a	atto	10 ⁻¹⁸

Table 6-2 Replaceable Parts

Reference Designation	HP Part Number	C D	Qty	Description	Mfr Code	Mfr Part Number
A1	37203-60001	5	1	BOARD ASSEMBLY	28480	37203-60001
A1C1	0160-0939	4	1	CAPACITOR-FXD 430PF +-5% 300VDC MICA	28480	0160-0939
A1C2	0160-2204	0	1	CAPACITOR-FXD 100PF +-5% 300VDC MICA	28480	0160-2204
A1C3	0160-2207	3	1	CAPACITOR-FXD 300PF +-5% 300VDC MICA	28480	0160-2207
A1C4	0160-2814	0	1	CAPACITOR-FXD 220UF+-20% 10VDC TA	28480	0160-2814
A1C5	0180-3029	1	3	CAPACITOR-FXD 2200UF 25V AL	28480	0180-3029
A1C6	0180-3029	1	1	CAPACITOR-FXD 2200UF 25V AL	28480	0180-3029
A1C7	0180-3029	1	1	CAPACITOR-FXD 2200UF 25V AL	28480	0180-3029
A1C8	0160-0550	7	2	CAPACITOR-FXD 330UF+100-10% 25VDC AL	28480	0160-0550
A1C9	0160-0550	7	2	CAPACITOR-FXD 330UF+100-10% 25VDC AL	28480	0160-0550
A1C10	0180-2662	6	2	CAPACITOR-FXD 10UF+-10% 10VDC TA	25088	D4R7881A10K
A1C11	0160-2662	6	2	CAPACITOR-FXD 10UF+-10% 10VDC TA	25088	D4R7881A10K
A1C12	0160-0576	5	15	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C13	0160-0576	5	15	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C14	0140-0196	3	3	CAPACITOR-FXD 150PF +-5% 300VDC MICA	72136	DM15F151J0300HV1CR
A1C15	0140-0196	3	3	CAPACITOR-FXD 150PF +-5% 300VDC MICA	72136	DM15F151J0300HV1CR
A1C16	0140-0196	3	3	CAPACITOR-FXD 150PF +-5% 300VDC MICA	72136	DM15F151J0300HV1CR
A1C17	0180-2816	2	4	CAPACITOR-FXD 68uF+-20% 10VDC TA	28480	0180-2816
A1C18	0180-2816	2	4	CAPACITOR-FXD 68uF+-20% 10VDC TA	28480	0180-2816
A1C19	0180-2816	2	4	CAPACITOR-FXD 68uF+-20% 10VDC TA	28480	0180-2816
A1C20	0180-2816	2	4	CAPACITOR-FXD 68uF+-20% 10VDC TA	28480	0180-2816
A1C21	0160-3879	7	2	CAPACITOR-FXD .01UF +-20% 100VDC CER	28480	0160-3879
A1C22	0160-2199	2	2	CAPACITOR-FXD 30PF +-5% 300VDC MICA	28480	0160-2199
A1C23	0140-0199	9	1	CAPACITOR-FXD 240PF +-5% 300VDC MICA	72136	DM15F241J0300HV1CR
A1C24	0160-3879	7	1	CAPACITOR-FXD .01UF +-20% 100VDC CER	28480	0160-3879
A1C25	0121-0061	1	1	CAPACITOR-V TRMR-CER 5.5-18PF 150V	52763	304322 5.5/18PF NPO
A1C26	0160-0418	6	2	CAPACITOR-FXD 1UF+-20% 35VDC TA	28480	0160-0418
A1C27	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C28	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C29	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C30	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C31	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C32	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C33	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C34	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C35	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C36	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C37	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C38	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C39	0160-0576	5	2	CAPACITOR-FXD .1UF +-20% 50VDC CER	28480	0160-0576
A1C40	0160-2199	2	2	CAPACITOR-FXD 30PF +-5% 300VDC MICA	28480	0160-2199
A1C41	0180-0418	6	2	CAPACITOR-FXD 1UF+-20% 35VDC TA	28480	0180-0418
A1CR1	1901-0673	5	2	DIODE-PWR RECT 100V 1.5A	28480	1901-0673
A1CR2	1901-0673	5	2	DIODE-PWR RECT 100V 1.5A	28480	1901-0673
A1CR3	1906-0096	7	1	DIODE-FW BRDG 200V 2A	04713	MDA202
A1CR4	1902-3002	3	1	DIODE-ZNR 2.37V 5% DO-7 PD=4W TC=+.074K	28480	1902-3002
A1CR5	1901-0539	3	2	DIODE-SCHOTTKY	28480	1901-0539
A1CR6	1901-0539	3	2	DIODE-SCHOTTKY	28480	1901-0539
A1CR7	1901-0040	1	1	DIODE-SWITCHING 30V 50MA 2N8 DO-35	28480	1901-0040
A1D81	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D82	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D83	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D84	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D85	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D86	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D87	1990-0486	6	7	LED-VISIBLE LUM-INT=1MCD IF=20MA-MAX	28480	5082-4684
A1D88	1990-0485	5	1	LED-VISIBLE LUM-INT=800UCD IF=30MA-MAX	28480	5082-4984
A1J1	1251-3263	1	3	CONNECTOR 24-PIN F MICRORIBBON	28480	1251-3263
A1J2	1251-3507	2	1	CONNECTOR-PC EDGE 24-CONT/ROW 2-ROWS	28480	1251-3507
A1L1	9140-0096	1	1	COIL-MLD 1UH 10% Q=50 .155DX,375LG-NOM	28480	9140-0096
A1L2	9140-0096	3	2	COIL-MLD 2.2UH 10% Q=33 .155DX,375LG-NOM	28480	9140-0096
A1L3	9140-0096	3	2	COIL-MLD 2.2UH 10% Q=33 .155DX,375LG-NOM	28480	9140-0096
A1Q1	1854-0215	1	4	TRANSISTOR NPN SI PD=350MW FT=300MHZ	04713	2N3904
A1Q2	1853-0015	7	2	TRANSISTOR PNP SI PD=200MW FT=500MHZ	28480	1853-0015
A1Q3	1853-0015	7	2	TRANSISTOR PNP SI PD=200MW FT=500MHZ	28480	1853-0015
A1Q4	1854-0215	1	4	TRANSISTOR NPN SI PD=350MW FT=300MHZ	04713	2N3904
A1Q5	1854-0215	1	4	TRANSISTOR NPN SI PD=350MW FT=300MHZ	04713	2N3904
A1Q6	1854-0215	1	4	TRANSISTOR NPN SI PD=350MW FT=300MHZ	04713	2N3904
A1Q7	1854-0019	3	2	TRANSISTOR NPN SI TO-18 PD=360MW	28480	1854-0019
A1Q8	1854-0019	3	2	TRANSISTOR NPN SI TO-18 PD=360MW	28480	1854-0019
A1R1	0757-0442	9	4	RESISTOR 10K 1% .125W F TC=0+-100	24546	C4=1/8-T0=1002-F
A1R2	0698-3156	2	1	RESISTOR 14.7K 1% .125W F TC=0+-100	24546	C4=1/8-T0=1472-F
A1R3	0698-3449	6	1	RESISTOR 28.7K 1% .125W F TC=0+-100	24546	C4=1/8-T0=2872-F
A1R4	0698-3447	4	8	RESISTOR 422 1% .125W F TC=0+-100	24546	C4=1/8-T0=422R-F
A1R5	0698-3447	4	8	RESISTOR 422 1% .125W F TC=0+-100	24546	C4=1/8-T0=422R-F

See introduction to this section for ordering information
 *Indicates factory selected value

Table 6-2 Replaceable Parts (continued)

Reference Designation	HP Part Number	C D	Qty	Description	Mfr Code	Mfr Part Number
A1R6	0698-3447	4		RESISTOR 422 1% .125W F TC0+-100	24546	C4-1/8-T0-422R-F
A1R7	0698-3447	4		RESISTOR 422 1% .125W F TC0+-100	24546	C4-1/8-T0-422R-F
A1R8	0698-3447	4		RESISTOR 422 1% .125W F TC0+-100	24546	C4-1/8-T0-422R-F
A1R9	0698-3447	4		RESISTOR 422 1% .125W F TC0+-100	24546	C4-1/8-T0-422R-F
A1R10	0698-3447	4		RESISTOR 422 1% .125W F TC0+-100	24546	C4-1/8-T0-422R-F
A1R11	0757-0442	9		RESISTOR 10K 1% .125W F TC0+-100	24546	C4-1/8-T0-1002-F
A1R12	0757-0438	3	2	RESISTOR 5.11K 1% .125W F TC0+-100	24546	C4-1/8-T0-5111-F
A1R13	0757-0438	3		RESISTOR 5.11K 1% .125W F TC0+-100	24546	C4-1/8-T0-5111-F
A1R14	0698-3444	1	3	RESISTOR 316 1% .125W F TC0+-100	24546	C4-1/8-T0-316R-F
A1R15	0757-0416	7	1	RESISTOR 511 1% .125W F TC0+-100	24546	C4-1/8-T0-511R-F
A1R16	0698-3444	1		RESISTOR 316 1% .125W F TC0+-100	24546	C4-1/8-T0-316R-F
A1R17	2100-0554	5	1	RESISTOR-TRMR 500 10% C TOP-ADJ 1-TRN	28480	2100-0554
A1R18	0698-3441	6	1	RESISTOR 215 1% .125W F TC0+-100	24546	C4-1/8-T0-215R-F
A1R19	0698-3444	1		RESISTOR 316 1% .125W F TC0+-100	24546	C4-1/8-T0-316R-F
A1R20	0757-0398	4	2	RESISTOR 75 1% .125W F TC0+-100	24546	C4-1/8-T0-75R0-F
A1R21	0698-3442	9	2	RESISTOR 237 1% .125W F TC0+-100	24546	C4-1/8-T0-237R-F
A1R22	0698-3434	9	2	RESISTOR 34.8 1% .125W F TC0+-100	24546	C4-1/8-T0-348R-F
A1R23	0698-3434	9	9	RESISTOR 34.8 1% .125W F TC0+-100	24546	C4-1/8-T0-348R-F
A1R24	0698-3442	9	4	RESISTOR 237 1% .125W F TC0+-100	24546	C4-1/8-T0-237R-F
A1R25	0757-0398	4	9	RESISTOR 75 1% .125W F TC0+-100	24546	C4-1/8-T0-75R0-F
A1R26	0757-0276	7	1	RESISTOR 61.9 1% .125W F TC0+-100	24546	C4-1/8-T0-6192-F
A1R27	0757-0427	0	1	RESISTOR 1.5K 1% .125W F TC0+-100	24546	C4-1/8-T0-1501-F
A1R28	0698-0084	9	3	RESISTOR 2.15K 1% .125W F TC0+-100	24546	C4-1/8-T0-2151-F
A1R29	0698-3447	4	4	RESISTOR 422 1% .125W F TC0+-100	24546	C4-1/8-T0-422R-F
A1R30	2100-3211	7	1	RESISTOR-TRMR 1K 10% C TOP-ADJ 1-TRN	28480	2100-3211
A1R31	0757-0442	9		RESISTOR 10K 1% .125W F TC0+-100	24546	C4-1/8-T0-1002-F
A1R32	0698-0084	9		RESISTOR 2.15K 1% .125W F TC0+-100	24546	C4-1/8-T0-2151-F
A1R33	0698-0084	9		RESISTOR 2.15K 1% .125W F TC0+-100	24546	C4-1/8-T0-2151-F
A1R34	0698-3439	4	3	RESISTOR 178 1% .125W F TC0+-100	24546	C4-1/8-T0-178R-F
A1R35	0757-0274	5	1	RESISTOR 1.21K 1% .125W F TC0+-100	24546	C4-1/8-T0-1213-F
A1R36	0757-0280	3	1	RESISTOR 1K 1% .125W F TC0+-100	24546	C4-1/8-T0-1001-F
A1R37	0698-3439	4		RESISTOR 178 1% .125W F TC0+-100	24546	C4-1/8-T0-178R-F
A1R38	0757-0442	9		RESISTOR 10K 1% .125W F TC0+-100	24546	C4-1/8-T0-1002-F
A1R39	0698-3439	4		RESISTOR 178 1% .125W F TC0+-100	24546	C4-1/8-T0-178R-F
A181	3101-1596	0	3	SWITCH-SL DPDT MINTR 1A 125VAC PC	28480	3101-1596
A182	3101-2158	2	1	SWITCH-SL 6-1A DIP-SLIDE-ASSY 1A 50VDC	28480	3101-2158
A183	3101-0493	4	1	SWITCH-SLIDE 8P3T S	28480	3101-0493
A184	3101-1596	0	0	SWITCH-SL DPDT MINTR 1A 125VAC PC	28480	3101-1596
A185	3101-1596	0	2	SWITCH-SL DPDT MINTR 1A 125VAC PC	28480	3101-1596
A1T11	37203-20002	2	1	TEST BOARD	28480	37203-20002
A1U1	1820-1669	4	4	IC UART TTL QUAD	04713	MC3446P
A1U2	1820-1689	4	4	IC UART TTL QUAD	04713	MC3446P
A1U3	1820-1689	4	4	IC UART TTL QUAD	04713	MC3446P
A1U4	1820-1689	4	4	IC UART TTL QUAD	04713	MC3446P
A1U5	1820-1199	1	4	IC INV TTL LS HEX 1-INP	01295	8N74LS04N
A1U6	1820-1199	1		IC INV TTL LS HEX 1-INP	01295	8N74LS04N
A1U7	1820-1199	1		IC INV TTL LS HEX 1-INP	01295	8N74LS04N
A1U8	1820-1144	6	5	IC GATE TTL LS NOR QUAD 2-INP	01295	8N74LS02N
A1U9	1820-1144	6		IC GATE TTL LS NOR QUAD 2-INP	01295	8N74LS02N
A1U10	1820-1144	6		IC GATE TTL LS NOR QUAD 2-INP	01295	8N74LS02N
A1U11	1820-1144	6		IC GATE TTL LS NOR QUAD 2-INP	01295	8N74LS02N
A1U12	1820-1201	6	4	IC GATE TTL LS AND QUAD 2-INP	01295	8N74LS08N
A1U13	1820-1197	9	8	IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U14	1820-1197	9		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U15	1820-1197	9		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U16	1820-1197	9		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U17	1820-1202	7	4	IC GATE TTL LS NAND TPL 3-INP	01295	8N74LS10N
A1U18	1820-1203	8	1	IC GATE TTL LS AND TPL 3-INP	01295	8N74LS11N
A1U19	1820-1212	9	3	IC FF TTL LS J-K NEG-EDGE-TRIG	01295	8N74LS112AN
A1U20	1820-1212	9		IC FF TTL LS J-K NEG-EDGE-TRIG	01295	8N74LS112AN
A1U21	1820-1212	9		IC FF TTL LS J-K NEG-EDGE-TRIG	01295	8N74LS112AN
A1U22	1820-1433	6	1	IC SHF-RGTR TTL LS R-S SERIAL-IN PRL-OUT	01295	8N74LS164N
A1U23	1820-1423	4	2	IC MV TTL LS MONOSTBL RETRIG DUAL	01295	8N74LS123N
A1U24	1820-1423	4		IC MV TTL LS MONOSTBL RETRIG DUAL	01295	8N74LS123N
A1U25	1820-1201	6		IC GATE TTL LS AND QUAD 2-INP	01295	8N74LS08N
A1U26	1820-1199	1		IC INV TTL LS HEX 1-INP	01295	8N74LS04N
A1U27	1820-1478	9	2	IC CNTR TTL LS BIN ASYNCHRD	01295	8N74LS93N
A1U28	1820-1478	9		IC CNTR TTL LS BIN ASYNCHRD	01295	8N74LS93N
A1U29	1820-1991	1	1	IC CNTR TTL LS DECD DUAL 4-BIT	01295	8N74LS390N
A1U30	1820-1191	3	1	IC FF TTL S D-TYPE POS-EDGE-TRIG COM	01295	8N74S175N
A1U31	1820-0685	8	1	IC GATE TTL S NAND TPL 3-INP	01295	8N74S16N
A1U32	1820-0681	4	2	IC GATE TTL S NAND QUAD 2-INP	01295	8N74S00N
A1U33	1820-0681	4		IC GATE TTL S NAND QUAD 2-INP	01295	8N74S00N
A1U34	1820-1197	6		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U35	1820-1276	5	9	IC SHF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN

See introduction to this section for ordering information
 *Indicates factory selected value

Table 6-2 Replaceable Parts (continued)

Reference Designation	HP Part Number	C D	Qty	Description	Mfr Code	Mfr Part Number
A1U36	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U37	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U38	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U39	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U40	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U41	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U42	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U43	1820-1276	5		IC 8MF-RGTR TTL LS R-S PRL-IN PRL-OUT	01295	8N74LS194AN
A1U44	1820-1730	6	2	IC FF TTL LS D-TYPE POS-EDGE-TRIG COM	01295	8N74LS273N
A1U45	1820-1730	6		IC FF TTL LS D-TYPE POS-EDGE-TRIG COM	01295	8N74LS273N
A1U46	1820-1207	2	1	IC GATE TTL LS NAND 8-INP	01295	8N74LS30N
A1U47	1820-1430	3	2	IC CNTR TTL LS BIN SYNCHRO POS-EDGE-TRIG	01295	8N74LS161AN
A1U48	1820-1430	3		IC CNTR TTL LS BIN SYNCHRO POS-EDGE-TRIG	01295	8N74LS161AN
A1U49	1820-1195	7	2	IC FF TTL LS D-TYPE POS-EDGE-TRIG COM	01295	8N74LS175N
A1U50	1820-1195	7		IC FF TTL LS D-TYPE POS-EDGE-TRIG COM	01295	8N74LS175N
A1U51	1820-1211	8	2	IC GATE TTL LS EXCL-OR QUAD 2-INP	01295	8N74LS86N
A1U52	1820-1211	8		IC GATE TTL LS EXCL-OR QUAD 2-INP	01295	8N74LS86N
A1U53	1820-1202	7		IC GATE TTL LS NAND TPL 3-INP	01295	8N74LS10N
A1U54	1820-1144	6		IC GATE TTL LS NOR QUAD 2-INP	01295	8N74LS02N
A1U55	1820-1197	9		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U56	1820-1197	9		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U57	1820-1202	7		IC GATE TTL LS NAND TPL 3-INP	01295	8N74LS10N
A1U58	1820-1204	9	1	IC GATE TTL LS NAND DUAL 4-INP	01295	8N74LS20N
A1U59	1820-1202	7		IC GATE TTL LS NAND TPL 3-INP	01295	8N74LS10N
A1U60	1820-1197	9		IC GATE TTL LS NAND QUAD 2-INP	01295	8N74LS00N
A1U61	1820-1201	6		IC GATE TTL LS AND QUAD 2-INP	01295	8N74LS08N
A1U62	1990-0429	7	2	OPTO-ISOLATOR LED-IC GATE IF=10MA-MAX	28480	1990-0429
A1U63	1990-0429	7		OPTO-ISOLATOR LED-IC GATE IF=10MA-MAX	28480	1990-0429
A1U64	1826-0181	1	1	IC V RGLTR TO-3	27014	LM321K
	37203-00004	2	1	HEAT SINK	28480	37203-00004
A1U65	1826-0122	0	1	IC 7905 V RGLTR TO-220	07263	7805UC
A1U66	1826-0445	0	1	IC 7905 V RGLTR TO-220	07263	UA7905UC
A1U67	1820-1201	6		IC GATE TTL LS AND QUAD 2-INP	01295	8N74LS08N
A1V1	0410-1217	4	1	CRYSTAL-24MHZ	28480	0410-1217

See introduction to this section for ordering information
 *Indicates factory selected value

Table 6-2 Replaceable Parts (continued)

Reference Designation	HP Part Number	C D	Qty	Description	Mfr Code	Mfr Part Number
37203A MAIN LIST						
A1	37203-60001		5	BOARD ASSEMBLY	28480	37203-60001
C1	0160-3561		4	CAPACITOR=FXD 1000PF/1000PF +100-0%	28480	0160-3561
C2	0150-0096		3	CAPACITOR=FXD .05UF +80-20% 100VDC CER	28480	0150-0096
F1	2110-0201		0	FUSE .25A 250V 1.25X.25 UL (115V OPERATION)	75915	313.250
F1	2110-0318		0	FUSE .125A 250V 1.25X.25 UL (250V OPERATION)	75915	313.125
MP1	37203-00001		9	FRONT PANEL	28480	37203-00001
MP2	37203-00003		1	FRONT PANEL-SUB	28480	37203-00003
MP3	5020-8813		8	FRAME-FRONT	28480	5020-8813
MP4	5040-7203		0	TRIM STRIP-TOP	28480	5040-7203
MP5	5020-8830		9	SIDE STRUT	28480	5020-8830
MP6	5060-9818		9	COVER-TOP	28480	5060-9818
MP7	5060-9963		5	COVER-BOTTOM	28480	5060-9963
MP8	5040-7201		6	FOOT (STANDARD)	28480	5040-7201
MP9	5001-0438		7	TRIM STRIP-SIDE	28480	5001-0438
MP10	37203-00002		0	REAR PANEL	28480	37203-00002
MP11	5020-8814		9	FRAME-REAR	28480	5020-8814
MP12	37203-00006		4	BRACKET=TRANSFORMER	28480	37203-00006
MP13	37203-00005		3	BRACKET=SWITCH	28480	37203-00005
RV1	0837-0204		6	VARIATOR=24V	28480	0837-0204
S1	3101-2216		3	SW-PB DPDT	28480	3101-2216
	5041-0268		5	KEY CAP	28480	5041-0268
T1	37203-80001		7	TRANSFORMER	28480	37203-80001
W1	37203-60011		7	CABLE ASSEMBLY=POWER	28480	37203-60011
W2	37203-60010		6	CABLE ASSEMBLY=COAX	28480	37203-60010
	00310-48801		0	INSULATOR=BNC	28480	00310-48801
MISCELLANEOUS PARTS						
	6960-0006		6	PLUG=HOLE DOME=MD FOR .25-D=HOLE STL	28480	6960-0006
	0380-0644		4	STANDOFF=HEX .327-IN=LG 6-32THD	00000	ORDER BY DESCRIPTION
OPTION 001 HAS THE SAME PARTS AS THE STANDARD - LESS THE FOLLOWING						
	6960-0006		6	PLUG=HOLE DOME=MD FOR .25-D=HOLE STL	28480	6960-0006
	37203-60001		5	BOARD ASSEMBLY	28480	37203-60001
PLUS THE FOLLOWING:						
	37203-60101		6	BOARD ASSEMBLY	28480	37203-60101
OPTION 001 BUS EXTENDER						
USE STANDARD REPLACEABLE PARTS LIST WITH THE FOLLOWING EXCEPTIONS						
A1	37203-60101		6	BOARD ASSEMBLY-OPTION 001	28480	37203-60101
A10P1	1005-0021		9	FIBRE OPTIC TRANSMITTER	28480	1005-0021
A10P2	1005-0005		8	FIBRE OPTIC RECEIVER	28480	1005-0005

See introduction to this section for ordering information
 *Indicates factory selected value

Table 6-3 Code List of Manufacturers

Mfr No.	Manufacturer Name	Address	Zip Code
00000	ANY SATISFACTORY SUPPLIER		
01295	TEXAS INSTR INC SEMICOND COMPNT DIV	DALLAS TX	75222
04713	MOTOROLA SEMICONDUCTOR PRODUCTS	PHOENIX AZ	85062
07263	FAIRCHILD SEMICONDUCTOR DIV	MOUNTAIN VIEW CA	94042
24546	CORNING GLASS WORKS (BRADFORD)	BRADFORD PA	16701
25088	SIEMENS CORP	ISELIN NJ	08830
27014	NATIONAL SEMICONDUCTOR CORP	SANTA CLARA CA	95051
28480	HEWLETT-PACKARD CO CORPORATE HQ	PALO ALTO CA	94304
52763	STETTNER-TRUSH INC	CAZENOVIA NY	13035
72136	ELECTRO MOTIVE CORP SUB IEC	WILLIMANTIC CT	06226
75915	LITTELFUSE INC	DES PLAINES IL	60016

SECTION VII MANUAL CHANGES

7-1 INTRODUCTION

7-2 This section normally contains information for adapting this manual to instruments for which the manual content does not apply directly. Since this manual does

apply directly to instruments having Serial Numbers listed on the title page, no change information is given here. Refer to INSTRUMENT AND MANUAL IDENTIFICATION in Section I for additional important information about Serial Numbers coverage.