


Agilent 81110A/'11A Performance Test

Introduction

Use the tests in this chapter if you want to check that the Agilent 81110A Pulse Generator Frame with the Agilent 81111A 165 MHz Output Channel(s) is working correctly. Before starting any testing allow all test equipment to warm up for at least 30 minutes.

Conventions Used

When referring to actions that you perform during the tests, the following conventions are used:

FUNCTION This indicates that a labelled button must be pressed

[**MODE/TRG**] This shows that a soft-key must be pressed. A soft-key is an unlabelled button whose label is shown on the display, and which can vary according to the job that the button is doing

CONTINUOUS PULSES This is an option shown on the display, and is selected by use of the vernier keys. It is shown in upper or lower case to match the case displayed.

Test Results Tables

Tables for entering the results of the tests are included at the end of this chapter. The tests are numbered and reference numbers for each Test Result (TR) are given in a small table at the end of each test. The reference number shows you where the actual results should be entered in the Test Results Tables.

The Test Results tables at the end of the chapter should be photocopied, and the Test Results entered on the copies. Then, if the tests need to be repeated, the tables can be copied again.

Agilent 81110A/'11A Performance Test

If Channel 2 has been fitted to your instrument, make an extra copy of the Test Results tables for entry of the results of tests on that channel. In this case, however, it is not necessary to repeat the Period tests, as these are common to both channels.

Agilent 81110A/'11A Performance Test

Recommended Test Equipment and Accessories

The following tables list the recommended test equipment you need to perform all the tests in this chapter. You can use alternative instruments if they meet the critical specifications given. The test set-ups and procedures assume you are using the recommended equipment.

Test Equipment	Model	Critical Specifications
Oscilloscope or	Agilent 54121T	20 GHz, 10 bit vertical resolution, Histogram
Oscilloscope	Agilent 54750A + Agilent 54751A	20 GHz, 15 bit vertical resolution, Histogram
Counter or	Agilent 5334B #010, 030	Period and Time Interval measurements Oven Osci, 1.3 GHz C-Channel
Counter	Agilent 53132A #001/010, 030	Frequency measurements > 150 MHz High-Stability Timebase, 3 GHz Channel
Digital Voltmeter	Agilent 3458A	DCV up to 20 V
Pulse Generator	Agilent 8110A	up to 150 MHz
Delay line	Agilent 54008A	22 ns

Accessories	Model	Critical Specifications
Digitizing Oscilloscopes Accessories Attenuators	33340C#020 33340C#006	20 dB 6 dB
Power Splitter	11667B	
SMA/SMA (m-m)Adapter	1250-1159	
SMA/BNC Adapter	1250-1700	
SMA Cable	8120-4948	

Agilent 81110A/'11A Performance Test

Accessories	Model	Critical Specifications
50 Ω Feedthrough Termination	10100C See Figure	2 W,1% 10 W,0.1%
Adapter	1251-2277	BNC to Banana
Cable Assemblies, BNC	8120-1839	
Torque Wrench	8710-1582	5/16 in, 5 lb-in (56 Ncm)


NOTE:

When you connect the test equipment for the first time, and whenever you change the setup during the course of these tests, use the 8710 - 1582 torque wrench to tighten and loosen SMA connectors. This will ensure that the connectors are at the correct tightness and give the best signal transfer.

Agilent 81110A/'11A Performance Test

50 Ohm, 0.1%, 10 W Feedthrough Termination

The following figure provides a schematic and a parts list except for the case. The case must provide shielding and maintain grounding integrity.

**50 Ohm, 0.1%, 10 W Feedthrough Termination**

The following parts are required:

1. R1 = 53.6 Ω , 1%, 10 W; Part Number: 0699-0146.
2. R2 = 200 Ω , 10%, 0.5 W, Variable trimmer; Part Number: 2100-3350.
3. R3 = 681 Ω ;, 1%, 0.5 W; Part Number: 0757-0816.
4. BNC (M): Part Number: 1250-0045.
5. BNC (F): Part Number: 1250-0083.

Getting Started

The Agilent 81110A is controlled by selecting options in a series of **pages** that are displayed on the instrument's screen. These options vary with the boards that are fitted in the instrument. When the Agilent 81110A is being tested, therefore, different situations can arise, depending on whether you have a standard instrument or one that has had additional boards fitted. The following examples illustrate this

Typical Examples of Displayed Screens

Per 1.000ms Normal				OFF	1	☒ MODIFY
Delay	0ps	Offset	+0.0mV	*OFF		
Width	100.0ns	Amplit	1.00V	ON		
LeadE	2.00ns	50W into	50.0W			
Traie	=LeadE					
MODE/TRG		OUTPUT		LIMITS		PATTERN

The OUTPUT Screen in a Standard Agilent 81110A

Agilent 81110A/'11A Performance Test

1	OFF	Per 1.000ms	OFF	2	MODIFY
Delay		0ps	Delay		0ps
Width		100.0ns	Width		100.0ns
LeadE		2.00ns	LeadE		2.00ns
Traie		=LeadE	Traie		=LeadE
				*Period Frequency	
MODE/TRG		TIMING		LEVELS	
				PATTERN	

The TIMING Screen in an Agilent 81110A with qty 2 of Agilent 81111A

1	ON	Normal	Normal	ON	2	MODIFY
Seperate Outputs						
High		+2.50V	High		+2.50V	
Low		+0.0mV	Low		+0.0mV	
50W into 50.0W		50W into 50.0W				
				Set TTL *High-Low Offs-Ampl Set ECL		
MODE/TRG		TIMING		LEVELS		
				PATTERN		

The LEVELS Screen in an Agilent 81110A with qty 2 of Agilent 81111A

Instrument Serial Numbers

You will need to write the serial numbers of the instrument at the top of the Test Reports. These can be found as follows:

Press HELP, [**SERIAL #**]

The Agilent 81110A display lists the instrument's products and serial number.

The display on your instrument should look similar to this:

FRAME : **81110A** **165 MHz**
Serial No : **DE38700135**

OUTPUTS

Ch1-Bd. : **81111A**
Ch2-Bd. : **81111A**

The serial number given for the **FRAME** applies to the Mainframe, the Power Supply, the Microprocessor Board, and the Timing Board. The number(s) available of the Output Channel(s) applies to the installed numbers of outputs and Model Number.

Agilent 81110A/'11A Performance Test

Initial Setup of the Agilent 81110A

In the majority of these tests the initial setting up of the instrument is identical. Therefore, it is described once here, and then referred-to where appropriate. In cases where the initial setup differs, an illustration of the settings is shown.

Set up the Agilent 81110A as follows:

1. Select [MODE/TRG]
 - **CONTINUOUS PULSES**
 - **Single-Pulses at Out 1** (plus **Single-Pulses at Out 2**, if second channel is installed)
 - **Pulse-Period:internal Osc**
2. If a second output channel is installed, select MORE [CONFIG] screen and set up as follows:

 GPIB Address: 10 Perform: Timing Calibration Group Params by: OUTPUT 1 / 2 PLL-Ref : Internal		☒ MODIFY Tim/Lev *Out 1/2
LIMITS	TRG-LEV	MEMCARD CONFIG

CONFIG Screen, Parameters grouped by OUTPUT

Agilent 81110A/'11A Performance Test

NOTE:

Set-ups are given in all the tests for [**OUTPUT 1**] and [**OUTPUT 2**]. If you are testing a single channel instrument set up the [**OUTPUT**] screen with the settings given for [**OUTPUT 1**].

Agilent 81110A/'11A Performance Test

Test 1: Period (PLL not active)**Test Specifications**


Range	6.06 ns to 999.5 s
Resolution	3.5 digits, best case 5 ps
Accuracy	$\pm 3\%$ typical $\pm 0.5\%$ after selfcal

Equipment Needed

Counter
Cable, 50 Ω , coaxial, BNC

Procedure

1. Connect the Agilent 81110A to the Counter as shown:



Connecting the Agilent 81110A to the Counter

2. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"

Agilent 81110A/'11A Performance Test

On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per	6.060ns	Normal	ON	1	☒ MODIFY
Delay	0ps	Offset	+0.0mV		6.060 ns
DtyCyc	50.00%	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
TraiE	=LeadE				
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output 1

Per	6.060ns	Normal	OFF	2	☒ MODIFY
Delay	0ps	Offset	+0.0mV		6.060 ns
DtyCyc	50.00%	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
TraiE	=LeadE	Separate	Out2		
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output 2

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. For the Period test you can switch OFF the channels that are not being tested.

3. Set the Counter to:

FUNCTION	Period A / Freq C
INPUT A	50 Ω
SENSE	On

4. Check the Agilent 81110A period at the following settings:

Period	Acceptable Range	TR entry
6.060 ns	<i>without selfcal!</i> 5.878 ns to 6.242 ns	1 - 1
9.990 ns	9.690 ns to 10.290 ns	1 - 2
10.00 ns	9.7 ns to 10.3 ns	1 - 3
50.00 ns	48.5 ns to 51.5 ns	1 - 4
99.90 ns	96.903 ns to 102.897 ns	1 - 5
100 ns	97 ns to 103 ns	1 - 6
500 ns	485 ns to 515 ns	1 - 7
1 μ s	970 ns to 1030 ns	1 - 8
500 μ s	485 μ s to 515 μ s	1 - 9
500 ms	485 ms to 515 ms	1 - 10

5. To perform the Timing Calibration (shown as selfcal) set up [CONFIG] page as shown in the following illustration:

HP-IB Address: 10		☒ MODIFY Selftest *Calibrate	
Perform: Timing Calibration			
Group Params by: OUTPUT 1 / 2			
PLL-Ref : Internal			
LIMITS	TRG-LEV	MEMCARD	CONFIG

6. Press ENTER and wait till the display shows TIMING CALIBRATION PASSED and gets back to the above shown display .

Test 2: PLL Period

NOTE: This test is only performed if PLL is switched on.

Test Specifications

Range 6.06 ns to 999.5 s
 Resolution 4 digits, best case 1 ps
 Accuracy $\pm 0.01\%$

Equipment Needed

Counter Agilent 53132A

Agilent 81110A/'11A Performance Test


Cable, 50 Ω , coaxial, BNC

NOTE:

The Agilent 53132A counter is used in frequency mode to meet the MIL CAL A uncertainty requirements for TAR (Test Accuracy Ratio) > 4:1.

Procedure

Connect the Agilent 81110A to the counter as follows:

**Connecting Agilent 81110A to the Counter**

7. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
8. Select the [MODE/TRG] screen on the Agilent 81110A and set up as follows:

Agilent 81110A/'11A Performance Test

CONTINUOUS PULSES		 MODIFY	
Single-Pulses at Out1			
Single-Pulses at Out2			
Pulse-Period: internal PLL			
		int. OSC	
		*int. PLL	
		CLK-IN	
MODE/TRG	TIMING	LEVELS	PATTERN

The MODE/TRG Screen Setup

9. On the Agilent 81110A set up **[OUTPUT 1]** and **[OUTPUT 2]** pages as shown in the test before!

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. For the Period test you can switch OFF the channels that are not being tested.

10. Set the Counter to measure the frequency at the chosen input
1 / 3

11. Check the Agilent 81110A PLL pulse period at the following settings:

Period	Frequency	Acceptable Range	TR Entry
6.061 ns	165.000MHz	164.9835 MHz to 165.0165 MHz	2 - 1
10.00 ns	100 MHz	99.990 MHz to 100.010 MHz	2 - 2
50.00 ns	20 MHz	19.9980 MHz to 20.0020 MHz	2 - 3
100 ns	10 MHz	9.9990 MHz to 10.0010 MHz	2 - 4
500 ns	2 MHz	1.9998 MHz to 2.0002 MHz	2 - 5
1 μ s	1 MHz	999.9 kHz to 1.0001 MHz	2 - 6
50 μ s	20 kHz	9.998 kHz to 20.002 kHz	2 - 7
5 ms	200 Hz	199.980 Hz to 200.020 Hz	2 - 8
500 ms	2 Hz	1.9998 Hz to 2.0002 Hz	2 - 9
5 s	0.2 Hz	0.19998 Hz to 0.20002 Hz	2 - 10

Test 3: Width

Test Specifications


Range	3.03 ns to (period - 3.03 ns)
Resolution	3.5 digits, best case 5 ps
Accuracy	$\pm 3\% \pm 250$ ps typical $\pm 0.5\% \pm 250$ ps after selfcal

Equipment Needed

Digitizing Oscilloscope with Accessories
Counter
Cable, 50 Ω , coaxial, BNC

Procedure

1. Connect Agilent 81110A to the Scope as shown:



Connecting Agilent 81110A to the Scope

2. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"

Agilent 81110A/'11A Performance Test

3. On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per	200 ns	Normal	ON	1	 MODIFY
Delay	0ps	Offset	+0.0mV		100.0 ns
Width	100.0ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
TraIE	=LeadE				
MODE/TRG OUTPUT 1 OUTPUT 2 PATTERN					

Configuring Output Screen 1

Per	200 ns	Normal	OFF	2	 MODIFY
Delay	0ps	Offset	+0.0mV		3.030 ns
Width	3.030ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
TraIE	=LeadE	Separate Out2			
MODE/TRG OUTPUT 1 OUTPUT 2 PATTERN					

Configuring Output Screen 2

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested.

If you then test the other channel:

- c. Switch ON the channel your are testing, and switch OFF the other channel.

4. Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
- Select the Display menu and set the Number of Averages to 32
- Select the delta V menu and turn the voltage markers On
- Set the preset levels to 50% -50% and press AUTO LEVEL SET
- Select the delta t menu and turn the time markers ON
- Set START ON EDGE = POS 1 and STOP ON EDGE = NEG1

5. Change the oscilloscope timebase to 1 ns/div

6. Change the Agilent 81110A Ch-1 Width to 3.03 ns

7. Center the pulse in the Scope display


8. Press the PRECISE EDGE FIND key for each new Width setting

Agilent 81110A/'11A Performance Test

9. Check the Agilent 81110A pulse width at the following settings:

Oscilloscope Timebase	Period	Width	Acceptable Range	TR Entry
1 ns/div	200 ns	<i>without selfcal!</i> 3.030 ns	2.689 ns to 3.371 ns	3 - 1
1 ns/div	200 ns	6.060 ns	5.528 ns to 6.492 ns	3 - 2
2 ns/div	200 ns	10.00 ns	9.450 ns to 10.55 ns	3 - 3
10 ns/div	200 ns	50.00 ns	48.25 ns to 51.75 ns	3 - 4
20 ns/	1 μ s	100.0 ns	96.75 ns to 103.25 ns	3 - 5
100 ns	1 μ s	500.0 ns	484.75 ns to 515.25 ns	3 - 6

10. Connect the Agilent 81110A to the Counter as shown:



Connecting Agilent 81110A to the Counter

11. Set the Counter to:

Agilent 81110A/'11A Performance Test

FUNCTION TI A → B
 SENSE On
 INPUT A 50 Ω
 COM A On
 INPUT B 50 Ω , negative slope

12. Check the Agilent 81110A width at the following settings:

Period	Width	Acceptable Range	TR Entry
100 μ s	50 μ s	48.5 μ s to 51.5 μ s	3 - 7
10 ms	5 ms	4.85 ms to 515ms	3 - 8
999 ms	500ms	485 ms to 515 ms	3 - 9

NOTE:

Repeat the entire test for the second channel, if it is installed

Test 4: Delay

Test Specifications

Range	Fixed typical Delay of EXT INPUT to TRIGGER OUT 12 ns TRIGGER OUT to OUTPUT 1/2 14 ns Variable Delay: 0 ns to period - 3.03ns
Resolution	3.5 digits, best case 5 ps
Accuracy	$\pm 3\%$ ± 0.5 ns typical $\pm 0.5\%$ ± 0.5 ns after selfcal


Equipment Needed

Digitizing Oscilloscope with Accessories
Pulse Generator
Counter
Cable, 50 Ω , coaxial, BNC

Procedure

Connect Agilent 81110A to the Scope as shown:

Agilent 81110A/'11A Performance Test



Connecting Agilent 81110A to the Scope

13. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"

14. Set the Pulse Generator to:

Period	1 μ s
Width	100 ns
Amplitude	1 V
Offset	+1.0 V
Output	Enable

15. Select the [MODE/TRG] screen on the Agilent 81110A and set up as follows:

Agilent 81110A/'11A Performance Test

TRIGGERED	PULSES	☒ MODIFY Continous *Triggered Gated Ext-Width
Single-Pulses at Out1 Single-Pulses at Out2		
Trg'd by: EXT-IN 		
MODE/TRG	TIMING	LEVELS
PATTERN		

The TRG MODE Screen Setup

16. On the Agilent 81110A press MORE and set up [TRIG-LEV] page as shown:

EXT-IN: Threshold	+1.0V 50W	☒ MODIFY Set TTL Set ECL *Voltage
CLK-IN: Threshold	+1.0V 50W	
TRIGGER-OUT: TTL STROBE-OUT : TTL		
LIMITS	TRG-LEV	MEMCARD
CONFIG		

The TRG-LEV Screen Setup

Agilent 81110A/'11A Performance Test

17. On the Agilent 81110A set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per	-----	Normal	ON	1	☒ MODIFY
Delay	0ps	Offset	+0.0mV		0 ps
Width	100ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
Traie	=LeadE				
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output Screen 1

Per	-----	Normal	OFF	2	☒ MODIFY
Delay	0ps	Offset	+0.0mV		0 ps
Width	100ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
Traie	=LeadE	Separate Out2			
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output Screen 2

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested.

If you then test the other channel:

- c. Switch ON the channel your are testing, and switch OFF the other channel.

Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
- Set timebase to TIME/DIV = 10 ns/div
- Center the positive-going edges of the two signals
- Select the Display menu and set the screen function to single; set the number of averages to 32
- Select the Delta V menu and turn the voltage markers ON and assign marker 1 to channel 3 and marker 2 to channel 4
- Set Preset levels to 50% - 50% and press AUTO LEVEL SET
- Select the Delta t menu and turn the time markers ON
- Set START ON EDGE= POS1 and STOP ON EDGE= POS 1
- Press the PRECISE EDGE FIND key

18. Check the Agilent 81110A delay at the following settings:


Agilent 81110A/'11A Performance Test

NOTE:

Record the value of the fixed delay and subtract it from the other readings.

Oscilloscope Timebase	Delay	Acceptable Range	TR Entry
10 ns/div	0 ps	fixed Delay of TRIG OUT to OUT 1/2: 14 ns typ.	4 - 1
10 ns/div	<i>without selfcal!</i> 5.000 ns	4.35 ns to 5.65 ns	4 - 2
20 ns/div	10.00 ns	9.200 ns to 10.80 ns	4 - 3
20 ns/div	50.00 ns	48.00 ns to 52.00 ns	4 - 4
50 ns/div	100.0 ns	96.50 ns to 103.50 ns	4 - 5
200 ns/div	500.0 ns	484.50 ns to 515.50 ns	4 - 6

19. Connect the Agilent 81110A to the Counter as follows:



Connecting Agilent 81110A to the Counter

20. Set Agilent 81110A to **Continuous-Pulses** on the MODE/TRG screen

21. Set the Counter to:

Agilent 81110A/'11A Performance Test

FUNCTION TI A $\rightarrow$ B
 SENSE On
 INPUT A 50 Ω
 INPUT B 50 Ω

22. Check the Agilent 81110A delay at the following settings:

NOTE:

Subtract the fixed delay from the other readings

Period	Delay	Acceptable Range	TR Entry
100 μ s	50 μ s	48.5 μ s to 51.5 μ s	4 - 7
10 ms	5 ms	4.85 ms to 515ms	4 - 8
999 ms	500ms	485 ms to 515 ms	4 - 9

NOTE:

Repeat the entire test for the second channel, if it is installed.

Test 5: Double Pulse Delay

Test Specifications


Range	6.06 ns to (period - width - 3.03 ns)
Resolution	3.5 digits, best case 5 ps
Accuracy	$\pm 3\% \pm 150$ ps typical $\pm 0.5\% \pm 150$ ps after selfcal

Equipment Needed

Digitizing Oscilloscope with Accessories
Counter
Cable, 50 Ω , coaxial, BNC

Procedure

1. Connect Agilent 81110A to the Scope as shown:



Connecting Agilent 81110A to the Scope

Agilent 81110A/'11A Performance Test

2. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
3. Select the [MODE/TRG] screen on the Agilent 81110A and set up Output 1 and Output 2 as follows:

CONTINUOUS PULSES		MODIFY Single * Double	
Double-Pulses at Out1			
Double-Pulses at Out2			
Pulse-Period: internal Osc			
MODE/TRG	TIMING	LEVELS	PATTERN

The MODE/TRG Screen Setup

4. On the Agilent 81110A set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Agilent 81110A/'11A Performance Test

Per	200.0ns	Normal	ON	1	☒ MODIFY 6.060 ns
Db1Del	6.060ns	Offset	+0.0mV		
Width	3.030ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
Tra1E	=LeadE				
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output Screen 1

Per	200.0ns	Normal	OFF	2	☒ MODIFY 6.060 ns
Db1Del	6.060ns	Offset	+0.0mV		
Width	3.030ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
Tra1E	=LeadE	Separate Out2			
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output Screen 2

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested.

If you then test the other channel:

- c. Switch ON the channel your are testing, and switch OFF the other channel.


Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
 - Center the double pulse signal
 - Select the Display menu and set the Number of Averages to 32
 - Select the Delta V menu and turn the Voltage markers On
 - Set Preset Levels = 50% -50% and press AUTO LEVEL SET
 - Select the Delta t menu and turn the Time markers On
 - Set START ON EDGE = POS1 and STOP ON EDGE = POS2
5. Press the PRECISE EDGE FIND key for each new Double Delay setting
 6. Check the Agilent 81110A double delay at the following settings:

Agilent 81110A/'11A Performance Test

Oscilloscope Timebase	Double Delay	Acceptable Range	TR Entry
2 ns/div	<i>without selfcal!</i> 6.060 ns	5.628 ns to 6.392 ns	5 - 1
2 ns/div	10.00 ns	9.550 ns to 10.45 ns	5 - 2
10 ns/div	50.00 ns	48.35 ns to 51.65 ns	5 - 3
20 ns/div	100.0 ns	96.85 ns to 103.15 ns	5 - 4

7. Connect the Agilent 81110A to the Counter as shown:



Connecting Agilent 81110A to the Counter

8. Set the Counter to:

FUNCTION	Period A
INPUT A	50 Ω
SENSE	On
(EXT ARM	
SELECT	a. Start (ST): leading edge b. Stop (SP): trailing edge)

9. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"

Agilent 81110A/'11A Performance Test

10. Select the **[MODE/TRG]** screen on the Agilent 81110A and set up as follows;

TRIGGERED	PULSES	MODIFY
Double-Pulses at Out1 Double-Pulses at Out2		
Trg'd by: MANKey 		*MAN Key EXT INPUT PLL
MODE/TRG	OUTPUT 1	OUTPUT 2
PATTERN		

The MODE/TRG Screen Setup

11. On the Agilent 81110A set up **[OUTPUT 1]** and **[OUTPUT 2]** pages as shown in the following illustrations:

Per -----	Normal	ON 1	MODIFY
DblDel	500.0ms	Offset	+0.0mV
Width	20.00ns	Amplit	1.00V
LeadE	2.00ns	50W into	50.0W
TraIE	=LeadE		
			500.0 ms
MODE/TRG	OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per	-----	Normal	OFF	2	 MODIFY
DblDel	100.0ns	Offset	+0.0mV		100.0 ns
Width	20.00ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
Traie	=LeadE	Separate	Out2		
MODE/TRG		OUTPUT 1		OUTPUT 2	PATTERN

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- Configure *both* channels.
- Switch OFF the channel that is not being tested.

If you then test the other channel:

- Switch ON the channel your are testing, and switch OFF the other channel.

12. Check the Agilent 81110A double pulse delay at the following settings:

Press MAN to check each new setting!

Agilent 81110A/'11A Performance Test

Double Delay	Acceptable Range	TR Entry
<i>after selfcal!</i>		
500 ms	485 ms to 515 ms	5 - 5
1 s	970.00 ms to 1030.00 ms	5 - 6

NOTE:

Repeat the entire test for the second channel, if it is installed.

Test 6: Jitter

The following tests are required:

1. Period Jitter
 - a. Internal Oscillator
 - b. Internal PLL
2. Width Jitter
3. Delay Jitter

Test 6.1a: Period Jitter, Internal Oscillator

Test Specifications

RMS-Jitter 0.01% + 15 ps


Equipment Needed

Digitizing Oscilloscope with Accessories
Delay Line (22 ns)
Power Splitter
Cable, 50 Ω , coaxial, BNC
Cable, SMA

Procedure

1. Connect Agilent 81110A to the Scope as shown:

Agilent 81110A/'11A Performance Test



Equipment Set-up for Jitter Test

- Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
- On the Agilent 81110A set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per	50.00ns	Normal	ON	1	MODIFY
Delay	0ps	Offset	+500mV		50.00
Width	25.00ns	Amplit	1.00V		ns
LeadE	2.00ns	50W into	50.0W		
Traie	=LeadE				
MODE/TRG		OUTPUT 1		OUTPUT 2	
				PATTERN	

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per 50.00ns Normal		OFF	2	 MODIFY
Delay	0ps	Offset	+500mV	50.00 ns
Width	25.00ns	Amplit	1.00V	
LeadE	2.00ns	50W into	50.0W	
Traie	=LeadE	Separate	Out2	
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 2

NOTE:

Do not change the Amplitude to >1V. This may damage the scope!

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- Configure *both* channels.
- Switch OFF the channel that is not being tested.

If you then test the other channel:

- Switch ON the channel your are testing, and switch OFF the other channel.

Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
- Select the Display menu and set the Number of Averages to 64
- Select the Channel menu and set the Attenuation factor of channel 2 to 2

Agilent 81110A/'11A Performance Test

- Set the VOLTS/DIV of channel 4 to 10 mV/div
 - Set OFFSET to 500 mV
 - Select the Timebase menu and set the TIME/DIV to 100 ps/div
 - Center the first positive-going edge of the signal (approximate Delay = 29ns)
 - Select the Delta V menu and turn the V markers On
 - Set the Marker 1 Position to 490 mV and the Marker 2 Position to 500 mV
 - Select the Delta t menu and turn the T Markers On
 - Set START ON EDGE = POS1 and STOP ON EDGE = POS1
 - Press the PRECISE EDGE FIND key
4. RECORD the delta t reading. This is the rise time of the reference signal within a 1% amplitude window of the signal connected to Input 2. This value is needed later to calculate the correct jitter.(delta.t.up)
 5. Select the Timebase menu and center the second positive-going edge of the signal(approximate Delay = 79 ns)
 6. Press MORE and HISTOGRAM
 - Select the Window submenu and set:
 - Source is channel 2
 - Choose the Time Histogram
 - Press WINDOW MARKER 1 and set it to 490 mV
 - Press WINDOW MARKER 2 and set it to 500 mV

Agilent 81110A/'11A Performance Test

7. Select the Acquire submenu, set the Number of Samples to 1000 and press START ACQUIRING
8. After the data for the time histogram has been acquired (# Samples = 100%), select the Result submenu.
9. Press MEAN and SIGMA. RECORD the values of sigma
10. The RMS-jitter is calculated as follows:

$$RMS - jitter = \frac{6\sigma - \Delta t_{up}}{6}$$

11. The RMS-jitter for period of 50 ns is 20 ps. Enter the result in the Test Report as TR entry 6.1a - 1
12. Set the Agilent 81110A period to 500 ns
13. Repeat steps 6 to 11

NOTE:

TIME/DIV = 200 ps/div; approximate Delay = 529 ns

14. The RMS-jitter for period of 500 ns is 65 ps. Enter the result in the Test Report as TR entry 6.1a - 2

Agilent 81110A/'11A Performance Test

Test 6.1b: Period Jitter, Internal PLL**Test Specifications**

RMS-Jitter 0.001% + 15 ps

Equipment Needed

Digitizing Oscilloscope with Accessories

Delay Line (22 ns)


Power Splitter

Cable, 50 Ω , coaxial, BNC

Cable, SMA

Procedure

1. Connect Agilent 81110A to the Scope as shown.



Equipment Set-up for Jitter Test using the Agilent 54750A + 54751A

Using the Agilent 54121T the Set-up is the same as before.

Agilent 81110A/'11A Performance Test

2. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
3. Select the [MODE/TRG] screen on the Agilent 81110A and set up as follows:

CONTINUOUS PULSES		☒ MODIFY int. OSC *int. PLL CLK-IN	
Single-Pulses at Out1			
Single-Pulses at Out2			
Pulse-Period: internal PLL			
MODE/TRG	TIMING	LEVELS	PATTERN

The TRG MODE Screen Setup

4. On the Agilent 81110A set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per	20.00ns	Normal	ON	1	☒ MODIFY 20.00 ns
Delay	0ps	Offset	+500mV		
Width	10.00ns	Amplit	1.00V		
LeadE	2.00ns	50W into	50.0W		
TraIE	=LeadE				
MODE/TRG	OUTPUT 1	OUTPUT 2	PATTERN		

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per	20.00ns	Normal	OFF	2	 MODIFY
Delay	0ps	Offset	+500mV	20.00	
Width	10.00ns	Amplit	1.00V	ns	
LeadE	2.00ns	50W into	50.0W		
Traie	=LeadE	Separate	Out2		
MODE/TRG	OUTPUT 1	OUTPUT 2	PATTERN		

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- Configure *both* channels.
- Switch OFF the channel that is not being tested.

If you then test the other channel:

- Switch ON the channel your are testing, and switch OFF the other channel.

5. Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
- Select the Display menu and set the Number of Averages to 64
- Select the Channel menu and set the Attenuation factor of channel 2 to 2

Agilent 81110A/'11A Performance Test

- Set the VOLTS/DIV of channel 2 to 10 mV/div
 - Set OFFSET to 500 mV
 - Select the Timebase menu and set the TIME/DIV to 100 ps/div
 - Center the first positive-going edge of the signal (approximate Delay = 29 ns)
 - Select the Delta V menu and turn the V markers On
 - Set the Marker 1 Position to 490 mV and the Marker 2 Position to 500 mV
 - Select the Delta t menu and turn the T Markers On
 - Set START ON EDGE = POS1 and STOP ON EDGE = POS1
 - Press the PRECISE EDGE FIND key
6. RECORD the delta t reading. This is the rise time of the reference signal within a 1% amplitude window of the signal connected to Input 2. This value is needed later to calculate the correct jitter. (delta.t.up)
7. Select the Timebase menu and center the second positive-going edge of the signal (approximate Delay = 49 ns)
8. Press MORE and HISTOGRAM
- Select the Window submenu and set:
 - Source is channel 2
 - Choose the Time Histogram
 - Press WINDOW MARKER 1 and set it to 490 mV
 - Press WINDOW MARKER 2 and set it to 500 mV

Agilent 81110A/'11A Performance Test

9. Select the Acquire submenu, set the Number of Samples to 1000 and press START ACQUIRING
10. After the data for the time histogram has been acquired (# Samples = 100%), select the Result submenu.
11. Press MEAN and SIGMA. RECORD the values of sigma
12. The RMS-jitter is calculated as follows:

$$RMS - jitter = \frac{6\sigma_{\Delta t}}{6}$$

13. The RMS-jitter for period of 20 ns is 15.2 ps. Enter the result in the Test Report as TR entry 6.1b - 1

NOTE:

See the Agilent54750A User's Guide / Service Guide to get the info needed to do the Jitter Test using this scope.

Agilent 81110A/'11A Performance Test

Test 6.2: Width Jitter (PLL not active)**Test Specifications**

RMS-Jitter 0.01% + 15 ps

Equipment Needed

Digitizing Oscilloscope with Accessories

Delay Line (22 ns)


Power Splitter

Cable, 50 Ω , coaxial, BNC

Cable, SMA

Procedure

1. Connect Agilent 81110A to the Scope as shown:

**Equipment Set-up for Jitter Test**

2. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"

Agilent 81110A/'11A Performance Test

3. On the Agilent 81110A set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per 1.000ms Normal				ON	1	☒ MODIFY 3.030 ns
Delay	0ps	Offset	+500mV			
Width	3.030ns	Amplit	1.00V			
LeadE	2.00ns	50W into	50.0W			
TraIE	=LeadE					
MODE/TRG		OUTPUT 1		OUTPUT 2		PATTERN

Configuring Output Screen 1

Per 1.000ms Normal				OFF	2	☒ MODIFY 3.030 ns
Delay	0ps	Offset	+500mV			
Width	3.030ns	Amplit	1.00V			
LeadE	2.00ns	50W into	50.0W			
TraIE	=LeadE	Separate	Out2			
MODE/TRG		OUTPUT 1		OUTPUT 2		PATTERN

Configuring Output Screen 2

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested.

If you then test the other channel:

- c. Switch ON the channel your are testing, and switch OFF the other channel.
-

Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
- Select the Display menu and set the Number of Averages to 128
- Select the Channel menu and set the Attenuation factor of channel 2 to 2
- Set the VOLTS/DIV of channel 2 to 10 mV/div
- Set OFFSET to 500 mV
- Select the Timebase menu and set the TIME/DIV to 10 ps/div
- Center the first negative-going edge of the signal (approximate Delay = 33.8 ns)
- Select the Delta V menu and turn the V markers On
- Set the Marker 1 Position to 500 mV and the Marker 2 Position to 490 mV
- Select the Delta t menu and turn the T Markers On
- Set START ON EDGE = NEG1 and STOP ON EDGE = NEG1
- Press the PRECISE EDGE FIND key

Agilent 81110A/'11A Performance Test

4. RECORD the delta t reading. This is the fall time of the reference signal within a 1% amplitude window of the signal connected to Input 2. This value is needed later to calculate the correct jitter. (delta.t.dn)
5. Set the Agilent 81110A Pulse Width to 50 ns
6. Select the Timebase menu and center the first negative-going edge of the signal (approximate Delay = 80.5 ns)
7. Press MORE and HISTOGRAM
8. Select the Window submenu and set:
 - Source is channel 2
 - Choose the Time Histogram
 - Press WINDOW MARKER 1 and set it to 500 mV
 - Press WINDOW MARKER 2 and set it to 490 mV
9. Select the Acquire submenu, set the Number of Samples to 1000 and press START ACQUIRING
10. After the data for the time histogram has been acquired (# Samples = 100%), select the Result submenu.
11. Press MEAN and SIGMA. RECORD the value of sigma
12. The RMS-jitter is calculated as follows:

Agilent 81110A/'11A Performance Test

$$\text{RMS - jitter} = \frac{6 \text{ sigma} - \text{delta.t.dn}}{6}$$

13. The RMS-jitter for pulse width of 50 ns is 20 ps. Enter the result in the Test Report as TR entry 6.2 - 1

14. Set the Agilent 81110A for pulse width of 500ns

15. Repeat steps 7 to 13

NOTE:

TIME/DIV = 100ps/div. Approximate delay = 530 ns

16. The RMS-jitter for pulse width of 500 ns is 65 ps. Enter the result in the Test Report as TR entry 6.2 - 2

NOTE:

Repeat the entire test for the second channel, if it is installed.

Agilent 81110A/'11A Performance Test

Test 6.3: Delay Jitter (PLL not active)**Test Specifications**


RMS-Jitter 0.01% + 15 ps

Equipment Needed

Digitizing Oscilloscope with Accessories

Procedure

1. Connect Agilent 81110A to the Scope as shown:

**Equipment Set-up for Delay Jitter Test**

2. For calculating the RMS-jitter, the rise time of the reference signal within a 1% amplitude window is required. If this value

Agilent 81110A/'11A Performance Test

is not already measured in the Period Jitter test, then perform the first 6 steps of the Period Jitter test.

3. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
4. On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per 1.000ms Normal		ON 1	 MODIFY
Delay	50.00ns	Offset	+500mV
Width	50.00ns	Amplit	1.00V
LeadE	2.00ns	50W into	50.0W
Traie	=LeadE		
		50.00 ns	
MODE/TRG	OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 1

Per 1.000ms Normal		OFF 2	 MODIFY
Delay	50.00ns	Offset	+500mV
Width	50.00ns	Amplit	1.00V
LeadE	2.00ns	50W into	50.0W
Traie	=LeadE	Separate	Out2
		50.00 ns	
MODE/TRG	OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 2

Agilent 81110A/'11A Performance Test

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested.

If you then test the other channel:

- c. Switch ON the channel your are testing, and switch OFF the other channel.
-

Set the Digitizing Oscilloscope Agilent 54121T:

- Press AUTOSCALE
 - Select the Display menu and set the Number of Averages to 64
 - Set the VOLTS/DIV = 10 mV/div
 - Set OFFSET to 500 mV
 - Select the Timebase menu and set the TIME/DIV to 100 ps/div
 - Center the first positive-going edge of the signal (approximate Delay = 64 ns)
5. Press MORE and HISTOGRAM
 6. Select the Window submenu and press WINDOW MARKER 1 and set it to 490 mV
 7. Press WINDOW MARKER 2 and set it to 500 mV
 8. Select the Acquire submenu, set the Number of Samples to 1000 and press START ACQUIRING
 9. After the delta for the time histogram has been acquired (# Samples = 100%), select the Result submenu.
 10. Press MEAN and SIGMA. RECORD the values of sigma!
 11. The RMS-jitter is calculated as follows:

Agilent 81110A/'11A Performance Test

$$\text{RMS - jitter} = \frac{6\sigma - \Delta t_{\text{up}}}{6}$$

12. The RMS-jitter for delay of 50 ns is 20 ps. Enter the result in the Test Report as TR entry 6.3 - 1
13. Set Agilent 81110A for delay of 500 ns
14. Repeat steps 9 to 12

NOTE:

TIME/DIV = 100 ps/div. Approximate delay = 514ns

15. The RMS jitter for delay of 500 ns is 65 ps. Enter the result in the Test Report as TR entry 6.3 - 2

NOTE:

Repeat the entire test for the second channel, if it is installed.

Test 7: High and Low Levels

The following tests are required:

1. High level from 50Ω into 50Ω
2. Low level from 50Ω into 50Ω
3. High level from $1K\Omega$ into 50Ω
4. Low level from $1K\Omega$ into 50Ω

Test Specifications

	Load Impedance 50Ω	
Source Impedance	50Ω	$1K\Omega$
High Level	-9.90 V to +10.0 V	-19.8 V to +20.0 V
Low Level	-10.0 V to +9.9 V	-20.0 V to +19.8 V
Amplitude	0.10 V _{pp} to 10.0 V _{pp}	0.20 V _{pp} to 20.0 V _{pp}
Level Resolution	10 mV	20 mV
Level Accuracy	$\pm 1\%$ of ampl ± 50 mV	$\pm 1\%$ of ampl ± 100 mV for amplitude ≤ 19 V


Equipment Needed

1. Digitizing Voltmeter (DVM)
2. 50Ω Feedthrough Termination, 0.1%, 10 W Adapter.
3. BNC to dual banana plug (1251-2277)
4. Cable, 50Ω , coaxial, BNC

Procedure

Connect Agilent 81110A to the DVM as shown:

Agilent 81110A/'11A Performance Test



Connecting the DVM for High and Low Levels Tests

Test 7.1: High Level, 50 Ohms into 50 Ohms

1. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
2. On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per 100.0ms Normal		ON	1	MODIFY
Delay	25.00ms High	+10.0V		
Width	50.00ms Low	+0.0mV		+10.0 V
LeadE	2.00ns 50W into 50.0W			
Traie	=LeadE			
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per 100.0ms Normal		OFF	2	 MODIFY
Delay	25.00ms High	+10.0V		+10.0 V
Width	50.00ms Low	+0.0mV		
LeadE	2.00ns 50W into 50.0W			
TraIE	=LeadE Separate Out2			
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- Configure *both* channels.
- Switch OFF the channel that is not being tested

If you then test the other channel:

- Switch ON the channel you are testing, and switch OFF the other channel.

3. Set the DVM Agilent 3458A to:

Function: DCV
 Trigger: TRIG EXT
 AD-Converter integration time NPLC: 0.1
 (Number of Power Line Cycles)

Agilent 81110A/'11A Performance Test

4. Check the Agilent 81110A high level at the following high level settings with the low level set to 0.0 V.

High Level	Acceptable Range	TR Entry
10.0 V	9.85 V to 10.15 V	7.1 - 1
5.0 V	4.90 V to 5.10 V	7.1 - 2
3.0 V	2.92 V to 3.08 V	7.1 - 3
1.0 V	0.94 V to 1.06 V	7.1 - 4
0.5 V	445 mV to 555 mV	7.1 - 5
0.1 V	49 mV to 151 mV	7.1 - 6

The low level may vary within $\pm 1\%$ of amplitude ± 50 mV

Test 7.2: Low Level, 50 Ohms into 50 Ohms

- Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
- On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per 100.0ms Normal ON 1				MODIFY
Delay	75.00ms	High	+0.0mV	-100 mV
Width	50.00ms	Low	-100mV	
LeadE	2.00ns	50W into	50.0W	
TraIE	=LeadE			
MODE/TRG OUTPUT 1 OUTPUT 2 PATTERN				

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per 100.0ms Normal				OFF	2	MODIFY
Delay	75.00ms	High	+0.0mV		- 100 mV	
Width	50.00ms	Low	-100mV			
LeadE	2.00ns	50W into 50.0W				
Traie	=LeadE Separate Out2					
MODE/TRG		OUTPUT 1		OUTPUT 2		PATTERN

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested

If you then test the other channel:

- c. Switch ON the channel you are testing, and switch OFF the other channel.

3. Check the Agilent 81110A low level at the following low level settings with the high level set to 0.0 V

Agilent 81110A/'11A Performance Test

Low Level	Acceptable Range	TR Entry
-0.1 V	-49 mV to -151 mV	7.2 - 1
-0.5 V	-445 mV to -555 mV	7.2 - 2
-1.0 V	-0.94 V to -1.06 V	7.2 - 3
-3.0 V	-2.92 V to -3.08 V	7.2 - 4
-5.0 V	-4.90 V to -5.10 V	7.2 - 5
-10.0 V	-9.85 V to -10.15 V	7.2 - 6

The high level 0.0 V may vary $\pm 1\%$ of amplitude ± 50 mV.

Test 7.3: High Level, 1K Ohms into 50 Ohms

1. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
2. On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per 100.0ms Normal		ON	1	MODIFY
Delay	25.00ms	High	+20.0V	+20.0 V
Width	50.00ms	Low	+0.0mV	
LeadE	2.00ns	1kW into	50.0W	
TraIE	=LeadE			
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per 100.0ms Normal				OFF	2	MODIFY
Delay	25.00ms	High	+20.0V		+20.0	
Width	50.00ms	Low	+0.0mV		V	
LeadE	2.00ns	1kW into	50.0W			
Traie	=LeadE	Separate	Out2			
MODE/TRG		OUTPUT 1		OUTPUT 2		PATTERN

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- Configure *both* channels.
- Switch OFF the channel that is not being tested

If you then test the other channel:

- Switch ON the channel you are testing, and switch OFF the other channel.

- Check the Agilent 81110A high level at the following high level settings with the low level set to 0.0 V.

Agilent 81110A/'11A Performance Test

High Level	Acceptable Range	TR Entry
19.0 V	18.71 V to 19.29 V	7.3 - 1
10.0 V	9.80 V to 10.20 V	7.3 - 2
5.0 V	4.85 V to 5.15 V	7.3 - 3
1.0 V	0.89 V to 1.11 V	7.3 - 4
0.2 V	98 mV to 302 mV	7.3 - 5

The low level 0.0 V may vary $\pm 1\%$ of amplitude ± 100 mV.

Test 7.4: Low Level, 1K Ohms into 50 Ohms

1. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
2. On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Per 100.0ms Normal		ON	1	☒ MODIFY -200 mV
Delay	75.00ms High	+0.0mV		
Width	50.00ms Low	-200mV		
LeadE	2.00ns 1kW into 50.0W			
TraIE	=LeadE			
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 1

Agilent 81110A/'11A Performance Test

Per 100.0ms Normal		OFF	2	 MODIFY
Delay	75.00ms High	+0.0mV		- 200 mV
Width	50.00ms Low	-200mV		
LeadE	2.00ns 1kW into 50.0W			
Traie	=LeadE Separate Out2			
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested

If you then test the other channel:

- c. Switch ON the channel you are testing, and switch OFF the other channel.

3. Check the Agilent 81110A low level at the following low level settings with the high level set to 0.0 V.

Agilent 81110A/'11A Performance Test

Low Level	Acceptable Range	TR Entry
-0.2 V	-98 mV to -302 mV	7.4 - 1
-1.0 V	-0.89 V to -1.11 V	7.4 - 2
-5.0 V	-4.85 V to -5.15 V	7.4 - 3
-10.0 V	-9.80 V to -10.20 V	7.4 - 4
-19.0 V	-18.71 V to -19.29 V	7.4 - 5

The high level 0.0 V may vary $\pm 1\%$ of amplitude ± 100 mV

NOTE:

Repeat the High and Low Level tests for the second channel, if it is installed.

Test 8: Transition Time

Test Specifications

Range	2.0 ns to 200 ms (measured between 10% and 90% of amplitude)
Minimum Transitions	≤ 2.0 ns (typical 1.4 ns for ECL levels measured between 20% and 80% of amplitude typical 5 ns for 1kOhm source impedance)
Accuracy	$\pm 10\%$ ± 200 ps
Linearity	typical $\pm 3\%$ for transitions > 100 ns

Equipment Needed

Digitizing Oscilloscope with Accessories
Cable, SMA

Procedure


Perform the tests as shown in the following sections:

Agilent 81110A/'11A Performance Test

Test 8.1a: Leading Edge Test

Minimum Leading Edge and Leading Edge ranges .

1. Connect Agilent 81110A to the Scope as shown:



Connecting Agilent 81110A to the Scope

NOTE:

When you connect the test equipment the first time, and whenever you change the setup during the following tests, use the torque wrench (8170-1582) to tighten and loosen the SMA connectors. This will ensure that the connectors are at the correct tightness and give the best signal transfer!

2. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
3. On the Agilent 81110A press MORE and set up [OUTPUT 1] and [OUTPUT 2] pages as shown in the following illustrations:

Agilent 81110A/'11A Performance Test

Per 500.0ms Normal				ON	1	 MODIFY
Delay	0ps	Offset	+0.0mV			
DtyCyc	50.00%	Amplit	5.00V			2.00 ns
LeadE	2.00ns	50W into	50.0W			
Traie	=LeadE					
MODE/TRG		OUTPUT 1		OUTPUT 2		PATTERN

Configuring Output Screen 1

Per 500.0ms Normal				OFF	2	 MODIFY
Delay	0ps	Offset	+0.0mV			
DtyCyc	50.00%	Amplit	5.00V			2.00 ns
LeadE	2.00ns	50W into	50.0W			
Traie	=LeadE	Separate	Out2			
MODE/TRG		OUTPUT 1		OUTPUT 2		PATTERN

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- Configure *both* channels.
- Switch OFF the channel that is not being tested

Agilent 81110A/'11A Performance Test

If you then test the other channel:

c. Switch ON the channel you are testing, and switch OFF the other channel.

-
4. Set the Digitizing Oscilloscope Agilent 54121T:
 - Press AUTOSCALE
 - Center one pulse on screen, e.g.:
 - TIME/DIV = 50 μ s/div, DELAY = 380 μ s,
 - Select the Display menu and set the Number of Averages to 32
 - Select the Channel menu and set the Attenuation factor to 10
 - Select the Delta V menu and turn the voltage markers On
 - Set the Preset Levels = 10-90% and press AUTO LEVEL SET
 - Select the Timebase menu and set TIME/DIV = 1 ns/div, DELAY = 19.5 ns
 - Select the Delta t menu and turn the markers On
 - Set START ON EDGE = POS1 and STOP ON EDGE = POS1
 5. Set period of Agilent 81110A to: Period = 1 μ s and change the Agilent 81110A Delay to center the leading edge of the first pulse on the screen
 6. After the averaging, while the oscilloscope is in the Delta t menu, Press the PRECISE EDGE FIND key

Agilent 81110A/'11A Performance Test

7. Check the Agilent 81110A rise times at the following leading edge settings:

Oscilloscope TIME/ DIV	Period	Leading Edge	Trailing Edge	Acceptable Range	TR Entry
2 ns/div	1 μ s	2.0 ns	2.0 ns	≤ 2 ns to 2.4 ns	8.1a - 1
5 ns/div	1 μ s	10 ns	10 ns	8.8 ns to 11.2 ns	8.1a - 2
10 ns/div	1 μ s	50 ns	50 ns	44.8 ns to 55.2 ns	8.1a - 3
100 ns/div	5 μ s	500 ns	500 ns	449.8 ns to 550.2 ns	8.1a - 4
1 μ s/div	50 μ s	5 μ s	5 μ s	4.4998 μ s to 5.5002 μ s	8.1a - 5
10 μ s/div	500 μ s	50 μ s	50 μ s	45 μ s to 55 μ s	8.1a - 6
100 μ s	5 ms	500 μ s	500 μ s	450 μ s to 550 μ s	8.1a - 7
10 ms/div	500 ms	50 ms	50 ms	45 ms to 55 ms	8.1a - 8

Programming down to 1.8 ns is allowed, to meet the minimum of ≤ 2 ns.

Use the Agilent 81110A Delay (scope delay) to center the leading edge on screen.

Test 8.1b: Trailing Edge Test

Minimum Trailing Edge and Trailing Edge range.

1. Connect Agilent 81110A to the Scope as shown in Test 8.1a Leading Edge Test.
2. Set up the Agilent 81110A as described in Test 8.1a Leading Edge Test.

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested

If you then test the other channel:

- c. Switch ON the channel you are testing, and switch OFF the other channel.

3. Set the digitizing oscilloscope Agilent 54121T:
 - Select the oscilloscopes Timebase menu and set TIME/DIV to 1 ns/div and DELAY to approximately 510 ns
 - Select the oscilloscopes Delta t menu and set START ON EDGE = NEG1 and STOP ON EDGE = NEG1
4. While the oscilloscope is in the Delta t menu, press the PRE-CISE EDGE FIND key
5. Check the Agilent 81110A output signal falls at the following trailing edge settings:

Agilent 81110A/'11A Performance Test

Oscilloscope TIME/DIV	Delay	Period	Trailing Edge	Leading Edge	Acceptable Range	TR Entry
2 ns/div	529 ns	1 μ s	2.0 ns	2 ns	≤ 2 ns to 2.4 ns	8.1b - 1
5 ns/div	529 ns	1 μ s	10 ns	5 ns	8.8 ns to 11.2 ns	8.1b - 2
10 ns/div	529 ns	1 μ s	50 ns	50 ns	44.8 ns to 55.2 ns	8.1b - 3
100 ns/div	25 μ s	5 μ s	500 ns	500 ns	449.8 ns to 550.2 ns	8.1b - 4
1 μ s/div	25 μ s	50 μ s	5 μ s	5 μ s	4.4998 μ s to 5.5002 μ s	8.1b - 5
10 μ s/div	250 μ s	500 μ s	50 μ s	50 μ s	45 μ s to 55 μ s	8.1b - 6
100 μ s/div	2.5 ms	5 ms	500 μ s	500 μ s	450 μ s to 550 μ s	8.1b - 7
10 ms/div	250 ms	500 ms	50 ms	50 ms	45 ms to 55 ms	8.1b - 8

Programming down to 1.8 ns is allowed, to meet the minimum of ≤ 2 ns.

Agilent 81110A/'11A Performance Test

Test 9: Pulse Aberration Test

The following tests are required:

Overshoot and Ringing
Preshoot

Test Specifications


Overshoot/Preshoot/Ringing
 $\pm 5\%$ of amplitude ± 20 mV

Equipment Needed

Digitizing Oscilloscope with Accessories

Procedure

6. Set up the Agilent 81110A as described in "Initial Setup of the Agilent 81110A"
1. Connect Agilent 81110A to the Scope as shown:



Connecting Agilent 81110A to the Scope

Agilent 81110A/'11A Performance Test

Per 500.0ms Normal			ON	1	 MODIFY
Delay	0ps	High	+5.0V	+5.00 V —	
DtyCyc	50.00%	Low	+0.0mV		
LeadE	2.00ns	50W into 50.0W			
TraIE	=LeadE				
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN	

Configuring Output Screen 1

Per 500.0ms Normal			OFF	2	 MODIFY
Delay	0ps	High	+5.0V	+5.00 V —	
DtyCyc	50.00%	Low	+0.0mV		
LeadE	2.00ns	50W into 50.0W			
TraIE	=LeadE				
MODE/TRG		OUTPUT 1	OUTPUT 2	PATTERN	

Configuring Output Screen 2

NOTE:

When you are testing instruments with 2 output channels it is necessary to:

- a. Configure *both* channels.
- b. Switch OFF the channel that is not being tested

If you then test the other channel:

- c. Switch ON the channel you are testing, and switch OFF the other channel.
-

Overshoot and Ringing

2. Set the digitizing oscilloscope Agilent 54121T:
 - Press AUTOSCALE
 - Select the Display menu and set the Number of Averages to 32
 - Select the Channel menu and set the Attenuation factor to 10
 - Center one pulse horizontally and vertically on screen
 - (e.g. TIME/DIV = 50 μ s/div, DELAY = 250 μ s)
 - Select the delta V menu and turn the voltage markers On
 - Set the VARIABLE LEVELS = 95% - 105% and press AUTO LEVEL SET
 - Select the channel menu and center vertically the top pulse (offset = 5 V)
 - Set the VOLTS/DIV = 200 mV/div
 - Select the Timebase menu and set TIME/DIV = 5 ns/div, DELAY = 16 ns (>> 500 ns)
3. Set the Agilent 81110A to period = 500 ns

Agilent 81110A/'11A Performance Test

4. Check that Overshoot and Ringing are within the $\pm 5\%$ of amplitude ± 20 mV window
5. Enter the result in the Test Report as TR entry 9 - 1

NOTE:

Take the oscilloscope's trace flatness error (GaAs input circuit) into account.

Preshoot

6. Set Agilent 81110A to:
 - Period = 500 μ s
 - High Level = 5 V
 - Low Level = 0 V
 - Delay = 10 ns
7. Set the digitizing oscilloscope, Agilent 54121T:
 - Press AUTOSCALE
 - Select the Display menu and set the Number of Averages to 32
 - Select the Channel menu and set the Attenuation factor to 10
 - Center one pulse horizontally and vertically on screen
 - (e.g. TIME/DIV = 50 μ s/div, DELAY = 265 μ s)
 - Select the delta V menu and turn the voltage markers On
 - Set the VARIABLE LEVELS = -5% to +5% and press AUTO LEVEL SET
 - Select the channel menu and center vertically the bottom of the pulse (offset = 0 V)
 - Set the VOLTS/DIV = 200 mV/div

Agilent 81110A/'11A Performance Test

- Select the Timebase menu and set TIME/DIV = 5 ns/div, DELAY = 16 ns
- 8. Set Agilent 81110A to period = 500 ns
- 9. Check that Preshoot is within the $\pm 5\%$ of amplitude ± 20 mV window.
- 10. Enter the result in the Test Report as TR entry 9 - 3

Agilent 81110A/'11A Performance Test

Agilent 81110A/'11A Performance Test

Agilent 81110A/'11A Performance Test Records

Test Facility:

_____	Report No. _____
_____	Date _____
_____	Customer _____
_____	Tested By _____

×

Model Agilent 81110A/'11A 165 MHz Pulse Generator

Serial No. _____

Options	_____	Ambient temperature	_____ °C
	_____	Relative humidity	_____ %

Firmware Rev. _____ Line frequency _____ Hz

Special Notes:

Agilent 81110A/'11A Performance Test

Test Equipment Used		Model No.	Trace No.	Cal. Due
Description	Date			
1. Oscilloscope		Agilent 54121T	_____	_____
2. Counter		Agilent 5334B	_____	_____
3. Digital Voltmeter		Agilent 3458A	_____	_____
4. Pulse Generator		Agilent 8110A	_____	_____
5. Delay Line		Agilent 54008A	_____	_____
6.	_____			
7.	_____			
8.	_____			
9.	_____			
10.	_____			
11.	_____			
12.	_____			
13.	_____			
14.	_____			

Agilent 81110A/'11A Performance Test

Test Results for Agilent 81110A Mainframe

Serial No. _____ Ambient temperature _____ °C

Customer _____ Relative humidity _____ %

CSO# _____ Line frequency _____ Hz

Tested by _____ Date _____

Comments

_____**Internal Oscillator Period**

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
1-1	6.06ns	5.878 ns	_____	6.242 ns	_____	_____
1-2	9.99ns	9.690 ns	_____	10.290 ns	_____	_____
1-3	10.0ns	9.7 ns	_____	10.3 ns	_____	_____
1-4	50.0ns	48.5 ns	_____	51.5 ns	_____	_____
1-5	99.9ns	96.903 ns	_____	102.897 ns	_____	_____

Agilent 81110A/'11A Performance Test

Counter Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
1-6	100 ns	97.0ns	_____	103.0 ns	_____	_____
1-7	500 ns	485.0 ns	_____	515.0 ns	_____	_____
1-8	1 μ s	970.0 ns	_____	1030.0 ns	_____	_____
1-9	5 00 μ s	485 μ s	_____	5 15 μ s	_____	_____
1-10	500 ms	485 ms	_____	515 ms	_____	_____

Agilent 81110A/'11A Performance Test

PLL Period
(Results measured as frequency by counter)

Counter Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
2-1	6.061 ns	164.9835MHz	_____	165.0165MHz	__	__
2-2	10.00 ns	99.990MHz	_____	100.010 MHz	__	__
2-3	50.00 ns	19.9980MHz	_____	20.0020MHz	__	__
2-4	100 ns	9.9990MHz	_____	10.0010MHz	__	__
2-5	500 ns	1.9998MHz	_____	2.0002MHz	__	__
2-6	1 μ s	999.9 kHz	_____	1.0001 MHz	__	__
2-7	50 μ s	19.998 kHz	_____	20.002 kHz	__	__
2-8	5 m	199.98 Hz	_____	200.02 Hz	__	__
2-9	500 m	1.9998 Hz	_____	2.0002 Hz	__	__
2-10	5 s	0.19998 Hz	_____	0.20002 Hz	__	__

Agilent 81110A/'11A Performance Test

Period Jitter

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
6.1a-1	50 ns		_____	20 ps	_____	_____
6.1a-2	500 ns		_____	65 ps	_____	_____
6.1b-1	20 ns		_____	15.2 ps	_____	_____

Agilent 81110A/'11A Performance Test

Test Results for Agilent 8111A Output Channel _____

Serial No. _____

Width

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
3-1	3.030 ns	2.689 ns	_____	3.371 ns	_____	_____
3-2	6.06ns	5.528 ns	_____	6.492 ns	_____	_____
3-3	10.0 ns	9.450ns	_____	10.550 ns	_____	_____
3-4	50.0 ns	48.25 ns	_____	51.75 ns	_____	_____
3-5	100 ns	96.75 ns	_____	103.25 ns	_____	_____
3-6	500 ns	484.75 ns	_____	515.25 ns	_____	_____
3-7	50 μ s	48.5 μ s	_____	51.5 μ s	_____	_____
3-8	5 ms	4.85 ms	_____	5.15 ms	_____	_____
3-9	500 ms	485 ms	_____	515 ms	_____	_____

Agilent 81110A/'11A Performance Test

Width Jitter

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
6.2-1	50 ns	_____	_____	20 ps	_____	_____
6.2-2	500 ns	_____	_____	65 ps	_____	_____

Agilent 81110A/'11A Performance Test

Delay

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
4-1	0.00 ns		_____	Fixed Delay	_____	_____
4-2	5.00 ns	4.35 ns	_____	5.65 ns	_____	_____
4-3	10 ns	9.20 ns	_____	10.80 ns	_____	_____
4-4	50.0 ns	48.0 ns	_____	52.0 ns	_____	_____
4-5	100 ns	96.5 ns	_____	103.5 ns	_____	_____
4-6	500 ns	484.5 ns	_____	515.5 ns	_____	_____
4-7	50 μ s	48.5 μ s	_____	51.5 μ s	_____	_____
4-8	5 ms	4.85 ms	_____	5.15 ms	_____	_____
4-9	500 ms	485 ms	_____	515 ms	_____	_____

Agilent 81110A/'11A Performance Test

Delay Jitter

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
6.3-1	50 ns		_____	20 ps	_____	_____
6.3-2	500 ns		_____	65 ps	_____	_____

Agilent 81110A/'11A Performance Test

Double Pulse Delay

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
5-1	6.06 ns	5.628 ns	_____	6.392 ns	_____	_____
5-2	10.0 ns	9.550 ns	_____	10.45 ns	_____	_____
5-3	50.0ns	48.35 ns	_____	51.65 ns	_____	_____
5-4	100ns	96.85 ns	_____	103.15 ns	_____	_____

Counter Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
5-5	500 ms	485 ms	_____	515 ms	_____	_____
5-6	1 s	970.0 ms	_____	1030.0 ms	_____	_____

Agilent 81110A/'11A Performance Test

High Level 50Ω-50Ω

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
7.1-1	10.0 V	9.85 V	_____	10.15 V	_____	_____
7.1-2	5.0 V	4.90 V	_____	5.10 V	_____	_____
7.1-3	3.0V	2.92 V	_____	3.08 V	_____	_____
7.1-4	1.0 V	0.94 V	_____	1.06 V	_____	_____
7.1-5	0.5 V	445 mV	_____	555 mV	_____	_____
7.1-6	0.1 V	49 mV	_____	151 mV	_____	_____

High Level 1KΩ-50Ω

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
7.3-1	19.0 V	18.71V	_____	19.29 V	_____	_____
7.3-2	10.0 V	9.80 V	_____	10.20 V	_____	_____
7.3-3	5.0 V	4.85 V	_____	5.15 V	_____	_____
7.3-4	1.0 V	0.89 V	_____	1.11V	_____	_____
7.3-5	0.2 V	98 mV	_____	302mV	_____	_____

Agilent 81110A/'11A Performance Test

Low Level 50Ω-50Ω

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
7.2-1	-0.1 V	-49 mV	_____	-151 mV	_____	_____
7.2-2	-0.5 V	-445 mV	_____	-555 mV	_____	_____
7.2-3	-1.0 V	-0.94 V	_____	-1.06 V	_____	_____
7.2-4	-3.0V	-2.92 V	_____	-3.08 V	_____	_____
7.2-5	-5.0V	-4.90 V	_____	-5.10 V	_____	_____
7.2-6	-10.0V	-9.85 V	_____	-10.15 V	_____	_____

Low Level 1KΩ-50Ω

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
7.4-1	-0.2V	-98 mV	_____	-302 mV	_____	_____
7.4-2	-1.0V	-0.89 V	_____	-1.11 V	_____	_____
7.4-3	-5.0V	-4.85V	_____	-5.15 V	_____	_____
7.4-4	-10.0V	-9.80 V	_____	-10.20 V	_____	_____
7.4-5	-19.0V	-18.71 V	_____	-19.29 V	_____	_____

Agilent 81110A/'11A Performance Test

Leading Edge

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
8.1a-1	2.0 ns	≤ 2 ns	_____	2.4 ns	_____	_____
8.1a-2	10 ns	8.8 ns	_____	11.2 ns	_____	_____
8.1a-3	50 ns	44.8 ns	_____	55.2 ns	_____	_____
8.1a-4	500 ns	449.8 ns	_____	550.2 ns	_____	_____
8.1a-5	5 μ s	4.4998 μ s	_____	5.5002 μ s	_____	_____
8.1a-6	50 μ s	45 μ s	_____	55 μ s	_____	_____
8.1a-7	500 μ s	450 μ s	_____	550 μ s	_____	_____
8.1a-8	50 ms	45 ms	_____	55 ms	_____	_____

Agilent 81110A/'11A Performance Test

Trailing Edge

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
8.1b-1	2.0 ns	≤ 2 ns	_____	2.4 ns	_____	_____
8.1b-2	10 ns	8.8 ns	_____	11.2 ns	_____	_____
8.1b-3	50 ns	44.8 ns	_____	55.2ns	_____	_____
8.1b-4	500 ns	449.8 n	_____	550.2 ns	_____	_____
8.1b-5	5 μ s	4.4998 μ s	_____	5.5002 μ s	_____	_____
8.1b-6	50 μ s	45 μ s	_____	55 μ s	_____	_____
8.1b-7	500 μ s	450 μ s	_____	550 μ s	_____	_____
8.1b-8	50 ms	45 ms	_____	55 ms	_____	_____

Agilent 81110A/'11A Performance Test

Overshoot and Ringing

Scope Uncertainty factor _____

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
9-1	5V		_____	$\pm 5\%$ of ampl. $\pm 20\text{mV}$	_____	_____
9-2	500 mV		_____	$\pm 5\%$ of ampl. $\pm 20\text{mV}$	_____	_____

Preshoot

TR Entry	Test	Limit Min	Actual Result	Limit Max	Pass	Fail
9-3	0 V		_____	$\pm 5\%$ of ampl. $\pm 20\text{mV}$	_____	_____

Publication Number: 5988-4853EN



Agilent Technologies