

HP E5100A/B Network Analyzer Programming Manual

SERIAL NUMBERS

This manual applies directly to instruments with serial number prefix "JP2KC" and above, and whose firmware is version 2.01 and above. For additional important information about serial numbers, read "Serial Number" in Appendix A of this Manual.



HP Part No. E5100-90047
Printed in JAPAN May 1997

Sixth Edition

Notice

The information contained in this document is subject to change without notice.

This document contains proprietary information that is protected by copyright. All rights are reserved. No part of this document may be photocopied, reproduced, or translated to another language without the prior written consent of the Hewlett-Packard Company.

Hewlett-Packard Japan, LTD.
Kobe Instrument Division
1-3-2, Murotani, Nishi-ku, Kobe-shi,
Hyogo, 651-22 Japan

The customer shall have the personal, non-transferable rights to use, copy, or modify SAMPLE PROGRAMS in this manual for the Customer's internal operations. The customer shall use the SAMPLE PROGRAMS solely and exclusively for their own purpose and shall not license, lease, market, or distribute the SAMPLE PROGRAMS or modification of any part thereof.

HP shall not be liable for the quality, performance, or behavior of the SAMPLE PROGRAMS. HP especially disclaims that the operation of the SAMPLE PROGRAMS shall be uninterrupted or error free. The SAMPLE PROGRAMS are provided AS IS.

HP DISCLAIMS THE IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE.

HP shall not be liable for any infringement of any patent, trademark, copyright, or other proprietary rights by the SAMPLE PROGRAMS or their use. HP does not warrant that the SAMPLE PROGRAMS are free from infringements of such rights of third parties. However, HP will not knowingly infringe or deliver software that infringes the patent, trademark, copyright, or other proprietary right of a third party.

MS-DOS[®] is a U.S. registered trademark of Microsoft Corporation.

© Copyright 1995, 1996, 1997 Hewlett-Packard Japan, LTD.

Manual Printing History

The manual printing date and part number indicate its current edition. The printing date changes when a new edition is printed. (Minor corrections and updates that are incorporated at reprint do not cause the date to change.) The manual part number changes when extensive technical changes are incorporated.

July 1995	First Edition
October 1995	Second Edition
November 1995	Third Edition
December 1995	Forth Edition
September 1996	Fifth Edition
May 1997	Sixth Edition

Typeface Conventions

Bold Boldface type is used when a term is defined. For example: **icons** are symbols.

Italics Italic type is used for emphasis and for titles of manuals and other publications.

Italic type is also used for keyboard entries when a name or a variable must be typed in place of the words in italics. For example: *copy filename* means to type the word *copy*, to type a space, and then to type the name of a file such as *file1*.

Computer Computer font is used for on-screen prompts and messages.

HARDKEYS Labeled keys on the instrument front panel are enclosed in **□**.

SOFTKEYS Softkeys located to the right of the LCD display are enclosed in **▣**.

Contents

1. Introduction	
How to Use This Manual	1-1
Target Reader	1-1
What's in This Manual?	1-1
How to Use the Program Modules	1-2
Building a Working Program Using Program Modules	1-2
Initializing Module	1-2
Example	1-3
HP-IB Overview	1-4
Controller	1-4
Device Selector	1-5
Preparation for Operation	1-6
Using Instrument BASIC as the Controller (except for Option UKR)	1-6
1. Connecting the HP-IB Cables For Instrument BASIC	1-6
2. Setting the HP-IB Address For Instrument BASIC	1-6
3. Setting Up Instrument BASIC	1-6
Using an External Controller	1-7
1. Connecting External HP-IB Cables	1-7
2. Setting the HP-IB Address	1-7
3. Preparing For HP BASIC Operation	1-7
Sample Program Disk	1-8
Loading a Program from Disk	1-8
Reading the Sample Program Disk on a PC	1-8
Related Documentation Information	1-9
2. Setup and Measurement Program	
Overview of HP-IB Control	2-1
Sending HP-IB Commands	2-1
Sending a Query and Reading the Response	2-2
Automating a Measurement Procedure	2-3
1. Setting the Active Channel	2-3
2. Setting the Measurement and Format	2-3
3. Sweep Setup	2-4
4. Calibration	2-4
Calibration Process	2-5
5. Device Connection	2-5
6. Triggering a Measurement	2-6
When You are Using an External Controller	2-6
7. Data Processing and Transfer	2-6
8. Exiting the Program	2-6
Sample Program -1: Basic Measurement	2-7
Column Parameters and Variables	2-8

3. Data Processing and Transfer	
Introduction	3-1
Data Arrays	3-1
Raw Data Array	3-4
Data Array	3-4
Unformatted Data Array	3-4
Formatted Data Array	3-5
Calibration Coefficient Array	3-5
Accessing Arrays	3-5
Arrays for Memory Trace	3-6
Accessing Memory Array	3-6
Accessing Trace Array	3-6
Data Transfer Methods	3-7
ASCII Transfer	3-7
Binary Transfer	3-8
Data Header	3-9
Getting Data from the Analyzer	3-10
Sample Program -2: Binary Data Transfer	3-11
4. Using the I/O Port	
Introduction	4-1
Output Port (Pin 5 to 28)	4-1
Data Output	4-1
Input Port (Pin 21 to 28)	4-2
Data Input	4-2
INPUT1 Input (Pin 2), OUTPUT1 Output (Pin 3), OUTPUT2 Output (Pin 4)	4-3
Examples of Data I/O	4-4
The Parallel I/O Mode A (Option 005 Only)	4-6
Data Output (4 Bit)	4-6
Data Input (8 Bit)	4-6
5. Using Status Reporting Function	
General Status Register Model	5-1
Event Register	5-2
Enable Register	5-2
Status Byte Register	5-2
Transition Filter and Condition Register	5-3
Status Register Structure	5-4
OSPT, OSNT	5-7
OSPT (Operational Status Positive Transition Filter)	5-7
OSNT (Operational Status Negative Transition Filter)	5-8
How to Use the Status Registers in a Program	5-9
Reading an Event Register Directly	5-9
SRQ and Interrupt	5-10
6. Programming Miscellaneous	
Introduction	6-1
Using an External Controller and the HP E5100A/B Together	6-1
Locking Out Local Operation (LOCAL LOCKOUT)	6-1
How To Pass Control (PASS CONTROL)	6-2
Controlling Instrument BASIC From an External Controller	6-3
Referring to and Transferring an Array	6-3
Referring to a Numeric Variable	6-3
Referring to a String Character Variable	6-3
Transferring a Numeric Variable	6-4

Transferring a String Character Variable	6-4
Referring to and Transferring an Array	6-4
Remote Processing a BASIC Command (PROG:EXEC)	6-4
Controlling the Process Status	6-5
Controlling the Process Status of Instrument BASIC	6-5
Checking the Process Status of an Instrument BASIC Program	6-5
Transferring a Program	6-6
Transferring from an External Controller to Instrument BASIC	6-6
Transferring from Instrument BASIC to an External Controller	6-7
Sample Program -3: Controlling Instrument BASIC from an External Controller	6-8
The Program for Instrument BASIC	6-9
The Program for the External Controller	6-10
Programming Techniques	6-12
Using a Disk	6-12
Using a Softkey Label (ON KEY LABEL)	6-13
Displaying the Softkey Label	6-13
Measuring the Processing Time (TIMEDATE)	6-14
Checking HP-IB Errors (OUTPERRO?)	6-14
Tips for Increasing Speed	6-14
Sample Program -4: Reading and Running a Program by a Softkey Operation	6-16
Using the Parallel Processing	6-18
Programming for Parallel Processing	6-18
Sample Program -4a :Parallel Processing	6-19
Notes for Parallel Processing	6-21
7. Application Sample Programs	
Sample Program -5: Setting Up and Performing the List Sweep	7-2
Saving and Resetting Calibration Data	7-4
Sample Program -6: Saving Calibration Data in an External Controller	7-4
Sample Program -7: Setting Up Calibration Data from an External Controller	7-5
Sample Program -8: Analyzing a Ceramic Resonator	7-8
Sample Program -9: Analyzing a Crystal Resonator	7-10
8. Command Reference	
Common Command	8-1
*CLS	8-1
*ESE <i>value</i>	8-1
*ESR?	8-1
*IDN?	8-1
*OPC	8-2
*PCB <i>value</i>	8-2
*RST	8-2
*SRE <i>value</i>	8-2
*STB?	8-3
*TRG	8-3
*TST?	8-3
*WAI	8-3
Command Reference	8-4
ADTOTRAC	8-4
ANAMODE {GAINP ZREFL ZTRAN}	8-4
ANAOCH{1-4}	8-4
ANAODATA	8-4
ANAOMEMO	8-4
ANARANG <i>value1</i> [<i>suffix</i>], <i>value2</i> [<i>suffix</i>]	8-5
ANARANGP□< <i>value1</i> >,< <i>value2</i> >	8-5

ANARFULL	8-5
AR	8-5
ATRC {1 2}	8-5
ATTI{A B C R } {0 25}	8-6
ATTIAUTO{OFF ON 0 1}	8-6
ATTIBAUTO {OFF ON 0 1}	8-6
ATTICAUTO {OFF ON 0 1}	8-7
ATTIRAUTO {OFF ON 0 1}	8-7
ATTW <i>value</i>	8-7
AUTO	8-7
BASL	8-8
BEEPDONE {ON OFF 0 1}	8-8
BEEPFAIL {ON OFF 1 0}	8-8
BEEPWARN {ON OFF}	8-8
BINSIZE{<value>}	8-9
BOTV{<value>}	8-9
CALCOPY{<value1>, <value2>}	8-9
CALI <i>parameter</i>	8-10
CALK{O S L }{LS RS CP} <value>	8-10
CENS <i>value1</i> [<i>suffix</i>], <i>value2</i> [<i>suffix</i>]	8-10
CENT <i>value</i> [<i>suffix</i>]	8-11
CHAD <i>string</i>	8-11
CHAN {1 2 3 4}	8-11
CIN	8-11
CIVAL{<value>}	8-11
CLASS11 {A B C}	8-12
CLEL	8-12
CLEM{1-4}	8-12
CLEMNU3	8-13
CLES	8-13
CONT	8-13
CONV {OFF ZTRA YTRA ZREF YREF}	8-13
CORR {ON OFF}	8-14
COUC {ON OFF}	8-14
COUT	8-14
CRED <i>string</i>	8-14
CURD?	8-15
CURMPOIN?	8-15
CWFREQ <i>value</i> [<i>suffix</i>]	8-15
DATI	8-15
DAYMYEAR	8-16
DELA	8-16
DELO	8-16
DELR{1-4}	8-16
DELRFIXM	8-16
DIN	8-17
DISA <i>parameter</i>	8-17
DISG {ON OFF}	8-17
DISP <i>parameter</i>	8-17
DIST {ON OFF}	8-18
DONE	8-18
DOUT	8-18
DSKEY	8-19
DUAC {ON OFF}	8-19
EDITDONE	8-19

EDITLIS1	8-19
EDITLIS2	8-20
EDITLIST	8-20
ELED <i>value</i> [s]	8-20
ENKEY	8-20
EQUC0? <i>value</i>	8-20
EQUCPARA5?	8-21
EQUCPARA?	8-21
EQUCPARS4?	8-22
EQUCPARS?	8-22
EQU <i>value</i>	8-23
ESB?	8-23
ESNB <i>value</i>	8-23
EXET <i>value</i>	8-23
EXPCL{ON OFF}	8-24
EXPZP {ON OFF 1 0}	8-24
EXTRLOCK?	8-24
EXTT <i>parameter</i>	8-24
FILC <i>string1,string2,string3,string4</i>	8-25
FMT <i>parameter</i>	8-25
FORM2	8-26
FORM3	8-26
FORM4	8-27
FORM5	8-27
GRAPFORM{PCL TIFF}	8-27
GRODAPER <i>value</i> [pct]	8-27
HOLD	8-27
IDN?	8-28
IFBW <i>value</i>	8-28
IFBWAUTO	8-28
INID	8-29
INP8IO	8-29
INPT?	8-29
INPUALC {01-03} <i>value</i>	8-29
INPUTDATA < <i>value1</i> >,< <i>value2</i> >, . . . ,< <i>valuen</i> >	8-30
INPUFORM <i>value</i>	8-30
INPUIFORM < <i>value1</i> >,< <i>value2</i> >, . . . ,< <i>valuen</i> >	8-30
INPUMEMO < <i>value1</i> >,< <i>value2</i> >, . . . ,< <i>valuen</i> >	8-30
INPURAW□< <i>value (1)</i> >,< <i>value (2)</i> >, . . . ,< <i>value (n)</i> >	8-31
INPUFORM < <i>value1</i> >,< <i>value2</i> >, . . . ,< <i>valuen</i> >	8-31
INPURMEM < <i>value (1)</i> >,< <i>value (2)</i> >, . . . ,< <i>value (n)</i> >	8-31
INPUTSTIM□< <i>value1</i> >,< <i>value2</i> >, . . . ,< <i>valuen</i> >	8-31
INPUTRAC□< <i>value1</i> >,< <i>value2</i> >,< <i>value3</i> >	8-32
INPUTRACB□< <i>value1</i> >,< <i>value2</i> >,< <i>value3</i> >	8-32
IOPO?	8-33
LIMILINE {ON OFF}	8-33
LIMITEST {ON OFF}	8-33
LINFREQ	8-34
LINM	8-34
LISDFBASE	8-34
LISDOBASE	8-34
LISFREQ	8-34
LISSLIS1	8-34
LISSLIS2	8-34
LMAX? <i>value</i>	8-35

LMIN? <i>value</i>	8-35
LOGM	8-35
LOGMD	8-35
LOGMP	8-35
LOWELIMI < <i>value1</i> >< <i>value2</i> >< <i>value3</i> > < <i>valuen</i> >	8-35
MARD {ON OFF}	8-36
MARK {1-4}	8-36
MARKBUCK <i>value</i>	8-37
MARKCENT	8-37
MARKCONT	8-37
MARKCOUP	8-37
MARKDISC	8-37
MARKFSTI <i>value</i> [<i>suffix</i>]	8-38
MARKFVAL <i>value</i> [<i>suffix</i>]	8-38
MARKL {ON OFF}	8-38
MARKODATA	8-39
MARKOFF	8-39
MARKMEMO	8-39
MARKREF	8-39
MARKSPAN	8-39
MARKSTAR	8-40
MARKSTOP	8-40
MARKTIME <input type="checkbox"/> {OFF ON}	8-40
MARKUNCO	8-40
MARKZERO	8-41
MAXPOIN?	8-41
MAXPORT?	8-41
MEAS1PT? < <i>value</i> >	8-41
MEAS <i>parameter</i>	8-42
MEASA	8-42
MEASR	8-42
MEASSTAT {ON OFF}	8-42
MENU3 <input type="checkbox"/> < <i>value</i> >, < <i>string1</i> >, < <i>string2</i> >, < <i>string3</i> >	8-43
MOHMSW {A B}, {ON OFF}	8-43
MONDYEAR	8-43
MOVADARY	8-43
MULC {OFF ON 0 1}	8-44
NEGL	8-44
NEXPK?	8-44
NUMC < <i>value</i> >	8-44
NUMG <i>value</i>	8-45
NUMLMAX?	8-45
NUMLMIN?	8-45
NUMLMINMAX?	8-45
OSE <i>value</i>	8-45
OSER?	8-46
OSNT	8-46
OSPT	8-46
OSR?	8-46
OUT1ENVH	8-47
OUT1ENVL	8-47
OUT1H	8-47
OUT1L	8-47
OUT2ENVH	8-47
OUT2ENVL	8-47

OUT2H	8-48
OUT2L	8-48
OUT8IO <i>value</i>	8-48
OUTAIO <i>value</i>	8-48
OUTBIO <i>value</i>	8-48
OUTCIO <i>value</i>	8-49
OUTDIO <i>value</i>	8-49
OUTEIO <i>value</i>	8-49
OUTFIO <i>value</i>	8-49
OUTGIO <i>value</i>	8-49
OUTHIO <i>value</i>	8-50
OUTPCALC{01-03}?	8-50
OUTPCERR?	8-50
OUTPCF2?□ <i><value1></i> , <i><value2></i> , <i><value3></i> , <i><value4></i> , <i><value5></i> ... , <i><valuen + 4></i>	8-51
OUTPCFIL? <i>value1,value2,value3,value4,value5,value6</i>	8-52
OUTPDATA?	8-52
OUTPDATAP <i><value></i>	8-53
OUTPDATAT?	8-53
OUTPDATTP? <i>value</i>	8-53
OUTPERRO?	8-53
OUTPFILT? <i>value</i>	8-54
OUTPFORM?	8-54
OUTPFORMP? <i>value</i>	8-54
OUTPIFORM?	8-55
OUTPINP8IO?	8-55
OUTPINPCIO?	8-55
OUTPINPDIO?	8-55
OUTPINPEIO?	8-56
OUTPIRFORM?	8-56
OUTPIRTMEM?	8-56
OUTPMARK?	8-56
OUTPMAX?	8-57
OUTPMEAN?	8-57
OUTPMEMO?	8-57
OUTPMEMOT?	8-58
OUTPMEMTP? <i>value</i>	8-58
OUTPMIN?	8-58
OUTPMINMAX?	8-58
OUTPMSTA?	8-59
OUTPMWID?	8-59
OUTPMWIL?	8-60
OUTPMWLF?	8-60
OUTPRAW?	8-60
OUTPRESF? <i>value1,value2</i>	8-61
OUTPRESO?	8-61
OUTPRESR?	8-61
OUTPRFORM?	8-62
OUTPRFORMP? <i><value></i>	8-62
OUTPRTMEM?	8-62
OUTPRTMEMP? <i><value></i>	8-63
OUTPSTIM?	8-63
OUTPTITL?	8-63
OUTPTRAC?□ <i><value></i>	8-63
OUTPTRACB?□ <i><value1></i> , <i><value2></i>	8-64

OUTPXF2? □ <value1>, <value2>, <value3>, ... , <valuen + 3>	8-64
OUTPXFIL? value1, value2, value3, value4, value5	8-65
PARS {OFF ON 0 1}	8-66
PEAK?	8-67
PEAKLIST? value	8-67
PHAO value [deg]	8-67
PHAS	8-67
PICIRC □ {1 0 ON OFF}	8-68
POIN value	8-68
POLE? value	8-68
POSL	8-69
POWE value [dBm]	8-69
POWS	8-69
POWU □ {DBM WATT AMP}	8-69
PRES	8-69
PRINALL	8-70
PRIR □ <string>	8-70
PTABORT □ {ON OFF}	8-70
PTFOVHD □ <value>	8-70
PTFR □ <value>	8-71
PTFRSR □ <value>	8-71
PTPARA □ <value>	8-71
PTRACK □ {OFF ON}	8-71
PTREPN □ <value>	8-72
PTSTAT?	8-72
PTTRGLMT □ <value>	8-72
PTTRGPHS □ <value>	8-73
PURG string	8-73
RAID	8-73
RAISOL	8-73
RAIRESP	8-73
RECD string	8-74
REFP value	8-74
REFV value [suffix]	8-74
RESAVD string	8-74
RESPDONE	8-75
REST	8-75
RFOUTSW {1 2}	8-75
RPLENV?	8-75
RPLHEI?	8-76
RPLLHEI?	8-76
RPLMEA?	8-76
RPLMM?	8-76
RPLPP?	8-77
RPLPPS?	8-77
RPLRHEI?	8-77
RPLVAL?	8-77
RPOS	8-78
SADD	8-78
SAVI	8-78
SAVCA {OFF ON 0 1}	8-78
SAVDA {ON OFF 1 0}	8-78
SAVDALL string	8-79
SAVDASC string	8-79
SAVDDAT string	8-79

SAVDGRAPL<string>	8-80
SAVDMNU3L<string>	8-80
SAVDSTA string	8-80
SAVFA {OFF ON 0 1}	8-80
SAVMA {ON OFF 1 0}	8-81
SAVRA {OFF ON 0 1}	8-81
SAVTA {OFF ON 0 1}	8-81
SAVTMA {OFF ON 0 1}	8-82
SCAFDATA	8-82
SCAFMEMO	8-82
SCAL value	8-82
SCAY {1 0}	8-83
SDEL	8-83
SDON	8-83
SEAL	8-83
SEALMAX	8-83
SEALMIN	8-84
SEAM {OFF MAX MIN TARG MEAN LMAX LMIN PPEAK}	8-84
SEAMAX	8-84
SEAMEAN	8-84
SEAMIN	8-85
SEAOFF	8-85
SEAPPEAK	8-85
SEAR	8-85
SEARSTOR	8-85
SEATARG value	8-86
SEDI value	8-86
SEET {ON OFF}	8-86
SETCDATE year,month,day	8-86
SETCTIME value 1,value 2,value 3	8-87
SETZ value [ohm]	8-87
SING?	8-88
SINSPEAK {ON OFF}	8-88
SMOO {ON OFF}	8-88
SMOOPER value [pct]	8-88
SPAN value [HZ KHZ MAHZ GHZ DBM]	8-89
SPLD {ON OFF 0 1}	8-89
SRCHFR?L{1 2 3 4},<value>	8-89
STANC	8-90
STAR value [suffix]	8-90
STAS value1[HZ KHZ MAHZ GHZ DBM], value2 [HZ KHZ MAHZ GHZ DBM]	8-90
STAW value	8-90
STIDROUT{1-16} value [suffix]	8-91
STIMROUT{1-16} value [suffix]	8-91
STOD{DISK MEMO}	8-91
STOMDISK	8-91
STOP value [suffix]	8-92
STPSIZE value [HZ KHZ MAHZ]	8-92
STR {ON OFF}	8-92
SWED {DOWN UP}	8-93
SWET value [s]	8-93
SWETAUTO	8-93
SWPT {LINF LOGF LIST POWE RAMPF}	8-93
TARL? value	8-94
TARR? value	8-94

THRR <i>value</i>	8-94
TIMOL{ON OFF 0 1}	8-95
TITL <i>string</i>	8-95
TOPVL< <i>value</i> >	8-95
TOTIMEL< <i>value</i> >	8-95
TRABGE	8-96
TRABLE	8-96
TRACK {OFF ON 0 1}	8-96
TRAFDATA	8-97
TRAFMEMO	8-97
TRAPL{OFF ON 0 1}	8-97
TRARL< <i>value1</i> >,< <i>value2</i> >	8-97
TRIGMEAS	8-98
TRIM {CONT HOLD SING}	8-98
UPDD {ON OFF}	8-98
UPPELIMI < <i>value1</i> >< <i>value2</i> >< <i>value3</i> > < <i>valuen</i> >	8-99
WIDT {ON OFF 0 1}	8-99
WIDV <i>value</i>	8-99
WRIT16	8-100
WRIT24	8-100
SCPI Command (PROGRAM sub-system command)	8-101
:PROGRAM:CATalog?	8-101
:PROGRAM[:SElected]:DEFineL< <i>block</i> >	8-101
:PROGRAM[:SElected]:DELeTe[:SElected]	8-102
:PROGRAM[:SElected]:DELeTe:ALL	8-103
:PROGRAM[:SElected]:EXECuteL< <i>string</i> >	8-103
:PROGRAM[:SElected]:NUMBerL< <i>string</i> >,< <i>numeric (1)</i> >[,< <i>numeric (2)</i> >[, . . .	
[,< <i>numeric (n)</i> >]	8-103
:PROGRAM[:SElected]:STATeL{RUN PAUSe STOP CONTinue}	8-103
:PROGRAM[:SElected]:STRingL< <i>string (varname)</i> >,< <i>string (value1)</i> >[,< <i>string</i>	
(<i>value2</i>)>[, . . . [,< <i>string (value n)</i> >	8-104
:PROGRAM[:SElected]:WAIT	8-104
A. Manual Changes	
Introduction	A-1
Manual Changes	A-1
Serial Number	A-2
Change 1	A-3
Change 2	A-3
B. Command Summary	
Meas/Format	B-1
Meas/Format Menu	B-1
Function Menu	B-1
Port Select Menu	B-2
Gain-Phase Format Menu	B-2
Z Format Menu	B-3
Display	B-3
Display Menu(1/3) (2/3) (3/3)	B-3
Linear Scale Menu	B-4
Log Scale Menu	B-4
Define Trace Menu	B-4
Basic Allocation Menu	B-4
Cal	B-5
Gain-Phase CAL Menu	B-5

Thru CAL Menu	B-5
Response & Isolation CAL Menu	B-5
Gain-Phase 3 Term CAL Menu	B-5
Gain-Phase CAL STD value menu	B-6
Z:Refl CAL Menu	B-6
Z: Refl CAL Menu	B-6
Z: Refl CAL STD value menu	B-6
Z: Trans CAL Menu	B-6
Z:Trans 3 Term CAL Menu	B-7
1 Term CAL Menu	B-7
Z: Trans CAL STD value menu	B-7
Marker	B-7
Marker Menu	B-7
Marker Serch Menu	B-7
Serch Range Menu	B-8
Marker Utility Menu	B-8
ΔMode Menu	B-8
Fixed Marker Position Menu	B-8
Marker Mode Menu	B-8
Sweep	B-9
Sweep Menu	B-9
Sweep Type Menu	B-9
Sweep Time Menu	B-9
Trigger	B-10
Trigger Menu	B-10
Start Stop Center Span	B-10
→ Function Menu	B-10
System	B-11
System Menu(1/3) (2/3) (3/3)	B-11
Clock Menu	B-11
Att Setting Menu	B-11
Save/Recall	B-11
Save/Recall Menu	B-11
Save Menu	B-12
File Utility Menu	B-12
Binary Define Save Data Menu	B-12
ASCII Define Save Data Menu	B-12
MISC Save Menu	B-13
Menu2	B-13
C. Data I/O Format	
Data Format	C-1
Form 2	C-1
Form 3	C-1
Form 4	C-2
Form 5	C-2
Internal Data Array	C-3

D. Waveform Analysis Commands

Conventions and Definitions	D-2
Waveform Analysis Setup Commands	D-3
ANAOC{1 2 3 4}	D-3
ANARANG	D-4
ANARANGP	D-4
ANARFULL	D-5
ANAODATA	D-5
ANAOMEMO	D-5
THRR	D-6
Maximum/Minimum/Mean Value Search Commands	D-7
OUTPMAX?	D-7
OUTPMIN?	D-7
OUTPMINMAX?	D-8
OUTPMEAN?	D-8
PEAK?	D-9
NEXPK?	D-9
NUMLMAX?	D-10
NUMLMIN?	D-10
LMAX?	D-11
LMIN?	D-11
TARR?	D-12
TARL?	D-12
Ripple Analysis Commands	D-13
RPLPP?	D-13
RPLHEI?	D-14
RPLRHEI?	D-15
RPLLHEI?	D-16
RPLENV?	D-17
RPLMEA?	D-18
RPLMM	D-19
RPLVAL?	D-20
POLE?	D-21
Filter and Resonator Analysis Commands	D-22
OUTPFILT?	D-23
OUTPXFIL?	D-25
OUTPXF2?	D-28
OUTPCFIL?	D-30
OUTPCF2?	D-33
OUTPRESO?	D-36
OUTPRESR?	D-38
OUTPRESF?	D-40
OUTPCERR?	D-42
Equivalent circuit analysis commands	D-43
EQUCPARA?	D-44
EQUCPARA5?	D-46
EQUCPARS?	D-47
EQUC0? <i>value</i>	D-48
EQUCPARS4?	D-50

E. Command HP 87510A v.s. HP E5100A	
COMMAND	E-1
F. Keyword Guide to Porting	
Trigger Function	F-2
Single Trigger (SRQ)	F-2
Single Trigger Mode (SING)	F-2
Calibration	F-2
Calibration Measurement (OPC)	F-2
Calibration Done (OPC)	F-2
Data Transfer Command	F-2
Binary Transfer Commands (FORM2, FORM3, FORM4)	F-2
Meas/Format	F-3
Input Port (AR, MEASA, MEASR)	F-3
Format (LOGM, PHAS, DELA, LINM, EXPP, LOGMP, LOGMD, REAL, IMAG, POLAR)	F-3
Impedance Conversion Function (CONVZTRA, CONVZREF, CONVYTRA, CONVYREF)	F-4
Analysis Commands	F-5
BIN sort (BINSLINE, BINS, BINOA, BINOB, EDITBINL)	F-5
Limit Line Edit (EDITLIS1, EDITLIS2)	F-5
Other Commands	F-5
KEY command	F-5
G. Option 022	
Overview	G-1
Using HP E5100A/B Option 022	G-2
Required Parameters	G-4
The Sample Program List	G-5
HP-IB Commands of Option 022	G-13
BINSIZE \square <value>	G-13
INPUTRAC \square <value1>,<value2>,<value3>	G-13
INPUTRACB \square <value1>,<value2>,<value3>	G-13
MEAS1PT? <value>	G-14
OUTPTRAC? \square <value>	G-14
OUTPTRACB? \square <value1>,<value2>	G-15
TIMO \square {ON OFF 0 1}	G-15
TOTIME \square <value>	G-15
TRABGE	G-15
TRABLE	G-16
TRAFDATA	G-16
TRAFMEMO	G-16
TRAP \square {OFF ON 0 1}	G-17
TRAR \square <value1>,<value2>	G-17
WRIT16	G-17
WRIT24	G-18

H. Option 023

Overview	H-1
Measuring Drive Level Characteristics using Phase Tracking Function	H-2
1. Measurement Settings for Phase Tracking	H-2
2. Setting for Using π Network Test Fixture	H-3
3. Settings for the Drive Level Measurement	H-3
4. Settings Parameters for Tracking Function	H-4
5. Searching for Resonant Point	H-4
6. Tracking Measurement	H-4
7. The Measurement Result	H-4
Sample Program List	H-5
Measuring Aging Characteristics	H-9
Compensation of Sweep Time for Aging Characteristics	H-9
Sub-routine for Measuring Aging Characteristics	H-9
Notes on the Phase Tracking	H-9
When the Phase Tracking Fails	H-9
Limitations on the Phase Tracking	H-10
HP-IB Commands for Option 023	H-10
FMT \square MAGZDF	H-10
PTABORT \square {ON OFF}	H-10
PTFOVHD \square <value>	H-10
PTFR \square <value>	H-11
PTFRSR \square <value>	H-11
PTPARA \square <value>	H-11
PTRACK \square {OFF ON}	H-11
PTREPN \square <value>	H-12
PTSTAT?	H-12
PTTRGLMT \square <value>	H-12
PTTRGPHS \square <value>	H-13
SRCHFR? \square {1 2 3 4},<value>	H-13

Index

Figures

1-1. HP-IB Device and Address	1-5
3-1. Simplified Data Processing Flow 1	3-2
3-2. Simplified Data Processing Flow 2	3-3
3-3. Simplified Internal Process of ASCII and Binary Transfer	3-7
3-4. IEEE 64-bit floating point format	3-8
3-5. IEEE 32-bit floating point format	3-9
3-6. Binary Data Header	3-9
4-1. The I/O Port Pin Layout	4-1
4-2. INPUT1, OUTPUT1, OUTPUT2 Timing Chart	4-4
5-1. General Status Register Model	5-1
5-2. Transition Filter and Condition Register	5-3
5-3. Status Reporting Structure	5-4
5-4. Example of Reading Status Byte (1)	5-5
5-5. Example of Reading Status Byte (2)	5-5
5-6. Example of Generating a Service ReQuest (SRQ)	5-7
5-7. SRQ Generation Sequence	5-11
6-1. Controlling Instrument BASIC from an External Controller (The Control Flow)	6-8
6-2. Parallel Processing	6-18
8-1. The Data Output Timing Chart	8-32
A-1. Serial Number Plate	A-2
C-1. Form 2 Data Transfer Format	C-1
C-2. Form 3 Data Transfer Format	C-1
C-3. Data Processing Flow Diagram	C-3
D-1. THRR	D-6
D-2. RPLPP?	D-13
D-3. RPLHEI?	D-14
D-4. RPLRHEI?	D-15
D-5. RPLLHEI?	D-16
D-6. RPLENV?	D-17
D-7. RPLMEA?	D-18
D-8. RPLVAL?	D-20
D-9. POLE?	D-21
D-10. OUTPFILT?	D-23
D-11. OUTPXFIL?	D-26
D-12. OUTPCFIL?	D-31
D-13. OUTPRESO?	D-36
D-14. OUTPRESR?	D-38
D-15. OUTPRESF?	D-40
D-16. OUTPCERR?	D-42
D-17. Six-Device Equivalent Circuit of Crystal Resonator	D-44
D-18. Four-Device Equivalent Circuit of Crystal Resonator	D-50
D-19. Admittance Characteristic Circle Diagram	D-51
G-1. The Trap Function	G-1
G-2. How HP E5100A/B option 022 Works with the Default Setup	G-2
G-3. The Displayed Measurement Result	G-4
G-4. The Data Output Timing Chart	G-14

H-1. Measuring Drive Level Characteristics	H-1
H-2. The Aging Characteristic Measurement	H-2
H-3. An Example of the Measurement Result	H-3
H-4. A Displayed Result	H-5

Tables

2-1. Setting the Active Channel	2-3
2-2. Setting the Measurement and Format	2-3
2-3. Setting the Sweep	2-4
2-4. Calibration	2-4
5-1. Status Bit Definitions of the Status Byte (STB)	5-5
5-2. Status Bit Definitions of the Event Status Register (ESR)	5-6
5-3. Status Bit Definitions of the Event Status Register B (ESB)	5-6
5-4. Status Bit Definitions of the Operational Status Register (OSR)	5-7
A-1. Manual Changes by Serial Number	A-1
A-2. Manual Changes by Firmware Version	A-1
C-1. Data Array Output Commands	C-3
C-2. Data Array Input Commands	C-3
F-1.	F-3
F-2.	F-4
F-3.	F-4

Introduction

How to Use This Manual

This manual introduces HP-IB programming for the HP E5100A/B. It provides additional information on how to write programs that might be difficult to understand when using only the HP-IB Command Reference. It also provides information, techniques, and examples of how to effectively control HP-IB instruments.

To use this manual effectively, you need one of the following HP-IB controllers:

- HP Instrument BASIC (The HP E5100A/B has this function unless Option UKR Delete HP Instrument BASIC is installed). That is an internal HP-IB controller in the HP E5100A/B.
- An HP 9000 Series 200 or 300 computer that has HP BASIC 3.0 or later. The HP BASIC 3.0 must have the HPIB, GRAPH, IO, KBD, and ERR binaries.

This manual helps you to learn how to write programs that control the HP E5100A/B. To help you learn quickly, many sample modules and programs are provided.

Target Reader

The target reader of this manual is a programmer who wants to control the HP E5100A/B through the HP-IB interface.

This manual explains HP-IB programming using HP BASIC. Therefore, you should have some experience using BASIC. If you have never written a program in BASIC, review the applicable documentation listed at the end of this chapter before starting this manual. This manual does not require extensive knowledge of BASIC programming.

This manual assumes you understand the operations and features of the HP E5100A/B. If you have never operated the HP E5100A/B, read the *User's Guide* to learn how to operate the HP E5100A/B.

What's in This Manual?

The following chapters are provided in this manual:

- Chapter 1 "Introduction" provides an introduction to this manual, how to use a sample program, an HP-IB overview, hardware preparation, and a description of the sample program disk. This chapter provides important information that is used throughout this manual. You should read this chapter first.
- Chapter 2 "Setup and Measurement Program" provides HP-IB command basics. It also shows how to build a measurement program including setups, compensating, triggering, and getting data. If you want to build an automated measurement program, read this chapter.
- Chapter 3 "Data Processing and Transfer" shows the data processing flow and the arrays of the analyzer, describes how to access an internal data array (including trace data or

calibration data). If you want to get measured trace data from the analyzer, read this chapter.

- Chapter 5 “Using Status Reporting Function” describes the status reporting system of the analyzer and how to use it. This chapter also describes an SRQ interrupt. If you want to obtain the analyzer’s status using a BASIC program, read this chapter.
- Chapter 4 “Using the I/O Port” provides information on how to use the I/O port on the rear panel. If you want to use the I/O port for communicating with an external instrument (such as a handler), read this chapter.
- Chapter 6 “Programming Miscellaneous” provides information not directly concerned with measurements, but useful for programming. This includes accessing the disk, controlling Instrument BASIC, or debugging a program.
- Chapter 7 “Sample Programs” describes sample programs for some applications.
- Chapter 8 “Command Reference” provides a summary of all available HP-IB commands.
- Appendix A “Manual Changes” shows revision information for this manual.

How to Use the Program Modules

This manual provides many sample program modules that are not in a complete program style. You can easily understand the module’s objective because the program module does not include unnecessary code. You can use these modules to build your own program by combining them.

The program modules are provided in the following style and typeface:

`SAMPLE CODE` *This is a comment for the sample code.*

As shown in the example above, a module has no line number, no initializing steps, and no END statement. All these are required for an executable BASIC program.

Building a Working Program Using Program Modules

To make a program that uses sample program modules, perform the following steps:

1. Add an initializing module at the beginning of your program.
2. Arrange the program modules.
3. Add an END statement on the last line executed by your program.

The line numbers are added automatically by the BASIC editor.

Initializing Module

The initializing module defines a hardware identifier as a variable to eliminate the difference between Instrument BASIC and HP BASIC. Usually, you can use the same program for Instrument BASIC and HP BASIC by changing the initializing module. The initializing module also initializes the HP-IB.

The following are typical initializing modules for a program:

ASSIGN @E5100 TO 800	<i>Assigning HP-IB address to 800.</i>
Scode=8	<i>Assigning interface select code to 8.</i>
ABORT Scode	<i>Get active control.</i>
CLEAR @E5100	<i>Preset the interface.</i>
OUTPUT @E5100;"DISAHIHB"	<i>Set display mode.</i>

Module 1-1. Instrument BASIC Initialization

ASSIGN @E5100 TO 717	<i>Assigning HP-IB address to 717.</i>
Scode=7	<i>Assigning interface select code to 7.</i>
ABORT Scode	<i>Get active control.</i>
CLEAR @E5100	<i>Preset the interface.</i>

Module 1-2. External Controller Initialization

Each module of this manual assumes that one of the initializing modules exists at the beginning of the program, and uses the following variables without additional explanation:

@E5100 Represents the device selector of the HP E5100A/B. 800 is for Instrument BASIC and 717 is for the external controller.

Scode Represents the interface select code to which the HP E5100A/B is connected. 8 is for Instrument BASIC and 7 is for the external controller.

Example

For example, a complete program using Module 2-2 in Chapter 2 and Instrument BASIC, is shown below:

```

10 ASSIGN @E5100 TO 800      ! Module 1-1
20 Scode=8                  !
30 ABORT Scode              !
40 CLEAR @E5100            !
50 OUTPUT @E5100;"DISAHIHB" !
60 !
70 OUTPUT @E5100;"PRES;CHAN2" ! Module 2-2
80 !
90 END

```

HP-IB Overview

The HP-IB is a general purpose digital interface system that is used to integrate the controller, measurement instruments, and peripherals into a system. HP-IB is Hewlett-Packard's implementation of the IEEE 488 Bus.

Controller

The controller is a device that can address an HP-IB device to talk (output data) or listen (receive data).

The active controller can control the other devices on the bus at that time (when multiple controllers are connected). Only one controller can be active at a time. The active controller can pass control to another controller by using the `PASS CONTROL` command.

Only one controller can be the system controller on the same bus. The system controller is the active controller when the system is turned on. When another controller is the active controller, the system controller can become the active controller at any time by executing `ABORT select-code`.

Device Selector

HP-IB device control is accomplished by sending commands from the active controller. The active controller can select the target device for the commands by specifying the device selector.




Figure 1-1. HP-IB Device and Address

Figure 1-1 shows the relationship between the HP-IB address and the device selector. For example, the device selector of the HP E5100A/B on HP-IB with an address of "17," is "717" on the HP-IB.

HP Instrument BASIC is connected in the HP E5100A/B internally by the internal interface. The interface select code of the internal interface is "8" to distinguish it from the external select code of "7."

You can use any address from "00" to "30" to specify the internally connected analyzer from Instrument BASIC, because only the analyzer is connected to the internal interface. This manual uses address "00," thus the device selector is "800."

Preparation for Operation

This section describes HP-IB cable connection and address setting for HP-IB control. You can choose one of the following two methods to control the analyzer:

- Using the Instrument BASIC controller (except for Option UKR).
- Using an external controller.

In both cases, you can use the following procedure to prepare the controller and the analyzer:

1. Connect the HP-IB cables.
2. Set the HP-IB Address.
3. Prepare the controller for use.

Using Instrument BASIC as the Controller (except for Option UKR)

The HP Instrument BASIC system is a BASIC controller that is built into the analyzer. HP Instrument BASIC is a subset of HP BASIC.

You can control the HP E5100A/B internally by using HP Instrument BASIC. HP Instrument BASIC has the capability to be a system controller. The other HP-IB instruments are also controllable through the HP-IB connector that is located on the rear panel of the HP E5100A/B.

1. Connecting the HP-IB Cables For Instrument BASIC

A connection between the analyzer and Instrument BASIC is not required because they are already connected via the internal interface in the analyzer. See Figure 1-1. However, if any other HP-IB instruments must be connected, see "Connecting External HP-IB Cables" on the next page.

2. Setting the HP-IB Address For Instrument BASIC

This is not required because the front panel setting of the HP-IB address does not affect the internal interface. You can use any address to specify the analyzer via the internal interface as described in "Device Selector".

3. Setting Up Instrument BASIC

To set up Instrument BASIC, perform the following steps:

1. Connect the DIN keyboard to the rear panel connector. (See the *User's Guide*.)
2. Turn the analyzer power on.
3. To allocate the Instrument BASIC output area on the LCD display for the PRINT statement, press **Display** **DISPLAY ALLOCATION HALF INSTR HALF BASIC**.
4. Press **System** **I-BASIC EDIT** to open the Instrument BASIC editor.

For more information on how to use the Instrument BASIC editor, see the *Using HP Instrument BASIC*.

If you connect the keyboard after turning on the analyzer, press **Preset** to enable key inputs.

Using an External Controller

You can use an external controller to control the analyzer by using the HP-IB interface on the rear panel. Connect the controller and the analyzer using an HP-IB cable.

1. Connecting External HP-IB Cables

Connect the analyzer and external instruments with HP-IB cables. The HP-IB connector is on the rear panel of the analyzer.

The rules for connecting HP-IB cables are as follows:

- The total number of HP-IB devices can be up to 15 instruments.
- The total length of all the cables used can be up to 20m or 2m for each instrument.

You can connect the HP-IB cables in any configuration (linear, star, or combination), as long as the above rules are satisfied.

Note

Do not use a screwdriver when connecting the HP-IB cables. The screwdriver slots in the lock screws are provided for REMOVAL only.



2. Setting the HP-IB Address

The analyzer has no hardware switch for setting the HP-IB address. You can only set it by front panel operation. To change the HP-IB address of the analyzer, perform the following steps:

1. Press **System** **HP-IB**.
2. Press **ADDRESS: E5100A** or **ADDRESS: E5100B**.
3. Enter the new address by using the numerical keys. (Avoid duplication with the HP-IB address of the external controller.)
4. Press **x1** to complete the operation.

When you want to control another HP E5100A/B, change the HP-IB address to avoid duplication of addresses on the same bus.

When the analyzer receives any HP-IB command from an external controller, **Local** is appeared on the display to indicate the analyzer is in the remote mode. In remote mode, front panel key operation is disabled. To cancel the remote mode, press the **Local**.

3. Preparing For HP BASIC Operation

To prepare HP BASIC for operation, see your HP BASIC system manual.

Sample Program Disk

The sample programs (not the program modules) in this manual are included on the furnished *Sample Program Disk* (HP Part Number E5100-61001).

Loading a Program from Disk

To use an IBASIC sample program, load it into Instrument BASIC and then run it.

1. Insert the *Sample Program Disk* into the internal disk drive that is below the display.
2. Press **Display** **DISPLAY ALLOCATION BASIC STATUS**.
3. MSI ": INTERNAL, 4"
4. Type, GET "*filename*" **Return**.

To use an external controller sample program, load it into an external controller and run it.

1. Set up HP BASIC on your external controller.
2. Insert the *Sample Program Disk* into the disk drive of the controller.
3. Set the current directory to the disk drive.
4. Type, GET "*filename*" **Return**.

The applicable *filename* is printed in front of the sample programs in this manual.

Reading the Sample Program Disk on a PC

Because the sample program disk is provided in the LIF (Logical Interchange Format) that is used on an HP controller, you cannot read this disk on a PC. If you want to convert the program into DOS format that can be read by a PC, perform the following steps:

1. Prepare a DOS formatted disk (3.5 inch).
2. Load the program into the Instrument BASIC from the sample program disk. See "Loading a Program from Disk" for the procedure.
3. Remove the sample program disk, then insert the DOS formatted disk.
4. Type, SAVE "*filename*" **Return**.

The analyzer supports either LIF or DOS format, and automatically detects which format disk is used.

Note You cannot initialize a disk in the LIF format with the HP E5100A/B.



Related Documentation Information

You can obtain more detailed information than provided by this manual by referring to the following documents.

The following manuals are provided with the HP E5100A/B.

- *HP E5100A/B Function Reference* explains all functions accessed from the front panel key.
- *HP E5100A/B User's Guide* to learn about the analyzer itself and its front panel key operation.

The following manual is also provided with the HP E5100A/B unless Option UKR (Delete HP Instrument BASIC) is installed.

- *HP Instrument BASIC User's Handbook Supplement* for Instrument BASIC information.

The following documents also provide related information:

- *HP BASIC Programming Guide* for learning HP BASIC programming. (Furnished with the HP BASIC system.)
- *Tutorial Description of the Hewlett-Packard Interface Bus* for an overview of the HP-IB and IEEE 488 standard (HP literature no. 5952-0156).

Setup and Measurement Program

This chapter describes a basic measurement program that include setups, user calibration, triggering, and getting data. This chapter discusses the following topics:

- Overview of HP-IB Control.
- Automating a Measurement Procedure.
- Sample Program: Basic Measurement Program.

Overview of HP-IB Control

Before starting to program, you should know how to send an HP-IB command to the analyzer. This section describes the following items:

- Sending HP-IB commands.
- Sending a query and reading the response.

If you have experience programming HP-IB instruments, you can skip this section and go to "Automating a Measurement Procedure".

Sending HP-IB Commands

HP BASIC and Instrument BASIC use the **OUTPUT** statement to send HP-IB commands that control HP-IB devices. For example:

```
OUTPUT @E5100;"PRES" Presets the analyzer.
```

Module 2-1. Presetting the Analyzer

This line sends the HP-IB command within the double quotes (**PRES**) to the HP-IB device at address @E5100. This command presets the analyzer. This is equivalent to pressing the **Preset** key.

You can send multiple commands in a single line by separating each HP-IB command with a semicolon (;).

```
OUTPUT @E5100;"PRES;DISAHIHB" Sends PRES and DISAHIHB by single line.
```

Module 2-2. Sending Multiple Commands in a Line

Sending a Query and Reading the Response

There are commands that return a response after being sent. These commands are called queries. A query has a question mark (?) at the end of the command.

You can retrieve the response by using the ENTER statement as shown below:

OUTPUT @E5100;"CENT?"	<i>This line queries center frequency setting.</i>
ENTER @E5100;Center	<i>This line retrieves the return value.</i>

Module 2-3. Querying Center Frequency Setting

You must retrieve the response into the correct type variable. In the example above, the query returns a value, depending on the current center frequency setting. Therefore, the second line retrieves the response into a numeric type variable (**Center**).

The response data type, numeric or string, for each command is shown in the reference section of the *HP-IB Command Reference*.

Automating a Measurement Procedure

This section describes the sample program modules and equivalent commands for setting up the analyzer using the following functions:

1. Setting the Active Channel
2. Setting the Measurement and Format
3. Sweep Setup
4. Calibration
5. Device Connection
6. Triggering a Measurement
7. Data Processing and Transfer
8. Exiting the Program

1. Setting the Active Channel

To begin setting the Measurement, Format, Calibration, Display, and Marker functions of the analyzer, specify the active channel first because it affects all these settings.

Table 2-1. Setting the Active Channel

Setups	Key Operations	HP-IB Commands
Active channel 1	<div style="border: 1px solid black; padding: 2px; display: inline-block;">Meas/Format</div> ACTIVE CH [CH2] ¹	CHAN1

1: Press this softkey to alter the label to [CH1] when the label is [CH2].

The equivalent program is as follows:

```
OUTPUT @E5100;"CHAN1" Sets the active channel to channel 1.
```

Module 2-4. Setting the Active Channel

2. Setting the Measurement and Format

Table 2-2. Setting the Measurement and Format

Setups	Key Operations	HP-IB Commands
Gain-Phase Mode	<div style="border: 1px solid black; padding: 2px; display: inline-block;">Meas/Format</div> FUNCTION[] GAIN-PHASE	ANAMODE GAINP
A/R Measurement	Meas A/R	MEAS AR
LOG MAG Format	FORMAT LOG MAG	FMT LOGM

The equivalent program for setting Active Channel, Measurement, and Format is as follows:

```
OUTPUT @E5100;"CHAN1;MEAS AR;FMT LOGM" Sets the Active Channel to 1, the Measurement to A/R, and the
Format to LOG MAG.
```

Module 2-5. Active Channel/Measurement/Format Setup

3. Sweep Setup

Table 2-3. Setting the Sweep

Setups	Key Operations	HP-IB Commands
Center Frequency to 70 MHz	CENTER 70 × M	CENT 70MHZ
Frequency Span to 100 kHz	SPAN 100 × k	SPAN 100KHZ

The equivalent program is as follows:

```
OUTPUT @E5100;"CENT 70MHZ" Sets the center frequency to 70 MHz.
OUTPUT @E5100;"SPAN 100KHZ" Sets the span to 100 kHz.
```

Module 2-6. Sweep Setup

Note



You must be careful with the order of commands when you execute them. The HP E5100A/B has 2 pairs of channel settings and 2 stimulus settings, and they can be set up separately. A setup is applied to the currently selected active channel or active stimulus. You must, therefore, set up the active channel/stimulus before performing other setups.

4. Calibration

The key operations and corresponding HP-IB commands are described below.

Table 2-4. Calibration

Setups	Key Operations	HP-IB Commands
Response Calibration Startup	CAL RESPONSE	CALI RESP
THRU Measurement	THRU	STANC
Completion of Calibration	DONE RESPONSE	RESPDONE

As an example, let's look at a frequency response calibration procedure and its corresponding HP-IB commands.

As described in the setup process, you must execute HP-IB commands in the correct order using **OUTPUT** statements to perform calibration. In addition, there is another important consideration: You must pause the program when the HP E5100A/B is measuring the calibration standard or is calculating the calibration coefficient.

The reason that you must pause the program is that the measurement modules of the HP E5100A/B and Instrument BASIC process data independently. Instrument BASIC, therefore, would try to execute the next command before the instrument measurement module completes

its job. This can occur when the measurement module is processing a time consuming job (for example, a sweep or a calibration coefficient calculation). The measurement module cannot process an HP-IB command if the module receives it during a process. Therefore, you must synchronize the measurement module and the process used by the controller to execute HP-IB commands successfully. The following modules show a way to synchronize these two procedures.

Calibration Process

1. Setup the calibration mode.

```
OUTPUT @E5100;"CALI RESP"  Sets the calibration mode to Response.
```

Module 2-7. Calibration Mode Setup

2. Connect a standard and perform a measurement.

You can use the STANC? Query to make a calibration measurement and to check to check for the completion of the sweep because the sweep process requires extra time. When you execute STANC?, it returns 1 as soon as the HP-IB command executed immediately before it is completed. The program does not proceed to the next step until the ENTER statement receives the return value from the Query.

```
INPUT "CONNECT THRU, THEN PRESS ANY KEY",A$  Shows the prompt message and waits for the key
entry.
OUTPUT @E5100;"STANC?"                        Performs a measurement.
ENTER @E5100;Tmp
```

Module 2-8. Connection and Measurement

3. The calibration coefficient is calculated when the measurement is completed. You can also pause the program during a calibration coefficient calculation. RESPDONE? is used to monitor for the completion of the calculation.

```
OUTPUT @E5100;"RESPDONE?"  Performs the calculation of calibration coefficient.
ENTER @E5100;Tmp
```

Module 2-9. Calculation of the Calibration Coefficient

The calibration is now complete and the HP E5100A/B is ready to make a measurement.

5. Device Connection

You must tell the operator to connect a device. This statement uses INPUT to pause the program until data is entered from the keyboard.

```
INPUT "CONNECT DUT, THEN PRESS [RETURN]",A$  Displays a message and waits for the data entry
```

Module 2-10. Device Connection

The program starts a measurement when a device is connected. You need to pause the program, as you needed to do for a calibration, until the measurement completes.

6. Triggering a Measurement

If the measurement sweeps only once, you can omit the process to monitor for the completion of the sweep by executing the following line:

```
EXECUTE "SING" Execute a single measurement.
```

Module 2-11. Triggering a Measurement

When the **SING** command is executed by the **EXECUTE** statement, the program pauses until the sweep completes.

Note



You cannot use this method from an external controller because only the HP E5100A/B Instrument BASIC supports the **EXECUTE** command. You must use the same method as you do for the multiple sweep to synchronize the completion of sweep and the program.

When You are Using an External Controller

When you are using an external controller, you can use the ***SING?** Query to make a single measurement and check the completion of a sweep.

```
OUTPUT @E5100;"SING?"
ENTER @E5100;Tmp
```

Module 2-12. Triggering Module For an External Controller

7. Data Processing and Transfer

You can use a Query command to extract the measurement data required.

The following process executes the command **OUTPMAX?** (which returns the maximum value within the display range).

```
OUTPUT @E5100;"OUTPMAX?" A Query to return the maximum value.
ENTER @E5100;Value
PRINT "Maximum value:",Value;"dB" Display the return value.
```

Module 2-13. Data Processing and Transfer

8. Exiting the Program

Exit the program. You can also repeat a measurement using the **GOTO** statement.

```
END Exits the program.
```

Module 2-14. Exiting the Program

Sample Program -1: Basic Measurement

The following is a list of the program described in this section. You can refer to the this program to check the explanations.

Disk



This program is contained in the attached sample disk with the file name **BASIC** (For Instrument BASIC). The program for an external controller has the name **BASIC_E**.

```

110 ASSIGN @E5100 TO 800          ! IBASIC INITIALIZATION
120 Scode=8                      ! [MODULE 1-1]
130 CLEAR @E5100                 !
140 OUTPUT @E5100;"DISAHIHB"     !
150 !                             !
160 OUTPUT @E5100;"PRES"        ! ANALYZER PRESETTING
170                               ! [MODULE 2-1]
180 !                             !
190 OUTPUT @E5100;"CHAN1;ANAMODE GAINP;MEAS AR;FMT LOGM"
                               ! CHAN/MEAS/FORMAT SETUP
                               ! [MODULE 2-5]
200                               !
210 !                             !
220 OUTPUT @E5100;"CENT 70MHZ"   ! SWEEP SETUP
230 OUTPUT @E5100;"SPAN 100KHZ" ! [MODULE 2-6]
240 !                             !
250 OUTPUT @E5100;"CALI RESP"    ! CAL MODE SETUP
260                               ! [MODULE 2-7]
270 !                             !
280 INPUT "CONNECT THRU, THEN PRESS [RETURN].",Dum$
                               ! CONNECTION & CAL MEASUREMENT
                               ! [MODULE 2-8]
290 OUTPUT @E5100;"STANC?"      !
300 ENTER @E5100;Tmp           !
310 !                             !
340 OUTPUT @E5100;"RESPDONE?"   ! CALCULATION OF CAL COEFFICIENT
350 ENTER @E5100;Tmp           ! [MODULE 2-9]
360 !                             !
400 DISP "RESPONSE CAL COMPLETED" ! DISPLAY MESSAGE
410 !                             !
420 INPUT "CONNECT DUT, THEN PRESS [RETURN].",Dum$
                               ! PROMPT DEVICE CONNECTION
                               ! [MODULE 2-12]
430                               !
440 !                             !
450 EXECUTE "SING"              ! TRIGGERING MEASUREMENT
460                               ! [MODULE 2-13]
470 !                             !
480 OUTPUT @E5100;"AUTO"        ! AUTO SCALING
490 !                             !
500 OUTPUT @E5100;"OUTPMAX?"    ! DATA PROCESSING & TRANSFER
510 ENTER @E5100;Value         ! [MODULE 2-14]
520 PRINT "MAXIMUM VALUE:",Value,"dB"
530 !                             !
540 END                          ! EXITING PROGRAM
550                               ! [MODULE 2-15]

```

Column Parameters and Variables

As explained in the previous section, an HP-IB command that requires a parameter is executed as follows:

```
OUTPUT @E5100;"CENT 70MHZ" Sets the center frequency to 70 MHz.
```

Module 2-15. Command with Fixed Value

In this case, the parameter value is fixed in the program. You can, however, use a variable to change the parameter depending on the situation.

```
OUTPUT @E5100;"CENT ";Center Sets the center frequency to Center.
```

Module 2-16. Command with Valuable

This statement passes the variable **Center** to the HP-IB command **CENT** as a parameter. You can place a parameter outside of double quotes and separate them with a semicolon to use a variable as a parameter. (You must always insert a space after a command.)

The following shows an HP-IB command that takes a character string parameter.

```
OUTPUT @E5100;"SAVDGRA ""filename"" Saves the graphic image as the filename.
```

Module 2-17. Command with Character String Parameter

Character strings are usually double quoted ("). In this case, however, you must double the double quotes (") for the inside double quotes because an HP-IB command itself is already double quoted.

You can also use single quotes(') instead of double double quotes("").

```
OUTPUT @E5100;"SAVDGRA 'filename'"
```

Module 2-18. Command with Character String Parameter

You can also use a variable as a string character parameter.

```
OUTPUT @E5100;"SAVDGRA ";filename$;"
```

Module 2-19. Command with Variable as String Character Parameter

This statement is basically the same as the numeric value variable example. You must, however, place a single quote before and after **filename\$** because a string character must be single quoted.

To close a double double quote, you must use a quadruple double quote. You must use a double double quote (") to close a single quote.

The following examples are both valid. (The command **PROG:EXEC""GET....** can be executed from an external controller only.)

```
OUTPUT @E5100;"PROG:EXEC ""GET ""Filename""""""
```

Module 2-20. Command With Double Quotes

```
OUTPUT @E5100;"PROG:EXEC 'GET ""Filename""'"
```

Module 2-21. Command With Single Quotes

Data Processing and Transfer

Introduction

This chapter describes data processing and how to access the internal data arrays.

The following information is covered in this chapter:

- Data arrays.
- Data transfer method.
- Sample program: compensation data transfer.

Data Arrays

The analyzer has data arrays that contain measurement data, error correction data, and stimulus data. You can read or write data to these arrays using HP-IB commands.

Figure 3-1 and Figure 3-2 show simplified diagrams of the data processing in the analyzer.





Figure 3-1. Simplified Data Processing Flow 1



C7209008

Figure 3-2. Simplified Data Processing Flow 2

Dotted line boxes represent data arrays that hold intermediate or processed data. The following section describes each of these data arrays.

Raw Data Array

The raw data array stores the results of all the preceding data processing operations including the correction by calibration data. The data is in a complex form (real/imaginary pairs) and read out with the commands **OUTPRAW?**. When you want to use your own compensation method for a measurement data, ① take out the raw data from the raw data array (see Module 3-1), ② apply your compensation method to this data, ③ enter the compensated data into the data array (Module 3-2).

The following example module queries for raw data and retrieves it:

DIM Dat(1:201,1:2)	<i>Define NOP×2 for receiving complex data.</i>
OUTPUT @E5100;"OUTPRAW?"	<i>Query raw data array.</i>
ENTER @E5100;Dat(*)	

Module 3-1. Getting Raw Data Array

■ Related HP-IB Commands

The following command is used for sending data to the raw data array:

INPURAW, data Sends data to the raw data array of the active channel.

Data Array

This is the raw data with error correction applied. The array is for the currently measured parameter, and is in a complex form (real/imaginary pairs). The data array data is read out with **OUTPDATA?** or **OUTPDATA?P?**. The **OUTPMEMO?** or **OUTPMEMOP?** query reads the trace memory if available (which is also error corrected). Neither raw nor data array data reflect post-processing functions such as electrical delay offset or trace math.

The following example module sets data for the data array:

OUTPUT @E5100;"INPUDATA?";Dat(*)	<i>Sets data to data array.</i>
----------------------------------	---------------------------------

Module 3-2. Setting Data Array Data

■ Related HP-IB Commands

The following command is used to query the data array data:

OUTPDATA, data Queries data in data array of the active channel.

Unformatted Data Array

This is the array of the complex number pairs that will be formatted in the next stage. *The unformatted data cannot be read out.*

Formatted Data Array

This is the array for the data being displayed. It reflects all post-processing functions such as electrical delay. The units of the array read out depend on the current display format.

This data array is generally the most useful because it is the same information as that seen on the display. When you want to use data with a selected parameter unit, use this array.

The following example module queries the data trace array and retrieves it:

DIM Dat(1:201,1:2)	Define <i>NOPx2</i> for receiving complex data.
OUTPUT @E5100;"OUTPFORM?"	Query data trace array.
ENTER @E5100;Dat(*)	<i>OUTPFORM?</i> outputs data in a complex format.

Module 3-3. Getting Data Trace Array

■ Related HP-IB Commands

The following commands are used for accessing the data trace array:

OUTPFORM? Outputs data trace array of the active channel.
OUTPRTMEM? Outputs memory trace array of the active channel.

Using the following commands, you can access a data trace array for the inactive channel. These commands are useful to get data from both traces without altering the active channel.

OUTPIFORM? Outputs data trace array of inactive channel.
OUTPRITMEM? Outputs memory trace array of the inactive channel.

Calibration Coefficient Array

The results of a calibration are stored as arrays of calibration coefficients that are used by the error correction routines. Each array corresponds to a specific error term in the error model. The calibration coefficients are read out with **OUTPCALC{01|03}?**.

For detailed information about the calibration and error model, see the *Function Reference*.

Accessing Arrays

If you want to enter calibration data from the controller to the calibration coefficient array, the analyzer must have previously done the calibration to enable the calibration data. To do this, perform the following steps:

1. Execute a dummy calibration to validate the correction.
2. Send the new calibration coefficients.

Note that the calibration coefficients are complex data.

■ Related HP-IB Commands

The following commands are used to access the calibration coefficient array:

OUTPCALC{01|02|03}? Outputs the specified calibration coefficient array data of the active channel.
INPUCALC{01|02|03} Sets the specified calibration coefficient array data of the active channel.

Arrays for Memory Trace

When you store the trace data into the trace memory by sending the **DATI** command or by pressing **DATA → MEM**, the data array data is stored into the memory array. Memory array data passes through the formatting process, and then is stored into the memory trace array that is being displayed on the LCD display. See Figure 3-1 and Figure 3-2.

Besides the 16-trace limitation, the number of traces that can be stored into trace memory at one time depends on the capacity of the system memory and the number of points in the traces. The analyzer always reserves memory for 3 traces. The remaining memory is shared with the Instrument BASIC graphics.

Accessing Memory Array

You can only read data for the memory array that is activated. The memory array is read only. The following example module reads data from the memory array:

```
DIM Dat(1:201,1:2)
OUTPUT @E5100;"OUTPTMEM?"  Queries an activated memory array.
ENTER @E5100;Dat(*)         Receives the query response.
```

Module 3-4. Getting Memory Array Data

Accessing Trace Array

You can read or write trace array data. The following example module sets data for the trace array. You have to execute a trace-to-memory store procedure before setting a memory to display:

```
OUTPUT @E5100;"DATI"        Store trace data into trace array memory to allocate a trace array area.
OUTPUT @E5100;"OUTPTMEM?"  Queries a trace array.
OUTPUT @E5100;Dat(*)       Receives the query response.
```

Module 3-5. Setting Trace Array Data

■ Related HP-IB Commands

The following commands are used to query the contents of the memory trace array:

OUTPTMEM?	Outputs the memory trace of the active channel.
OUTPITMEM?	Outputs the memory trace of the inactive channel.

Data Transfer Methods

This section describes the data transfer methods. When you get or send the data array's data, there are two methods for data transfer: ASCII and binary.




Figure 3-3. Simplified Internal Process of ASCII and Binary Transfer

Because the ASCII transfer passes through the formatted process, the program does not care about the data format. On the other hand, the binary transfer directly passes the data, but you have to indicate what data is transferred using the data header. The binary transfer is faster than the ASCII transfer.

ASCII Transfer

The ASCII format transfer is the easiest way to transfer array data between the analyzer and the controller. You do not have to worry about the data format because the analyzer and the controller automatically handles the formatting of the transferred data in this format. You can just send or retrieve array data by using the **OUTPUT** and **ENTER** statements.

The ASCII transfer format is sent as a 14-character (data) or 22-character (stimulus) string for each data point. This string includes a digit, sign or decimal point. Therefore, the data length of 201 points of complex data is 6030 bytes. (Including data delimiter "LF" for each data.)

To retrieve data from the analyzer using the ASCII format transfer, the following procedure is used:

1. Define a data array that is the same size as the data to be retrieved.

2. Specify the data transfer format is ASCII.
3. Send the data query command.
4. Retrieve the data.

DIM Dat(1:201,1:2)	Define the data array for receiving.
OUTPUT @E5100;"FORM4"	Specify the ASCII format.
OUTPUT @E5100;"OUTPDAT?"	Query the data trace.
ENTER @E5100;Dat(*)	Retrieve the data.

Module 3-6. Retrieving Data from the Analyzer Using ASCII Transfer

Sending data to the analyzer is easy, just specify a format, then send the data:

OUTPUT @E5100;"FORM4"	Specify the ASCII format.
OUTPUT @E5100;"INPUDATA ";Dat(*)	Send command and data.

Module 3-7. Sending Data to Analyzer by ASCII Transfer

Binary Transfer

For a faster data transfer, use the binary format. There are three formats for binary transfer. The following list shows the data format that the analyzer outputs when you query the data:

■ IEEE 64-bit Floating Point Format

Figure 3-4 shows the data transfer format of the IEEE 64-bit floating point format. Data is stored internally in the 200/300 series computer with the IEEE 64-bit floating point format, eliminating the need for any reformatting by the computer. In this mode, each number takes 8 bytes.




Figure 3-4. IEEE 64-bit floating point format

■ IEEE 32-bit Floating Point Format

Figure 3-5 shows the data transfer format the IEEE 32-bit floating point format. In this mode, each data point is 4 bytes. The difference from the 64-bit format is a significant digit. The 64-bit format has double the precision of this format.




Figure 3-5. IEEE 32-bit floating point format

■ MS-DOS[®] Format

This mode is a modification of the IEEE 32-bit floating point format with the byte order reversed. The MS-DOS format also has a four-byte header that must be read to maintain the data order. In this mode, a PC can store the data internally without reformatting it.

Data Header

As shown in Figure 3-4 and Figure 3-5, the data header always precedes the data itself in binary format transfer. When you use a binary transfer, you must handle the data header with the data body.

When you query data in binary format, the analyzer outputs a fixed length (8 byte) data header. You can handle the data header as 8-byte strings for this purpose.

When you send the data to data array using binary transfer, you must prepare the data header for the data you send. The data header indicates the size of the transferred data. The data header consists of the following three parts: sharp (#), Number of bytes of "Data Size", and data size.




Figure 3-6. Binary Data Header

For example, the data size of 201 points of complex data in the 64-bit format is 3216 byte (=201×2×8). The "3216" is 4 digit (4 byte) number. Thus, the data header is "#43216". The queried data header that is generated from the analyzer is a fixed length header of 8 bytes that is obtained by adding "0" before "Data Size". For example, the data header above becomes "#6003216" as an 8-byte string. You can use either type of header to send data to the array.

Getting Data from the Analyzer

To get a data from the analyzer using the binary transfer method, the following procedure is used:

1. Assign a binary data path. (Specifying format off.)
2. Specify the data transfer format as binary.
3. Define a data array that is the same size as the data that will be retrieved.
4. Send the data query command.
5. Retrieve the data header.
6. Retrieve the data.
7. Retrieve the terminator.
8. Set the transfer format to ASCII mode if binary transfer is finished.

The binary data is sent in a mixed format of an ASCII header and a binary data body as shown in Figure 3-4 and Figure 3-5. To retrieve data correctly, you must retrieve the data header and data itself independently.

The following is a sample module for receiving data using the IEEE 64-bit format:

ASSIGN @Dt TO 800;FORMAT OFF	<i>Binary path must turn off the formatting. Use 717 instead of 800 for the external controller.</i>
OUTPUT @E5100;"FORM3"	<i>Specify the format as IEEE 64-bit format.</i>
DIM Dat(1:201,1:2)	<i>Assume that the receiving data size is 201x2.</i>
OUTPUT @E5100;"OUTPUTDATA?"	<i>Query the data trace array.</i>
ENTER @E5100 USING "#,8A";Header\$	<i>Data header is 8 byte character.</i>
ENTER @Dt;Dat(*)	<i>Receiving data via binary path.</i>
ENTER @E5100;End\$	<i>Reading terminator.</i>
OUTPUT @E5100;"FORM4"	<i>Set ASCII mode if binary transfer is finished.</i>

Module 3-8. Getting Data from Analyzer Using Binary Transfer

Note Binary data transfer to the analyzer is not allowed. Use ASCII transfer for sending data to the analyzer.



■ Related HP-IB Commands

The following commands are used to specify the data transfer format.

FORM2	Selects IEEE 32-bit floating point format.
FORM3	Selects IEEE 64-bit floating point format.
FORM4	Selects ASCII format.
FORM5	Selects MS-DOS format.

Sample Program -2: Binary Data Transfer

This is a sample program for the trace data transfer that uses the IEEE 64 bit floating decimal point format. This program transfers the data at each measured trace point in the binary format and displays them on the screen.

Disk



This program is contained in the attached sample disk with the file name **BINARY** (For Instrument BASIC). The program for an external controller has the name **BINARY_E**.

```

110 ASSIGN @E5100 TO 800;FORMATON! IBASIC INITIALIZATION
120 ASSIGN @Dt TO 800;FORMAT OFF !
130 Scode=8 !
140 CLEAR @E5100 !
150 !
160 DIM Dat(1:201),Stim(1:201) ! VARIABLE DECLARATION
170 !
180 OUTPUT @E5100;"PRES" ! MEASUREMENT SETUP
190 OUTPUT @E5100;"CHAN1;ANAMODE GAINP;MEAS AR;FMT LOGM"
200 INPUT "Enter center frequency (MHz)",F_cent
210 INPUT "Enter frequency span (kHz)",F_span
220 OUTPUT @E5100;"CENT ";F_cent*1.E+6
230 OUTPUT @E5100;"SPAN ";F_span*1000
240 !
250 EXECUTE "SING" ! TRIGGERING MEASUREMENT
260 ! [MODULE 2-12]
270 !
280 OUTPUT @E5100;"POIN?" ! GETTING DATA USING BINARY TRANSFER
290 ENTER @E5100;Nop !
300 OUTPUT @E5100;"FORM3" !
310 OUTPUT @E5100;"OUTPRFORM?" !
320 ENTER @E5100 USING "#,8A";Header$
330 ENTER @Dt;Dat(*) !
340 ENTER @E5100 USING "#,1A";End$
350 !
360 OUTPUT @E5100;"OUTPSTIM?" ! TRANSFER STIMULUS VALUE IN BINARY
370 ENTER @E5100 USING "#,8A";Header$
380 ENTER @Dt;Stim(*) !
390 ENTER @E5100 USING "#,1A";End$
400 !
410 ASSIGN @Dt TO * ! DISPLAY TRANSFERED DATA
420 FOR I=1 TO Nop !
430 PRINT "POINT";I,Stim(I);"[Hz]",Dat(I);"[dB]"
440 NEXT I !
450 !
460 END ! END OF PROGRAM

```


Using the I/O Port

Introduction

The HP E5100A/B has a 24 bit I/O port on the rear panel. This section explains how to use this I/O port from BASIC.

The 24 bit I/O port has the following pin layout. See Appendix D 'Parallel I/O (Standard and Option 006)' of the *User's Guide*.




Figure 4-1. The I/O Port Pin Layout

Following is the description of how to use lines that can be controlled by Instrument BASIC.

Output Port (Pin 5 to 28)

The HP E5100A/B can select 8 types of output ports depending on the output data width. Note that port C (Pin 21 to 24) and port D (Pin 25 to 28) can also function as input ports.

There are 2 ways to output data to an I/O port.

- Use the **WRITEIO** statement of Instrument BASIC.
- Use **OUTAIO** to **OUTHIO** of the HP-IB command.

Data Output

Use the following procedure to output data:

1. If you want to use output ports other than A, B or F, you must specify ports C and D as output ports.

Ports Used	HP-IB Commands Executed
Port C (4 Bit)	COUT
Port D (4 Bit)	DOUT
Port E (8 Bit)	COUT,DOUT
Port G (20 Bit)	COUT
Port H (24 Bit)	COUT,DOUT

2. Output data from a controller through the I/O ports. Use the following commands to output data.

Ports Used	HP-IB Commands	Instrument BASIC
Port A (8 Bit)	OUTPAIO <i>data</i>	WRITEIO 15,0; <i>data</i>
Port B (8 Bit)	OUTPBIO <i>data</i>	WRITEIO 15,1; <i>data</i>
Port C (4 Bit)	OUTPCIO <i>data</i>	WRITEIO 15,2; <i>data</i>
Port D (4 Bit)	OUTPDIO <i>data</i>	WRITEIO 15,3; <i>data</i>
Port E (8 Bit)	OUTPEIO <i>data</i>	WRITEIO 15,4; <i>data</i>
Port F (16 Bit)	OUTPFIO <i>data</i>	WRITEIO 15,5; <i>data</i>
Port G (20 Bit)	OUTPGIO <i>data</i>	WRITEIO 15,6; <i>data</i>
Port H (24 Bit)	OUTPHIO <i>data</i>	WRITEIO 15,7; <i>data</i>

If you output data to port C or D (including ports E, G and H) when they are specified as input ports, a warning message **CAUTION:WRONG I/O PORT DIRECTION** will be displayed.

When data is outputted, a negative pulse will be outputted to the write strobe output (pin 31) of the output port. Monitor this output to synchronize the transfer when you are reading the data outputted to the I/O port from an external device.

Input Port (Pin 21 to 28)

As described in the previous section, input ports can also function as output ports. If you want to input data, you must set ports C and D to the input mode before processing the data. If ports C and D are set to the input mode, their I/O status line (pins 29 and 30) will be set to LOW. Be sure to check the I/O status lines (pins 29 and 30) of ports C and D to make sure they are set to the input mode when you want to import data from an outside source.

There are 2 ways to read data from an I/O port.

- Use the **READIO** statement of Instrument BASIC.
- Use **OUTPINPCIO?** to **OUTPINPEIO?** of the HP-IB command.

Data Input

Use the following procedure to input data from an external device:

1. Set ports C and D as input ports.

Ports Used	HP-IB Commands
Port C (4 Bit)	CIN
Port D (4 Bit)	DIN
Port E (8 Bit)	CIN,DIN

- Monitor the I/O status lines (pins 29 and 30) of ports C and D to make sure they are set to LOW. (If positive logic is specified by the **POSL** command, the input mode will be HIGH.)
- Input data to the input ports.
- Execute the following command from the controller to read the data inputted to the I/O ports.

Ports Used	HP-IB Commands	Instrument BASIC
Port C (4 Bit)	OUTPINPCIO? <i>data</i>	Data=READIO(15,2)
Port D (4 Bit)	OUTPINPDIO? <i>data</i>	Data=READIO(15,3)
Port E (8 Bit)	OUTPINPEIO? <i>data</i>	Data=READIO(15,4)

The INPUT1 Input (Pin 2) is provided to inform the HP E5100A/B of a data input. See the "INPUT1 Input (Pin 2), OUTPUT1 Output (Pin 3), OUTPUT2 Output (Pin 4)" commands for additional information.

INPUT1 Input (Pin 2), OUTPUT1 Output (Pin 3), OUTPUT2 Output (Pin 4)

To synchronize the HP E5100A/B and an external device at a measurement, the status lines are assigned to pins 2 to 4. Their functions are described in the following paragraphs.

INPUT1 Input (Pin 2) Used to send timing information from an outside source. If there is input to this pin, OUTPUT1 and OUTPUT2 are set to the specified status.

OUTPUT1 Output (Pin 3), Used to send the status of the HP E5100A/B to an external device.
OUTPUT2 Output (Pin 4) This is set to the status specified at the INPUT1 input. You can also set this to any status using an HP-IB command.

The status of OUTPUT1 and OUTPUT2 at the INPUT1 input is set by the following commands.

Commands to Set the Status of OUTPUT1 and OUTPUT2 at the INPUT1 Input

	LOW	HIGH
OUTPUT1	OUT1ENVL	OUT1ENVH
OUTPUT2	OUT2ENVL	OUT2ENVH

You can also use the following commands to set OUTPUT1 and OUTPUT2 to any status.

	LOW	HIGH
OUTPUT1	OUT1L	OUT1H
OUTPUT2	OUT2L	OUT2H

Examples of Data I/O

Let's look at an example of data I/O that uses the status line. OUTPUT1 and OUTPUT2 are used as the status indicators.

OUTPUT1 Set to HIGH when the HP E5100A/B completes a measurement and data process and is waiting for a trigger.

OUTPUT2 Set to HIGH while the HP E5100A/B is performing a measurement.

The following is a timing chart to describe the status.




Figure 4-2. INPUT1, OUTPUT1, OUTPUT2 Timing Chart

1. Specify the status of OUTPUT1 and OUTPUT2 at the INPUT1 input. See Figure 4-2 to set OUTPUT1 to LOW and set OUTPUT2 to HIGH.

```
OUTPUT @E5100;"OUT1ENVL"  
OUTPUT @E5100;"OUT2ENVH"
```

2. Start a measurement. Set a trigger at the INPUT1 input to perform a measurement. HIGH is set to OUTPUT1 to go back to the beginning of the loop to wait for a trigger when the data process is completed. You can check the timing of data input to the INPUT1 port by using the INPT? query.

```
ON ERROR GOTO Finished
LOOP
OUTPUT @E5100;"OUT1H"      Set OUTPUT1 to HIGH.
REPEAT
OUTPUT @E5100;"INPT?"     Wait for an INPUT1 input.
ENTER @E5100;Inpt         OUTPUT1 is set to LOW and OUTPUT2 is set to HIGH by the INPUT1 input.
UNTIL Inpt=1
:
:                          Set a trigger and perform a measurement.
:
OUTPUT @E5100;"OUT2L"     Set OUTPUT2 to LOW at the completion of measurement.
:
:                          Process the measured data.
:
END LOOP
Finished:!
```

Module 5-1. Using Data I/O

The Parallel I/O Mode A (Option 005 Only)

The HP E5100A/B Option 005 has an Input-4 Bit/Output-8 Bit I/O port. The following is a description of how to use the Option 005 I/O port from a BASIC program.

Data Output (4 Bit)

You can use the following methods to output data:

- For Instrument BASIC, use the **WRITEIO** statement.

```
WRITEIO 15,2;data
```

- For HP BASIC use an HP-IB command **OUT8IO**.

```
OUTPUT @E5100;"OUT8IO ";data
```

Data Input (8 Bit)

You can use the following methods to read inputted data:

- For Instrument BASIC, use the **READIO** statement.

```
Data=READIO(15,2)
```

- For HP BASIC use the HP-IB command **INP8IO?** or **OUTPINP8IO?**.

<pre>OUTPUT @E5100;"INP8IO?"</pre>	<i>Output data from the I/O port directly.</i>
<pre>ENTER @E5100;Data</pre>	<i>The return value is 4 bit data.</i>

Module 5-2. Reading Input Data

Or

<pre>OUTPUT @E5100;"INP8IO"</pre>	<i>Read data from the I/O port to the memory and output it from the memory.</i>
<pre>OUTPUT @E5100;"OUTPINP8IO?"</pre>	
<pre>ENTER @E5100;Data</pre>	<i>The return value is 8 bit data with 0s in its upper 4 bits.</i>

Module 5-3. Reading Input Data

Using Status Reporting Function

The analyzer has status registers that report system conditions. The register contents are changed depending on the particular condition of the analyzer. By reading this register, you can determine the specific analyzer status.

This chapter provides the following information:

- General status register model.
- Status register structure of the HP E5100A/B.
- How to use status register in a program.
- Sample program: performing calibration.

General Status Register Model

The analyzer has a status reporting system to report the condition of the analyzer.




Figure 5-1. General Status Register Model

The status reporting system has a hierarchical structure as shown in Figure 5-1. When the analyzer condition satisfies the particular condition, the corresponding bit of the event register is set "1". Therefore, you can check the analyzer condition by reading the event register.

When the event register bit is set to "1", and corresponding enable register bit is also "1", the summary bit of the status byte register is set to "1". You can read the status byte register by using the serial poll.

If the corresponding bit of the service request enable register is "1", the service request (SRQ) is generated with the positive transition of the status byte register bit. By generating the SRQ, you can notify the controller that the analyzer is requesting service.

Event Register

Reflects the correspondent analyzer condition as a bit status. These bits monitor the changing analyzer state continuously and change the bit status as required.

You cannot change bit status by HP-IB command.

The analyzer has the following event registers:

- Instrument Event Status Register.
- Standard Event Status Register.
- Operation Status Event Register.

Enable Register

The enable register selects which event register bits can set the bit in the summary bit of the status byte register that is connected to SRQ generation. The register bits work like mask bits. When you want to set a bit in the status byte register by a specific register condition, set the corresponding enable register to 1. This sets a 1 bit in the status byte register with a corresponding event register bit.

Use this register to select which event register bits generate the SRQ.

All event registers have a corresponding enable register for each bit.

Status Byte Register

If enabled event register is set to 1, the corresponding bit of the status byte register is set to 1. This register also indicates the output queue and SRQ status.

The value of the status byte register can be read by using the **SPOLL** statement or the ***STB?** query from the controller. **SPOLL** reads the status byte register value directly without the analyzer being set to remote. Therefore, you can continue to operate front panel keys while a controller is reading the status byte register. On the other hand, the ***STB?** query sets the analyzer to remote mode. Reading the status byte register by either command does not affect the contents of the status byte register, except that **SPOLL** clears the RQS bit.


A serial poll initiated by using the **SPOLL** command reads bit 6 of the status byte register as the RQS bit. The ***STB?** command reads bit 6 as the MSS bit.

SRQ (Service Request) can be generated by the status byte register by setting the service request enable register. For more information about SRQ, see "SRQ and Interrupt" in this chapter.

Transition Filter and Condition Register

The transition filter allows you to select which transitions of the analyzer condition will set a bit in the event register.

When the status register has a transition filter, there is a lower register called a condition register under the event register. The transition filter is between the event register and the condition register. The transition filter enables you to select a positive and/or negative transition of the condition register bit to set a bit in the corresponding event register. For example, if you set the negative transition filter, a 1 is set in the event register by changing from 1 to 0 in the event register.



C6304002

Figure 5-2. Transition Filter and Condition Register

For the HP E5100A/B, only the “Program Running” bit of the operation status register has a transition filter. By using the transition filter, you can generate an SRQ either at the start or the end of the program execution.

Status Register Structure

Status byte totals three status registers which indicate the internal condition of an instrument. Figure 5-3 shows the status reporting structure of the HP E5100A/B.




Figure 5-3. Status Reporting Structure

The HP E5100A/B has a status reporting system to report the condition of the instrument. Status bytes consist of 8-bit registers, each bit represents specific instrument conditions. The value of the status byte can be read by using SPOLL(717) statement from an external controller. This command reads value directly from the HP E5100A/B without being set to remote. So, you can operate front panel keys while a controller is reading the status byte. Contents of the status byte can also be read by using the *STB? command. Reading the status byte has no effect on the contents of the status byte. Table 5-1 shows contents of status byte.

Table 5-1. Status Bit Definitions of the Status Byte (STB)

Bit	Name	Description
2	Check event status register B	One of the enabled bits in event status register B has been set.
4	Message Available	"1" is set when Output Queue has data and "0" is set when Output Queue has no data.
5	Summary bit of event status register	One of the enabled bits in the event status register has been set.
6	MSS (Master Summary Status Bit)	One of the enabled status byte bits is causing an SRQ.
7	Operational status summary bit	One of the enabled bits in the operational status register has been set.

For example, to read the contents of Message in the output queue,

```
10 Stat=SPOLL(717)
20 Stb4=BIT(Stat,4)
30 PRINT Stb4
40 END
```

Figure 5-4. Example of Reading Status Byte (1)

or,

```
10 ASSIGN @E5100; TO 717
20 OUTPUT @E5100;"*STB?"
30 ENTER @E5100;Stat
40 Stb4=BIT(Stat,4)
50 PRINT Stb4
60 END
```

Figure 5-5. Example of Reading Status Byte (2)

The Event Status Register (ESR), Event Status register B (ESB), and Operational Status Register (OSR) are subordinate to the status byte. Each register is set a bit with condition which is watched by status bit. Status bit is cleared when is read by query or *CLS command is executed.

Table 5-2. Status Bit Definitions of the Event Status Register (ESR)

Bit	Name	Description
0	Operation complete	A command for which OPC has been enabled, and completed an operation.
2	Query error	<ol style="list-style-type: none"> 1. The HP E5100A/B has been addressed to talk, but there is nothing in the output queue to transmit. 2. Data in the Output Queue has been lost.
3	Device dependent error	An error, other than a command error, a query error, and an execution error has occurred.
4	Execution error	<ol style="list-style-type: none"> 1. A program data element following a header exceeded its input range, or is inconsistent with the HP E5100A/B's capabilities. 2. A valid program message could not be properly executed due to some instrument condition.
5	Command error	<ol style="list-style-type: none"> 1. An IEEE 488.2 syntax error has been occurred. Possible violations include, a data element violated the HP E5100A/B listening formats or a data element type is unacceptable to the HP E5100A/B. 2. A semantic error which indicates that an unrecognized header was received has occurred. Unrecognized headers include incorrect device-specific headers and incorrect or unimplemented IEEE 488.2 common commands.
7	Power on	A power-on sequence has occurred.

Table 5-3. Status Bit Definitions of the Event Status Register B (ESB)

Bit	Name	Description
0	Sweep or group complete, or cal std. complete	A single sweep or group has been completed since the last read of the register. Operates in conjunction with SING or NUMG.
3	Limit failed, 4	Limit test failed on trace 4.
4	Limit failed, 3	Limit test failed on trace 3.
5	Limit failed, 2	Limit test failed on trace 2.
6	Limit failed, 1	Limit test failed on trace 1.
10	Limit Passed, 4	Limit test passed on trace 4.
11	Limit Passed, 3	Limit test passed on trace 3.
12	Limit Passed, 2	Limit test passed on trace 2.
13	Limit Passed, 1	Limit test passed on trace 1.

Table 5-4.
Status Bit Definitions of the Operational Status Register (OSR)

Bit	Name	Description
14	Program running	An HP Instrument BASIC program is running.

Each status register has a register mask which enables generating Service ReQuest (SRQ) with condition of a status bit. For instance, to generate an SRQ when the HP E5100A/B completes the specified number of sweep, enable ESNB bit 1 which is the mask register for ESB 0 ("SING, NUMG, or Cal Std. Complete") which shows sweep completion and SRE bit 2. This makes a path from ESB bit 0 to an SRQ. This example is listed as a program listing:

```

10 ASSIGN @E5100 TO 717
20 !
30 OUTPUT @E5100;"CLES" ! Clears status registers
40 OUTPUT @E5100;"ESNB 1" ! Enables mask register of "SING. NUMG. or
50 ! ! Cal Std. Complete" of ESB
60 OUTPUT @E5100;"*SRE 4" ! Enables mask register of "Event Status
70 ! ! Register B" of STB
80 !
90 ON INTR 7 GOTO End ! Declare SRQ interrupt
100 ENABLE INTR 7;2
100 OUTPUT @E5100;"SING" ! Execute single sweep
110 GOTO 110 ! Endless loop
120 !
130 End: ! Exit from loop when sweep is completed
140 END

```

Figure 5-6. Example of Generating a Service ReQuest (SRQ)

OSPT, OSNT

OSPT (Operational Status Positive Transition Filter)

Sets the positive transition filter. Setting a bit in OSPT will cause a 0 to 1 transition in the corresponding bit of the associated operational status register (OSR) to cause a 1 to be written in the associated bit of corresponding operational status event register (OSER).

Because only bit 17 of the HP E5100A/B's OSR is used to show program status, when bit 17 of OSPT is set to 1, starting a program causes a 1 to be written in bit 17 of OSER. (And then a 1 is written in bit 7 of STB.)

OSNT (Operational Status Negative Transition Filter)

Sets the negative transition filter. Setting a bit in the negative transition filter will cause a 1 to 0 transition in the corresponding bit of the associated operational status register to cause a 1 to be written in the associated bit of corresponding operational status event register.

Because only bit 17 of the HP E5100A/B's OSR the is used to show program status, when bit 17 of OSNT is set to 1, stopping a program causes a 1 to be written in bit 17 of OSER. (And then a 1 is written in bit 7 of STB.)

How to Use the Status Registers in a Program

You can use the status registers to determine the specific analyzer status in the program. To determine the contents of the status register, the following methods are used:

- Read an event register directly.
- Use the Service Request (SRQ).

Reading an Event Register Directly

You can read the contents of the event register directly to determine the specific analyzer condition. Use this method if you do not need to know the timing of the event register changes. The following procedure reads the register directly:

1. Query the event register contents.
2. Retrieve a return value.
3. Check the bit condition using the BASIC **BIT** function.

OUTPUT @E5100;"ESB?"	<i>Queries instrument event status register contents.</i>
ENTER @E5100;Esb	<i>Retrieve return value.</i>
IF BIT(Esb,4) THEN	<i>If bit 4 of the instrument event status register is set to 1, the limit</i>
DISP "LIMIT TEST FAILED AT Ch 1."	<i>test failed on channel 1.</i>
END IF	

Module 4-1. Reading an Event Register

- Related HP-IB Commands. The following query commands can be used to read the contents of an event register directly.

*STB?	Returns Status Byte Register contents.
*ESR?	Returns Event Status Register contents.
ESB?	Returns Instrument Event Status Register contents.
OSR?	Returns Operation Status Register contents.

SRQ and Interrupt

You can initialize your program to enable interrupt processing by the Service Request (SRQ) from the analyzer. The analyzer generates an SRQ when the specified condition is satisfied.

The SRQ itself does not contain information on the SRQ source. However, the Request Service (RQS) bit in the Status Byte Register of the SRQ source device is set to 1. If multiple devices are connected on the bus, you can check the RQS bit (bit 6) of the analyzer by using a serial poll, **SPOLL**.

Use the SRQ interrupt if you want to determine when the condition changes. The following procedure is used to set the SRQ interrupt:

1. Define the branch for the interruption. (Use **ON INTR** statement.)
2. Set the enable register for the correspondent event register bit to 1.
3. Set the service request enable register bit for the correspondent status byte register bit to 1.
4. Clear the status register before enabling the SRQ interruption.
5. Enable the SRQ interruption. (Use **ENABLE INTR** statement.)
6. Start the event.
7. Wait for the SRQ. Usually, the program waits within an endless loop.
8. If multiple devices that can generate an SRQ exist on the HP-IB, you should check bit 6 of the status byte register of the target device. If the SRQ is generated from the target device, the status byte register bit 6 is set to 1.

The following example uses an SRQ interruption for detecting the end of sweep. Bit 0 of the instrument event status register is used for this purpose.

ON INTR Code GOTO Sweep_end	<i>When the SRQ has occurred, jumps to label, "Sweep_end."</i>
OUTPUT @E5100;"ESNB 1"	<i>Set bit 0 of the instrument event status enable register to 1. ($2^0 = 1$)</i>
OUTPUT @E5100;"*SRE 4"	<i>Set bit 2 of the service request enable register to 1. ($2^2 = 4$)</i>
OUTPUT @E5100;"*CLS"	<i>Clears the event register.</i>
OUTPUT @E5100;"*OPC?"	<i>Confirms the clear operation is completed.</i>
ENTER @E5100;0pc	
!	
OUTPUT @E5100;"HOLD"	<i>Sets the trigger mode HOLD.</i>
ENABLE INTR Code;2	<i>Enables the SRQ interruption just before triggering.</i>
OUTPUT @E5100;"SING"	<i>Trigger a measurement.</i>
Waiting: GOTO Waiting	<i>Waits until SRQ is generated.</i>
Sweep_end:!	<i>When on SRQ, program jumps to this label.</i>
IF NOT BIT(SPOLL(@E5100),6) THEN	<i>Check the SRQ is generated from the target device by checking status byte register bit 6.</i>
ENABLE INTR Code;2	<i>If not, enable the SRQ again, then</i>
GOTO Waiting	<i>returns to the endless looping.</i>
END IF	


Module 4-2. Detecting Sweep End Using SRQ and Interrupt

Note



*CLS clears only the event registers and the status byte register. The enable register and transition filter settings are not altered by executing the *CLS command. To clear the enable register and transition filter, use the **PRES** command.

Figure 5-7 shows the SRQ generation sequence of the example above.



C7204005

Figure 5-7. SRQ Generation Sequence

■ Related HP-IB Commands

The following HP-IB commands are used for setting the SRQ generation:

- *SRE *decimal* Sets the service request enable register.
- *ESE *decimal* Sets the enable register for event status register.
- ESNB *decimal* Sets the enable register for instrument event status register.
- OSE *decimal* Sets the enable register for operation status register.
- OSPT *decimal* Sets the transition filter to positive for operation status register.
- OSNT *decimal* Sets the transition filter to negative for operation status register.

Programming Miscellaneous

Introduction

This chapter describes methods to use the HP E5100A/B, Instrument BASIC and an external controller together and some notes on the operation. The later sections introduce techniques for programming and hints to increase process speed.

This chapter describes the following topics:

- Using an External Controller and the HP E5100A/B Together
- Controlling Instrument BASIC from an External Controller
- Programming Techniques

Using an External Controller and the HP E5100A/B Together

The HP E5100A/B has the ability to function as a controller to control other HP-IB devices. If you want to use an external controller, you can also use both of them on the same bus. This section explains how to use the two controllers on the same bus.

Locking Out Local Operation (LOCAL LOCKOUT)

If you are controlling the HP E5100A/B with an external controller, the HP E5100A/B is in the remote mode. In the remote mode, the front panel key operations are not accepted. You can, however, press the **LOCAL** key to cancel the remote mode. To prevent the operator from pressing the **LOCAL** key by mistake and changing the device setup, you can make all keys (including the **LOCAL** key itself) unavailable. This status is called Local Lockout.

You can set local lockout as follows.

ASSIGN Scode TD 7 LOCAL LOCKOUT Scode
--

Module 6-1. Setting Local Lockout

You can cancel local lockout by sending the **LOCAL** command.

LOCAL Scode	<i>Cancel the all devices on the bus.</i>
LOCAL @E5100	<i>Cancel any one device.</i>

Module 6-2. Canceling Local Lockout

How To Pass Control (PASS CONTROL)

Though more than one controller can be on the same bus, only one controller can actually control the bus at any one time. The controller that has the right to control the bus is called the Active Controller.

The controller that has control at system startup is called the System Controller. The system controller is specified physically and it cannot be controlled and changed by a program. If you want to use more than one controller, you must pass control to the other controller. The **PASS CONTROL** statement of BASIC is used to transfer control to another controller.

```
PASS CONTROL @E5100
```

Module 6-3. Pass Control

Only the active controller can execute this statement.

The system controller can become the active controller at any time by executing the ABORT statement.

```
ABORT 7
```

Module 6-4. Abort

Only the system controller can execute this statement.

Controlling Instrument BASIC From an External Controller

The HP E5100A/B has a command called the Program Subsystem command that enables an external controller to control Instrument BASIC. The program subsystem command enables you to execute a program written in Instrument BASIC or refer to the variables used by Instrument BASIC from an external controller. This section will describe the following topics related to the program subsystem command usage.

- Referring to and Transferring an Array
- Remote Processing a BASIC Command
- Transferring a Program

Note The following commands and programs are all executed on an external controller.



Note The program subsystem command is based on SCPI (Standard Commands for Programmable Instrument).



Referring to and Transferring an Array

You may need to exchange data between programs when you are using an external controller and Instrument BASIC together. This section describes how to use the program subsystem command to exchange data.

Referring to a Numeric Variable

Execute the following statements to refer to numeric variable data in Instrument BASIC from an external controller:

```
OUTPUT @E5100;"PROG:NUMB? 'Dat'"
ENTER @E5100;Dat
```

Module 6-5. Referring to a Numeric Variable

Referring to a String Character Variable

Similarly, refer to a string character variable as follows:

```
OUTPUT @E5100;"PROG:STR? 'String$'"
ENTER @E5100;String$
```

Module 6-6. Referring to a String Character Variable

Transferring a Numeric Variable

Execute a command with a variable name and data as a parameter to transfer numeric data to a variable in Instrument BASIC.

```
OUTPUT @E5100;"PROG:NUMB 'Center',100000000"
INPUT "ENTER CENTER FREQUENCY",Center
OUTPUT @E5100;"PROG:NUMB 'Center',";Center
```

Module 6-7. Transferring a Numeric Variable

Transferring a String Character Variable

Execute a command with a variable name and data as a parameter to transfer string character data to a variable in Instrument BASIC.

```
OUTPUT @E5100;"PROG:STR 'File$','TEST1'"
File$="TEST1"
OUTPUT @E5100;"PROG:STR 'File$','';File$;''"
```

Module 6-8. Transferring a String Character Variable

Referring to and Transferring an Array

Similarly, you can refer to the contents of an array.

```
DIM Dat_array(1:201,1:2)           Define an array of the same size as the array you are
                                   referring to.
OUTPUT @E5100;"PROG:NUMB? 'Dat_array'"
ENTER @E5100;Dat_array(*)
```

Module 6-9. Transferring Array

You cannot refer to an array by each element (Ex. **PROG:NUMB? 'Dat_array(1,1)'**).

Execute a command with an array name and data as a parameter to transfer array data.

```
OUTPUT @E5100;"PROG:NUMB ""Dat_array"",";Dat_array(*)
```

Module 6-10. Referring to Array

In fact, the program is more complex because you have to synchronize two programs. See "Sample Program -3: Controlling Instrument BASIC from an External Controller".

Remote Processing a BASIC Command (PROG:EXEC)

You can remote process a BASIC command supported by the HP E5100A/B's BASIC command lines.

Following is an example of how to use the **PROG:EXEC** command.

```

OUTPUT @E5100;"PROG:EXEC 'EDIT'"           Start up the BASIC editor.
OUTPUT @E5100;"PROG:EXEC 'MSI "":INTERNAL,4""'" Select the internal disk drive.
INPUT "Enter File Name:",File$
OUTPUT @E5100;"PROG:EXEC 'GET """;File$;"'" Read a file into the editor.

```

Module 6-11. PROG:EXEC Command

Controlling the Process Status

You can check or set the process status of a BASIC program from an external controller.

Controlling the Process Status of Instrument BASIC

You can use the **PROG:STAT** command to control the process status of Instrument BASIC from an external controller.

```

OUTPUT @E5100;"PROG:STAT RUN" Execute a program.

```

Module 6-12. PROG:STAT Command

See the *HP-IB Command Reference* for detail.

Checking the Process Status of an Instrument BASIC Program

The following are examples of how to remote check the process status of Instrument BASIC.

- Use the **PROG:STAT?** Query.
Depending on the return value, you can see if the process is in the **RUN**, **PAUSE** or **STOP** status.

```

OUTPUT @E5100;"PROG:STAT?" Use Query to get the process status.
ENTER @E5100;Status$      The return value is character string.
DISP Status$

```

Module 6-13. PROG:STAT? Query

- Check the bit 14 of OSR (Operation Status Register).
B 14 of the OSR is set depending on the process status of Instrument BASIC. A 1 is set if the process is running, or a 0 is set if it is stopped (including the PAUSE status).

```

OUTPUT @E5100;"OSR?"
ENTER @E5100;Osr
BIT(Osr,14)

```

Module 6-14. Checking Operation Status Register

When you want to generate an SRQ according to the process status of Instrument BASIC, you can select whether you want to generate it at the start of program or at the completion of program by specifying OSPT or OSNT.

<pre>OUTPUT @E5100;"OSPT 16384" OUTPUT @E5100;"OSE 16384; *SRE 32"</pre>	<i>Generate an SRQ at the start of program.</i>
--	---

Module 6-15. Generating SRQ

Change **OSPT** to **OSNT** to generate an SRQ at the completion of program.

Note

CLES or ***CLS** cannot clear **OSPT** or **OSNT**. Execute **OSPT 0** or **OSNT 0** to clear the setup.



- Use the **PROG:WAIT?** command to pause an external controller until a program completes. 1 is returned at the completion of running Instrument BASIC if you execute the **PROG:WAIT?** query. If you read the return value by the **ENTER** statement, the program will pause until the value is returned.

<pre>OUTPUT @E5100;"PROG:WAIT?" ENTER @E5100;Wt</pre>	<i>Wait for a program to complete. 1 is returned if the program is completed.</i>
---	---

Module 6-16. PROG:WAIT Command

- Use Instrument BASIC to set the variable to indicate the program status and refer to this variable by the **PROG:NUMB?** or **PROG:STR?** query from an external controller. This is useful when you want to know the actual process status of an Instrument BASIC program. Ex. Check if a calibration is completed. You can have controllers to cooperate and communicate with each other in detail if you use the **CASE** statement in an external controller to branch to appropriate process depending on the contents of this variable. A sample program that uses this method is in "Sample Program -3: Controlling Instrument BASIC from an External Controller".

Transferring a Program

You can transfer a program between an external controller and Instrument BASIC.

Transferring from an External Controller to Instrument BASIC

The following example transfers a program on the disk in an external controller to the Instrument BASIC editor through HP-IB.

```

OUTPUT @E5100;"PROG:DEL:ALL"  Clear the editor before a transfer.
ASSING @File TO Filename$
ON ERROR GOTO Done
DIM LINE$[100]
OUTPUT @E5100;"PROG:DEF #0"  Send the header.
LOOP
  Line$=""
  ENTER @File;Line$          Read and transfer a program line by line.
  OUTPUT @E5100;Line$
END LOOP
DONE: !                      If the control gets to the end of file, it jumps to this label by an error
                              interrupt.

OFF ERROR
OUTPUT @E5100;"END"         Send a terminator and end the transfer.

```

Module 6-17. Transferring a Program to the Instrument BASIC Editor

Instrument BASIC checks the transferred program and comments out lines that have grammatical errors.

Transferring from Instrument BASIC to an External Controller

You can also transfer a program from the Instrument BASIC editor to an external controller. The following example stores each Instrument BASIC line to a character string array **PROG\$**.

```

DIM PROG$(1:3000)[100]      Up to 3000 line program can be stored.
!
ON ERROR GOTO Finished
ENABLE INTR Scode;2
OUTPUT @E5100;"PROG:DEF?"
ENTER @E5100 USING "#,8A";Heade$
ENTER @E5100;PROG$(*)
!
Finished: OFF ERROR        In case a program does not have exactly 3000 lines, use an error
                              interrupt to prevent the program from abnormally exiting.


```

Module 6-18. Transferring a Program From the Instrument BASIC Editor

Sample Program -3: Controlling Instrument BASIC from an External Controller

This program runs an external controller and Instrument BASIC at the same time to synchronize the function of both sides. To check the process status of Instrument BASIC from an external controller, you must set up a variable (**Stat\$**) to indicate the program status of Instrument BASIC. In the external controller, use the **SELECT ... CASE** statement to process data according to the contents of **Stat\$**.

Actually, Instrument BASIC handles processes such as setups, calibrations, measurements and data analysis and then transfers the results to the external controller. The external controller requests the process at each step.



013,0301

Figure 6-1. Controlling Instrument BASIC from an External Controller (The Control Flow)

The Program for Instrument BASIC

Disk



This program is contained in the attached sample disk with the file name
IB_CTRL.

```

110  ASSIGN @E5100 TO 800          ! IBASIC INITIALIZATION
120  Scode=8                      !
130  CLEAR @E5100                !
140  !                            !
150  Stat$="SETTING"             !
160  Cal$=""                     !
170  Center=0                   !
180  Span=0                      !
190  !                            !
200  OUTPUT @E5100;"DISAHIHB"    ! MEASUREMENT SETUP
210  OUTPUT @E5100;"PRES"       !
220  OUTPUT @E5100;"CHAN1;ANAMODE GAINP;MEAS AR;FMT LOGM"
230  OUTPUT @E5100;"HOLD"       !
240  !                            !
250  EXECUTE "ANAOCH1"          ! SETUP WAVEFORM ANALYSIS
260  EXECUTE "ANARFULL"        !
270  EXECUTE "ANAODATA"        !
280  !                            !
290  PAUSE                      !
300  !                            !
310  OUTPUT @E5100;"CENS ";Center,Span
320  !                            !
330  Stat$="CALIBRATION"        !
340  PAUSE                      !
350  !                            !
360  IF Cal$="Y" OR Cal$="y" THEN GOSUB R_cal
370  !                            !
380  LOOP                        !
390  Stat$="CONNECT DEVICE"     !
400  DISP "CONNECT DEVICE"     !
410  PAUSE                      !
420  !                            !
430  Stat$="MEASUREMENT"       !
440  DISP "NOW MEASURING..."  !
450  EXECUTE "SING"            !
460  !                            !
470  WRITEIO 8,0;-3            !
480  EXECUTE "OUTPFILT?"       !
490  DIM Ret(5)                 !
500  FOR I=0 TO 5              !
510  Ret(I)=READIO(8,I)        !
520  NEXT I                     !
530  !                            !
540  PAUSE                      !
550  END LOOP                  !
560  !                            !
570  STOP                      !

```

```

580      !
590 R_cal:                               !
600      Stat$="CONNECT R"              !
610      DISP "Connect THRU"           !
620      PAUSE                          !
630      DISP                            !
640      OUTPUT @E5100;"CALI RESP"     !
670      OUTPUT @E5100;"STANC?"       !
680      ENTER @E5100;Tmp              !
690      !                               !
730      OUTPUT @E5100;"RESPDONE"     !
740      ENTER @E5100;Tmp              !
760      RETURN                         !
770      !                               !
780      END                            ! END OF PROGRAM

```

The Program for the External Controller

Disk



This program is contained in the attached sample disk with the file name **IBCTRL E.**

```

110  ASSIGN @E5100 TO 717                ! EXT CONTROLLER INITIALIZATION
120  Scode=7                             !
130  ABORT Scode                         !
140  !                                    !
150  INPUT "Center Frequency (MHz) ?",Center ! SWEEP CONDITION ENTRY
160  Center=Center*1.E+6                 !
170  INPUT "SPAN (kHz) ?",Span          !
180  Span=Span*1000.                    !
190  !                                    !
200  OUTPUT @E5100;"CLES"               !
210  OUTPUT @E5100;"*OPC?"             !
220  ENTER @E5100;Opc                   !
230  !                                    !
240  OUTPUT @E5100;"PROG:STAT RUN"      !
250  !                                    !
260  OUTPUT @E5100;"OSNT 16384;
      OSPT 0"                            !
270  OUTPUT @E5100;"OSE 16384;*SRE 128" !
280  ON INTR Scode GOTO Paused          !
290  !                                    !
300  Begin:                             !
310  WAIT .5                             !
320  ENABLE INTR Scode;2                !
330  LOOP                               !
340  END LOOP                           !
350  Paused:                             !
360  !                                    !
370  OUTPUT @E5100;"PROG:STR? 'Stat$'"  !
380  ENTER @E5100;Stat$                 !
390  !                                    !
400  SELECT Stat$                       !

```

```

410 CASE ""SETTING"" !
420 OUTPUT @E5100;"PROG:NUMB'CENTER'," ;Center
430 OUTPUT @E5100;"PROG:NUMB'SPAN'," ;Span
440 GOSUB Continue !
450 !
460 CASE ""CALIBRATION"" !
470 INPUT "CAL?",Cal$ !
480 OUTPUT @E5100;"PROG:STR 'Cal$'," ;Cal$;"'"
490 GOSUB Continue !
500 !
510 CASE ""CONNECT R"" !
520 INPUT "Connect THRU, Then Press [ENTER]",$
530 GOSUB Continue !
540 !
550 CASE ""CONNECT DEVICE"" !
560 INPUT "Connect Device, Then Press [ENTER]",$
570 GOSUB Continue !
580 !
590 CASE ""MEASUREMENT"" !
600 DIM Ret(5) !
610 OUTPUT @E5100;"PROG:NUMB? 'RET'"
620 ENTER @E5100;Ret(*) !
630 CLEAR SCREEN !
640 PRINT "LOSS:",Ret(0)," [dB],BW:",Ret(1)," [Hz]"
650 PRINT "CENT Freq:",Ret(2)," [Hz],Q:",Ret(3)
660 PRINT "D.LF:",Ret(4)," [Hz],D.RF:",Ret(5)," [Hz]"
670 GOSUB Continue !
680 !
690 END SELECT !
700 GOTO Begin !
710 !
720 Continue: !
730 OUTPUT @E5100;"CLES" !
740 OUTPUT @E5100;"*OPC?" !
750 ENTER @E5100;Opc !
760 OUTPUT @E5100;"PROG:STAT CONT"
770 RETURN !
780 END ! END OF PROGRAM

```

Programming Techniques

This section provides information that you may find useful when you write a program to control the HP E5100A/B using BASIC.

This section provides the following information.

- Using a Disk
- Using a Softkey Label
- Measuring the Processing Time
- Checking HP-IB Errors
- Tips for Increasing Speed

Using a Disk

The HP E5100A/B has 2 storage devices: a disk drive and a RAM disk drive. There are 2 ways to access these storage devices from a controller.

- Accessing Via the Measurement Module

Use this method to save or read information, such as the setup of the HP E5100A/B, by using the HP-IB command.

The following program saves measurement setup data to the internal disk with a specified file name.

```

10  ASSIGN @E5100 TO 717
20  INPUT "Enter File Name (without extension)",File_name$
30  OUTPUT @E5100;"STODDISK"
40  OUTPUT @E5100;"SAVDSTA """;File_name$;"""""
50  END

```

Use STODMEMO for the RAM disk.

Module 6-19. Saving Measurement Setup Data

The following program loads the saved setup data to the HP E5100A/B.

```

10  ASSIGN @E5100 TO 717
20  INPUT "Enter File Name (with extension)",File_name$
30  File_name$=UPC$(File_name$)
40  OUTPUT @E5100;"STODDISK"
50  OUTPUT @E5100;"RECD """;File_name$;"""""
60  END

```

You must specify a file name in capital letters.

Module 6-20. Loading Measurement Setup Data

- Accessing Directly from Instrument BASIC

This method accesses a file directly without going through the HP E5100A/B. You use this method when you are handling file data itself. See Chapter 7 "Data Storage and Retrieval" of *HP Instrument BASIC Programming Technique* for more information.

The following program stores trace data on the internal disk into an array **Dat** in the program.

```

10 ASSIGN @E5100 TO 717
20 INPUT "Enter File Name (without extension)",File_name$
30 MSI ":INTERNAL,4" Use MSI ":MEMORY,0" for the RAM disk.

40 DIM Dat(1:201,1:2)
50 File_name$=UPC$(File_name$)+"_D"
60 !
70 ASSIGN @File TO File_name$
80 ENTER @File USING "17X,#" Read the header.
90 ENTER @File;Nop
100 ENTER @File USING "4X,#" Store the data into an array.
110 ENTER @File;Dat(*)
120 ASSIGN @File TO *
130 PRINT Dat(*)
140 END

```

Module 6-21. Storing Trace Data

Note



The setups **STODDISK/STODMEMO** and **MSI** are stored independently. This causes the RAM disk to be accessed by an HP disk access command, as long as **STODMEMO** is executed, even though the internal disk drive is selected by **MSI**.

Using a Softkey Label (ON KEY LABEL)

You can use the **ON KEY LABEL** statement to display your own softkey label.

```

ON KEY 1 LABEL "ORIGINAL LABEL" GOSUB Jump1
!
LOOP
END LOOP

```

Module 6-22. Using a Softkey Label

The label is displayed only when the program is running. You must, therefore, use an infinite loop to keep the programming running.

Up to a 20-character softkey label can be displayed.

Displaying the Softkey Label

The softkey label displayed by **ON KEY LABEL** is displayed when **SYSTEM** **I-BASIC** **ON KEY LABELS** are pressed.

You can use the HP-IB command **BASL** to display the softkey label.

```

OUTPUT @E5100;"BASL" Display the softkey label.

```

Module 6-23. Using Softkey Label

See "Sample Program -4: Reading and Running a Program by a Softkey Operation" which uses the softkey label.

Measuring the Processing Time (TIMEDATE)

You can use the real time clock of the HP E5100A/B to time the processing of a program.

You can time a process as follows:

```

Start_time=TIMEDATE           The start time.
:
[The processes to time.]
:
End_time=TIMEDATE           The end time.
Total=End_time-Start_time    Calculate the elapsed time.
PRINT Total,"seconds of processing"

```

Module 6-24. Measuring Processing Time

The real time clock measures in seconds [s].

Checking HP-IB Errors (OUTPERRO?)

You can check if an HP-IB error occurred in a program by using the **OUTPERRO?**Query. You can use this function to easily find bugs when you are debugging a program. You can check for an HP-IB error as follows:

```

DIM Err$(50)
!
OUTPUT @E5100;"OUTPERRO?"
ENTER @E5100;Err$
DISP Err$

```

Module 6-25. Checking HP-IB Errors

For example, the following message will be displayed if you use the above procedure to check an error occurred when you send a wrong HP-IB command.

```
-113,"Undefined header"
```

You can look at "Error Messages" in the *HP-IB Command Reference* to see if the command sent is supported by the HP E5100A/B.

An error message is stored in the error queue. Execute ***CLS** or **CLES** to clear the error queue.

Tips for Increasing Speed

This section describes how to increase the processing speed of a program. Please modify your program if you find that you can use some of these suggestions.

- Try not to use markers for a data analysis. If you have to use them, use the **MARD OFF** command to erase the marker display.
- Use **EXECUTE** (only for Instrument BASIC).

Note



The **EXECUTE** process may not complete successfully if an external controller sends an HP-IB command while Instrument BASIC is processing another command using the **EXECUTE** statement. The HP E5100A/B may even stop in certain circumstances. To prevent this, do not send a new command from an external controller while the HP E5100A/B is processing a command.

See "Sample Program -3: Controlling Instrument BASIC from an External Controller" for an actual method.

- For Instrument BASIC, do not use the **ON INTR** interrupt. Do not use an interrupt process but refer to the status register directly.

```
REPEAT
  OUTPUT @E5100;"*ESB?"
  ENTER @E5100;Esb
UNTIL BIT(Esb,0)           Monitor the completion of sweep.
```

Module 6-26. Referring to Status Register

- Try not to switch active channels. Turn dual channels on if you need switch channels.

The following commands are valid for inactive channels.

OUTPIFORM? Output the format array data on the inactive channel.
OUTPIRFORM? Output the actual part of format array data on the inactive channel.
OUTPITMEM? Output the sub trace data on the inactive channel.
OUTPIRTMEM? Output the actual part of sub trace data on the inactive channel.

The trigger command is valid for the active channel. You can, however, use the **REST** command (**MEASUREMENT RESTART** of **MENU**) to remeasure both the active channel and the inactive channel without switching.

- Use subroutines instead of subprograms.
- Comment out or delete unnecessary lines.

Sample Program -4: Reading and Running a Program by a Softkey Operation

This program displays the file names contained in the inserted disk on the softkey labels and executes the selected file as a program. With this program, you need only softkey operation to easily execute a program.

If you rename this file as **AUTOST**, it will be automatically executed at the startup. This is convenient when no keyboard is connected to the instrument.

Disk



This program is contained in the attached sample disk with the file name **LABEL** (For Instrument BASIC). The program for an external controller is not provided.

```

110 ASSIGN @E5100 TO 800          ! INITIALIZATION
120 DIM Dir$(1:200)[80],File$(1:200)[10]
130 !
140 CAT TO Dir$(*)              ! STORE FILENAMES ON DISK
150 !
160 File_end=0                  ! READ FILENAMES
170 File_number=1              !
180   WHILE File_end=0 AND File_number<200
190     File$(File_number)=Dir$(File_number+7)[1,10]
200     IF File$(File_number)="" THEN
210       File_end=1            !
220       File_number=File_number-1 !
230     ELSE                    !
240       File_number=File_number+1 !
250     END IF                  !
260   END WHILE                !
270 !
280 Max_page=INT(File_number/6)+1 ! SET ON-KEY-LABEL DISPLAY
290 Npage=1                    !
300 OUTPUT @E5100;"BASL"      !
310 !
320 Head:                      !
330 Page=(Npage-1)*6
340 !
350   ON KEY 1 LABEL File$(Page+1) GOSUB Jump1 ! DISPLAY FILENAMES ON LABELS
360   ON KEY 2 LABEL File$(Page+2) GOSUB Jump2
370   ON KEY 3 LABEL File$(Page+3) GOSUB Jump3
380   ON KEY 4 LABEL File$(Page+4) GOSUB Jump4
390   ON KEY 5 LABEL File$(Page+5) GOSUB Jump5
400   ON KEY 6 LABEL File$(Page+6) GOSUB Jump6
410   ON KEY 7 LABEL "NEXT PAGE" GOTO Jump7
420   ON KEY 8 LABEL "PREV.PAGE" GOTO Jump8
430 !
440   LOOP                      ! WAIT KEY ENTRY
450   END LOOP                  !
460 !
470 Jump1:GET File$(Page+1)    ! GET SELECTED FILE
480 Jump2:GET File$(Page+2)    !
490 Jump3:GET File$(Page+3)    !

```

```
500 Jump4:GET File$(Page+4)      !
510 Jump5:GET File$(Page+5)      !
520 Jump6:GET File$(Page+6)      !
530 Jump7:Npage=Npage+1         !
540   IF Npage>Max_page THEN Npage=Max_page
550   GOTO Head                   !
560 Jump8:Npage=Npage-1         !
570   IF Npage<=1 THEN Npage=1   !
580   GOTO Head                   !
590   !                           !
600 END                           ! END OF PROGRAM
```

Using the Parallel Processing

Because the HP 87510A performs measurement and analysis sequentially, the total throughput time is summary of the measurement time plus the time required to analyze the measurement data. However, the HP E5100A/B can perform the measurement and analyze the data simultaneously. This parallel processing improves the total throughput as shown in the following figure:




Figure 6-2. Parallel Processing

The parallel processing flow is as follows: The digital signal processor (DSP) measures the DUT, and at the same time the CPU processes the previous measurement's data in parallel with the current measurement. Parallel processing is useful for a production line that successively measures and analyzes many devices and performs GO/NO-GO testing. It cannot be used to make a single measurement and analyze the data at the same time.

Programming for Parallel Processing

The following commands are provided for parallel processing:

TRIGMEAS : Triggers the measurement for parallel processing. The DSP starts measuring the DUT.

MOVADARY : Enters measurement data triggered by the TRIGMEAS command to a data array. The CPU then analyzes this data in parallel with the next measurement. This command must be executed for each measurement.

ADTOTRAC : Starts data process into the data trace array. This command must be executed for each measurement.

The following program example shows a parallel processing program that uses these commands:

```

OUTPUT @E5100;"TRIGMEAS" ! 1st. measurement
LOOP
OUTPUT @E5100;"MOVADARY" ! Data transfer
OUTPUT @E5100;"TRIGMEAS" ! Start measuring
OUTPUT @E5100;"ADTOTRAC" ! Start parallel processing
!
OUTPUT @E5100;"PEAK?" ! Analysis command in parallel processing
ENTER @E5100;F_pk,G_pk ! Detects two peaks
!
OUTPUT @E5100;"NXPEAK?" !
ENTER @E5100;F_nxpk,G_nxpk!
!
END LOOP

```

Module 6-27. Parallel Processing

The following commands must be programmed as one group.

```

OUTPUT @E5100;"MOVADARY" ! Data transfer
OUTPUT @E5100;"TRIGMEAS" ! Start measuring
OUTPUT @E5100;"ADTOTRAC" ! Start parallel processing

```

Commands that are executed during parallel processing should be put after ADTOTRAC.

Sample Program -4a :Parallel Processing

The following is a program listing that shows how to use parallel processing. This program executes the analysis commands during the time the DSP measures the DUT. As a result, the total throughput time is improved.

Disk



This program is contained in the attached sample disk with the file name **PARA** (for Instrument BASIC). The program for an external controller has the name **PARA_E**.

```

110 ! CONCURRENT PROCESS
120 !
130 !
140 ASSIGN @E5100 TO 800
150 CLEAR @E5100
160 DIM Pnext(1:10),Fnext(1:10)
170 Ll=5
180 !
190 GOSUB Setting
200 GOSUB THRUICAL
210 GOSUB Measure
220 STOP
230 !
240 Setting: !
250 Fc=7.E+7 !70MHZ
260 Fspan=100000 !100kHz
270 Ifbw=8000 !8kHz
280 Nop=201

```

```

290  !
300  OUTPUT @E5100;"PRES"
310  OUTPUT @E5100;"POIN ";Nop
320  OUTPUT @E5100;"CHAN1;ANAMODE GAINP" MEAS AR;FMT LOGM"
330  OUTPUT @E5100;"HOLD"
340  OUTPUT @E5100;"CENT ";Fc;";SPAN ";Fspan
350  OUTPUT @E5100;"IFBW ";Ifbw
360  RETURN
370  !
380  ThruCal:  !
390  OUTPUT @E5100;"CALI RESP"  !Select and execute THRU CAL
400  OUTPUT @E5100;"STANC?"
410  ENTER @E5100;Tmp
420  OUTPUT @E5100;"RESPDONE?"
430  ENTER @E5100;Tmp
440  RETURN
450  !
460  Measure:  !
470  EXECUTE "ANAOCH1"  ! Select channel for analysis and set analysis range
480  EXECUTE "ANARFULL"  !
490  EXECUTE "ANAODATA"  !
500  !
510  N=100
520  INPUT "Enter loop N :","N
530  Tstart=TIMEDATE
540  EXECUTE "TRIGMEAS"  !1st. measurement
550  I=1
560  WHILE I<=N          !Repeat measurement and analysis
570  EXECUTE "MOVADARY"  !Enters measurement data for parallel processing
580  EXECUTE "TRIGMEAS"  ! Triggers measurement for parallel processing
590  EXECUTE "ADTOTRAC"  !Starts parallel processing
600  EXECUTE "PEAK?"    ! Commands executed in parallel processing
610  Pmax=READIO(8,0)   !
620  Fmax=READIO(8,1)   !
630  FOR K=1 TO L1      !
640  EXECUTE "NEXPK?"    !
650  Pnext(K)=READIO(8,0) !
660  Fnext(K)=READIO(8,1) !
670  NEXT K
680  I=I+1
690  END WHILE
700  Tstop=TIMEDATE
710  !
720  PRINT "PEAK(MAX) :";Pmax;"[dB] @";Fmax/1.E+6;"[MHz] "
730  FOR K=1 TO L1
740  PRINT "PEAK(NEXT[";K;"]):";Pnext(K);"[dB] @";Fnext(K)/1.E+6;"[MHz] "
750  NEXT K
760  PRINT "LOOP :";N;" NEXT[K] :";L1
770  PRINT "TIME(AVE.) : ";(Tstop-Tstart)/N*1000;"[msec] "
780  RETURN
790  !
800  END

```

Notes for Parallel Processing

- Use the TRIGMEAS command to trigger the measurement in parallel processing.
- The following command sequence is required:
 1. TRIGMEAS (Trigger 1st. measurement)
 2. LOOP
 3. MOVADARY (Enter data)
 4. TRIGMEAS (Trigger measurement)
 5. ADTOTRAC (Start parallel processing)
 6. Analysis commands
- The parallel processing command sequence is used from the program.
- There is no advantage in using parallel processing when the analyzer is doing very high speed measurements. This is true because the parallel processing is processed during the measurement. Also, there is no advantage if very complex analysis is done in parallel processing because it takes long time to analyze the data. Parallel processing is not recommended under the following conditions:
 1. When using a short measurement time (less than 0.2msec/point).
 2. When using filter and resonator analysis commands (such as OUTPXFIL?, OUTPRESO?).
 3. When data is transferred through HP-IB.
 4. When a single measurement and analysis is done.

Application Sample Programs

This chapter provides you with sample programs that you can use for actual measurements.

- Setting Up and Performing the List Sweep
- Saving and Setting Up Calibration Data
- Analyzing a Ceramic Resonator
- Analyzing a Crystal Resonator

Some of the sample programs listed in this chapter do not include subroutines. In these cases, the required subroutines and their function are listed at the beginning of each sample program.

The programs in the sample program disk contain all the necessary subroutines.

Sample Program -5: Setting Up and Performing the List Sweep

This program sets up the list sweep table and performs a list sweep.

When you run this program, the program will ask you the number of segments, the segment range and the number of displayed points so that you can easily set up the list sweep table.

Disk



This program is contained in the attached sample disk with the file name **LIST_SWP** (for Instrument BASIC). The program for an external controller has the name **LISTSWP_E**.

```

110 ASSIGN @E5100 TO 800           ! IBASIC INITIALIZATION
120 CLEAR @E5100                   !
130 OUTPUT @E5100;"DISAHIHB"      !
140 !
150 DIM Table(1:31,1:3)           ! VARIABLE DECLARATION
160 !
170 INPUT "Number of segments?",Numb ! ENTER AND DISPLAY NUMBER OF SEGMENTS
180 !
190 PRINT USING "10A,11A,11A,20A";"Segment","Start(MHz)","Stop(MHz)"
    ,"Number of points"          !
200 !
210 FOR I=1 TO Numb                ! GOTO TABLE SETUP ROUTINE
220   GOSUB Loadpoin              !
230 NEXT I                          !
240 !
250 LOOP                           ! ASK IF EDIT NEEDED
260   INPUT "Do you want to edit? (Y/N)",An$
270   EXIT IF An$="N" OR An$="n"   !
280   INPUT "Segment Number?",I   !
290   GOSUB Loadpoin              !
300 END LOOP                       !
310 !
320 OUTPUT @E5100;"PRES"          ! MEASUREMENT SETUP
330 OUTPUT @E5100;"CHAN1;ANAMODE GAINP;MEAS AR;FMT LOGM" !
340 !
350 OUTPUT @E5100;"EDITLIST"     ! TABLE SETUP
360 OUTPUT @E5100;"CLEL"         !
370 FOR I=1 TO Numb              !
380   OUTPUT @E5100;"SADD"        !
390   OUTPUT @E5100;"STAR ";Table(I,1)*1.E+6 !
400   OUTPUT @E5100;"STOP ";Table(I,2)*1.E+6 !
410   OUTPUT @E5100;"POIN ";Table(I,3)      !
420   OUTPUT @E5100;"SDON"       !
430 NEXT I                        !
440 OUTPUT @E5100;"EDITDONE"     !
450 OUTPUT @E5100;"SWPT LIST"    !
460 OUTPUT @E5100;"LISDOBASE"   !
470 !
480 OUTPUT @E5100;"SING?"        ! MAKE A MEASUREMENT
490 ENTER @E5100;Tmp

```

```
500 OUTPUT @E5100;"AUTO"           ! AUTO SCALING
510 STOP                             ! STOP PROGRAM
520 !
530 Loadpoin:                        ! TABLE PARAMETER SETUP SUBROUTINE
540   INPUT "Enter start frequency (MHz)",Table(I,1)
550   INPUT "Enter stop frequency (MHz)",Table(I,2)
560   INPUT "Enter number of points",Table(I,3)
570   IF Table(I,3)=1 THEN Table(I,2)=Table(I,1)
580   PRINT TABXY(0,I+1);RPT$(" ",58)
590   PRINT TABXY(0,I+1);I;TAB(11);Table(I,1);TAB(22);Table(I,2);
TAB(35);Table(I,3)
600   RETURN                          !
610   !
620   END                             ! END OF PROGRAM
```

Saving and Resetting Calibration Data

This program saves and resets calibration data.

This program transfers error correction data in the calibration coefficient array memory of the HP E5100A/B to an external controller, saves the data and uses that data in the file as calibration data.

Calibration coefficient arrays corresponding to each error item are stored in three memories as complex numbers. This sample transfers the error data of a frequency response calibration (stored in the array 1) and resets it. See Appendix D of the *HP-IB Command Reference* for each error item and its corresponding calibration coefficient array.

You can use either of the following two ways to save calibration data:

1. Use the **SAVDSTA** command to save the data with the measurement setup data to the HP E5100A/B's internal disk or the RAM disk memory.
2. Transfer calibration data to a controller and save it to the disk in the controller.

This section lists the program for example 2. See "Using a Disk" of Chapter 3 for example 1.

Sample Program -6: Saving Calibration Data in an External Controller

This program transfers calibration data to an external controller and saves it to a file.

Disk



This program is contained in the attached sample disk with the file name **CALSTR_E** (for the external controller). The program for Instrument BASIC is not provided.

```

110 ASSIGN @E5100 TO 717;FORMAT ON           !  INITIALIZATION
120 ASSIGN @Dt TO 717;FORMAT OFF           !
130 Scode=7                                !
140 CLEAR @E5100                           !
150 CLEAR @Dt                               !
160 !                                       !
170 DIM Dat(1:201,1:2)                     !
180 Nop=201                                 !
190 !                                       !
200 GOSUB Setup                             !  CALL SUBROUTINES
210 GOSUB Cal                               !
220 GOSUB Get_cal_data                     !
230 GOSUB Save_cal_data                   !
240 GOTO Ending                             !
250 !                                       !
260 Setup:                                 !  MEASUREMENT SETUP
270 INPUT "CENTER? [MHz]",F_cent           !
280 F_cent=F_cent*1.E+6                    !
290 INPUT "SPAN? [kHz]",F_span            !
300 F_span=F_span*1000                    !
310 OUTPUT @E5100;"PRES"                  !
320 OUTPUT @E5100;"CHAN1;ANAMODE GAINP;MEAS AR;FMT LOGM"
330 OUTPUT @E5100;"HOLD"                  !
340 OUTPUT @E5100;"POIN ";Nop             !

```

```

350 OUTPUT @E5100;"CENS ";F_cent,F_span      !
360 RETURN                                  !
370 !                                       !
380 Cal:                                    ! CALIBRATION
390 INPUT "CONNECT THRU,THEN PRESS [Return]",Dum$
400 OUTPUT @E5100;"CALI RESP"              !
410 OUTPUT @E5100;"STANC?"                 !
420 ENTER @E5100;Tmp                       !
430                                         !
440 Sweep_end:DISP "SWEEP COMPLETED"      !
450 OUTPUT @E5100;"RESPDONE?"             !
460 ENTER @E5100;Tmp                       !
470                                         !
480 DISP "THRU CAL. COMPLETED."          !
490 RETURN                                  !
500 !                                       !
510 Get_cal_data:                          ! GETTING CALIBRATION DATA
520 INPUT "PRESS [RETURN] TO GET CAL DATA FROM
    HP E5100A/B.",Dum$                      !
530 OUTPUT @E5100;"FORM3"                 !
540 OUTPUT @E5100;"OUTPCALCO1?"          !
550 ENTER @E5100 USING "#,8A";Header$     !
560 ENTER @Dt;Dat(*)                       !
570 ENTER @E5100 USING "#,A";End$         !
580 RETURN                                  !
590 !                                       !
600 Save_cal_data:                         ! SAVING CALIBRATION DATA
610 INPUT "Enter File Name (with extension)",File$
620 CREATE BDAT File$,Nop,16               !
630 ASSIGN @File TO File$                 !
640 OUTPUT @File;Dat(*)                   !
650 ASSIGN @File TO *                     !
660 RETURN                                  !
670 !                                       !
680 Ending:                                ! ENDING PROGRAM
690 ASSIGN @Dt TO *                       !
700 !                                       !
710 END                                    !

```

Sample Program -7: Setting Up Calibration Data from an External Controller

This program transfers the calibration data saved in an external controller by **CALSTOR_E** to the HP E5100A/B and sets it up again.

To set the calibration function on again, you must measure the standard to calculate the calibration coefficient. You cannot, however, set the calibration function on if you transferred calibration data only. This means that a measurement was not performed. This sample program, therefore, performs a dummy calibration measurement and transfers calibration data to set the calibration function on. (The subroutine **Dummy_cal**)

If you change the setup (for example, the frequency range) of the HP E5100A/B after transferring calibration coefficient, the transferred calibration coefficient becomes invalid. You must, therefore, set up the HP E5100A/B before transferring calibration data.

Disk



This program is contained in the attached sample disk with the file name **CALSET_E** (for the external controller). The program for Instrument BASIC is not provided.

```

110 ASSIGN @E5100 TO 717;FORMAT ON           ! INITIALIZATION
120 ASSIGN @Dt TO 717;FORMAT OFF           !
130 Scode=7                                 !
140 CLEAR @E5100                             !
150 CLEAR @Dt                                 !
160 !                                         !
170 DIM Dat(1:201,1:2)                       !
180 Nop=201                                   !
190 !                                         !
200 GOSUB Setup                               ! CALL SUBROUTINES
210 GOSUB Load_cal_data                       !
220 GOSUB Dummy_cal                           !
230 GOSUB Send_cal_data                       !
240 GOTO Ending                               !
250 !                                         !
260 Setup:                                   ! ANALYZER SETUP
270 INPUT "CENTER? [MHz]",F_cent              !
280 F_cent=F_cent*1.E+6                       !
290 INPUT "SPAN? [kHz]",F_span                !
300 F_span=F_span*1000                         !
310 OUTPUT @E5100;"PRES"                       !
320 OUTPUT @E5100;"CHAN1;ANAMODE GAINP;MEAS AR;FMT LOGM" !
330 OUTPUT @E5100;"HOLD"                       !
340 OUTPUT @E5100;"POIN ";Nop                  !
350 OUTPUT @E5100;"CENS ";F_cent,F_span        !
360 RETURN                                     !
370 !                                         !
380 Dummy_cal:                                ! DUMMY CALIBRATION
390 OUTPUT @E5100;"CALI RESP"                  !
400 OUTPUT @E5100;"STANC?"                    !
420 ENTER @E5100;Tmp                          !
430 !                                         !
440 Cal_end:                                  ! CALIBRATION END
450 OUTPUT @E5100;"RESPDONE?"                 !
470 ENTER @E5100;Tmp                          !
480 RETURN                                     !
490 !                                         !
500 Load_cal_data:                            ! LOADING CAL DATA
510 INPUT "Enter File Name (with extension)",File$ !
520 ASSIGN @File TO File$                      !
530 ENTER @File;Dat(*)                        !
540 ASSIGN @File TO *                          !
550 RETURN                                     !
560 !                                         !
570 Send_cal_data:                            ! SENDING CAL DATA
580 INPUT "PRESS [RETURN] TO SEND CAL DATA FROM HP E5100A/B.",Dum$ !
590 OUTPUT @E5100;"FORM4"                     !
600 OUTPUT @E5100;"INPUCALC01 ";Dat(*)        !

```

```
610 RETURN !
620 !
630 Ending: ! ENDING PROGRAM
640 ASSIGN @Dt TO * !
650 !
660 END !
```

Sample Program -8: Analyzing a Ceramic Resonator

This program analyses the resonant impedance and ripple of a ceramic resonator.

A wave form analysis command **OUTPCERR?** is used for the measurement. The parameters to be measured are Z_r , f_r , Z_a , f_a , Rpl_1 , Rpl_2 , and Rpl_3 . See "OUTPCERR?" of the *HP-IB Command Reference* for more information.

Measurement Conditions:

- The Log Amplitude & The Phase Format
- Impedance Conversion ON
- Frequency Response Calibration

Items to Prepare:

- A device (a ceramic resonator)
- A cable (or a fixture) to connect the device and the HP E5100A/B
- THRU calibration standard

Disk



This program is contained in the attached sample disk with the file name **CER_RES** (for Instrument BASIC). The program for an external controller is not provided.

```

110 ASSIGN @E5100 TO 800                ! IBASIC INITIALIZATION
120 CLEAR @E5100                        !
130 !
140 OUTPUT @E5100;"PRES"                ! MEASUREMENT SETUP
150 OUTPUT @E5100;"CHAN1;HOLD"          !
160 OUTPUT @E5100;"ANAMODE ZTRAN;MEAS AR;FMT MAGZP;SCAY 1"
170 OUTPUT @E5100;"COUC OFF"           !
180 OUTPUT @E5100;"DUAC OFF"           !
190 OUTPUT @E5100;"POIN 201"          !
200 OUTPUT @E5100;"IFBW 4KHZ"         !
210 OUTPUT @E5100;"POWE 0"            !
240 OUTPUT @E5100;"DISAHIHB"          !
250 !
260 INPUT "Center Frequency (kHz)",Center ! SWEEP CONDITION ENTRY
270 Center=Center*1000.                 !
280 INPUT "Frequency Span (kHz)",Span   !
290 Span=Span*1000.                     !
300 OUTPUT @E5100;"CENS ";Center,Span  !
310 !
320 GOSUB Thru_cal                       ! GOTO CAL ROUTINE
330 !
340 EXECUTE "ANAOCH1"                   ! WAVEFORM ANA SETUP
350 EXECUTE "ANAODATA"                  !
360 EXECUTE "ANARFULL"                  !
370 !
380 WRITEIO 8,0;.05                     ! MEASUREMENT
390 EXECUTE "THRR"                      !
400 LOOP                                !

```

```

410  DISP "Connect Device and Press Continue."
420  PAUSE
430  DISP "MEASURING"
440  EXECUTE "SING"
450  EXECUTE "OUTPCERR?"
460  Rr=READIO(8,0)
470  Fr=READIO(8,1)
480  Ra=READIO(8,2)
490  Fa=READIO(8,3)
500  Rpl_left=READIO(8,4)
510  Rpl_center=READIO(8,5)
520  Rpl_right=READIO(8,6)
530  GOSUB Printing
540  END LOOP
550  STOP
560  !
570  Printing:
580  PRINT USING "5A,6D.3D,6A";"Fr ";Fr/1000.;"(kHz)"
590  PRINT USING "5A,6D.3D,6A";"Rr ";Rr;" (ohm)"
600  PRINT USING "5A,6D.3D,6A";"Fa ";Fa/1000.;"(kHz)"
610  PRINT USING "5A,6D.3D,7A";"Ra ";Ra/1000;" (kohm)"
620  PRINT USING "12A,3D.3D,5A";"Ripple LEFT ";Rpl_left;" (dB)"
630  PRINT USING "14A,3D.3D,5A";"Ripple CENTER ";Rpl_center;" (dB)"
640  PRINT USING "13A,3D.3D,5A";"Ripple RIGHT ";Rpl_right;" (dB)"
650  RETURN
660  !
670  Thru_cal:
680  DISP "Connect THRU standard and Press Continue."
690  PAUSE
700  DISP
710  OUTPUT @E5100;"CALI RESP"
750  OUTPUT @E5100;"STANC?"
780  ENTER @E5100;Tmp
800  OUTPUT @E5100;"RESPDONE?"
810  ENTER @E5100;Tmp
830  RETURN
840  !
850  END

```

Sample Program -9: Analyzing a Crystal Resonator

This program measures the resonant frequency (Phase 0°), the resonant impedance (Phase 0°) and the equalizing circuit constant of a crystal resonator.

The measured parameters are as follows:

The resonance parameter (**OUTPRESO?** is Z_r, F_r used.)

The equalizing circuit constant (**EQUCO?** $C_0, C_1, L_1, R_1, G_0, R_0$ and **EQUCPARA?** are used.)

Measurement conditions are as follows:

- A π network fixture is used.
- The drive level [μW] of a crystal at resonance and the resonant impedance [Ω] are used to calculate and set the power level [dBm] of the HP E5100A/B.
- Uses the list sweep to perform two types of measurements: the parallel capacity measurement point and the measurement near the resonant point.

Required Subroutines

Power_set: The drive level [μW] of a crystal at resonance and the resonant impedance [Ω] are used to calculate the power level [dBm].

List_set: Set up the list sweep table.

Pi_cal: A calibration subroutine for a π fixture.

Printing: Display analysis data on the screen.

Disk



This program is contained in the attached sample disk with the file name **FRCL_EQV** (for Instrument BASIC). The program for an external controller is not provided.

```
110 DIM Startf(1:2),Stopf(1:2),Ifbw(1:2),Nop(1:2)
120 !
130 ASSIGN @E5100 TO 800
140 CLEAR @E5100
150 !
160 OUTPUT @E5100;"PRES"
170 OUTPUT @E5100;"CHAN1;ANAMODE ZTRAN;MEAS AR;FMT MAGZP;SCAY 1"
180 OUTPUT @E5100;"HOLD"
190 OUTPUT @E5100;"DUAC OFF"
200 OUTPUT @E5100;"COUC OFF"
210 OUTPUT @E5100;"DISG OFF"
220 OUTPUT @E5100;"DISAHIHB"
230 !
240 INPUT "Center frequency ? (MHz)",Center
250 Center=Center*1.E+6
260 INPUT "Span frequency ? (kHz)",Span
270 Span=Span*1000.
280 !
290 GOSUB Power_set
310 OUTPUT @E5100;"POWE ";Power
320 !
330 GOSUB List_set
340 !
380 !
390 GOSUB Pi_cal
400 !
410 EXECUTE "ANAOCH1"
420 EXECUTE "ANARFULL"
430 EXECUTE "ANAODATA"
440 !
450 DISP "CONNECT DEVICE, and PRESS CONTINUE."
460 PAUSE
470 DISP
480 !
490 !     MEASUREMENT
500 !
510 LOOP
520 !
530     EXECUTE "SING"
540 !
550     EXECUTE "OUTPRESO?"
560     Ci=READIO(8,0)
570     Fr=READIO(8,1)
580 !
590     WRITEIO 8,0;Center*.9
600     EXECUTE "EQUCO?"
610     CO=READIO(8,0)
620 !
630     EXECUTE "EQUCPARA?"
640 !     CO=READIO(8,0)
650     C1=READIO(8,1)
660     L1=READIO(8,2)
670     R1=READIO(8,3)
680     GO=READIO(8,4)
```

```
690      RO=READIO(8,5)
700      !
710      ! EXECUTE "EQUCPARS4?"
720      ! ! CO=READIO(8,0)
730      ! C1=READIO(8,1)
740      ! L1=READIO(8,2)
750      ! R1=READIO(8,3)
760      !
770      GOSUB Printing
780      !
790      DISP "CONNECT NEXT DEVICE and Press Continue."
800      PAUSE
810      !
820      END LOOP
830      !
840      STOP
850      !
860      Power_set: !
870      !
880      DATA 10
890      DATA 15
900      !
910      READ W
920      READ R
930      !
940      INPUT "POWER (uW)",W
950      W=W/1.E+6
960      !
970      INPUT "RESISTANCE (OHM)",R
980      !
990      ! Power level calculation
1000     !
1010     Pi_r3=12.5
1020     R0=50
1030     R1=50
1040     R2=50
1050     R3=83.3
1060     R4=159
1070     R5=66.2
1080     R6=14.2
1090     Pi_r2=R6*(R3*R4+R5*(R3+R4))/(R3*R4+(R3+R4)*(R5+R6))
1100     VO_a=(Pi_r2+R+Pi_r3)*SQRT(W/R)
1110     Vp=(R3*R4+(R3+R4)*(R5+R6))/(R4*R6)*VO_a
1120     VO=(R0+R1+R2)/(R1+R2)*Vp
1130     V1=VO/4
1140     Power=10*LGT(1000*V1^2/50)
1150     !
1160     RETURN
1170     !
1180     !
1190     Pi_cal: !
1200     OUTPUT @E5100;"CALIS111"
1210     !
1220     DISP "Connect SHORT, and press [Continue]"
1230     PAUSE
```

```

1240 OUTPUT @E5100;"CLASS11B?"
1260 ENTER @E5100;Tmp
1270 !
1280 DISP "Leave the terminal OPEN, and press [Continue]"
1290 PAUSE
1300 OUTPUT @E5100;"CLASS11A?"
1320 ENTER @E5100;Tmp
1330 !
1340 DISP "Connect LOAD, and press [Continue]"
1350 PAUSE
1360 OUTPUT @E5100;"CLASS11C?"
1380 ENTER @E5100;Tmp
1390 !
1400 OUTPUT @E5100;"SAV1?"
1420 ENTER @E5100;Tmp
1430 !
1440 RETURN
1450 !
1460 !
1470 Printing: !
1480 !
1490 CLEAR SCREEN
1500 !
1510 PRINT TABXY(1,1);"RESULT"
1520 PRINT USING "25A,3D.8D,6A";"RESONANCE FREQUENCY (Fr) ";Fr/1.E+6;" (MHz)"
1530 PRINT USING "14A,15X,5D.2D,6A";"CI (Zr) VALUE ";Ci;" (ohm)"
1540 PRINT
1550 PRINT USING "10A,5X,5D.3D,5A";"CO VALUE ";CO*1.E+12;" (pF)"
1560 PRINT USING "10A,5X,5D.5D,5A";"C1 VALUE ";C1*1.E+12;" (pF)"
1570 PRINT USING "10A,5X,5D.3D,5A";"L1 VALUE ";L1*1000;" (mH)"
1580 PRINT USING "10A,5X,5D.3D,6A";"R1 VALUE ";R1;" (ohm)"
1590 PRINT USING "10A,5X,5D.3D,5A";"GO VALUE ";GO*1000;" (mS)"
1600 PRINT USING "10A,5X,5D.3D,6A";"RO VALUE ";RO;" (ohm)"
1610 !
1620 RETURN
1630 !
1640 List_set: !
1650 Startf(1)=Center*.9
1660 Stopf(1)=Center*.9
1670 Nop(1)=1
1680 Ifbw(1)=1000
1690 Startf(2)=Center-Span/2
1700 Stopf(2)=Center+Span/2
1710 Nop(2)=200
1720 Ifbw(2)=1000
1730 !
1740 OUTPUT @E5100;" ;EDITLIST;EDITLIS1;CLEL"
1750 FOR I=1 TO 2
1760     OUTPUT @E5100;"SADD"
1770     OUTPUT @E5100;"STAR ";Startf(I)
1780     OUTPUT @E5100;"STOP ";Stopf(I)
1790     OUTPUT @E5100;"POIN ";Nop(I)
1800     OUTPUT @E5100;"IFBW ";Ifbw(I)
1810     OUTPUT @E5100;"SDON"
1820 NEXT I

```

```
1830 OUTPUT @E5100;"EDITDONE"  
1840 OUTPUT @E5100;"SWPT LIST"  
1850 OUTPUT @E5100;" ;LISDOBASE"  
1860 RETURN  
1870 !  
1880 END
```

Command Reference

Common Command

*CLS

Clears the error queue, the status byte register, the event register of the standard operation status register structure, and the standard event status register. (no query)

Examples	OUTPUT @E5100;"*CLS"
----------	----------------------

*ESE *value*

Sets the enable bits of the standard status register.

Parameter Range	0 ~ 255 (decimal expression of enable bits of the operation status register)
Query Response	<value>
Examples	OUTPUT @E5100;"*ESE 1" OUTPUT @E5100;"*ESE?" ENTER @E5100;A

*ESR?

Outputs the contents of the standard event status register. (Query only)

Query Response	<value>
Examples	OUTPUT @E5100;"*ESR?" ENTER @E5100;A

*IDN?

Outputs the HP E5100A/B ID. (Query only)

***IDN?**

Query Response	<pre><manufacturer> <model> <serial no.> <firmware rev.> <manufacturer> : HEWLETT-PACKARD <model> : E5100A <serial no.> : serial number, such as JP3KA00101 <firmware rev.> : firmware revision number, such as REV1.00</pre>
Examples	<pre>OUTPUT @E5100;"*IDN?" ENTER @E5100;A\$</pre>

***OPC**

Tells the analyzer to set bit 0 (Operation Complete bit) in the Standard Event Status Register when it completes all pending operations.

*OPC? query places an ASCII character 1 into the analyzer's output queue when all pending operations have been completed.

Query Response	{1}
Examples	<pre>OUTPUT 717;"*OPC" OUTPUT 717;"*OPC?" ENTER 717;A</pre>

***PCB value**

Specifies the address of a controller that is temporarily passing HP-IB control to the HP E5100A/B.

Parameter Range	0 to 30
Query Response	<value>
Examples	<pre>OUTPUT @E5100;"*PCB 0"</pre>

***RST**

Resets the HP E5100A/B to its initial settings. See *Function Reference* for more information on default settings. (No Query)

***SRE value**

Sets the enable bits of the status byte register.

Parameter Range	0 to 255 (decimal expression of enable bits of the status byte register)
Query Response	<value>
Examples	<pre>OUTPUT @E5100;"*SRE 1" OUTPUT @E5100;"*SRE?" ENTER @E5100;A</pre>

***STB?**

Reads the status byte by reading the master summary status bit.

Query Response	<value>
Examples	<pre>OUTPUT @E5100;"*STB?" ENTER @E5100;A</pre>

***TRG**

Triggers the HP E5100A/B when the trigger mode is set to EXTERNAL trigger. (No Query)

Examples	<p>To start measurement</p> <pre>OUTPUT @E5100;"*TRG"</pre>
----------	---

***TST?**

Executes a power-on self test and returns the test result. (No Query)

Query Response	{0 1}
	0 : No error
Examples	<pre>OUTPUT @E5100;"*TST?" ENTER @E5100;A</pre>

***WAI**

Makes the HP E5100A/B wait until all previously sent commands are completed. (No Query)

Command Reference

ADTOTRAC

Starts data process. This command is only used for the parallel processing using with TRIGMEAS and MOVADARY. (No Query)

ANAMODE {GAINP|ZREFL|ZTRAN}

Select measurement function.

Equivalent Key Sequence	Meas/Format FUNCTION <input type="checkbox"/> GAIN-PHASE, IMPEDANCE:Refl, Trans
Parameter Description	GAINP : Gain-Phase measurement ZREFL : Reflection impedance measurement ZTRAN : Transmission impedance measurement
Query Response	{ GAINP ZREFL ZTRAN }
Examples	OUTPUT @E5100;"ANAMODE GAINP" OUTPUT @E5100;"ANAMODE?" ENTER @E5100;A\$

ANAOCH{1-4}

Selects channel for waveform analysis. For details, refer to Appendix D. (No Query)

Examples	OUTPUT 717;"ANAOCH1"
----------	----------------------

ANAODATA

Selects a data trace for waveform analysis. For details, refer to Appendix D. (No Query)

Examples	OUTPUT 717;"ANAODATA" OUTPUT 717;"ANAODATA?" ENTER 717;A
----------	--

ANAOMEMO

Selects a sub-trace for waveform analysis. For details, refer to Appendix D. (No Query)

ANARANG *value1*[*suffix*],*value2*[*suffix*]

Sets the waveform analysis stimulus range by entering the START (*value1*) and STOP (*value2*) values. For details, refer to Appendix D.

Parameter Range	<i><value1></i> : 10×10^3 (=10k) to 300×10^6 (=300M) Hz <i><value2></i> : 10×10^3 (=10k) to 300×10^6 (=300M) Hz
Query Response	{ <i>value1</i> }{ <i>value2</i> }
Examples	OUTPUT 717;"ANARANG 1000,10000"

ANARANGP □*<value1>*,*<value2>*

Sets the waveform analysis stimulus range by entering the point number of START and point number of STOP values.

Parameter Description	<i><value1></i> : The point number of START (1 to NOP) <i><value2></i> : The point number of STOP (1 to NOP)
Query Response	{ <i>value1</i> }{ <i>value2</i> }
Examples	OUTPUT @E5100;"ANARANGP 10,100" OUTPUT @E5100;"ANARANGP?" ENTER @E5100;A, B

ANARFULL

Sets the analysis range equal to the full stimulus range. For details, refer to Appendix D. (No Query)

AR

This command equals MEAS AR. See MEAS.

ATRC {1|2}

Selects the active trace for the marker function.

Equivalent Key Sequence	Marker ACTIVE TRC []
Parameter Description	1: Main trace 2: Sub trace
Query Response	{0 1}
Examples	OUTPUT @E5100;"ATRC" OUTPUT @E5100;"ATRC?" ENTER @E5100;A

ATTI{A|B|C|R} {0|25}

Selects the attenuator value at an input port.

Equivalent Key Sequence	System ATTENUATOR, PROT:A, PROT:B, PROT:C, PROT:R 0dB, 25dB
Parameter Description	0 : 0 dB 25 : 25 dB
Query Response	{0 25}
Examples	<pre>OUTPUT @E5100;"ATTIA 25" OUTPUT @E5100;"ATTIA?" ENTER @E5100;A</pre>

ATTIAAUTO□{OFF|ON|0|1}

Select the attenuator auto mode at the port A ON or OFF. The auto mode automatically selects the proper attenuator value.

Equivalent Key Sequence	System ATTENUATOR PORT:A AUTO
Parameter Description	OFF' or 0 : AUTO OFF ON or 1 : AUTO ON
Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"ATTIAAUTO ON" OUTPUT @E5100;"ATTIAAUTO?" ENTER @E5100;A</pre>

ATTIBAUTO {OFF|ON|0|1}

Select the attenuator auto mode at the port B ON or OFF. The auto mode automatically selects the proper attenuator value.

Equivalent Key Sequence	System ATTENUATOR PROT:B AUTO
Parameter Description	OFF' or 0 : AUTO OFF ON or 1 : AUTO ON
Query Response	{0 1}

ATTICAUTO {OFF|ON|0|1}

Select the attenuator auto mode at the port C ON or OFF. The auto mode automatically selects the proper attenuator value.

Equivalent Key Sequence	System ATTENUATOR PROT:C AUTO
Parameter Description	OFF' or 0 : AUTO OFF ON or 1 : AUTO ON
Query Response	{0 1}

ATTIRAUTO {OFF|ON|0|1}

Select the attenuator auto mode at the port R ON or OFF.

Equivalent Key Sequence	System ATTENUATOR PORT:R AUTO
Parameter Description	OFF' or 0 : AUTO OFF ON or 1 : AUTO ON
Query Response	{0 1}

ATTW value

Sets the waiting time when the attenuator switch is changed at the power sweep mode. (Option 010 only, No warning will be shown even if this command is executed with the HP E5100A/B with no option 010.) The power-on default setting is 1 ms.

Parameter Range	0 to 100 sec.
Query Response	{value} (sec)
Examples	OUTPUT 717;"ATTW 10" OUTPUT 717;"ATTW?" ENTER 717;A

AUTO

Selects the scale/div value automatically to fit the trace data to the display.

Equivalent Key Sequence	Display AUTO SCALE
Examples	OUTPUT 717;"AUTO"

BASL

Displays the softkey lable defined by HP Instrument BASIC. (No Query)

Equivalent Key Sequence	System I-BASIC ON KEY LABEL
Examples	OUTPUT @E5100;"BASL"

BEEPDONE {ON|OFF|0|1}

Sets the operation completion beeper ON or OFF.

Equivalent Key Sequence	DISPLAY MORE BEEP DONE ON off
Parameter Description	OFF and 0 : operation completion beeper OFF ON and 1 : operation completion beeper ON
Query Response	{0 1}
Examples	OUTPUT 717;"BEEPDONE ON" OUTPUT 717;"BEEPDONE?" ENTER 717;A

BEEPFAIL {ON|OFF|1|0}

Sets the limit fail beeper ON or OFF.

Equivalent Key Sequence	System LIMIT MENU BEEP FAIL ON off
Parameter Description	OFF and 0 : limit fail beeper OFF ON and 1 : limit fail beeper ON
Query Response	{0 1}
Examples	OUTPUT 717;"BEEPFAIL ON" OUTPUT 717;"BEEPFAIL?" ENTER 717;A

BEEPWARN {ON|OFF}

Sets the warning beeper ON or OFF.

Equivalent Key Sequence	DISPLAY MORE BEEP WARN ON off
Parameter Description	OFF and 0 : warning beeper OFF ON and 1 : warning beeper ON
Query Response	{0 1}
Examples	OUTPUT 717;"BEEPWARN ON" OUTPUT 717;"BEEPWARN?" ENTER 717;A

BINSIZE \square $\langle value \rangle$

Specify the number of continuous data outputted to the I/O port by INPUTRACB.
 Specify the maximum value in the second parameter of INPUTRACB. (Option 022 only)

Parameter Range	0 to 6
Query Response	{value}
Examples	OUTPUT @E5100;"BINSIZE 2" OUTPUT @E5100;"BINSIZE?" ENTER @E5100;A

8. Command Reference

BOTV \square $\langle value \rangle$

Sets the value at the bottom line of the graticule.

Equivalent Key Sequence	Display BOTTOM VALUE
Parameter Range	-10^9 to 10^9
Query Response	{value}
Examples	OUTPUT @E5100;"BOTV 100" OUTPUT @E5100;"BOTV?" ENTER @E5100;A

CALCOPY \square $\langle value1 \rangle$, $\langle value2 \rangle$

Copy CAL data and CAL settings between two channels. (No Query)

Parameter Description	$\langle value1 \rangle$: Source channel $\langle value2 \rangle$: Destination channel
Examples	OUTPUT @E5100;"CALCOPY 1,2"

CALI *parameter*

Selects the measurement calibration type. (Query)

Equivalent Key Sequence	Cal CALIBRATE:NONE, RESPONSE, RESPONSE & ISOLATION, 3 TERM
Parameter Description	NONE : No calibration RESP : Response measurement calibration RAI : Response and isolation measurement calibration ONEP : 1-port 3-term measurement calibration
Query Response	{NONE RESP RAI ONEP}
Examples	OUTPUT @E5100;"CALI NONE" OUTPUT @E5100;"CALI?" ENTER @E5100;A\$

CALK{O|S|L}{LS|RS|CP} <value>

Sets the value of calibration kit of the active measurement mode.

Equivalent Key Sequence	Cal MODIFY CAL KIT OPEN STD, SHORT STD, LOAD STD, Rs, Ls, Cp
Parameter Range	O : OPEN S : SHORT L : LOAD LS : L _s RS : R _s CP : C _p
Query Response	{value}
Examples	OUTPUT @E5100;"CALKLS,49.9" OUTPUT @E5100;"CALKLS?" ENTER @E5100;A

CENS *value1* [*suffix*], *value2* [*suffix*]

Sets center and span stimulus value.

Parameter Description	<value1> : Center value <value2> : Span value
Parameter Range	<value1> : 10×10^3 (=10k) to 300×10^6 (=300M) Hz <value2> : 0 to 2.9999×10^8 (=299.99M) [Hz]
Examples	OUTPUT @E5100;"CENS 199.95MHZ,100KHZ"

CENT *value* [*suffix*]

Sets the center stimulus value.

Equivalent Key Sequence	Center
Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"CENT 99.95MAHZ" OUTPUT @E5100;"CENT?" ENTER @E5100;A</pre>

CHAD *string*

Changes the current directory.

Equivalent Key Sequence	Save/Recall FILE UTILITY CHANGE DIRECTORY
Parameter Description	<i><string></i> :Directory path name
Examples	<pre>OUTPUT @E5100;"CHAD "".."""</pre>

CHAN {1|2|3|4}

Selects the active measurement channel. (No Query)

Equivalent Key Sequence	Meas/Format ACTIVE CH []
Examples	<pre>OUTPUT 717;"CHAN 1"</pre>

CIN

Set port C of the 24-bit I/O port to be an input port.

Examples	<pre>OUTPUT 717;"CIN"</pre>
----------	-----------------------------

CIVAL *<value>*

Enters the CI value.

CIVAL *<value>*

Equivalent Key Sequence	MENU2 USER CT
Parameter Range	10 ⁻¹⁵ to 1 G [Ω]
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"CIVAL 10" OUTPUT @E5100;"CIVAL?" ENTER @E5100;A</pre>

CLASS11 {**A|B|C**}

Selects port 1 (S11) one-port calibration standard class: S11A (open), S11B (short), or S11C (load). (No Query)

Equivalent Key Sequence	Cal [S11] : OPEN , SHORT , LOAD
Parameter Description	<pre>A : OPEN B: SHORT C: LOAD</pre>
Examples	<pre>OUTPUT @E5100;"CLASS11A"</pre>

CLEL

Clears the current frequency list. (No Query)

Equivalent Key Sequence	Sweep
Examples	<pre>OUTPUT @E5100;"CLEL"</pre>

CLEM{**1-4**}

Clears the marker. (No Query)

Equivalent Key Sequence	Marker CLEAR MARKER
Examples	<pre>OUTPUT @E5100;"CLEM1"</pre>

CLEMNU3

Clears the softkey definitions of **Menu3**. (No Query)

Examples

OUTPUT @E5100;"CLEMNU3"

CLES

Clears the status byte, the event status register, the event status register B, and the operational status register.

(No Query)

Examples

OUTPUT @E5100;"CLES"

CONT

Continuous trigger. (No Query)

Equivalent Key Sequence

Trigger CONTINUOUS

Examples

OUTPUT @E5100;"CONT"

CONV {OFF|ZTRA|YTRA|ZREF|YREF}

Selects the measurement data conversion setting. This command is available when the ANAMODE is set to Gain-Phase mode.

Equivalent Key Sequence

Meas/Format FUNCTION []

Parameter Description

OFF : Conversion OFF
 ZTRA : Z:transmission
 YTRA : Y:transmission
 ZREF : Z:Reflection
 YREF : Y:Reflection

Query Response

{OFF|ZTRA|YTRA|ZREF|YREF}

Examples

OUTPUT @E5100;"CONV ZTRA"

 OUTPUT @E5100;"CONV?"
 ENTER @E5100;A\$

CORR {ON|OFF}

Sets the error correction function ON or OFF.

Equivalent Key Sequence	Cal CORRECTION ON off
Query Response	{0 1}
Parameter Range	OFF and 0 : error correction function OFF ON and 1 : error correction function ON
Examples	<pre>OUTPUT @E5100;"CORR OFF" OUTPUT @E5100;"CORR?" ENTER @E5100;A</pre>

COUC {ON|OFF}

Sets the channel coupling of stimulus values ON or OFF.

Equivalent Key Sequence	Sweep COUPLED CH ON off
Parameter Description	OFF and 0 : channel coupling OFF ON and 1 : channel coupling ON
Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"COUC OFF" OUTPUT @E5100;"COUC?" ENTER @E5100;A</pre>

COUT

Sets port C of the 24-bit I/O port to be an output port.

Examples	<pre>OUTPUT @E5100;"COUT"</pre>
----------	---------------------------------

CRED *string*

Create a directory (only MS-DOS format). (No Query)

Equivalent Key Sequence	Save/Recall FILE UTILITY CREATE DIRECTORY
Parameter Description	<string> :Up to 8 characters for directory name and up to 3 characters for extension
Examples	<pre>OUTPUT @E5100;"CRED ""DATA"""</pre>

CURD?

Outputs current directory. (Query only)

Equivalent Key Sequence	Save/Recall FILE UTILITY CURRENT DIRECTORY
Query Response	{string}
Examples	OUTPUT @E5100;"CURD?" ENTER @E5100;A\$

CURMPOIN?

Outputs the latest measurement point number and measurement data. (Query only)

Query Response	{numeric1}{numeric2} {numeric1} : measurement point number {numeric2} : Formatted data (complex value , data format: real, imaginary)
Examples	OUTPUT @E5100;"CURMPOIN?" ENTER @E5100;A,B,C

CWFREQ *value* [*suffix*]

Sets the frequency for power sweep.

Equivalent Key Sequence	Sweep CWFREQ
Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz
Query Response	{value}
Examples	OUTPUT @E5100;"CWFREQ 100KHZ" OUTPUT @E5100;"CWFREQ?" ENTER @E5100;A

DATI

Saves the data trace to the memory trace. (No Query)

Equivalent Key Sequence	Display DEFINE TRACE DATA->MEM
Examples	OUTPUT @E5100;"DATI"

DAYMYEAR

Sets the displayed date mode to day/month/year order. (No Query)

Equivalent Key Sequence	System MORE SET CLOCK DayMonYear
Examples	OUTPUT @E5100;"DAYMYEAR"

DELA

This command equals FMT DELA. See FMT.

DELO

Sets the delta marker mode OFF. (No Query)

Equivalent Key Sequence	Marker Δ MODE MENU Δ MODE OFF
Examples	OUTPUT @E5100;"DELO"

DELR{1-4}

Selects the delta reference marker.

Equivalent Key Sequence	ΔMODE MENU AREF MKR ΔREF=1 to AREF=4
Query Response	{value}
Examples	OUTPUT @E5100;"DEL1" OUTPUT @E5100;"DEL1?" ENTER @E5100;A

DELRFIXM

Sets the user-specified fixed reference marker. (No Query)

Equivalent Key Sequence	Marker ΔMODE MENU Δ REF=A FIXED MKR
Examples	OUTPUT @E5100;"DELRFIXM"

DIN

Set port D of the 24-bit I/O port to be an input port. (No Query)

Examples	OUTPUT @E5100;"DIN"
----------	---------------------

DISA parameter

Selects the display allocation mode.

Equivalent Key Sequence	Display DISPLAY ALLOCATION
Parameter Description	ALLI : All instrument HIHB : Half instrument half BASIC ALLB : All BASIC BASS : BASIC status
Query Response	{ALLI HIHB ALLB BASS}
Examples	OUTPUT @E5100;"DISA HIHB" OUTPUT @E5100;"DISA?" ENTER @E5100;A\$

DISG {ON|OFF}

Sets the graticule display on or off.

Equivalent Key Sequence	Display MORE GRATICULE ON off
Parameter Range	{OFF ON 0 1}
Query Response	{0 1}
Examples	OUTPUT @E5100;"DISG ON" OUTPUT @E5100;"DISG?" ENTER @E5100;A

DISP parameter

Selects the display trace type.

DISP parameter

Equivalent Key Sequence	Display MORE MORE TRACE:DATA, MEMORY, DATA and MEMORY, DATA-MEM, DATA/MEM
Parameter Range	DATA : Data only MEMO : Memory only DATM : Data and memory DDM : Data divided by memory DMM : Data minus memory
Query Response	{DATA MEMO DATM DDM DMM}
Examples	OUTPUT @E5100;"DISP DATA" OUTPUT @E5100;"DISP?" ENTER @E5100;A\$

DIST {ON|OFF}

Sets the trace display on or off.

(TRACE on off under **DISPLAY**; Query)

Parameter Range	OFF or 0 : trace display OFF ON or 1 : trace display ON
Query Response	{0 1}
Examples	OUTPUT @E5100;"DIST ON"

DONE

Completes the measurement of the selected standard calibration. (No Query)

Equivalent Key Sequence	Cal DONE
Examples	OUTPUT @E5100;"DONE"

DOUT

Sets D port of the 24-bit I/O port to output port. (No Query)

Examples	OUTPUT @E5100;"DOUT"
----------	----------------------

DSKEY

Disables the front panel keys and the rotary knob. To enable the keys and knob again, send the ENKEY command.

Examples	<code>OUTPUT @E5100;"DSKEY"</code>
-----------------	------------------------------------

DUAC {ON|OFF}

Selects the dual (ON) or single (OFF) channels display.

Equivalent Key Sequence	<code>(Display) MULTI CHAN ON off</code>
Parameter Range	OFF and 0 : Active channel only ON and 1 : Dual channel
Query Response	{0 1}
Examples	<code>OUTPUT @E5100;"DUAC ON"</code> <code>OUTPUT @E5100;"DUAC?"</code> <code>ENTER @E5100;A</code>

EDITDONE

Completes editing the frequency list for the list sweep. (No Query)

Equivalent Key Sequence	<code>(Sweep) LIST DONE</code>
Examples	<code>OUTPUT 717;"EDITDONE"</code>

EDITLIS1

Selects list 1 for editing. (No Query)

Equivalent Key Sequence	<code>(Sweep) LIST NO [] EDIT LIST</code>
Examples	<code>OUTPUT @E5100;"EDITLIS1"</code>

EDITLIS2

Selects list 2 for editing. (No Query)

Equivalent Key Sequence	Sweep LIST NO [] EDIT LIST
Examples	OUTPUT @E5100;"EDITLIS2"

EDITLIST

Begins editing the frequency list. (No Query)

Equivalent Key Sequence	Sweep EDIT LIST
Examples	OUTPUT @E5100;"EDITLIST"

ELED *value* [s]

Sets the electrical delay.

Equivalent Key Sequence	Display MORE ELECTRICAL DELAY
Parameter Range	-10 to 10 [sec.]
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"ELED 0" OUTPUT @E5100;"ELED?" ENTER @E5100;A

ENKEY

Re-enables the front panel keys and the rotary knob which have been disabled by the DSKEY command. (No Query)

EQUC0? *value*

Returns C_0 of the equivalent circuit of the resonator at specified frequency. C_0 is calculated by using the following equation:

$$C_0 = \frac{B_s}{\omega_s}$$

Where,

B_s	Imaginary part on f_s
ω_s	$2 \times \pi \times f_s$
f_s	Frequency which is specified as command parameter

If Z-conversion is selected, C_0 is calculated by using following equations:

$$C_0 = \frac{-1}{B_s \times \omega_s}$$

This command is only available when LOG MAG & Phase format is selected. If another format is selected, 0 will be returned. If the specified frequency is out of analysis range, 0 will be returned.

If B_s is 0 when the Z-conversion is activated, EQUCCO? returns 0.

(Query only)

Parameter Range	1000 (= 1k) to 3.0×10^8 (= 300M) [Hz]	
Query Response	{value(C ₀)}	
Examples	<pre>OUTPUT @E5100;"EQUCCO? 100MHZ" ENTER @E5100;C0 PRINT "C0=",C0</pre>	<p><i>Query C₀ at 100 MHz.</i> <i>Receive the returned C₀.</i> <i>Display C₀ on the CRT.</i></p>

EQUCPARA5?

Executes the equivalent circuit analysis for a resonator using the same equivalent circuit of EQUCPARA?, but does not output G₀. EQUCPARA5? does not display the warning even an anti-resonance frequency is not in the analysis range. (Query only)

Query Response	<pre>{value1},{value2},{value3},{value4},{value5} {value1}:C0 {value2}:C1 {value3}:L1 {value4}:R1 {value5}:R0</pre>	
Examples	<pre>OUTPUT @E5100;"EQUCPARA5?" ENTER @E5100;A,B,C,D,E</pre>	

EQUCPARA?

Executes four element analysis of a crystal resonator, and outputs parameters, C₀, C₁, L₁, and R₁. For more information, refer to "EQUCPARA?" in Appendix D. (Query only)

Equivalent Key Sequence	Menu 2 EQUIVALENT CKT 6	
Query Response	<pre>{numeric1}, {numeric2}, {numeric3}, {numeric4}, {numeric5}, {numeric6} <numeric1> : C₀ <numeric2> : C₁ <numeric3> : L₁ <numeric4> : R₁ <numeric5> : G₀ <numeric6> : R₀</pre>	
Examples	<pre>OUTPUT @E5100;"EQUCPARA?" ENTER @E5100;A,B,C,D</pre>	

EQUCPARS4?

Executes four element analysis of a crystal resonator, and outputs parameters, C_0 , C_1 , L_1 , R_1 , f_s , f_a , f_r , f_1 , f_2 . (Query only)

Equivalent Key Sequence	Menu 2 EQUIVALENT CKT 4
Query Response	<pre>{numeric1}, {numeric2}, {numeric3}, {numeric4}, {numeric5}, {numeric6}, {numeric7}, {numeric8}, {numeric9} <numeric1> : C₀ <numeric2> : C₁ <numeric3> : L₁ <numeric4> : R₁ <numeric5> : f_s <numeric6> : f_a <numeric7> : f_r <numeric8> : f₁¹ <numeric9> : f₂¹</pre>
Examples	<pre>OUTPUT @E5100;"EQUCPARS4?" ENTER @E5100;A,B,C,D,E,F,G,H,I</pre>

¹ $f_1 < f_2$

EQUCPARS?

Executes four elements analysis of a crystal resonator, and outputs parameters, C_0 , C_1 , L_1 , R_1 , f_s , f_a , f_r , f_1 , and f_2 . For more information, refer to "EQUCPARS?" in Appendix D. (Query only)

Equivalent Key Sequence	SPCL FNCT EQUIVALENT CKT
Parameter Range	
Query Response	<pre>{value1}, {value2}, {value3}, {value4}, {value5}, {value6}, {value7}, {value8}, {value9}, {value10}, {value11} <value1> : C₀ <value2> : C₁ <value3> : L₁ <value4> : R₁ <value5> : f_s <value6> : f_a <value7> : f_r <value8> : f₁¹ <value9> : f₂¹ <value10> : G₀ <value11> : R₀</pre>
Examples	<pre>OUTPUT 717;"EQUCPARS?" ENTER 717;A,B,C,D,E,F,G,H,I</pre>

¹ $f_1 < f_2$

EQU *value*

Specifies how many points are used for an approximation of a circle for EQUCPARA? and EQUCPARS? command. The default value is 8. For a detail information about EQU, refer to "EQUCPARA?" in Appendix D.

Parameter Range	2 to 801
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"EQU 30" OUTPUT @E5100;"EQU?" ENTER @E5100;A</pre>

ESB?

Outputs the event status register B value. (Query only)

Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"ESB?" ENTER @E5100;A</pre>

ESNB *value*

Specifies the bits of event status register B.

Parameter Range	0 to 32767 ($=2^{15}-1$) (decimal expression of the event status register)
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"ESNB 0" OUTPUT @E5100;"ESNB?" ENTER @E5100;A</pre>

EXET *value*

Executes the service test. (No Query)

Equivalent Key Sequence	System EXECUTE TEST
Parameter Description	{ <i>value</i> } : test number
Examples	<pre>OUTPUT @E5100;"EXET 11"</pre>

EXPC{ON|OFF}

Sets the interpolated error correction ON or OFF.

Parameter Description	OFF and 0 : interpolation OFF ON and 1 : interpolation ON
Query Response	{0 1}
Examples	OUTPUT @E5100;"EXPC OFF" OUTPUT @E5100;"EXPC?" ENTER @E5100;A

EXPZP {ON|OFF|1|0}

Selects the expanded phase format for the impedance measurement.

Equivalent Key Sequence	Meas/Format FORMAT EXP PHASE ON OFF
Parameter Description	ON, 1 : expanded phase format ON OFF, 0 : expanded phase format OFF
Query Response	{1 0}
Examples	OUTPUT @E5100;"EXPZP ON" OUTPUT @E5100;"EXPZP" ENTER @E5100;A

EXTRLOCK?

Outputs the state of the external reference (locked or unlocked). (Query only)

Parameter Range	0 : unlocked 1 : locked
Query Response	{0 1}
Examples	OUTPUT @E5100;"EXTRLOCK?" ENTER @E5100;A

EXTT *parameter*

Selects the external trigger mode.

Equivalent Key Sequence	Trigger TRIG EVENT [], INT. TRIG ON SWEEP
Parameter Range	OFF : External trigger OFF (internal trigger mode ON) ONSWEE : External trigger on sweep
Query Response	{OFF ONSWEE}
Examples	OUTPUT @E5100;"EXTT OFF" OUTPUT @E5100;"EXTT?" ENTER @E5100;A\$

FILC *string1,string2,string3,string4*

Copies file on flexible and RAM disks. (No Query)

Equivalent Key Sequence	SAVE FILE UTILITY COPY FILE
Parameter Description	< <i>string1</i> > : Source file name. String. < <i>string2</i> > : Source device name. "MEMORY" or "DISK" ¹ < <i>string3</i> > : Destination file name. < <i>string4</i> > : Destination device name. ¹
Examples	OUTPUT 717;"FILC ""DAT1.TXT"" ""MEMORY"" ""DAT1.TXT"" ""DISK"""

¹ Internal flexible disk drive:DISK; Internal RAM disk drive:MEMORY

FMT parameter

Selects the display format.

FMT parameter

Equivalent Key Sequence	Meas/Format) FFORMAT, LOG MAG & PHASE, LOG MAG & DELAY, LIN MAG & PHASE, LIN MAG & DELAY, REAL & IMAGINARY, LOG MAG, LIN MAG, PHASE, DELAY, REAL, IMAGINARY, EXPANDED PHASE, Z & PHASE z, Y & PHASE y, R-X, G-B, Z , Y , PHASE z, PHASE y, R, G, B, X
Parameter Range	LOGM : Log magnitude format PHAS : Phase format DELA : Delay format LINM : Linear magnitude format REAL : Real format IMAG : Imaginary format LOGMP : Log magnitude and phase format LOGMD : Log magnitude and delay format LINMP : Linear magnitude and phase format LINMD : Linear magnitude and delay format EXPP : Expanded phase format MAGZP : Z & phase format MAGYP : Y & phase format IMPRX : R-X format ADMGB : G-B format MAGZ : Z format MAGY : Y format PHAZ : Phase _z format PHAY : Phase _y format IMPR : R format IMPX : X format ADMG : G format ADMB : B format MAGZDF : Z-Δ format for tracking measurement (option 023 only)
Query Response	{LOGM PHAS DELA LINM REAL IMAG LOGMP LOGMD LINMP LIMMD EXPP MAGZP MAGYP IMPRX ADMGP MAGZ MAGY PHAZ PHAY IMPR IMPX ADMG ADMB MAGZDF}
Examples	OUTPUT 717;"FMT LOGM" OUTPUT 717;"FMT?" ENTER 717;A\$

FORM2

Sets the IEEE 32-bit floating point format to transfer trace data and waveform analysis (Refer to Appendix E) data via HP-IB. (No Query)

FORM3

Sets the IEEE 64-bit floating point format to transfer the trace data and waveform analysis (Refer to Appendix E) data via HP-IB. (No Query)

FORM4

Sets the ASCII transfer format to transfer the trace data and waveform analysis (Refer to Appendix E) data via HP-IB. (No Query)

FORM5

Sets MS-DOS format to transfer the trace data and waveform analysis (Refer to Appendix E) data via HP-IB. (No Query)

GRAPFORM {PCL|TIFF}

Selects the graphic format to be save the display image on the disk drive.

Equivalent Key Sequence	Save/Recall GRAPH FMT []
Parameter Description	{PCL TIFF} PCL : PCL TIFF : TIFF
Query Response	{PCL TIFF}
Examples	OUTPUT @E5100;"GRAPFORM TIFF" OUTPUT @E5100;"GRAPFORM?" ENTER @E5100;A\$

GRODAPER *value* [pct]

Sets the group delay aperture.

Equivalent Key Sequence	Meas/Format GROUP DELY APERTURE
Parameter Range	1 to 200 (% of span) [%]
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"GRODAPER 30" OUTPUT @E5100;"GRODAPER?" ENTER @E5100;A

HOLD

Holds the present measurement.

HOLD

Equivalent Key Sequence	Trigger HOLD
Query Response	{1 0} 0 : Sweep mode 1 : Hold mode
Examples	OUTPUT @E5100;"HOLD" OUTPUT @E5100;"HOLD?" ENTER @E5100;A

IDN?

Outputs the analyzer ID.

Query Response	HEWLET-PACKARD,E5100A,JP1KCmmmm,n.nn mmmm: serial number n.nn: revision number
Examples	OUTPUT @E5100;"IDN?" ENTER @E5100;A\$

IFBW value

Sets the bandwidth value for IF bandwidth reduction.

Equivalent Key Sequence	Sweep IF BW
Parameter Range	2, 20, 200, 1000, 4000, 8000 [Hz]
Query Response	{value}
Examples	OUTPUT @E5100;"IFBW 200HZ" OUTPUT @E5100;"IFBW?" ENTER @E5100;A

IFBWAUTO

Automatically selects the proper IF bandwidth for each measurement point. (No Query)

Equivalent Key Sequence	Sweep IF BW [] IF BW AUTO
Examples	OUTPUT @E5100;"IFBWAUTO"

INID

Initializes the disk in the built-in flexible disk drive. (No Query)

Equivalent Key Sequence	Save/Recall FILE UTILITY INITIALIZE DISK
Examples	<pre>OUTPUT @E5100;"STDDISK" OUTPUT @E5100;"INID"</pre>

INP8IO

Inputs data from the 4-bit parallel input port to the HP E5100A/B. (option 005 only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"INP8IO" OUTPUT @E5100;"OUTPINP8IO?" ENTER @E5100;A OUTPUT @E5100;"INP8IO?" ENTER @E5100;A</pre>

INPT?

Outputs value which tells whether there is pulse input at the Input1 port of the 24-bit I/O port. (Query only)

Query Response	{1 0} 0 : There is not pulse input at the Input1. 1 : there is pulse input at the Input1. Once INPT? returns 1, next INPT? query returns 0 until the next pulse input has occurred at Input1.
Examples	<pre>OUTPUT @E5100;"INPT?" ENTER @E5100;A</pre>

INPUCALC {01-03} value

Stores the measurement calibration error coefficient set real/imaginary pairs input via HP-IB into instrument memory. The data transfer format must be used ASCII format when this command. (No Query)

Parameter Range	<value> : Complex number (Data format: real, imaginary)
Examples	<pre>DIM A(1:201,1:2) NOP: 201 !Set calibration error coefficient OUTPUT @E5100;"INPUCALC01 ";A(*)</pre>

INPUDDATA *<value1>*,*<value2>*, ... ,*<valuen>*

Enters data to the data array of the active channel. (No Query)

Parameter Range	<i><value></i> : Complex value (data format:Real, Imaginary)	
Examples	DIM A(1:201,1:2)	<i>NOP: 201</i>
	! Set data to enter memory array	
	OUTPUT @E5100;"INPUDDATA";A(*)	

INPUFORM *value*

Inputs formatted data. (No Query)

Parameter Range	<i><value></i> : Complex value (Data format: real, imaginary)	
Examples	DIM A(1:201,1:2)	<i>NOP: 201</i>
	! Set formatted data	
	OUTPUT @E5100;"INPUFORM ";A(*)	

INPUIFORM *<value1>*,*<value2>*, ... ,*<valuen>*

Enters data to the formatted data of the inactive channel. (No Query)

When the number of channel is 3 or 4, this command enters data of the following channel:

Active Channel	Channel the INPUIFORM? command enters data
1	2
2	1
3	4
4	3

Parameter Range	<i><value></i> : Complex value (data format:Real, Imaginary)	
Examples	DIM A(1:201,1:2)	<i>NOP: 201</i>
	! Set data to enter memory array	
	OUTPUT @E5100;"INPUIFORM";A(*)	

INPUMEMO *<value1>*,*<value2>*, ... ,*<valuen>*

Enters data to the memory array of the active channel. (No Query)

Parameter Range	<i><value></i> : Complex value (data format:Real, Imaginary)	
Examples	DIM A(1:201,1:2)	<i>NOP: 201</i>
	! Set data to enter memory array	
	OUTPUT @E5100;"INPUMEMO";A(*)	

INPUSTIM \square $\langle value1 \rangle, \langle value2 \rangle, \dots, \langle valuen \rangle$

INPURAW \square $\langle value (1) \rangle, \langle value (2) \rangle, \dots, \langle value (n) \rangle$

Enters data to the raw data array of the active channel. (No Query)

Parameter Range	$\langle value \rangle$: Complex value (data format: real, imaginary)
Examples	<pre>DIM A(1:201,1:2) NOP: 201 ! Set data to enter to raw array OUTPUT @E5100;"INPURAW";A(*)</pre>

INPUFORM $\langle value1 \rangle, \langle value2 \rangle, \dots, \langle valuen \rangle$

Enters data to the format data array of the active channel. (No Query)

Parameter Range	$\langle value \rangle$: Complex value (data format : real, imaginary)
Examples	<pre>DIM A(1:201,1:2) NOP: 201 ! Set data to enter to format array OUTPUT @E5100;"INPUFORM ";A(*)</pre>

INPURTMEM $\langle value (1) \rangle, \langle value (2) \rangle, \dots, \langle value (n) \rangle$

Enters data to the sub-trace array of the active channel. The sub-trace is available when the number of channels is set to 2. (No Query)

Parameter Description	$\langle value \rangle$: real value
Examples	<pre>DIM A(1:201) NOP: 201 ! Set data to enter to sub-trace OUTPUT @E5100;"INPURTMEM";A(*)</pre>

INPUSTIM \square $\langle value1 \rangle, \langle value2 \rangle, \dots, \langle valuen \rangle$

Enters data to the stimulus data array of the active channel. (No Query)

Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz (frequency sweep) -9 to +11 dBm (-48 to +22 dBm option 010 only) (power sweep)
Examples	<pre>DIM A(1:201) NOP: 201 ! Set stimulus data OUTPUT @E5100;"INPUSTIM";A(*)</pre>

INPUTRAC□<value1>,<value2>,<value3>

Pass the phase value to trap at a measured point and specify whether data is outputted to the I/O port or not when the measured value reaches the phase value. The data to be outputted to the I/O port is specified by INPUTRACB. (No Query,Option 022 only)

Parameter Description	<value1> : The measured point number <value2> : The trap phase value <value3> : I/O output ON/OFF
Examples	OUTPUT @E5100;"INPUTRAC 10, 0, ON"

INPUTRACB□<value1>,<value2>,<value3>

Pass the data outputted to the I/O port when the measured value of phase reaches the limit value specified by the INPUTRAC command. Multiple data up to 6 data can be outputted continuously to the I/O port. (No Query,Option 022 only)

Parameter Description	<value1> : The measured point number <value2> : The order of output I/O data (1 to 6) <value3> : Data Query Response
Examples	OUTPUT @E5100;"INPUTRACB 5, 1, ";DVAL("1010",2) OUTPUT @E5100;"INPUTRACB 5, 2, ";DVAL("0100",2) OUTPUT @E5100;"INPUTRACB 5, 3, ";DVAL("0000",2)

If HP E5100A/B OPT. 022 is set up by the above commands, the data described in the following figure will be outputted to the I/O port when the measured phase reaches the target value at the measured point 5.




Figure 8-1. The Data Output Timing Chart

IOP0?

Returns the installed option number of the I/O port of the rear panel. If not option is installed, IOP0? returns string, "STD". If Option 005 is installed, IOP0? returns "005", and Option 006 is installed, IOP0? returns "006". (Query only)

Query Response	{STD 005 006} STD : Standard I/O port 005 : Option 005 I/O port 006 : Option 006 I/O port
Examples	OUTPUT @E5100;"IOP0?" ENTER @E5100;A\$

LIMILINE {ON|OFF}

Sets limit line display ON or OFF.

Equivalent Key Sequence	SYSTEM LIMIT LINE ON off
Parameter Range	OFF : Limit lines OFF ON : Limit lines ON
Query Response	{1 0} 0 : Limit lines OFF 1 : Limit lines ON
Examples	OUTPUT 717;"LIMILINE ON" OUTPUT 717;"LIMILINE?" ENTER 717;A

LIMITEST {ON|OFF}

Sets the limit testing ON or OFF.

Equivalent Key Sequence	SYSTEM LIMIT TEST ON off
Parameter Range	OFF : Limit testing OFF ON : Limit testing ON
Query Response	{1 0}
Examples	OUTPUT @E5100;"LIMITEST ON" OUTPUT @E5100;"LIMITEST?" ENTER @E5100;A

LINFREQ

This command equals SWPT LINF. See SWPT.

LINM

This command equals FMT LINM. See FMT.

LISDFBASE

Display trace on frequency base when the frequency list sweep is used. (No Query)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU MORE LIST DISP: FREQ BASE
Examples	OUTPUT @E5100;"LISDFBASE"

LISDOBASE

Displays the measured data on order base.

Equivalent Key Sequence	Sweep SWEEP TYPE MENU MORE ORDER BASE
Examples	OUTPUT @E5100;"LISDOBASE"

LISFREQ

This command equals SWPT LIST. See SWPT.

LISLIS1

Activates LIST 1 for the list sweep. (No Query)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU MORE LIST []
Examples	OUTPUT @E5100;"LISLIS1"

LISLIS2

Activates LIST 2 for the list sweep. (No Query)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU MORE LIST []
Examples	OUTPUT @E5100;"LISLIS2"

LOWELIMI <value1><value2><value3> ... <valuen>

LMAX? *value*

Outputs the *n*th local maximum value from the left of range which is set by the ANARANG command. (Query only)

Parameter Range	1 to
Query Response	{value} 3.40282347E+38 will be output when no appropriate points are found.
Examples	<pre>OUTPUT @E5100;"LMAX 1" OUTPUT @E5100;"LMAX?" ENTER @E5100;A</pre>

LMIN? *value*

Outputs the *n*th local minimum value from the left of range which is set by the ANARANG command. (Query only)

Parameter Range	1 to
Query Response	{value} 3.40282347E+38 will be output when no appropriate points are found.
Examples	<pre>OUTPUT @E5100;"LMIN 1" OUTPUT @E5100;"LMIN?" ENTER @E5100;A</pre>

LOGM

This command equals FMT LOGM. See FMT.

LOGMD

This command equals FMT LOGMD. See FMT.

LOGMP

This command equals FMT LOGMP. See FMT.

LOWELIMI <value1><value2><value3> ... <valuen>

Sets the lower limit values of the limit line. The lower value can be set at each measurement point.

LOWELIMI <value1><value2><value3> ... <valuen>

Parameter Description	n = Number of points <value> : Lower value of limit line
Query Response	{value1} {value2} {value3} ... {valuen}
Examples	<pre>DIM A(1:201) NOP: 201 Set lower value of limit line OUTPUT @E5100;"LOWELIMI";A(*) OUTPUT @E5100;"LOWELIMI?" ENTER @E5100;A(*)</pre>

MARD {ON|OFF}

Displays (ON) or does not display (OFF) markers and the marker information on the screen.

Parameter Range	OFF : Marker display OFF ON : Marker display ON
Query Response	{1 0} 0 : Marker display OFF 1 : Marker display ON
Examples	<pre>OUTPUT @E5100;"MARD OFF" OUTPUT @E5100;"MARD?" ENTER @E5100;A</pre>

MARK {1-4}

Selects the active marker and sets the marker stimulus value.

Equivalent Key Sequence	Marker ACTIVE MARKER 1 to 4
Parameter Range	<value> : 10×10^3 (=10k) to 300×10^6 (=300M) Hz <value> : -9 to +11 dBm (-48 to +22 dBm option 010 only) (Option 010, frequency sweep only)
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"MARK1 20MHZ" OUTPUT @E5100;"MARK1?" ENTER @E5100;A</pre>

MARKBUCK *value*

Moves the active marker to specified data point number. (No Query)

Parameter Range	1 to "number of points"
Examples	OUTPUT @E5100;"MARKBUCK 20"

MARKCENT

Changes the stimulus center value to the active marker value. (No Query)

Equivalent Key Sequence	Center, Span, Start, Stop, MARKER-CENTER
Examples	OUTPUT @E5100;"MARKCENT"

MARKCONT

Interpolates between measured points to allow the markers to be placed at any point on the trace.

Equivalent Key Sequence	Marker MKR MODE MENU CONTINUOUS
Examples	OUTPUT @E5100;"MARKCONT"

MARKCOUP

Couples the marker stimulus values for the two display channels.

Equivalent Key Sequence	Marker MKR MODE MENU MARKERS: COUPLED
Examples	OUTPUT @E5100;"MARKCOUP"

MARKDISC

Places markers only on measured trace points determined by the stimulus settings.

Equivalent Key Sequence	Marker MKR MODE MENU MARKERS: DISCRETE
Examples	OUTPUT @E5100;"MARKDISC"

MARKFSTI *value* [*suffix*]

Sets the fixed marker stimulus value offset.

Equivalent Key Sequence	Marker AMODE MENU FIXED MKR POSITION FIXED MKR STIMULUS
Parameter Range	<value> : 10 k to 300 MHz <value> : -64 to 18 dBm
Query Response	{value}
Examples	OUTPUT @E5100;"MARKFSTI" OUTPUT @E5100;"MARKFSTI?" ENTER @E5100;A

MARKFVAL *value* [*suffix*]

Sets the fixed marker position value offset.

Equivalent Key Sequence	Marker AMODE MENU FIXED MKR POSITION FIXED MKR VALUE
Parameter Range	-500×10^5 to 500×10^5
Query Response	{value}
Examples	OUTPUT @E5100;"MARKFVAL" OUTPUT @E5100;"MARKFVAL?" ENTER @E5100;A

MARKL {ON|OFF}

Displays (ON) or does not display (OFF) the list of stimulus values and response values of all markers.

Equivalent Key Sequence	MKR MKR LIST on OFF
Parameter Range	OFF : Marker list OFF ON : Marker list ON
Query Response	0 : Marker list OFF 1 : Marker list ON
Examples	OUTPUT @E5100;"MARKL OFF" OUTPUT @E5100;"MARKL?" ENTER @E5100;A

MARKODATA

Enables the marker to move on the measurement data trace.

Examples	OUTPUT @E5100;"MARKODATA"
-----------------	---------------------------

MARKOFF

Turns off all the markers and the delta reference marker.

Equivalent Key Sequence	Marker ALL MKR OFF
Examples	OUTPUT @E5100;"MARKOFF"

MARKOMEMO

Enables the marker to move on the memory data trace.

Examples	OUTPUT @E5100;"MARKOMEMO"
-----------------	---------------------------

MARKREF

Changes the reference value to the active marker's response value, without changing the reference position. (No Query)

Examples	OUTPUT @E5100;"MARKREF"
-----------------	-------------------------

MARKSPAN

Changes the start and stop values of the stimulus span to the active marker and the delta reference marker.

Equivalent Key Sequence	Center, Span, Start, Stop, MARKER SPAN
Examples	OUTPUT 717;"MARKSPAN"

MARKSTAR

Changes the start value of the stimulus to the active marker value. (No Query)

Equivalent Key Sequence	Center, Span, Start, Stop, MARKER → START
Examples	OUTPUT 717;"MARKSTAR"

MARKSTOP

Changes the stop value of the stimulus to the active marker value. (No Query)

Equivalent Key Sequence	Center, Span, Start, Stop, MARKER → STOP
Examples	OUTPUT 717;"MARKSTOP"

MARKTIME {OFF|ON}

Sets the x-axis marker readout to the sweep time (ON), or cancels the setting (OFF).

Equivalent Key Sequence	Mkr MKR TIME on OFF
Parameter Range	OFF: Marker time OFF ON : Marker time ON
Query Response	{1 0}
Examples	OUTPUT @E5100;"MARKTIME ON" OUTPUT @E5100;"MARKTIME?" ENTER @E5100;A

MARKUNCO

Allows the marker stimulus values to be controlled independently on each channel.

Equivalent Key Sequence	Marker MKR MODE MENU UNCOUPLE
Examples	OUTPUT @E5100;"MARKUNCO"

MARKZERO

Puts a fixed reference marker at the present active marker position, and makes the fixed marker stimulus and response values at that position equal to zero.

Equivalent Key Sequence	Marker <code>AMODE MENU MKR ZERO AREP=A</code>
Examples	<code>OUTPUT @E5100;"MARKZERO"</code>

MAXPOIN?

Outputs the maximum number of the measurement points. (Query only)

Query Response	{401 801 1601}
Examples	<code>OUTPUT @E5100;"MAXPOIN?"</code> <code>ENTER @E5100;A</code>

MAXPORT?

Outputs the number of ports (receivers). (Query only)

Query Response	{1 2 3 4}
Examples	<code>OUTPUT @E5100;"MAXPORT?"</code> <code>ENTER @E5100;A</code>

MEAS1PT? <value>

Outputs measurement value at the point number specified by the parameter.

When the trap function is turn on (TRAP ON), MEAS1PT waits to start a measurement until the phase value is in the condition specified by INPUTRAC command, and returns the query response.

Parameter Range	1 to NOP
Query Response	{value}
Examples	<code>OUTPUT @E5100;"MEAS1PT? 10000000"</code> <code>OUTPUT @E5100;"MEAS1PT?"</code> <code>ENTER @E5100;A</code>

MEAS *parameter*

Selects the parameters or inputs to be measured.

Equivalent Key Sequence	MEAS A/R, R, A
Parameter Description	AR : A/R measurement BR : A/R measurement CR : A/R measurement RA : R/A measurement BA : B/A measurement CA : C/A measurement RB : R/B measurement AB : A/B measurement CB : C/B measurement RC : R/C measurement AC : A/C measurement BC : B/C measurement R : R measurement A : A measurement B : B measurement C : C measurement
Query Response	{AR BR CR RA BA CA RB AB CB RC AC BC R A B C}
Examples	OUTPUT @E5100;"MEAS AR" OUTPUT @E5100;"MEAS?" ENTER @E5100;A\$

MEASA

This command equals MEAS A. See MEAS.

MEASR

This command equals MEAS R. See MEAS.

MEASTAT {ON|OFF}

Calculates and displays the mean, standard deviation, and peak-to-peak values among the search range (ON), or does not display them (OFF).

Equivalent Key Sequence	Marker UTILITY MENU STATISTICS ON OFF
Parameter Range	OFF : Marker statistic OFF ON : Marker statistic ON
Query Response	{1 0} 0 : Marker statistic OFF 1 : Marker statistic ON
Examples	OUTPUT @E5100;"MEASTAT ON" OUTPUT @E5100;"MEASTAT?" ENTER @E5100;A

MENU3 \square <value>, <string1>, <string2>, <string3>

Defines the user softkey menu of **Menu3**. (No Query)

Parameter Description	<value> : Softkey position (1 to 8) <string1> : Upper softkey label (Up to ten characters) <string2> : Lower softkey label (Up to ten characters) <string3> : HP-IB command to be executed
Examples	OUTPUT @E5100;"MENU3 1,""START"",""100 MHz"",""STAR 100MA"""

MOHMSW {A|B},{ON|OFF}

Sets 1 MΩ input ON or OFF. (option 101 only)

Equivalent Key Sequence	System INPUT Z PORT A \square , PROT B \square
Parameter Description	A : port A B : port B ON : 1 MΩ input OFF : 50 Ω input
Query Response	{ON OFF}
Examples	OUTPUT @E5100;"MOHMSW A,ON" OUTPUT @E5100;"MOHMSW? A" ENTER @E5100;A\$

8. Command Reference

MONDYEAR

Changes the displayed date to the "month:day:year" format. (No Query)

Equivalent Key Sequence	System MORE SET CLOCK DATE MODE: MonDayYear
Examples	OUTPUT @E5100;"MONDYEAR"

MOVADARY

Enters measurement data triggered by TRIGMEAS command. This command is only used for the parallel processing using with TRIGMEAS and MOVADARY. (No Query)

MULC {OFF|ON|0|1}

Selects the multi (ON) or single (OFF) channels display.

Equivalent Key Sequence	Display MULTI CH ON OFF
Parameter Description	OFF and 0 : Active channel only ON and 1 : Multi-channel
Query Response	{0 1}
Examples	OUTPUT @E5100;"MULC ON" OUTPUT @E5100;"MULC?" ENTER @E5100;A

NEGL

Sets the output of the 24-bit I/O port to the negative logic. (No Query)

Examples	OUTPUT @E5100;"NEGL"
----------	----------------------

NEXPK?

Outputs the maximum local maximum value and its stimulus next to the value last found by the PEAK?, or NEXPK? commands. For more information, refer to Appendix D. (Query only)

Query Response	{value1}{value2} <value1> : Maximum local maximum value <value2> : Stimulus value
Examples	OUTPUT @E5100;"NEXPK?" ENTER @E5100;A,B

NUMC <value>

Sets the number of channel.

Equivalent Key Sequence	Meas/Format NUM of CH []
Parameter Range	1 to 4
Query Response	{value}
Examples	OUTPUT @E5100;"NUMC 3" OUTPUT @E5100;"NUMC?" ENTER @E5100;A

NUMG value

Triggers a user-specified number of sweeps, and returns to the HOLD mode. (No Query)

Parameter Range	Greater than 0, integer only
Examples	OUTPUT @E5100;"NUMG 10"

NUMLMAX?

Outputs the number of local maximums within the range set by the ANARANG command. (Query only)

Query Response	{value} (number of local maximums)
Examples	OUTPUT @E5100;"NUMLMAX?" ENTER @E5100;A

NUMLMIN?

Outputs the number of local minimum within the range set by the ANARANG command. (Query only)

Query Response	{value} (number of local minimums)
Examples	OUTPUT @E5100;"NUMLMIN?" ENTER @E5100;A

NUMLMINMAX?

Outputs the total number of the local maximum and local minimum points within the range specified by the ANARANG command. (Query only)

Query Response	{value}
Examples	OUTPUT @E5100;"NUMLMINMAX?" ENTER @E5100;A

OSE value

Enables the operational status register.

Parameter Range	0 to 65535 ($=2^{16}-1$, decimal expression of enable bits of the the operational status register)
Query Response	{value}
Examples	OUTPUT @E5100;"OSE 1"

OSER?

Outputs the current value in the event register of an operational status register. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OSER?" ENTER @E5100;A</pre>

OSNT

Sets the negative transition filter of an operational status register.

Parameter Range	0 to 65535 ($=2^{16}-1$, decimal expression of the operational status register)
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OSNT 1" OUTPUT @E5100;"OSNT?" ENTER @E5100;A</pre>

OSPT

Sets the positive transition filter of an operational status register.

Parameter Range	0 to 65535 ($=2^{16}-1$, decimal expression of the operational status register)
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OSPT 1" OUTPUT @E5100;"OSPT?" ENTER @E5100;A</pre>

OSR?

Outputs the operational status register value. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OSR?" ENTER @E5100;A</pre>

OUT1ENVH

Sets OUTPUT1 set to HIGH when a pulse input has occurred at INPUT1. (No Query)

Examples

```
OUTPUT @E5100;"OUT1ENVH"
```

OUT1ENVL

Sets OUTPUT1 set to LOW when a pulse input has occurred at INPUT1. (No Query)

Examples

```
OUTPUT @E5100;"OUT1ENVL"
```

OUT1H

Sets OUTPUT1 to HIGH. (No Query)

Examples

```
OUTPUT @E5100;"OUT1H"
```

OUT1L

Sets OUTPUT1 to LOW. (No Query)

Examples

```
OUTPUT @E5100;"OUT1L"
```

OUT2ENVH

Sets OUTPUT2 set to HIGH when a pulse input has occurred at INPUT1 of the 24 bit I/O port. (No Query)

Examples

```
OUTPUT @E5100;"OUT2ENVH"
```

OUT2ENVL

Sets OUTPUT2 set to LOW when a pulse input has occurred at INPUT1 of the 24 bit I/O port. (No Query)

Examples

```
OUTPUT @E5100;"OUT2ENVL"
```

OUT2H

Sets OUTPUT2 to HIGH. (No Query)

Examples	OUTPUT @E5100;"OUT2ENVL"
-----------------	--------------------------

OUT2L

Sets OUTPUT2 to LOW. (No Query)

Examples	OUTPUT @E5100;"OUT2L"
-----------------	-----------------------

OUT8IO value

Outputs the data to the 8-bit parallel output port. (Option 005 only, No Query)

Parameter Range	0 to 255
Examples	OUTPUT @E5100;"OUT8IO 0"

OUTAIO value

Output decimal data specified as the parameter to port A (8 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 255
Examples	OUTPUT @E5100;"OUTAIO 0"

OUTBIO value

Output decimal data specified as the parameter to port B (8 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 255
Examples	OUTPUT @E5100;"OUTBIO 0"

OUTCIO *value*

Output decimal data specified as the parameter to port C (4 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 15
Examples	OUTPUT @E5100;"OUTCIO 0"

OUTDIO *value*

Output decimal data specified as the parameter to port D (4 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 15
Examples	OUTPUT @E5100;"OUTDIO 0"

OUTEIO *value*

Output decimal data specified as the parameter to port E (8 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 255
Examples	OUTPUT @E5100;"OUTEIO 0"

OUTFIO *value*

Output decimal data specified as the parameter to port F (16 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 65535
Examples	OUTPUT @E5100;"OUTFIO 0"

OUTGIO *value*

Output decimal data specified as the parameter to port G (20 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 1048575
Examples	OUTPUT @E5100;"OUTGIO 0"

OUTHIO *value*

Output decimal data specified as the parameter to port H (24 bit) of the 24-bit I/O port. (No Query)

Parameter Range	0 to 16777215
Examples	OUTPUT @E5100;"OUTHIO 0"

OUTPCALC{01-03}?

Outputs the active calibration set array of the active channel. (Query only)

Parameter Range	
Query Response	{ <i>value</i> (1)} { <i>value</i> (2)} ... { <i>value</i> (n)} n: Number of points <i>value</i> : Complex value (data format : real, imaginary)
Examples	DIM A(1:201,1:2) <i>NOP: 201</i> OUTPUT @E5100;"OUTCALC01?" ENTER @E5100;A(*)

OUTPCERR?

Outputs the ceramic resonator parameters within the range specified by the ANARANGE command. (Query only)

Query Response	{ <i>value</i> 1}, { <i>value</i> 2}, { <i>value</i> 3}, { <i>value</i> 4}, { <i>value</i> 5}, { <i>value</i> 6}, { <i>value</i> 7} { <i>value</i> 1} : Gain at f_r (G_r) { <i>value</i> 2} : Resonant frequency (f_r) { <i>value</i> 3} : Gain at (f_a) { <i>value</i> 4} : Anti-resonant frequency f_a (G_a) { <i>value</i> 5} : Maximum ripple height in left side of resonant point { <i>value</i> 6} : Maximum ripple height between resonant and anti-resonant point { <i>value</i> 7} : Maximum ripple height in right side of anti-resonant point If OUTPCERR? could not find any ripples, the query returns 0. This command is only available when the "LOG MAG & Phase", "LOG MAG & Delay" or "LOG MAG" formats are selected. If another format is selected, the query returns 0. If Z-conversion is selected, then the impedance at f_r (Z_r) instead of the G_r and the impedance at f_a (Z_a) instead of the G_a are returned.
Query Response	{ <i>value</i> 1}, { <i>value</i> 2}, ... , { <i>value</i> 7}
Examples	OUTPUT @E5100;"OUTPCERR?" <i>Query the ceramic resonator parameters.</i> ENTER @E5100;Gr,Fr,Ga,Fa,Rp11,Rp12,Rp13 <i>Recieve the all return value.</i>

OUTPCF2? □ <value1>, <value2>, <value3>, <value4>, <value5> ... , <valuen + 4>

OUTPCF2? □ <value1>, <value2>, <value3>, <value4>, <value5> ... , <valuen + 4>

Outputs filter parameters within the range specified by the ANARANG command. This command outputs the same parameters as OUTPCFIL and outputs up to 20 sets of frequency offsets from center frequency (f_c) to left and right cutoff points ($\Delta f_{left\ n}$, $\Delta f_{right\ n}$).

Where,

- If two cutoff points which are x_1 dB below the maximum peak are not found, zeros will be returned for all parameters.
- If two cutoff points which are X_n dB below the maximum value are not found, zeroes will be returned for $\Delta f_{left\ n}$ and $\Delta f_{right\ n}$.
- If no peaks are found, zeroes will be returned for Pole $_{x1}$, Pole $_{stim1}$, Pole $_{x2}$, Pole $_{stim2}$.*

Parameter Description	<p><value1> : Nominal frequency (f_c) <value2> : Difference vale from the maximum value (D) <value3> : Stop frequency of rejection band (f_1) <value4> : Start frequency of sprious level range (f_2) <value5> : Relative offset value from maximum peak value to determine the cutoff point (x_1 [dB]) <value6> : Relative offset value from maximum peak value to determine the cutoff point (x_2 [dB]) ... <valuen + 4> : Relative offset value from maximum peak value to determine the cutoff point (x_n [dB]) ($2 \leq n \leq 20$)</p>
Query Response	<p>{value1}{value2} ... {value(2n + 1)}</p> <p>{value1} : insertion loss {value2} : constant loss (Loss$_c$) {value3} : x_1 dB bandwidth {value4} : center freequency (f_{cent}) {value5} : Q {value6} : Ripple value at the passband {value7} : Rejection {value8} : Spurious level {value9} : Left pole (negative peak in the left side of the maximum value) (Pole$_{x1}$) {value10} : Stimulus value of Pole$_{x1}$ (Pole$_{stim1}$) {value11} : Right pole (negative peak in the right side of the maximum value) (Pole$_{x2}$) {value12} : Stimulus value of Pole$_{x2}$ (Pole$_{stim2}$) {value13} : difference between f_c and the left point of cutoff frequency (Δf_{left1}) {value14} : difference between f_c and the right point of cutoff frequency (Δf_{right1}) {value15} : Δf_{left2} {value16} : Δf_{right2} ... {value2n + 10} : $\Delta f_{left\ n}$ {value2n + 11} : $\Delta f_{right\ n}$</p>
Examples	<p>OUTPUT @E5100;"OUTPCF2? 90M,0.1,90.2,3,10,12,20" ENTER @E5100;A(10),B(*)</p>

8 Command Reference

OUTPCFIL? *value1,value2,value3,value4,value5,value6*

Outputs filter parameters within the range specified by the ANARANG command. Command parameter sets nominal frequency, the offset of $x1$ dB and $x2$ dB to the maximum peak value to determine the cutoff points, same parameter with POLE?, and f_1 and f_2 for determining rejection level and spurious level respectively. For details, refer to Appendix D. (Query only)

Parameter Description	<p><value1> : Nominal frequency (f_c) <value2> : Relative offset value from maximum peak value to determine the cutoff point (x_1) [dB] <value3> : Relative offset value from maximum peak value to determine the cutoff point (x_2) [dB] <value4> : POLE? parameter <value5> : Stop frequency of rejection band (f_1) <value6> : Start frequency of spurious level range (f_2)</p>
Query Response	<p>{value1}, {value2}, ... , {value16}</p> <p>{value1} : insertion loss {value2} : constant loss {value3} : xdB bandwidth {value4} : center freequency {value5} : Q {value6} : difference between the middle point of the analysis range and the left point of cutoff frequency {value7} : difference between the middle point of the analysis range and the right point of cutoff frequency {value8} : difference between the middle point of the analysis range and the left point of cutoff frequency {value9} : difference between the middle point of the analysis range and the right point of cutoff frequency {value10} : Ripple value at the passband {value11} : Rejection {value12} : Spurious level {value13} : Left pole (negative peak in the left side of the maximum value) {value14} : Stimulus value of the left pole {value15} : Right pole (negative peak in the right side of the maximum value) {value16} : Stimulus value of the right pole</p>
Examples	<pre>DIM ANS(1:16) OUTPUT @E5100;"OUTPCFIL?";7.0E6,-10,-20,-40,1E3,1E3 ENTER @E5100;ANS(*)</pre>

OUTPDATA?

Outputs data trace value. (Query only)

Query Response	<p>{numeric (1)} {numeric (2)} ... {numeric (n)}</p> <p>n : Number of points numeric : Complex value (data format: real, imaginary)</p>
Examples	<pre>DIM A(1:201,1:2) NOP: 201 OUTPUT @E5100;"OUTPDATA?" ENTER @E5100;A(*)</pre>

OUTPDATAP *<value>*

Outputs the data at the specified point from the data array.

Parameter Range	<i><value></i> : number of points :(1 to NOP)
Query Response	{ <i>value</i> } <i>value</i> : Complex data (data format: real, imaginary)
Examples	OUTPUT @E5100;"OUTPDATAP? 1" ENTER @E5100;A

OUTPDATAT?

Outputs data trace value on 16 points stimulus which is set by the STIDROUT command. If there are points that is not set by the STIDROUT command, the OUTPDATAT? returns the value at 100 kHz. (Query only)

Query Response	{ <i>value1</i> } { <i>value2</i> } ... { <i>value16</i> }
Examples	DIM A(1:16) OUTPUT @E5100;"STIDROUT1";Freq OUTPUT @E5100;"OUTPDATAT?" ENTER @E5100;A(*)

OUTPDATTP? *value*

Outputs the data-trace data at the specified point. (Query only)

Parameter Range	<i>value</i> : point number <i>n</i> :(1 to Number of Points)
Query Response	{ <i>numeric</i> }
Examples	OUTPUT @E5100;"OUTPDATTP ";1 ENTER @E5100;A

OUTPERRO?

Outputs the error message in the error queue. (Query only)

Query Response	{ <i>value</i> }, { <i>string</i> } <i><value></i> : Error number <i><string></i> : Error string
Examples	OUTPUT @E5100;"OUTPERRO?" ENTER @E5100;A,ERR\$

OUTPFILT? *value*

Outputs filter parameters within the range specified by the ANARANG command. Command parameter sets the offset of *x*dB to the maximum peak value to determine the cutoff points. For details, refer to Appendix D. (Query only)

Parameter Description	<value> : Relative offset value from maximum <i>x</i> [dB]
Query Response	{value1}, {value2}, ... , {value6} {value1} : Insertion loss {value2} : <i>x</i> dBbandwidth {value3} : Center frequency {value4} : Q {value5} : Difference between the middle value of analysis range and the left cutoff frequency (Δ L.F.) {value6} : Difference between the middle value of analysis range and the right cutoff frequency (Δ R.F.)
Examples	OUTPUT @E5100;"OUTPFILT? -3" ENTER @E5100;ANS(*)

OUTPFORM?

Outputs the formatted trace data. (Query only)

Query Response	{value (1)} {value (2)} ... {value (n)} n: Number of point value: complex value (data format : Real, Imaginary)
Examples	DIM A(1:201,1:2) NOP: 201 OUTPUT @E5100;"OUTPFORM?" ENTER @E5100;A(*)

OUTPFORMP? *value*

Outputs the formatted trace data at the specified point. (Query only)

Parameter Range	1 to "number of points"
Query Response	{value} value : complex value (data format : real, imaginary)
Examples	OUTPUT @E5100;"OUTPFORMP?" ENTER @E5100;A,B

OUTPIFORM?

Outputs the formatted data from the inactive channel. (Query only)

When the number of channel is 3 or 4, this command outputs data from the following channel:

Active Channel	Channel the OUTPIFORM? command outputs
1	2
2	1
3	4
4	3

Query Response	{value (1)} {value (2)} ... {value (n)}
	n: number of points value : Complex value (data format: real, imaginary)
Examples	<pre>DIM A(1:201,1:2) NOP: 201 OUTPUT @E5100;"OUTIFORM?" ENTER @E5100;A(*)</pre>

OUTPINP8IO?

Outputs the data entered from the 4-bit parallel input port. (option 005 only; Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OUTPINP8IO?" ENTER @E5100;A</pre>

OUTPINPCIO?

Outputs the data entered from port C (4 bit) of the 24-bit I/O port. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OUTPINPCIO?" ENTER @E5100;A</pre>

OUTPINPDIO?

Outputs the data entered from port D (4 bit) of the 24-bit I/O port. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OUTPINPDIO?" ENTER @E5100;A</pre>

OUTPINPEIO?

Outputs the data entered from port E (8 bit) of the 24-bit I/O port. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OUTPINPEIO?" ENTER @E5100;A</pre>

OUTPIRFORM?

Outputs the real part of the formatted data from the inactive channel. (Query only)

Query Response	{value (1)} {value (2)} ... {value (n)}
	n: Number of points
Examples	<pre>DIM A(1:201) NOP: 201 OUTPUT @E5100;"OUTPIRFORM?" ENTER @E5100;A(*)</pre>

OUTPIRTMEM?

Outputs the real part of the trace memory data from the inactive channel. (Query only)

When the number of channel is 3 or 4, this command outputs data from the following channel:

Active Channel	Channel the OUTPIRFORM? command outputs data
1	2
2	1
3	4
4	3

Query Response	{value (1)} {value (2)} ... {value (n)}
	n : Number of points
Examples	<pre>DIM A(1:201) NOP: 201 OUTPUT @E5100;"OUTPIRTMEM?" ENTER @E5100;A(*)</pre>

OUTPMARK?

Outputs the active marker values. (Query only)

Query Response	{ <i>numeric1</i> }, { <i>numeric2</i> }, { <i>numeric3</i> } { <i>numeric1</i> } : marker value { <i>numeric2</i> } : marker aux. value { <i>numeric3</i> } : stimulus
Examples	OUTPUT @E5100;"OUTPMARK?" ENTER @E5100;A,B,C

OUTPMAX?

Outputs the maximum value within the range specified with the ANARANG command. For details, refer to Appendix D. (Query only,)

Query Response	{ <i>numeric1</i> }, { <i>numeric2</i> } { <i>numeric1</i> } : maximum value { <i>numeric2</i> } : stimulus value
Examples	OUTPUT @E5100;"OUTPMAX?" ENTER @E5100;A,B

OUTPMEAN?

Outputs the mean value within the range specified with the ANARANG command. For details, refer to Appendix D. (Query only)

Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"OUTPMEAN?" ENTER @E5100;A

OUTPMEMO?

Outputs memory trace value. (Query only)

Query Response	{ <i>numeric (1)</i> } { <i>numeric (2)</i> } ... { <i>numeric (n)</i> } n : Number of points <i>numeric</i> : Complex value (data format: real, imaginary)
Examples	DIM A(1:201,1:2) <i>NOP: 201</i> OUTPUT @E5100;"OUTPMEMO?" ENTER @E5100;A(*)

OUTPMEMOT?

Outputs memory trace value on 16 points stimulus which is set by the STIMROUT command. If there are points that is not set by STIMROUT command, OUTPMEMOT? returns the value at 100 kHz. (Query only)

Query Response	{ <i>numeric1</i> } { <i>numeric2</i> } ... { <i>numeric16</i> }
Examples	<pre>DIM A(1:16) OUTPUT @E5100;"STIMROUT1";Freq OUTPUT @E5100;"OUTPMEMOT?" ENTER @E5100;A(*)</pre>

OUTPMEMTP? value

Outputs the memory data at a specified point. (Query only)

Parameter Description	<i>value</i> : point number <i>n</i> :(1 to Number of Points)
Query Response	<pre>{<i>numeric1</i>},{<i>numeric2</i>} {<i>numeric1</i>} : real part {<i>numeric2</i>} : imaginary part</pre>
Examples	<pre>OUTPUT @E5100;"OUTPMEMTP? ";1 ENTER @E5100;A,B</pre>

OUTPMIN?

Outputs the minimum value within the range specified with the ANARANG command. For details, refer to Appendix D. (Query only)

Query Response	<pre>{<i>numeric1</i>}, {<i>numeric2</i>} {<i>numeric1</i>} : minimum value {<i>numeric2</i>} : stimulus value</pre>
Examples	<pre>OUTPUT @E5100;"OUTPMIN?" ENTER @E5100;A,B</pre>

OUTPMINMAX?

Outputs the maximum and minimum values within the range specified with the ANARANG command. For details, refer to Appendix D. (Query only)

Equivalent Key Sequence	Memu2 RESDNANT (MINMAX)
Query Response	{ <i>numeric1</i> }, { <i>numeric2</i> }, { <i>numeric3</i> }, { <i>numeric4</i> } { <i>numeric1</i> } : minimum value { <i>numeric2</i> } : stimulus value { <i>numeric3</i> } : maximum value { <i>numeric4</i> } : stimulus value
Examples	OUTPUT @E5100;"OUTPMINMAX?" ENTER @E5100;A,B,C,D

OUTPMSTA?

Outputs the marker statistics within the specified range by the ANARANG command. (Query only)

Equivalent Key Sequence	Marker UTILITY MENU STATISTICS ON off
Query Response	{ <i>numeric1</i> }, { <i>numeric2</i> }, { <i>numeric3</i> } { <i>numeric1</i> } : mean { <i>numeric2</i> } : standard deviation { <i>numeric3</i> } : peak to peak
Examples	OUTPUT @E5100;"OUTPMSTA?" ENTER @E5100;A,B,C

Note

Marker UTILITY MENU STATISTICS ON displays the maker statistics within the entire display range.



OUTPMWID?

Outputs the results of the bandwidth search. (Query only)

Equivalent Key Sequence	Marker UTILITY MENU WIDTH ON OFF
Query Response	{ <i>numeric1</i> }, { <i>numeric2</i> }, { <i>numeric3</i> } { <i>numeric1</i> } : bandwidth { <i>numeric2</i> } : center { <i>numeric3</i> } : Q
Examples	OUTPUT @E5100;"WIDV -5" OUTPUT @E5100;"WIDT ON" OUTPUT @E5100;"OUTPMWID?" ENTER @E5100;A,B,C

OUTPMWIL?

Outputs the results of the bandwidth search with the bandwidth, center value, Q, and the insertion loss. (Query only)

Equivalent Key Sequence	Marker UTILITY MENU WIDTH ON OFF
Query Response	<pre>{value1}, {value2}, {value3}, {value4} {value1} : Bandwidth {value2} : Center value {value3} : Q {value4} : Insertion loss</pre>
Examples	<pre>OUTPUT @E5100;"WIDV -5" OUTPUT @E5100;"MARK1" OUTPUT @E5100;"SEAMAX" OUTPUT @E5100;"WIDT ON" OUTPUT @E5100;"OUTPMWIL?" ENTER @E5100;A,B,C,D</pre>

OUTPMWLF?

Outputs the results of the bandwidth search with the insertion loss, the difference between the center frequency and the lower cutoff frequency ($\Delta L.F$), and the difference between the center frequency and the upper cutoff frequency ($\Delta R.F$) values. (Query only)

Equivalent Key Sequence	Marker UTILITY MENU WIDTH ON OFF
Query Response	<pre>{numeric1}, {numeric2}, ... , {numeric6} {numeric1} > : bandwidth {numeric2} > : center {numeric3} > : Q {numeric4} > : insertion loss {numeric5} > : $\Delta L.F$ {numeric6} > : $\Delta R.F$</pre>
Examples	<pre>OUTPUT @E5100;"WIDV -5" OUTPUT @E5100;"WIDT ON" OUTPUT @E5100;"OUTPMWLF?" ENTER @E5100;A,B,C,D,E,F</pre>

OUTPRAW?

Outputs the raw data array of the active channel. (Query only)

Query Response	<pre>{value (1)} {value (2)} ... {value (n)} n = Number of points value: complex value(data format : Real, Imaginary)</pre>
Examples	<pre>DIM A(1:201,1:2) NOP: 201 OUTPUT @E5100;"OUTPRAW?" ENTER @E5100;A(*)</pre>

OUTPRESF? *value1,value2*

Returns the stimulus of the maximum local-maximum and its x_1 dB below points of both sides, and the stimulus of minimum local-minimum and its x_2 dB above points of both sides. For more details, refer to Appendix D. (Query only)

Parameter Description	$\langle \text{numeric1} \rangle$: down vale from the maximum peak x_1 [dB] $\langle \text{numeric2} \rangle$: down value from the minimum peak x_2 [dB]
Query Response	$\{ \text{numeric1} \}, \{ \text{numeric2} \}, \dots, \{ \text{numeric6} \}$ $\{ \text{numeric1} \}$: center frequency between f_{s1} and f_{s2} $\{ \text{numeric2} \}$: center frequency between f_{p1} and f_{p2} $\{ \text{numeric3} \}$: left point of x_1 dB down points from the maximum peak : f_{s1} $\{ \text{numeric4} \}$: right point of x_1 dB down points from the maximum peak : f_{s2} $\{ \text{numeric5} \}$: left point of x_2 dB down points from the maximum peak : f_{p1} $\{ \text{numeric6} \}$: right point of x_2 dB down points from the maximum peak : f_{p2}
Examples	DIM ANS(1:6) OUTPUT @E5100;"OUTPRESF? " ; -3,3 ENTER @E5100;ANS(*)

OUTPRESO?

Outputs the series resonant (Resonant) and parallel resonant (Anti-Resonant) parameters, 0° phase point frequency f_r (Resonant frequency) and f_a (Anti-Resonant frequency), and the corresponding gain values G_r and G_a . For details, refer to Appendix D. (Query only)

Equivalent Key Sequence	Menu2 RESONANT (0° PHASE)
Query Response	$\{ \text{numeric1} \}, \{ \text{numeric2} \}, \dots, \{ \text{numeric4} \}$ $\{ \text{numeric1} \}$: resonance impedance (G_r) $\{ \text{numeric2} \}$: resonance frequency (f_r) $\{ \text{numeric3} \}$: anti-resonance impedance (G_a) $\{ \text{numeric4} \}$: anti-resonance frequency (f_a)
Examples	DIM ANS(1:4) OUTPUT @E5100;"OUTPRESO?" ENTER @E5100;ANS(*)

OUTPRESR?

Outputs same parameter as OUTPRESO? and maximum difference, $rpl1$ of local maximum and its left local minimum on left of resonant point, maximum difference, $rpl2$ of local maximum and its right local minimum between resonant and anti-resonant points, and the maximum difference, $rpl3$ of the local maximum and its left local minimum on the right of the anti-resonant point. For details, refer to Appendix D. (Query only)

OUTPRESR?

Query Response	<pre>{numeric1}, {numeric2}, ... , {numeric7} {numeric1} : resonance impedance (G_r) {numeric2} : resonance frequency (f_r) {numeric3} : anti-resonance impedance (G_a) {numeric4} : anti-resonance frequency (f_a) {numeric5} : maximum difference of local maximum and its left local minimum on left of resonant point (rpl1) {numeric6} : maximum difference of local maximum and its right local minimum between resonant and anti-resonant points (rpl2) {numeric7} : maximum difference of the local maximum and its left local minimum on the right of the anti-resonant point (rpl3)</pre>
Examples	<pre>DIM ANS(1:7) OUTPUT @E5100;"OUTPRESR?" ENTER @E5100;ANS(*)</pre>

OUTPRFORM?

Outputs the main trace array of the active channel. (Query only)

Query Response	<pre>{value (1)} {value (2)} ... {value (n)} n = Number of points value : Real value</pre>
Examples	<pre>DIM A(1:201) NOP: 201 OUTPUT @E5100;"OUTPRFORM?" ENTER @E5100;A(*)</pre>

OUTPRFORMP? <value>

Outputs the specified points of the main trace array of the active channel. (This command equals OUTPDATTP?) (Query only)

Parameter Description	<i><value></i> : Point number to output : (1 to Number of points)
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"OUTPRFORMP? 2" ENTER @E5100;A</pre>

OUTPRTMEM?

Outputs the sub-trace array of the active channel. (Query only)

Query Response	<pre>{value (1)} {value (2)} ... {value (n)} n = Number of points value : Real value</pre>
Examples	<pre>DIM A(1:201) NOP: 201 OUTPUT @E5100;"OUTPRTMEM?" ENTER @E5100;A(*)</pre>

OUTPRTMEMP? <value>

Outputs the specified points of the sub-trace array of the active channel. (This command equals OUTPMEMTP?) The sub-trace is available when the number of channels is set to 2. (Query only)

Parameter Description	<value>: Point number : (1 to Number of points)
Query Response	{value}
Examples	OUTPUT @E5100;"OUTPRTMEMP? 2" ENTER @E5100;A

OUTPSTIM?

Outputs the stimulus array data from the active channel. (Query only)

Query Response	{numeric (1)} {numeric (2)} ... {numeric (n)} (n=number of points)
Examples	DIM A(1:201) NOP: 201 OUTPUT @E5100;"OUTPSTIM?" ENTER @E5100;A(*)

OUTPTITL?

Outputs the display title for the active channel. (Query only)

Query Response	{string}: less than 54 characters
Examples	OUTPUT @E5100;"OUTPTITL?" ENTER @E5100;A\$

OUTPTRAC? □ <value>

Outputs the phase value to trap at a measured point and the setting of I/O port at the point specified by the parameter. The phase value and the I/O port setting are specified by INPUTRAC command. (Query only, Option 022 only)

Parameter Description	{value} : measurement point number
Query Response	{value} {0 1} {value} : Phase value {1 0} : I/O port setting (ON/OFF)
Examples	OUTPUT @E5100;"OUTPTRAC? 10" ENTER @E5100;A,B

OUTPTRACB? \square \square \langle value1 \rangle , \langle value2 \rangle

Outputs the data outputted to the I/O port when the measured value of phase reaches the limit value. The data outputted to the I/O port is specified by INPUTRACB command. (Query only, Option 022 only)

Parameter Description	\langle value1 \rangle : Measurement point number \langle value2 \rangle : Output port number (1 to 6)
Query Response	{value} {value} : Data
Examples	OUTPUT @E5100;"INPUTRACB? 10,1" ENTER @E5100;A

OUTPXF2? \square \square \square \langle value1 \rangle , \langle value2 \rangle , \langle value3 \rangle , ... , \langle valuen + 3 \rangle

Outputs filter parameters within the range specified by the ANARANG command. This command outputs the same parameters as OUTPXFIL and outputs up to 20 sets of frequency offsets from center of the analysis range to left and right cutoff points. (Query only)

Where,

- If two cutoff points which are x_1 dB below the maximum peak are not found, zeros will be returned for all parameters.
- If two cutoff points which are X_n dB below the maximum value are not found, zeroes will be returned for $\Delta f_{left\ n}$ and $\Delta f_{right\ n}$.
- If no peaks are found, zeroes will be returned for Pole $_{x1}$, Pole $_{stim1}$, Pole $_{x2}$, Pole $_{stim2}$.

OUTPXFIL? *value1,value2,value3,value4,value5*

Parameter Description	<p><<i>value1</i>> : difference from the maximum value <<i>value2</i>> : stop frequency in rejection band <<i>value3</i>> : start frequency of spurious level range <<i>value4</i>> : offset to the maximum peak value to determine the cutoff points x_1 [dB] <<i>value5</i>> : offset to the maximum peak value to determine the cutoff points x_2 [dB] : <<i>valuen+3</i>> : offset to the maximum peak value to determine the cutoff points x_n [dB] ($2 \leq n \leq 20$)</p>
Query Response	<p>{<i>value1</i>}{<i>value2</i>}{<i>value2</i>}{<i>value2</i>}{<i>value2</i>}{<i>value2n+11</i>}</p> <p>{<i>value1</i>}:insertion loss {<i>value2</i>}:x_1dB bandwidth {<i>value3</i>}:center frequency {<i>value4</i>}:Q {<i>value5</i>}:passband ripple {<i>value6</i>}:rejection level {<i>value7</i>}:spurious level {<i>value8</i>}:negative peak in the left range (pole$_{x1}$) {<i>value9</i>}:stimulus value of negative peak in the left range (pole$_{stim1}$) {<i>value10</i>}:negative peak in the right range (pole$_{x2}$) {<i>value11</i>}:stimulus value of negative peak in the right range (pole$_{stim2}$) {<i>value12</i>}:frequency offset from center of the analysis range to left cutoff point (Δf_{left1}) {<i>value13</i>}:frequency offset from center of the analysis range to right cutoff point (Δf_{right1}) {<i>value14</i>}:Δf_{left2} {<i>value15</i>}:Δf_{right2} : {<i>value2n+10</i>}:$\Delta f_{left n}$ {<i>value2n+11</i>}:$\Delta f_{right n}$</p>
Examples	<pre>OUTPUT @E5100;"OUTPXF2? 3,90M,95M,3,10,12,15,20" ENTER @E5100;A(10),B(10)</pre>

OUTPXFIL? *value1,value2,value3,value4,value5*

Outputs filter parameters within the range specified by the ANARANG command. Command parameter sets the offset of x_1 dB and x_2 dB to the maximum peak value to determine the cutoff points, same parameter as POLE?, and f_1 and f_2 for determining the rejection level and the spurious level respectively. For details, refer to Appendix D. (Data format:loss, bandwidth, center frequency, Q, $\Delta L.F1$, $\Delta R.F1$, $\Delta L.F2$, $\Delta R.F2$, passband ripple, rejection level, spurious level, pole $_{x1}$, pole $_{stimulus1}$, pole $_{x2}$, pole $_{stimulus2}$) (Query only)

OUTPXFIL? *value1,value2,value3,value4,value5*

Parameter Range	<p><numeric1> : offset to the maximum peak value to determine the cutoff points x_1[dB]</p> <p><numeric2> : offset to the maximum peak value to determine the cutoff points x_2[dB]</p> <p><numeric3> : difference from the maximum value (same parameter as POLE?)</p> <p><numeric4> : stop frequency in rejection band</p> <p><numeric5> : start frequency of spurious level range</p>
Query Response	<p>{numeric1}, {numeric2}, ... , {numeric15}</p> <p>{numeric1} : insertion loss</p> <p>{numeric2} : x_1dB bandwidth</p> <p>{numeric3} : center frequency</p> <p>{numeric4} : Q</p> <p>{numeric5} : frequency offset from center of the analysis range to left cutoff point ($\Delta L.F1$)</p> <p>{numeric6} : frequency offset from center of the analysis range to right cutoff point ($\Delta R.F1$)</p> <p>{numeric7} : frequency offset from center of the analysis range to left cutoff point ($\Delta L.F2$)</p> <p>{numeric8} : frequency offset from center of the analysis range to right cutoff point ($\Delta R.F2$)</p> <p>{numeric9} : passband ripple</p> <p>{numeric10} : rejection level</p> <p>{numeric11} : spurious level</p> <p>{numeric12} : negative peak in the left range (pole_x1)</p> <p>{numeric13} : stimulus value of negative peak in the left range (pole_stimulus1)</p> <p>{numeric14} : negative peak in the right range (pole_x2)</p> <p>{numeric15} : stimulus value of negative peak in the right range (pole_stimulus2)</p>
Examples	<pre>DIM ANS(1:15) OUTPUT @E5100;"OUTPXFIL? ";-10,-20,-40,1E3,1E3 ENTER @E5100;ANS(*)</pre>

PARS {OFF|ON|0|1}

Sets the partial search of the marker search function on or off.

Equivalent Key Sequence	Marker MKR SEARCH <input type="checkbox"/> SEARCH RANGE PART SRCH ON OFF
Parameter Range	<p>OFF or 0 : partial search ON</p> <p>ON or 1 : partial search OFF</p>
Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"PARS ON" OUTPUT @E5100;"PARS?" ENTER @E5100;A</pre>

PEAK?

Outputs the maximum local maximum value and its stimulus within range which is set by the ANARANG command. For more information, refer to Appendix D. (Data format: maximum Local-maximum value, stimulus)

Query Response	{ <i>numeric1</i> } { <i>numeric2</i> } { <i>numeric1</i> } : maximum value { <i>numeric2</i> } : stimulus
Examples	OUTPUT @E5100;"PEAK?" ENTER @E5100;A,B

PEAKLIST? *value*

Outputs the *n* peak values in order of the peak value within the range specified by the ANARANG command.

Parameter Range	1 to NOP
Query Response	{ <i>value</i> }
Examples	DIM A(1:5,1:2) OUTPUT @E5100;"PEAKLIST? 5" ENTER @E5100;A(*)

PHAO *value* [deg]

Adds or subtracts a phase offset.

Equivalent Key Sequence	<input type="button" value="Display"/> PHASE OFFSET
Parameter Range	-360 to +360 [°]
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"PHAO 90" OUTPUT @E5100;"PHAO?" ENTER @E5100;A

PHAS

This command equals FMT PHAS. See FMT.

PICIRC {1|0|ON|OFF}

Selects to use π network test fixture. When PICIRC is set to ON, the analyser sets the test signal power level so that the power level applied to the DUT is equal to the power setting value, even the power unit is selected as W or A. This command uses the CI value set by CIVAL. The power unit can be selected by POWU.

Equivalent Key Sequence	(System) PI CIRCUIT ON off
Parameter Description	ON or 1: π network test fixture is used OFF or 0: π network test fixture is not used
Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"PICIRC ON" OUTPUT @E5100;"PICIRC?" ENTER @E5100;A</pre>

POIN *value*

Sets the number of the data points per sweep.

Equivalent Key Sequence	(Sweep) NUMBER OF POINTS
Parameter Range	2 to 1601
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"POIN 201" OUTPUT @E5100;"POIN?" ENTER @E5100;A</pre>

POLE? *value*

Outputs the first found local minimum for both side from the maximum point below the value which is the subtracted parameter from the maximum value. For example, to specify as -10 dB down, a command parameter becomes a -10. For more information, refer to Appendix D. (Data format: left local minimum, stimulus, right local minimum, stimulus) (Query only)

Parameter Range	TBD
Query Response	<pre>{<i>numeric 1</i>} {<i>numeric 2</i>} {<i>numeric 3</i>} {<i>numeric 4</i>}</pre> <p>{<i>numeric 1</i>} : local minimum (left) {<i>numeric 2</i>} : stimulus {<i>numeric 3</i>} : local minimum (right) {<i>numeric 4</i>} : stimulus</p>
Examples	<pre>OUTPUT @E5100;"POLE? -50" ENTER @E5100;A,B,C,D</pre>

POSL

Sets output of the 24-bit I/O port to the positive logic. (No Query)

Examples	OUTPUT @E5100;"POSL"
----------	----------------------

POWE *value* [dBm]

Sets the source output level.

Equivalent Key Sequence	Sweep POWER
Parameter Range	-9 to +11 dBm (-48 to +22 dBm option 010 only)
Query Response	{ <i>numeric</i> }
Examples	<pre>OUTPUT @E5100;"POWE 0" OUTPUT @E5100;"POWE?" ENTER @E5100;A</pre>

POWS

This command equals SWPT POWE. See SWPT.

POWU {DBM|WATT|AMP}

Selects the unit for the test signal power input.

Equivalent Key Sequence	System SRC UNIT []
Parameter Description	DBM : dBm WATT : W AMP : A
Query Response	{DBM WATT AMP}
Examples	<pre>OUTPUT @E5100;"POWU WATT" OUTPUT @E5100;"POWU?" ENTER @E5100;A\$</pre>

PRES

Presets the instrument state.

Equivalent Key Sequence	Preset
Examples	OUTPUT @E5100;"PRES"

PRINALL

Copies the measurement display to the printer according to plotting options.

Equivalent Key Sequence	System PRINT
Examples	OUTPUT @E5100;"PRINALL"

PRIR \square \langle string \rangle

Displays the query value of the HP-IB command given as a parameter. The query value is displayed on the Instrument BASIC print area.

Parameter Description	\langle string \rangle :Query command
Examples	OUTPUT @E5100;"MENU3 1, ""START"", ""FREQ"", ""PRIR 'STAR?'"""

PTABORT \square {ON|OFF}

Sets the measurement abort ON/OFF when the phase value is over the limit during the tracking. (Option 023 only)

Parameter Description	ON : Abort a measurement when the tracking is failed OFF : Continue a measurement even the tracking is failed
Query Response	{1 0}
Examples	OUTPUT @E5100;"PTABORT ON" OUTPUT @E5100;"PTABORT?" ENTER @E5100;A

PTFOVHD \square \langle value \rangle

Input the parameters required to display the time base at the temperature characteristic measurement (the aging measurement). Refer to "Compensation of Sweep Time for Aging Characteristics" in Appendix H for inputting. (Option 023 only)

Parameter Range	0 to 1 sec
Query Response	{value}

PTFR □ <value>

Input the start frequency for tracking. (Option 023 only)

Parameter Range	10 kHz to 300 MHz
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"PTFR 199 kHz" OUTPUT @E5100;"PTFR?" ENTER @E5100;A</pre>

PTFRSR □ <value>

Sets the range for searching for F_r . (Option 023 only)

Parameter Range	0 Hz to 100 kHz
Query Response	{value} (Hz)
Examples	<pre>OUTPUT @E5100;"PTFRSR 5000" OUTPUT @E5100;"PTFRSR?" ENTER @E5100;A</pre>

PTPARA □ <value>

Sets the tracking parameter. The tracking parameter is given by the SRCHFR? command query. (Option 023 only)

Parameter Range	-1000 to 1000
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"SRCHFR?" ENTER @E5100;Fr,Ci,Param OUTPUT @E5100;"PTPARA";Param OUTPUT @E5100;"PTPARA?" ENTER @E5100;A</pre>

PTRACK □ {OFF|ON}

Set the phase tracking ON/OFF. (Option 023 only)

PTRACK \square {OFF|ON}

Parameter Description	OFF : The phase tracking is OFF ON : The phase tracking is ON
Query Response	{0 1}
Examples	OUTPUT @E5100;"PTRACK ON" OUTPUT @E5100;"PTRACK?" ENTER @E5100;A

PTREPN \square <value>

Sets the number of tracking on each point. (Option 023 only)

Parameter Range	1 to 1,000,000
Query Response	{value}
Examples	OUTPUT @E5100;"PTREPN 5" OUTPUT @E5100;"PTREPN?" ENTER @E5100;A

PTSTAT?

Returns the status of the phase tracking. (Query only, Option 023 only)

Query Response	{0 1} 0 : Error encountered during phase tracking 1 : The phase tracking was successful.
Examples	OUTPUT @E5100;"PTSTAT?" ENTER @E5100;A

PTTRGLMT \square <value>

Defines the range value for a phase, which is used for the phase tracking. (Option 023 only)

Parameter Range	0° to 180°
Query Response	{value}
Examples	OUTPUT @E5100;"PTTRGLMT 8" OUTPUT @E5100;"PTTRGLMT?" ENTER @E5100;A

PTTRGPHS \square \langle value \rangle

Defines the phase at the resonant point. (Option 023 only)

Parameter Range	-180° to 180°
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"PTTRGPHS 0" OUTPUT @E5100;"PTTRGPHS?" ENTER @E5100;A</pre>

PURG *string*

Removes a file saved on the disk in the built-in flexible disk drive. (No Query)

Equivalent Key Sequence	Save/Recall FILE UTILITY PURGE FILE
Parameter Description	\langle string \rangle File name, up to 10 characters including the extension
Examples	<pre>OUTPUT @E5100;"PURG ""TEST_S"""</pre>

RAID

Completes the response and isolation calibration. (No Query)

Equivalent Key Sequence	Cal DONE:
Examples	<pre>OUTPUT @E5100;"RAID"</pre>

RAISOL

Executes the isolation measurement for the response and isolation calibration. (No Query)

Equivalent Key Sequence	Cal ISOL'N STD
Examples	<pre>OUTPUT @E5100;"RAISOL"</pre>

RAIRESP

Selects the response class for the response and isolation calibration. (No Query)

Equivalent Key Sequence	Cal RESPONSE
Examples	<pre>OUTPUT @E5100;"RAIRESP"</pre>

RECD *string*

Loads the instrument states or data from the disk in the built-in flexible disk drive. (No Query)

Equivalent Key Sequence	Save/Recall RECALL
Parameter Range	File name, Up to 10 characters including the extension
Query Response	{ <i>string</i> }
Examples	OUTPUT @E5100;"RECD ""TEST_S"""

REFP *value*

Sets the position of the reference line on the graticule of a Cartesian format.

Equivalent Key Sequence	Display SCALE MENU REFERENCE POSITION
Parameter Range	0 to 10 [Div]
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"REFP 0" OUTPUT @E5100;"REFP?" ENTER @E5100;A

REFV *value [suffix]*

Changes the value of the reference line, moving the measurement trace correspondingly.

Equivalent Key Sequence	Display Scale Menu REFERENCE VALUE
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"REFV 0" OUTPUT @E5100;"REFV?" ENTER @E5100;A

RESAVD *string*

Updates an already saved file on the disk in the built-in flexible disk drive. (No Query)

Equivalent Key Sequence	Save/Recall SAVE RE-SAVE
Parameter Range	File name up to 10 characters including the extension
Examples	OUTPUT @E5100;"RESAVD ""TEST_S"""

RESPDONE

Completes the response calibration. (No Query)

Equivalent Key Sequence	Cal DDNE:
Examples	OUTPUT @E5100;"RESPDONE"

REST

Aborts the sweep in progress, then restarts the measurement. (No Query)

Equivalent Key Sequence	Trigger MEASURE RESTART
Examples	OUTPUT @E5100;"REST"

RFOUTSW {1|2}

Changes the RF OUTPUT port. (option 003 only)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU RF OUTPUT []
Parameter Description	1 : RF OUT 1 2 : RF OUT 2
Query Response	{0 1}
Examples	OUTPUT @E5100;"RFOUTSW 1" OUTPUT @E5100;"RFOUTSW?" ENTER @E5100;A

RPLENV?

Searches all sets of neighboring peaks and their included valleys for the maximum perpendicular height from the valley minimum point included between neighboring peaks, to the intersection of an imaginary slope line drawn between the maximum peak points of the neighboring peaks in range specified by ANARANG, and outputs the resultant data via HP-IB. For details, refer to "RPLENV?" in Appendix D in Appendix D. (Query only)

Query Response	{value}
Examples	OUTPUT @E5100;"RPLENV?" ENTER @E5100;A

RPLHEI?

Searches for the maximum height between neighboring ripple peaks and outputs the resultant data via HP-IB. For details, refer to "RPLHEI?" in Appendix D in Appendix D. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"RPLHEI?" ENTER @E5100;A</pre>

RPLLHEI?

Searches for the maximum height between neighboring ripple peaks (measured from the ripple maximum peak point to the valley minimum point to the left of the ripple peak) and outputs the resultant data via HP-IB. For details, refer to "RPLLHEI?" in Appendix D in Appendix D. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"RPLLHEI?" ENTER @E5100;A</pre>

RPLMEA?

Averages all heights between neighboring local maximums and minimums within a specified range and outputs the result by HP-IB. If no ripple is detected, a zero is returned. For details, refer to Appendix D. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"RPLMEA?" ENTER @E5100;A</pre>

RPLMM?

Outputs the difference value between the maximum and minimum values within the range specified with the ANARANG command. (The maximum and minimum values are same as ones OTUPMINMAX? outputs.) (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"RPLMM?" ENTER @E5100;A</pre>

RPLPP?

Searches for the maximum ripple peak to peak value and outputs the resultant data via HP-IB. For details, refer to "RPLPP?" in Appendix D in Appendix D. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"RPLPP?" ENTER @E5100;A</pre>

RPLPPS?

Searches for the maximum ripple peak to peak value and outputs the resultant data and stimulus values at these points. (Query only)

Query Response	<pre>{numeric1} {numeric2} {numeric3} {numeric1} : maximum ripple peak {numeric2} : stimulus value at local maximum point {numeric3} : stimulus value at local minimum point</pre>
Examples	<pre>OUTPUT @E5100;"RPLPPS?" ENTER @E5100;A,B,C</pre>

RPLRHEI?

Searches for the maximum height between neighboring ripple peaks (measured from the ripple peak to the valley point to the right of the ripple peak) and outputs the resultant data via HP-IB. For details, refer to "RPLRHEI?" in Appendix D in Appendix D. (Query only)

Query Response	{value}
Examples	<pre>OUTPUT @E5100;"RPLRHEI?" ENTER @E5100;A</pre>

RPLVAL?

Outputs the maximum sum of the difference between the local minimum and the both sides local maximum, and the stimulus of the corresponding local minimum within range which is specified by ANARANG command. For more information, refer to Appendix D. (Data format: sum, stimulus) (Query only)

Query Response	<pre>{numeric1}{numeric2} {numeric1} : Maximum value of sum {numeric2} : stimulus value of local minimum</pre>
Examples	<pre>OUTPUT @E5100;"RPLVAL?" ENTER @E5100;A,B</pre>

RPOS

This command equals REFP. See REFP.

SADD

Adds a new segment to a list sweep table. (No Query)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU EDIT LIST
Examples	OUTPUT @E5100;"SADD"

SAV1

Saves the 3 term calibration results. (No Query)

Equivalent Key Sequence	CAL DONE:
Examples	OUTPUT @E5100;"SAV1"

SAVCA {OFF|ON|0|1}

Selects whether or not the calibration coefficients arrays are to be saved.

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA CAL ARRAY ON OFF
Parameter Description	{OFF ON 1 0} OFF or 0 : calibration coefficients arrays are not saved ON or 1 : calibration coefficients arrays are saved
Query Response	{ 1 0}
Examples	OUTPUT @E5100;"SAVCA ON" OUTPUT @E5100;"SAVCA?" ENTER @E5100;A

SAVDA {ON|OFF|1|0}

Sets the data array to be saved (ON) or not (OFF).

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA DATA ARRAY ON OFF
Parameter Description	{OFF ON 1 0} OFF or 0 : OFF ON or 1 : ON
Query Response	{1 0}
Examples	OUTPUT @E5100;"SAVDA ON" OUTPUT @E5100;"SAVDA?" ENTER @E5100;A

SAVDALL string

Saves the instrument states, the data array, and the memory array to the disk in the built-in flexible disk drive. (No Query)

Equivalent Key Sequence	Save/Recall SAVE ALL
Parameter Range	File name, up to 8 characters
Examples	OUTPUT @E5100;"SAVDALL ""TEST""

SAVDASC string

Save the current measurement data in ASCII file format. (No Query)

Equivalent Key Sequence	Save/Recall SAVE DATA ONLY (ASCII SAVE)
Parameter Range	File name, up to 8 characters
Examples	OUTPUT @E5100;"SAVDASC ""DATASCII""

SAVDDAT string

Saves the internal data arrays. (No Query)

Equivalent Key Sequence	Save/Recall SAVE DATA ONLY (BINARY SAVE)
Parameter Range	File name up to 8 characters
Query Response	{string}
Examples	OUTPUT @E5100;"SAVDDAT ""DATABIN""

SAVDGRAP \square *<string>*

Saves the display image to the disk in the built-in flexible disk drive or the RAM disk drive. Two graphic file formats, TIFF and PCL, are available. GRAPFORM command is used to select the format. (No Query)

Equivalent Key Sequence	Save/Recall SAVE GRAPHICS
Parameter Range	Up to 8 characters
Examples	OUTPUT @E5100;"SAVDGRAP ""RESULT1"""

SAVDMNU3 \square *<string>*

Save the user definition of softkey accessed from **Menu3**. (No Query)

Equivalent Key Sequence	Save/Recall SAVE MENU3
Parameter Range	Up to 8 characters
Examples	OUTPUT @E5100;"SAVDMNU3 ""USERMENU"""

SAVDSTA *string*

Saves only the instrument states and the calibration coefficients to the disk in the built-in flexible disk drive. (No Query)

Equivalent Key Sequence	Save/Recall SAVE STATE ONLY
Parameter Range	File name up to 8 characters
Examples	OUTPUT @E5100;"SAVDSTA ""STA1"""

SAVFA {OFF|ON|0|1}

Sets the formatted arrays to be saved (ON) or not (OFF).

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA FORMD ARRAY ON OFF
Query Response	{1 0} OFF or 0 : formatted arrays are not saved ON or 1 : formatted arrays are saved
Examples	OUTPUT @E5100;"SAVFA ON" OUTPUT @E5100;"SAVFA?" ENTER @E5100;A

SAVMA {ON|OFF|1|0}

Sets the memory array to be saved (ON) or not (OFF).

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA MEM ARRAY ON OFF
Parameter Description	{OFF ON 1 0} OFF or 0 : OFF ON or 1 : ON
Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"SAVMA ON" OUTPUT @E5100;"SAVMA?" ENTER @E5100;A</pre>

SAVRA {OFF|ON|0|1}

Sets the raw data arrays to be saved (ON) or not (OFF).

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA RAW ARRAY ON OFF
Query Response	{1 0} OFF or 0 : raw arrays are not saved ON or 1 : raw arrays are saved
Examples	<pre>OUTPUT @E5100;"SAVRA ON" OUTPUT @E5100;"SAVRA?" ENTER @E5100;A</pre>

SAVTA {OFF|ON|0|1}

Sets the trace arrays to be saved (ON) or not (OFF).

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA TRACE ARRAY ON OFF
Query Response	{1 0} OFF or 0 : trace arrays are not saved ON or 1 : trace arrays are saved
Examples	<pre>OUTPUT @E5100;"SAVTA ON" OUTPUT @E5100;"SAVTA?" ENTER @E5100;A</pre>

SAVTMA {OFF|ON|0|1}

Sets the memory trace arrays to be saved (ON) or not (OFF).

Equivalent Key Sequence	Save/Recall SAVE DEFINE SAVE DATA SUB ARRAY ON OFF
Query Response	{1 0} OFF or 0 : sub trace arrays are not saved ON or 1 : sub trace arrays are saved
Examples	OUTPUT @E5100;"SAVTMA ON" OUTPUT @E5100;"SAVTMA?" ENTER @E5100;A

SCAFDATA

Selects the data trace to be scaled. (No Query)

Equivalent Key Sequence	Display SCALE MENU SCALL FOR []
Examples	OUTPUT @E5100;"SCAFDATA"

SCAFMEMO

Selects the sub-trace to be scaled. (No Query)

Equivalent Key Sequence	Display SCALE MENU SCALL FOR []
Examples	OUTPUT @E5100;"SCAFMEM"

SCAL value

Changes the response value scale per graticule division.

Equivalent Key Sequence	Display SCALE MENU SCALE/DIV
Parameter Description	0.001 to 500 : (Log mag format) 0.01 to 500 : (Phase format) 1×10^{-14} to 10 : (Delay format) 1×10^{-11} to 10000 : (Lin mag, Real, and Imaginary formats)
Query Response	{value}
Examples	OUTPUT @E5100;"SCAL 1" OUTPUT @E5100;"SCAL?" ENTER @E5100;A

SCAY {1|0}

Selects Y-axis scale from log scale and linear scale.

Equivalent Key Sequence	Meas/Format FORMAT Y-AXIS[]
Parameter Description	0 : Linear scale 1 : LOG scale
Query Response	{0 1}
Examples	OUTPUT @E5100;"SCAY 1"

SDEL

Deletes a segment from a list sweep table. (No Query)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU EDIT LIST
Examples	OUTPUT @E5100;"SDEL"

SDON

Completes editing a segment of a list sweep table. (No Query)

Equivalent Key Sequence	Sweep SWEEP TYPE MENU EDIT LIST
Examples	OUTPUT @E5100;"SDON"

SEAL

Searches the trace for the next occurrence of the target value to the left of the marker. (No Query)

Examples	OUTPUT @E5100;"SEAL"
----------	----------------------

SEALMAX

Moves the active marker to the maximum peak point on the trace in the search range. (No Query)

Examples	OUTPUT @E5100;"SEALMAX"
----------	-------------------------

SEALMIN

Moves the active marker to the minimum peak point on the trace in the search range. (No Query)

Examples

OUTPUT @E5100;"SEALMIN"

SEAM {OFF|MAX|MIN|TARG|MEAN|LMAX|LMIN|PPEAK}

Selects the marker search function. (Query)

Equivalent Key Sequence	Marker MKR SEARCH[] SEARCH:MAX, MIN, TARGET
Parameter Description	OFF : Marker search function OFF MAX : Maximum MIN : Minimum TARG : Target MEAN : Mean LMAX : Local maximum LMIN : Local minimum PPEAK : Peak to peak
Query Response	{OFF MAX MIN TARG MEAN LMAX LMIN PPEAK}
Examples	OUTPUT @E5100;"SEAM PEAK" OUTPUT @E5100;"SEAM?" ENTER @E5100;A\$

SEAMAX

Moves the active marker to the maximum point on the trace. (No Query)

Equivalent Key Sequence	Marker MKR SEARCH[] SEARCH:MAX
Examples	OUTPUT @E5100;"SEAMAX"

SEAMEAN

Moves the active marker to the mean point on the trace. (No Query)

Examples

OUTPUT @E5100;"SEAMEAN"

SEAMIN

Moves the active marker to the minimum point on the trace. (No Query)

Equivalent Key Sequence	Marker MKR SEARCH [] MIN
Examples	OUTPUT @E5100;"SEAMIN"

SEAOFF

Turns off the marker search function. (No Query)

Equivalent Key Sequence	Marker MKR SEARCH [] TRACKING ON OFF
Examples	OUTPUT @E5100;"SEAOFF"

SEAPPEAK

Moves the active marker and the delta reference marker to the maximum peak point and the minimum peak point on the trace in the search range. (No Query)

Examples	OUTPUT @E5100;"SEAPPEAK"
----------	--------------------------

SEAR

Searches the trace for the next occurrence of the target value to the right of the marker. (No Query)

Examples	OUTPUT @E5100;"SEAR"
----------	----------------------

SEARSTOR

Stores the search range, which is defined between the active marker and the delta reference marker. (No Query)

Equivalent Key Sequence	Marker MKR SEARCH [] SEARCH RANGE SERCH RNG STORE
Examples	OUTPUT @E5100;"SEARSTR"

SEATARG *value*

Places the active marker at a specified target point on a trace.

Equivalent Key Sequence	Marker MKR SEARCH[] TARGET
Parameter Range	-5×10^5 to 5×10^5
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"SEATARG 0" OUTPUT @E5100;"SEATARG?" ENTER @E5100;A</pre>

SEDI *value*

Determines a segment of a list sweep table to be modified.

Equivalent Key Sequence	Sweep SWEEP TYPE MENU EDIT LIST
Parameter Range	1 to 31
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"SEDI" OUTPUT @E5100;"SEDI?" ENTER @E5100;A</pre>

SEET {ON|OFF}

SEET ON makes the analyzer display both of text displayed using PRINT statement of HP Instrument BASIC and measurement traces.

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"SEET ON" OUTPUT @E5100;"SEET?" ENTER @E5100;A</pre>

SETCDATE *year,month,day*

Changes date of the internal clock.

Equivalent Key Sequence	(System) SET CLOCK
Parameter Range	< <i>numeric (year)</i> > : 1901 to 2059 < <i>numeric (month)</i> > : 1 to 12 < <i>numeric (day)</i> > : 1 to 31
Query Response	{ <i>numeric(year)</i> } { <i>numeric(month)</i> } { <i>numeric(day)</i> }
Examples	OUTPUT @E5100;"SETCLOCK 1993,1,1" OUTPUT @E5100;"SETCLOCK?" ENTER @E5100;A,B,C

SETCTIME *value 1,value 2,value 3*

Changes time of the internal clock.

Equivalent Key Sequence	(System) SET CLOCK
Parameter Range	< <i>value 1 (hour)</i> > : 0 to 23 < <i>value 2 (minute)</i> > : 0 to 59 < <i>value 3 (second)</i> > : 0 to 59
Query Response	{ <i>numeric1</i> } { <i>numeric2</i> } { <i>numeric3</i> } { <i>numeric1</i> } : hour { <i>numeric2</i> } : minute { <i>numeric3</i> } : second
Examples	OUTPUT @E5100;"SETCTIME 10,30,0" OUTPUT @E5100;"SETCTIME?" ENTER @E5100;A,B,C

SETZ *value* [ohm]

Sets the characteristic impedance used by the HP E5100A/B in calculating measured impedance.

Equivalent Key Sequence	(Cal) SET Z0
Parameter Range	0.1 to 5×10^6 [Ω]
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"SETZ 75" OUTPUT @E5100;"SETZ?" ENTER @E5100;A

SING?

Makes a single measurement sweep and returns 1 when the sweep is completed. (Query Only)

Equivalent Key Sequence	Trigger SINGLE
Query Response	1
Examples	<pre>OUTPUT @E5100;"SING?" ENTER @E5100A</pre>

SINSPEAK {ON|OFF}

SINSPEAK ON makes the analyzer search the maximum or minimum point with each sweep.

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"SINSPEAK ON" OUTPUT @E5100;"SINSPEAK?" ENTER @E5100;A</pre>

SMOO {ON|OFF}

Sets the smoothing function to ON or OFF.

Equivalent Key Sequence	Display SMOOTHING on OFF
Parameter Range	OFF or 0 : smoothing OFF ON or 1 : smoothing ON
Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"SMOO ON" OUTPUT @E5100;"SMOO?" ENTER @E5100;A</pre>

SMOOAPER *value* [pct]

Changes the value of the smoothing aperture as a percent of the span.

Equivalent Key Sequence	Display SMOOTHING APERTURE
Parameter Range	0.05 to 100 [%]
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"SMOOAPER 10" OUTPUT @E5100;"SMOOAPER?" ENTER @E5100;A</pre>

SPAN value [HZ|KHZ|MAHZ|GHZ|DBM]

Sets the frequency span of a segment about a specified center frequency.

Equivalent Key Sequence	SPAN
Parameter Range	0 to 299.999 × 10 ⁶ (=299.999 M) : Hz (frequency) 0 to 70 : dBm (power)
Query Response	{value}
Examples	OUTPUT @E5100;"SPAN 100MHZ" OUTPUT @E5100;"SPAN?" ENTER @E5100;A

SPLD {ON|OFF|0|1}

Sets the multi channel display mode.

Equivalent Key Sequence	Display SPLIT DISP ON off
Parameter Description	OFF or 0 : a full-screen single graticule display ON or 1 : a split display with two half-screen graticules
Query Response	{1 0}
Examples	OUTPUT @E5100;"SPLD ON" OUTPUT @E5100;"SPLD?" ENTER @E5100;A

SRCHFR?□{1|2|3|4},<value>

Search for the resonance frequency (F_r). (Query only, Option 023 only)

Parameter Description	Search mode: 1 : Rough (High speed) 2 : Normal 3 : Fine 4 : Finer (Slow) <value> : Waiting time during searching (sec)
Query Response	{value1}, {value2}, {value3} {value1} : F _r [Hz] {value2} : CI [0] {value3} : Tracking parameter
Examples	OUTPUT @E5100;"SRCHFR? 2,0" ENTER @E5100;A,B,C

STANC

Measures the calibration standard in the THRU. (No Query)

Equivalent Key Sequence	Cal THRU
Examples	OUTPUT @E5100;"STANA"

STAR *value* [*suffix*]

Defines the start value of the stimulus.

Equivalent Key Sequence	START
Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz -9 to +11 dBm (-48 to +22 dBm option 010 only)
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"STAR 100KHZ" OUTPUT @E5100;"STAR?" ENTER @E5100;A

STAS *value1*[HZ|KHZ|MAHZ|GHZ|DBM], *value2* [HZ|KHZ|MAHZ|GHZ|DBM]

Sets start and stop stimulus values.

Equivalent Key Sequence	START STOP
Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz -9 to +11 dBm (-48 to +22 dBm option 010 only)
Examples	OUTPUT @E5100;"STAS 100KHZ,1MHZ"

STAW *value*

Sets the wait time for sweep.

Parameter Range	0 to 100 (sec.)
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"STAW 1" OUTPUT @E5100;"STAW?" ENTER @E5100;A

STIDROUT{1-16} value [suffix]

Sets stimulus of data trace up to 16 for OUTPDATAT? query. To execute STIDROUT? query, pass a number as the parameter.

Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz -9 to +11 dBm (-48 to +22 dBm option 010 only)
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"STIDROUT1 100KHZ" OUTPUT @E5100;"STIDROUT1?" ENTER @E5100;A</pre>

STIMROUT{1-16} value [suffix]

Sets stimulus of memory trace up to 16 for OUTPDATAT? query. To execute STIDROUT? query, pass a number as the parameter.

Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz -9 to +11 dBm (-48 to +22 dBm option 010 only)
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"STIMROUT1 100KHZ" OUTPUT @E5100;"STIMROUT1?" ENTER @E5100;A</pre>

STOD{DISK|MEMO}

Selects mass storage device. (No Query)

Equivalent Key Sequence	Save/Recall FILE UTILITY STOR DEV []
Parameter Range	STODDISK : internal flexible disc STODMEMO : internal RAM disk memory
Examples	<pre>OUTPUT @E5100;"STODDISK" OUTPUT @E5100;"STODMEMO"</pre>

STOMDISK

Stores the all files in the RAM disk to the FLASH disk. (No Query)

Equivalent Key Sequence	Save/Recall BACK UP MEMO DISK
Examples	<pre>OUTPUT @E5100;"STOMDISK"</pre>

STOP *value* [*suffix*]

Defines the stop value of the stimulus.

Equivalent Key Sequence	STOP
Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz -9 to +11 dBm (-48 to +22 dBm option 010 only)
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"STOP 100MHZ" OUTPUT @E5100;"STOP?" ENTER @E5100;A</pre>

STPSIZE *value* [**HZ**|**KHZ**|**MAHZ**]

Specifies the frequency step for a list sweep table.

Equivalent Key Sequence	SWEEP SWEEP TYPE MENU EDIT LIST STEP SIZE
Parameter Range	10×10^3 (= 10k) to 300×10^6 (= 300M) Hz
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"STPSIZE 1MHZ" OUTPUT @E5100;"STPSIZE?" ENTER @E5100;A</pre>

STR {**ON**|**OFF**}

Sets the trace storage ON or OFF.

Equivalent Key Sequence	Display STORAGE ON OFF
Parameter Range	ON : Display storage ON OFF : Display storage OFF
Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"STR ON"</pre>

SWED {DOWN|UP}

Sets the sweep direction.

Equivalent Key Sequence	Sweep SWEEP TYPE MENU SWEEP DIR []
Parameter Description	UP : sweep from START to STOP DOWN : sweep from STOP to START
Query Response	{DOWN UP}
Examples	OUTPUT @E5100;"SWED DOWN" OUTPUT @E5100;"SWED?" ENTER @E5100;A\$

SWET value [s]

Manually sets the sweep time.

Equivalent Key Sequence	Sweep SWEEP TIME []
Parameter Range	6.0×10^{-4} to 86400^1 [sec]
Query Response	{value}
Examples	OUTPUT @E5100;"SWET 1" OUTPUT @E5100;"SWET?" ENTER @E5100;A

¹ depends on stimulus settings**SWETAUTO**

Automatically sets the sweep time. (No Query)

Equivalent Key Sequence	Sweep SWEEP TIME [] SWEEP TIME AUTO
Examples	OUTPUT @E5100;"SWETAUTO ON"

SWPT {LINF|LOGF|LIST|POWE|RAMPF}

Selects the sweep type.

SWPT {LINF|LOGF|LIST|POWE|RAMPF}

Equivalent Key Sequence	Sweep SWEEP TYPE MENU LIN FREQ POWER LIST
Parameter Description	LINF : Linear frequency sweep LIST : Frequency list sweep POWE : Power sweep RAMPF : Ramp sweep
Query Response	{LINF LIST POWE RAMPF}
Examples	OUTPUT @E5100;"SWPT LINF" OUTPUT @E5100;"SWPT?" ENTER @E5100;A\$

TARL? *value*

Output stimulus of the first fund point which has a value specified by the parameter of this command for left direction from the right edge of analysis range which is set by the ANARANGE command. For more information, see Appendix D. (Data format: stimulus) (Query only)

Parameter Range	-5.0×10^5 to 5.0×10^5
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"TARL? -10" ENTER @E5100;A

TARR? *value*

Outputs stimulus of the first found point which has value specified by parameter of this command for right direction from left edge of analysis range which is set by ANARANG command. For more information, refer to Appendix D. (Query only)

Parameter Range	-5.0×10^5 to 5.0×10^5
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"TARR? -10" ENTER @E5100;A

THRR *value*

Specifies threshold height of peak for waveform analysis command. Waveform analysis commands ignore ripples which has less height than specified value.

Parameter Range	
Query Response	{ <i>value</i> }
Examples	OUTPUT @E5100;"THRR -50"

TIMO□{ON|OFF|0|1}

Sets the time limit for the trapping on/off. (Option 022 only)

Parameter Description	ON or 1 : ON OFF or 0 : OFF
Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"TIMO ON" OUTPUT @E5100;"TIMO?" ENTER @E5100;A</pre>

TITL *string*

Sends the string to the title area on the display.

Equivalent Key Sequence	Display TITL
Parameter Range	up to 53 characters
Query Response	{ <i>string</i> }
Examples	<pre>OUTPUT @E5100;"TITL ""COMMENT"""" OUTPUT @E5100;"TITL?" ENTER @E5100;A\$</pre>

TOPV□<value>

Sets the value at the top line of the graticule.

Equivalent Key Sequence	Display TOP VALUE
Parameter Range	-10 ⁹ to 10 ⁹
Query Response	{ <i>value</i> }
Examples	<pre>OUTPUT @E5100;"TOPV 100" OUTPUT @E5100;"TOPV?" ENTER @E5100;A</pre>

TOTIME□<value>

Sets the limit time for the trapping. (Option 022 only)

TOTIME `[] <value>`

Parameter Range	0 to
Query Response	{value} (ms)
Examples	<pre>OUTPUT @E5100;"TOTIME 2000" OUTPUT @E5100;"TOTIME?" ENTER @E5100;A</pre>

TRABGE

Sets the condition so that the measurement is advanced to the next point when a measurement value (phase value) is greater than or equal to the threshold value specified by the INPUTRAC command. (Option 022 only)

Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"TRABGE" OUTPUT @E5100;"TRABGE?" ENTER @E5100;A</pre>

TRABLE

Sets the condition so that the measurement is advanced to the next point when a measurement value (phase value) is less than or equal to the threshold value specified by the INPUTRAC command. (Option 022 only)

Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"TRABLE" OUTPUT @E5100;"TRABLE?" ENTER @E5100;A</pre>

TRACK `{OFF|ON|0|1}`

Tracks the search at the specified target value with each new sweep.

Equivalent Key Sequence	<code>Marker MKR_SEARCH [] TRACKING ON OFF</code>
Query Response	{value} {0 1}
Examples	<pre>OUTPUT @E5100;"TRACK ON" OUTPUT @E5100;"TRACK?" ENTER @E5100;A</pre>

TRAFDATA

Set the trap function on the data trace. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"TRAFDATA" OUTPUT @E5100;"TRAFDATA?" ENTER @E5100;A</pre>

TRAFMEMO

Set the trap function on the sub trace. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"TRAFMEMO" OUTPUT @E5100;"TRAFMEMO?" ENTER @E5100;A</pre>

TRAP□{OFF|ON|0|1}

Set the trap function on/off. (Option 022 only)

Parameter Description	<p>OFF or 0 : Set the trap function off. ON or 1 : Set the trap function on.</p>
Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"TRAP ON" OUTPUT @E5100;"TRAP?" ENTER @E5100;A</pre>

TRAR□<value1>,<value2>

Set the start and the end points for the partial sweep for the trap function.

If the start point and the end point are not between 1 to the value specified by **Number of Point**, HP E5100A/B option 022 will perform a sweep within the possible range. Use this function to reserve wider sweep range so that you can change the sweep range by simply changing this setup. This way, you can save more measurement time than changing the whole sweep setup. (Option 022 only)

TRAR \square \langle value1 \rangle , \langle value2 \rangle

Parameter Description	\langle value1 \rangle : The start point for the partial sweep \langle value2 \rangle : The end point for the partial sweep
Query Response	{value1}{value2}
Examples	<pre>OUTPUT @E5100;"TRAR 10, 20" OUTPUT @E5100;"TRAR?" ENTER @E5100;A,B</pre>

TRIGMEAS

Triggers DSP to start measurement and get measurement data into DSP. This command is only used for the parallel processing using with MOVADARY and ADTOTRAC command. (No Query)

TRIM {CONT|HOLD|SING}

Selects the trigger mode.

Equivalent Key Sequence	Trigger HOLD, SINGLE, CONTINUOUS
Parameter Description	CONT : continuous sweep HOLD : hold SING : single sweep
Query Response	{CONT HOLD}SING}
Examples	<pre>OUTPUT @E5100;"TRIM SING" OUTPUT @E5100;"TRIM?" ENTER @E5100;A\$</pre>

UPDD {ON|OFF}

Sets the refresh of the display on or off. When UPDD is turned OFF, the operating speed to measure or setup will be faster. It is recommended to use this command with ALL BASIC to avoid to make a misreading because the status display on the LCD may not coincide with a current status when UPDD is turned OFF. This command is not effect to the limit table.

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"UPDD ON" OUTPUT @E5100;"UPDD?" ENTER @E5100;A</pre>

UPPELIMI <value1><value2><value3> ... <valuen>

Sets the upper limit values of the limit line. The upper value can be set at each measurement point.

Parameter Description	n = Number of points <value> : Upper value of limit line
Query Response	{value1} {value2} {value3} ... {valuen}
Examples	<pre>DIM A(1:201) NOP: 201 Set upper value of limit line OUTPUT @E5100;"UPPELIMI";A(*) OUTPUT @E5100;"UPPELIMI?" ENTER @E5100;A(*)</pre>

WIDT {ON|OFF|0|1}

Sets the bandwidth search feature (ON) or not (OFF).

Equivalent Key Sequence	Marker UTILITY MENU WIDTHS ON OFF
Parameter Description	OFF or 0 : Bandwidth search OFF ON or 1 : Bandwidth search ON (display center stimulus value, bandwidth, Q, insertion loss, frequency difference between center and cut off points)
Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"WIDT ON" OUTPUT @E5100;"WIDT?" ENTER @E5100;A</pre>

WIDV value

Sets the amplitude parameter that defines the start and stop points for a bandwidth search.

Equivalent Key Sequence	Marker UTILITY MENU WIDTH VALUE
Parameter Range	-5×10^5 to 5×10^5
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"WIDV 0" OUTPUT @E5100;"WIDV?" ENTER @E5100;A</pre>

WRIT16

Set the bit width of the data outputted to the I/O port while performing the trap function to 16 bit. The port F is used. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"WRIT16" OUTPUT @E5100;"WRIT16?" ENTER @E5100;A</pre>

WRIT24

Set the bit width of the data outputted to the I/O port while performing the trap function to 24 bit. The port H is used. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"WRIT24" OUTPUT @E5100;"WRIT24?" ENTER @E5100;A</pre>

SCPI Command (PROGram sub-system command)

:PROGram:CATalog?

Returns all the defined program names. The program name is always "PROG", because the analyzer's HP Instrument BASIC only executes a single program at a time. This command can be used from an external controller only. (Query only)

Query Response	{"PROG"}
Examples	<pre>OUTPUT @E5100;":PROG:CAT?" ENTER @E5100;A\$</pre>

:PROGram[:SElected]:DEFine□<block>

Creates and downloads programs. The DEFine query uploads programs. This command can be used from an external controller only.

:PROGRAM[:SElected]:DEFine[<block>

Equivalent Key Sequence	
Parameter Description	<p><block> : block data of program</p> <p>The <block> must be arbitrary block program data containing the lines of program code. The first line of <block> must be a header, which shows the program size. There are two formats for the header as follows:</p> <p>#0 : Allows the OUTPUT statement to send program line until END is specified in the OUTPUT statement.</p> <p>* #NMM... M : Specifies the program size.</p> <p>N specifies the number of digits that define the program size</p> <p>M... M is program size in byte (N digits)</p> <p>Each line of the program must be separated by <CR> or <CR> <LF>. When the size of the <block> exceeds the amount of available memory in the instrument, the program lines are saved up to the point of memory overflow.</p> <p>In the response to the DEFine query, the selected program and its size are returned. The selected program must be in either the paused or stopped state for the program to be uploaded. The <block> is uploaded as definite length arbitrary block response data. The program size is returned in the first line as the header, then program lines are returned.</p>
Query Response	{block}
Examples	<pre>OUTPUT @E5100;":PROG:DEF #0" OUTPUT @E5100;"10 PRINT ""HELLO!"" OUTPUT @E5100;"20 END" OUTPUT @E5100;" " END DIM A\$[100000] OUTPUT @E5100;":PROG:DEF?" ENTER @E5100 USING "%,2A";HEAD\$! Gets the header. B=VAL(HEAD\$[2]) FOR I=1 TO B ENTER @E5100 USING "%,A";HEAD\$! NEXT I ENTER @E5100 USING "-K";A\$! Gets the program.</pre>

:PROGRAM[:SElected]:DELeTe[:SElected]

Deletes the program in the BASIC editor of the analyzer. This command can be used from an external controller only. (No query)

Examples

```
OUTPUT @E5100;":PROG:DEL"
```

:PROG*ram***[:SE***lected***]:STAT***e***□{RUN|PAUSE|STOP|CONT***inue***}**

:PROG*ram***[:SE***lected***]:DEL***ete***:ALL**

Deletes the program in the BASIC editor of the analyzer. This command can be used from an external controller only. (No query)

Examples

OUTPUT @E5100;":PROG:DEL:ALL"

:PROG*ram***[:SE***lected***]:EXE***Cute***□<***string***>**

Executes the program command. The program must be in either paused or stopped before the EXECute command is allowed. This command can be used from an external controller only. (No query)

Parameter Description	< <i>string</i> > :Legal program command
Examples	OUTPUT @E5100;":PROG:EXEC ""STEP""

:PROG*ram***[:SE***lected***]:NUM***ber***□<***string***>,<***numeric (1)***>[,<***numeric (2)***>[, ... [,<***numeric (n)***>]**

Sets or queries the contents of numeric program variables and arrays in the program on the BASIC editor of the analyzer. This command can be used from an external controller only.

Parameter Description	< <i>string</i> > : Name of an existing variable in the selected program (either character data or string data) < <i>numeric</i> > : Variable value
Query Response	{ <i>numeric (1)</i> } [{ <i>numeric (2)</i> } [... [{ <i>numeric (n)</i> }]
Examples	OUTPUT @E5100;":PROG:NUMB A,1"

:PROG*ram***[:SE***lected***]:STAT***e***□{RUN|PAUSE|STOP|CONT***inue***}**

Sets or queries the state of the program in the BASIC editor of the analyzer. The table below defines the affect of setting the state to the specified state from each of the possible current states. This command can be used from an external controller only.

Desired State	Current State		
	RUN	PAUSE	STOP
RUN	error (-221)	RUN	RUN
CONT	error (-221)	RUN	error (-221)
PAUSE	PAUSE	PAUSE	STOP
STOP	STOP	STOP	STOP

:PROGram[:SElected]:STATe␣{RUN|PAUSe|STOP|CONTInue}

Query Response	{ "RUN" "PAUS" "STOP" }
Examples	<pre>OUTPUT @E5100;":PROG:STAT ""STOP"" OUTPUT @E5100;":PROG:STAT?" ENTER @E5100;A\$</pre>

:PROGram[:SElected]:STRing␣<string (varname)>, <string (value1)> [, <string (value2)> [, ... [, <string (value n)>

Sets or queries the contents of string program variables and arrays in the program in the BASIC editor of the analyzer. If a string value is too long it is truncated when stored in the program's variable. This command can be used from an external controller only.

Parameter Description	<p><string(varname)> : Name of an existing variable in the selected program (either character data or string data).</p> <p><string(value)> : Variable value</p>
Query Response	{string (1)} [{string (2)} [... [{string (n)}]
Examples	<pre>OUTPUT @E5100;":PROG:STR ""A\$"" , ""TEST"" OUTPUT @E5100;":PROG:STR? ""A\$"" ENTER @E5100;B\$ OUTPUT @E5100;":PROG:STR? 'A\$' ENTER @E5100;B\$</pre>

:PROGram[:SElected]:WAIT

Causes no further commands or queries to be executed until the defined program exits from the RUN state. That is, the program is either stopped or paused. This command can be used from an external controller only.

Query Response	{1} (1 is returned when the program is either stopped or paused.)
Examples	<pre>OUTPUT @E5100;":PROG:WAIT" OUTPUT @E5100;":PROG:WAIT?" ENTER @E5100;A</pre>

Manual Changes

Introduction

This appendix contains the information required to adapt this manual to earlier versions or configurations of the analyzer than the current printing date of this manual. The information in this manual applies directly to the HP E5100A/B Network Analyzer serial number prefix listed on the title page of this manual.

Manual Changes

To adapt this manual to your HP E5100A/B, see Table A-1 and Table A-2, and make all the manual changes listed opposite your instrument's serial number and firmware version.

Instruments manufactured after the printing of this manual may be different from those documented in this manual. Later instrument versions will be documented in a manual changes supplement that will accompany the manual shipped with that instrument. If your instrument's serial number is not listed on the title page of this manual or in Table A-1, it may be documented in a *yellow MANUAL CHANGES* supplement.

In addition to change information, the supplement may contain information for correcting errors (Errata) in the manual. To keep this manual as current and accurate as possible, Hewlett-Packard recommends that you periodically request the latest *MANUAL CHANGES* supplement.

For information concerning serial number prefixes not listed on the title page or in the *MANUAL CHANGE* supplement, contact the nearest Hewlett-Packard office.

Turn on the line switch or execute the *IDN? command by HP-IB to confirm the firmware version. See the *HP-IB Command Reference* manual for information on the *IDN? command.

Table A-1. Manual Changes by Serial Number

Serial Prefix or Number	Make Manual Changes
JPIKC	Change 1

Table A-2. Manual Changes by Firmware Version

Version	Make Manual Changes
Rev. 1.xx	Change 1 and 2
Rev. 2.00	Change 2

Serial Number

Hewlett-Packard uses a two-part, ten-character serial number that is stamped on the serial number plate (see Figure A-1) attached to the rear panel. The first five digits and the letter are the serial prefix and the last five digits are the suffix.




Figure A-1. Serial Number Plate

Change 1

The firmware revision 1.xx does not support the following commands. Please delete the descriptions about these commands in this manual.

ANARANGP
CALCOPY
CIVAL
CLEMNU3
EQUCPARA
GRAPFORM
INPUSTIM
MARKTIME
MENU3
OUTPCF2
OUTPXF2
PICIRC
PRIR
POWU
RPLMM
SAVDGRAP
SAVDMNU3

Change 2

The firmware revision 1.xx and 2.00 do not support the following commands. Please delete the descriptions about these commands in this manual.

MAXPOIN
MAXPORT

Command Summary

Meas/Format

Meas/Format Menu

Key Label	HP-IB Command
FUNCTION <input type="checkbox"/>	ANAMODE
MEAS	MEAS
FORMAT	FMT
GROUP DELY APERTURE	GRODAPER
NUM of CH	NUMC
ACTIVE CH	CHAN

Function Menu

Key Label	HP-IB Command
GAIN-PHASE	ANAMODE GAINP
IMPEDANCE:Refl	ANAMODE ZREFL
Trans	ANAMODE ZTRAN

Port Select Menu

Key Label	HP-IB Command
A/R	MEAS AR
B/R	MEAS BR
C/R	MEAS CR
R/A	MEAS RA
B/A	MEAS BA
C/A	MEAS CA
R/B	MEAS RB
A/B	MEAS AB
C/B	MEAS CB
R	MEAS R
A	MEAS A
B	MEAS B
C	MEAS C

Gain-Phase Format Menu

Key Label	HP-IB Command
LOG MAG & PHASE	FMT LOGMP
LOG MAG & DELAY	FMT LOGMD
LIN MAG & PHASE	FMT LINMP
LIN MAG & DELAY	FMT LINMD
REAL & IMAGINARY	FMT RIMAG
LOG MAG	FMT LOGM
LIN MAG	FMT LINM
PHASE	FMT PHAS
DELAY	FMT DELA
REAL	FMT REAL
IMAGINARY	FMT IMAG
EXPANDED PHASE	FMT EXPP

Z Format Menu

Key Label	HP-IB Command
Z & PHASE z	FMT MAGZP
Y & PHASE y	FMT MAGYP
R-X	FMT IMPRX
G-B	FMT ADMGB
Z	FMT MAGZ
Y	FMT MAGY
PHASE z	FMT PHAZ
PHASE y	FMT PHAY
R	FMT IMPR
X	FMT IMPX
G	FMT ADMG
B	FMT ADMB
Y-AXIS <input type="checkbox"/>	SCAY
EXPANDED PHASE on OFF	EXPZP

Display

Display Menu(1/3) (2/3) (3/3)

Key Label	HP-IB Command
AUTOSCALE	AUTO
MULTI CH on OFF	MULC
SPLIT DISP on OFF	SPLD
SMOOTHING on OFF	SMOO
SMOOTHING APERTURE	SMOOAPER
ELECTRICAL DELAY	ELED
PHASE OFFSET	PHAO
TITLE	TITL
STORAGE on OFF	STR
GRATICULE on OFF	DISG

Linear Scale Menu

Key Label	HP-IB Command
SCALE/DIV	SCAL
REFERENCE POSITION	REFP
REFERENCE VALUE	REFV
MARKER -> REFERENCE	MARKREF
ACTIVE TRC <input type="checkbox"/>	ATRC

Log Scale Menu

Key Label	HP-IB Command
TOP VALUE	TOPV
BOTTOM VALUE	BOTV
ACTIVE TRC <input type="checkbox"/>	ATRC

Define Trace Menu

Key Label	HP-IB Command
TRACE: DATA	DISPDATA
MEMORY	DISPMEMO
DATA and MEMORY	DISPDATM
DATA-MEM	DISPDMM
DATA/MEM	DISPDMM
DATA—MEMORY	DATI

Basic Allocation Menu

Key Label	HP-IB Command
ALL INSTRUMENT	DISAALLI
HALF INSTR HALF BASIC	DISAHIHB
ALL BASIC	DISAALLB
BASIC STATUS	DISABASS

Cal

Gain-Phase CAL Menu

Key Label	HP-IB Command
CORRECTION on OFF	CORR
CALIBRATE: NONE	CALI NONE
RESPONSE	CALI RESP
RESPONSE & ISOL'N	CALI RAI
1-PORT 3-TERM	CALI ONE
SET ZO	SETZ

Thru CAL Menu

Key Label	HP-IB Command
THRU	STANC
DONE:	RESPDONE

Response & Isolation CAL Menu

Key Label	HP-IB Command
RESPONSE	RAIRESP
ISOL'N STD	RAIISOL
DONE:	RAID

Gain-Phase 3 Term CAL Menu

Key Label	HP-IB Command
OPEN	CALSS11A
SHORT	CLASS11B
LOAD	CLASS11C
DONE:	SAV1

Gain-Phase CAL STD value menu

Key Label	HP-IB Command
OPEN STD	CALKO{RS LS CP}
SHORT STD	CALKS{RS LS CP}
LOAD STD	CALKL{RS LS CP}

Z: Refl CAL Menu

Key Label	HP-IB Command
CORRECTION on OFF	CORR
CALIBRATE: NONE	CALI NONE
1-PORT 3-TERM	CALI ONP
SET ZO	SETZ

Z: Refl CAL Menu

Key Label	HP-IB Command
OPEN	CLASS11A
SHORT	CLASS11B
LOAD	CLASS11C
DONE:	SAV1

Z: Refl CAL STD value menu

Key Label	HP-IB Command
OPEN STD	CALKO{RS LS CP}
SHORT STD	CALKS{RS LS CP}
LOAD STD	CALKL{RS LS CP}

Z: Trans CAL Menu

Key Label	HP-IB Command
CORRECTION on OFF	CORR
CALIBRATE: NONE	CALI NONE
3 TERM	CALI ONEP
1 TERM	CALI RESP
SET ZO	SETZ

Z:Trans 3 Term CAL Menu

Key Label	HP-IB Command
OPEN	CALSS11A
SHORT	CALSS11B
LOAD	CALSS11C
DONE:	SAV1

1 Term CAL Menu

Key Label	HP-IB Command
THRU	STANC
DONE:	RESPDONE

Z: Trans CAL STD value menu

Key Label	HP-IB Command
OPEN STD	CALKO{RS LS CP}
SHORT STD	CALKS{RS LS CP}
LOAD STD	CALKL{RS LS CP}

Marker

Marker Menu

Key Label	HP-IB Command
ACTIVE MARKER	MARK{1 4}
CLEAR MARKER	CLEM{1 4}

Marker Serch Menu

Key Label	HP-IB Command
SEARCH:MAX	SEAMAX
MIN	SEAMIN
TARGET	SEATARG
TRACKING on OFF	TRACK

Serch Range Menu

Key Label	HP-IB Command
SEARCH RNG STORE	SEARSTOR
PART SRCH on OFF	PARS

Marker Utility Menu

Key Label	HP-IB Command
STATISTICS on OFF	MEASTAT
WIDTHS on OFF	WIDT
WIDTH VALUE	WIDV
MKR LIST on OFF	MARKL

 Δ Mode Menu

Key Label	HP-IB Command
Δ REF MARKER	DELR{1 4}
Δ REF= Δ FIXED MKR	DELRFIXM
Δ MODE OFF	DELO
MKR ZERO	MARKZERO

Fixed Marker Position Menu

Key Label	HP-IB Command
FIXED MKR STIMULUS	MARKFSTI
FIXED MKR VALUE	MARKFVAL

Marker Mode Menu

Key Label	HP-IB Command
MARKERS: DISCRETE	MARKDISC
CONTINUOUS	MARKCONT
MARKERS: COUPLED	MARKCOUP
UNCOUPLED	MARKUNCO
MKR TIME on OFF	MARKTIME

Sweep

Sweep Menu

Key Label	HP-IB Command
SWEEP TYPE MENU	SWPT
SWEEP TIME	SWET
NUMBER of POINTS	POIN
POWER	POWE
CW FREQ	CWFREQ
IF BW	IFBW
COUPLED CH on OFF	COUC
ACTIVE CH [CH1]	CHAN

Sweep Type Menu

Key Label	HP-IB Command
LIN FREQ	SWPT LINF
POWER SWEEP	SWPT POWE
LIST FREQ	SWPT POWE
LIST NO. <input type="checkbox"/>	LISLIS{1 2}
LIN FREQ [STEP]	SWPT RAMPF
LIN FREQ [RAMP]	SWPT LINF
SWEEP DIR <input type="checkbox"/>	SWED
LIST DISP: FREQ BASE	LISDFBASE
ORDER BASE	LISDOBASE

Sweep Time Menu

Key Label	HP-IB Command
SWEEP TIME AUTO	SWETAUTO

Trigger

Trigger Menu

Key Label	HP-IB Command
HOLD	TRIM HOLD
SINGLE	TRIM SING
CONTINUOUS	TRIM CONT
TRIG EVENT <input type="checkbox"/>	EXTT{ON OFF}
MEASURE RESTART	REST

Start

Stop

Center

Span

Key Label	HP-IB Command
Start	STAR
Stop	STOP
Center	CENT
Span	SPAN

→ Function Menu

Key Label	HP-IB Command
MKR→START	MARKSTAR
MKR→STOP	MARKSTOP
MKR→CENTER	MARKCENT
MKR→SPAN	MARKSPAN
MKR→REFRENCE	MARKREF

System

System Menu(1/3) (2/3) (3/3)

Key Label	HP-IB Command
PRINT	PRINALL

Clock Menu

Key Label	HP-IB Command
TIME HH:MM:SS	SETCTIME
DATE MM/DD/YY	SETCDATE
DATE MODE: MonDayYear	MONDYEAR
DayMonYear	DAYMYEAR

Att Setting Menu

Key Label	HP-IB Command
AUTO	ATTI{R A B C}AUTO ON
0 dB	ATTI{R A B C}AUTO OFF;ATTI{R A B C} 0
25 dB	ATTI{R A B C}AUTO OFF;ATTI{R A B C} 25

Save/Recall

Save/Recall Menu

Key Label	HP-IB Command
Recall	RECD
BACKUP MEMO DISK	STOMDISK
STOR DEV []	STOD{DISK MEMO}

Save Menu

Key Label	HP-IB Command
ALL	SAVALL
STATE ONLY	SAVDSTA
DATA ONLY(BINARY)	SAVDDAT
DATA ONLY(ASCII)	SAVDASC
RE-SAVE FILE	RESAVD

File Utility Menu

Key Label	HP-IB Command
PURGE FILE	PURG
CREATE DIRECTORY	CRED
CHANGE DIRECTORY	CHAD
COPY FILE	FILC
INITIALIZE	INID

Binary Define Save Data Menu

Key Label	HP-IB Command
RAW ARRAY on OFF	SAVCA
CAL ARRAY on OFF	SAVRA
DATA ARRAY on OFF	SAVDA
MEM ARRAY on OFF	SAVMA
FORMD ARRAY on OFF	SAVFA
MAIN ARRAY on OFF	SAVTA
SUB ARRAY on OFF	SAVTMA

ASCII Define Save Data Menu

Key Label	HP-IB Command
MAIN ARRAY on OFF	SAVTA
SUB ARRAY on OFF	SAVTMA

MISC Save Menu

Key Label	HP-IB Command
SAVE GRAPHICS	SAVDGRAP
GRAPH []	GRAPFORM
SAVE MENU3	SAVDMNU3
RE-SAVE	RESAVD

Menu2

Key Label	HP-IB Command
USER CI []	CIVAL

Key Label	HP-IB Command
SRC UNIT []	POWU
PICIRCUIT ON OFF	PICIRC

C

Data I/O Format

Data Format

The HP E5100A/B can output data over HP-IB in four different formats. The type of format affects what kind of data array is declared (real or integer), since the format determines what type of data is transferred.

Form 2

IEEE 32-bit floating point format. In this mode, each number takes 4 bytes. This means that a 201-point transfer takes 1,608 bytes. Figure C-1 shows the data transfer format of Form 2.




Figure C-1. Form 2 Data Transfer Format

Form 3

IEEE 64-bit floating point format. In this mode, each number takes 8 bytes. This means that a 201-point transfer takes 3,216 bytes. Data is stored internally in the 200/300 series computer with the IEEE 64-bit floating point format, eliminating the need for any reformatting by the computer. Figure C-2 shows the data transfer format of Form 3.




Figure C-2. Form 3 Data Transfer Format

Form 4

ASCII data transfer format. In this mode, each number is sent as a 24 character string, each character being a digit, sign, or decimal point. Use this format, when HP E5100A/B enters data to an internal array using INPUxxx command.

Form 5

MS-DOS[®] personal computer format. This mode is a modification of IEEE 32-bit floating point format with the byte order reversed. Form 5 also has a four byte header which must be read in so that data order is maintained. In this mode, an MS-DOS[®] PC can store data internally without reformatting it.

Internal Data Array

The data is stored in data arrays, denoted by double-line boxes in Figure C-3.




Figure C-3. Data Processing Flow Diagram

The following tables list HP-IB commands to output and enter data from/to the internal array.

Table C-1. Data Array Output Commands

Array Name	Array Output	One Point Output
Raw Data Array	OUTPRAW?	-
Cal. Coef. Array	OUTPCALC{01 02 03}	-
Data Array	OUTPDATA?	-
Memory Array	OUTPMEMO?	-
Formatted Array	OUTPFORM?	OUTPFORMP?
Main Array	OUTPRFORM?	OUTPRFORMP?
		(OUTPDATTP?)
Sub Array	OUTPRTMEM?	OUTPRTMEMP?
		(OUTPMEMTP?)

Table C-2. Data Array Input Commands

Array Name	Array Input
Raw Data Array	INPURAW?
Cal. Coef. Array	INPUCALC{01 02 03}
Data Array	INPUDATA?
Memory Array	INPUMEMO?
Formatted Array	INPUFORM?
Main Array	INPURFORM?
Sub Array	INPURTMEM?

D

Waveform Analysis Commands

The HP E5100A/B has added a command set that can be used to analyze waveforms of specific devices. The waveform analysis commands analyze and output the results using only a single command. This appendix provides information about the added waveform analysis commands.

The commands are divided into five groups as follows:

- Waveform analysis setup commands
- Maximum/Minimum/Mean search commands
- Ripple analysis commands
- Filter and Resonator analysis commands
- Equivalent circuit analysis commands

Conventions and Definitions

This section describes the conventions and definitions that are used to describe the waveform analysis commands.

- ① → **ANARANG**
- ② → Sets the stimulus range for the waveform. . . .
- ③ → **Syntax** **ANARANG start,stop**
- ④ → Where,

0	<i>start</i>	Start value of the analysis range
1	<i>stop</i>	Stop value of the analysis range
- ⑤ → **Query**
Response
- ⑥ → **Semantics**
- ⑦ → **Note**
- ⑧ → **Examples**

①	Command name.
②	Command description.
③	<p>Command syntax</p> <p>This part shows the syntax of the command. You must put a space between the command and the parameters.</p>
④	<p>Command parameter description</p> <p>The first column of the table lists the register number that is used by the EXECUTE command. You must put the parameter in the indicated register before using the EXECUTE command. For example (in the above case):</p> <pre style="margin-left: 40px;">WRITEIO 15,0;Start Put "Start" in register 0. WRITEIO 15,1;Stop Put "Stop" in register 1. EXECUTE "ANARANG" Execute "ANARANG".</pre> <p>The second column lists the parameter name that is shown in the Syntax area. The third column describes the parameters.</p>
⑤	<p>Query response.</p> <p>This part shows what values will be returned as the query response. The description of the query response is similar to the description of the Syntax area shown above.</p>
⑥	<p>Semantics</p> <p>This part describes how the command obtains the values for the query response.</p>
⑦	<p>Note</p> <p>This part describes the required conditions or limitations when using the command.</p>
⑧	<p>Examples</p> <p>This part shows examples of how to use the command. Examples are provided for both HP BASIC on an external controller and Instrument BASIC on the analyzer.</p>

Waveform Analysis Setup Commands

The following commands are used for setting up the conditions for waveform analysis:

- ANAOCH{1|2|3|4}
- ANARANG
- ANARANGP
- ANARFULL
- ANAODATA
- ANAOMEMO
- THRR

The settings are effective for all of the waveform analysis commands.

ANAOCH{1|2|3|4}

Selects channel for waveform analysis.

Syntax ANAOCH{1|2|3|4}

Query *boolean*

Response Where,

Register	Parameter	Description
0	<i>boolean</i>	1 or 0. Channel 1 is selected (1) or is not selected (0) for waveform analysis.

Note ■ The ANAOCH{1|2|3|4} channel setting is independent of the active channel setting.

ANARANG

Sets the stimulus range for waveform analysis commands by start and stop value.

Syntax ANARANG *start, stop*

Where,

Register	Parameter	Description
0	<i>start</i>	Start value of the analysis range.
1	<i>stop</i>	Stop value of the analysis range.

Query *start, stop*

Response

- Note**
- The waveform analysis range is independent of the marker search range.
 - You can set the range for each channel independently. Therefore, you need to set the analysis channel using ANAOCH1 or ANAOCH2 before using ANARANG.
 - The waveform analysis range will be truncated to fit the displayed stimulus range if the setting is exceeded.
 - If the displayed stimulus range is changed, the waveform analysis range is set equal to the displayed range.
 - Store the waveform analysis range setting using **SAVE ALL** or **STATE ONLY**.
 - The waveform analysis range is set to equal to the displayed stimulus range when the power is turned on.

Examples

```
INPUT "Enter Start for Analysis Range.",Start
INPUT "Enter Stop for Analysis Range.",Stop
OUTPUT @E5100;"ANARANG ";Start,Stop
```

ANARANGP

Sets the waveform analysis stimulus range by entering the point number of START and point number of STOP values.

Syntax ANARANG *value 1, value 2*

Where,

Register	Parameter	Description
0	<i>value 1</i>	Start point number of the analysis range.
1	<i>value 2</i>	Stop point number of the analysis range.

Query *start point number, stop point number*

Response

- Note**
- The power on default setting and other actions are same as that of ANARANG.

Examples

```
INPUT "Enter Start Point Number for Analysis Range.",Startp
INPUT "Enter Stop Point Number for Analysis Range.",Stopp
OUTPUT @E5100;"ANARANGP ";Startp,Stopp
```

ANARFULL

Sets the waveform analysis range equal to the displayed stimulus range. (No Query)

Syntax ANARFULL

Note ■ You can set the range for each channel independently. Therefore, you need to set the analysis channel using ANAOCH1 or ANAOCH2 before using ANARFULL.

ANAODATA

Selects the data trace for waveform analysis.

Syntax ANAODATA

Query *boolean*

Response Where,

Register	Parameter	Description
0	<i>boolean</i>	1 or 0. Data trace is selected (1) or is not selected (0) for waveform analysis.

Note ■ You can select the trace for each channel independently. Therefore, you need to set the analysis channel using ANAOCH1 or ANAOCH2 before using ANAODATA.

ANAOMEMO

Selects the data trace for waveform analysis.

Syntax ANAOMEMO

Query *boolean*

Response Where,

Register	Parameter	Description
0	<i>boolean</i>	1 or 0. Sub-trace is selected (1) or is not selected (0) for waveform analysis.

Note ■ You can select the trace for each channel independently. Therefore, you need to set the analysis channel using ANAOCH1 or ANAOCH2 before using ANAOMEMO.

THRR

Sets threshold ripple height for waveform analysis commands.

Syntax THRR *height*

Where,

Register	Parameter	Description
0	<i>height</i>	(Peak height) - (negative peak height)




Figure D-1. THRR

Query *height*

Response

- Semantics**
- Ripple height is defined as the difference between the positive peak and the negative peak.
 - Waveform analysis commands search only for ripples greater than the threshold value, any others are ignored.

Note ■ Default threshold value is 0.

Examples

```

INPUT "Enter Pos. Peak Gain [dB].",Local_max
INPUT "Enter Neg. Peak Gain [dB].",Local_min
Height=Local_max-Local_min
OUTPUT @E5100;"THRR ";Height

```

Maximum/Minimum/Mean Value Search Commands

The following commands return the maximum, minimum, and mean value of a trace within the range specified using the ANARANG command.

- OUTPMAX?
- OUTPMIN?
- OUTPMINMAX?
- OUTPMEAN?
- PEAK?
- NEXPK?
- NUMLMAX?
- NUMLMIN?
- LMAX?
- LMIN?
- TARR?
- TARL?

OUTPMAX?

Returns the maximum point value and its stimulus within the specified range. (Query only)

Syntax OUTPMAX?

Query *MAX, f_{max}*

Response Where,

Register	Parameter	Description
0	<i>MAX</i>	Maximum value
1	<i>f_{max}</i>	Stimulus at maximum point (Frequency or Power)

Examples OUTPUT @E5100;"OUTPMAX?"
 ENTER @E5100;Max_value,F_max
 PRINT Max_value,F_max

OUTPMIN?

Returns the minimum point value and its stimulus within the specified range. (Query only)

Syntax OUTPMIN?

Query *MIN, f_{min}*

Response Where,

Register	Parameter	Description
0	<i>MIN</i>	Minimum value
1	<i>f_{min}</i>	Stimulus at minimum point (Frequency or Power)

OUTPMINMAX?

Returns the maximum and minimum values and their stimulus values within the specified range. (Query only)

Syntax OUTPMINMAX?

Query *MIN, f_{min}, MAX, f_{max}*

Response Where,

Register	Parameter	Description
0	<i>MIN</i>	Minimum value
1	<i>f_{min}</i>	Stimulus at minimum point (Frequency or Power)
2	<i>MAX</i>	Maximum value
3	<i>f_{max}</i>	Stimulus at maximum point (Frequency or Power)

Examples OUTPUT @E5100;"OUTPMINMAX?"
 ENTER @E5100;Min_value,F_min,Max_value,F_max
 PRINT "MIN:",Min_value,F_min
 PRINT "MAX:",Max_value,F_max

OUTPMEAN?

Returns the mean value within the specified range. (Query only)

Syntax OUTPMEAN?

Query *mean*

Response Where,

Register	Parameter	Description
0	<i>mean</i>	Mean value.

Examples OUTPUT @E5100;"OUTPMEAN?"
 ENTER @E5100;Mean
 PRINT Mean

PEAK?

Returns maximum peak and its stimulus within the specified range. (Query only)

Syntax PEAK?

Query MAX_{peak} , $f_{maxpeak}$

Response Where,

Register	Parameter	Description
0	MAX_{peak}	Maximum peak value
1	$f_{maxpeak}$	Stimulus at maximum peak

Semantics ■ The analyzer defines the searched value and point as a reference point for the next NEXPK? command. The reference point is stored using **SAVE** **ALL** or **STATE ONLY**.

Note ■ If the search fails, the analyzer returns 0,0.

Examples

```
OUTPUT @E5100;"PEAK?"
ENTER @E5100;Peak,F_maxpeak
PRINT "Peak:",Peak,"[dB]","F_maxpeak,"[Hz]"
```

NEXPK?

Returns the maximum peak having a value less than the value that was found using last PEAK? or NEXPK? command within the specified range. It also returns the corresponding stimulus value. (Query only)

Syntax NEXPK?

Query $Peak$, f_{Peak}

Response Where,

Register	Parameter	Description
0	$Peak$	Searched peak value
1	f_{Peak}	Searched stimulus

Note ■ The analyzer defines the searched value and point as a reference point for the next NEXPK? command. The reference point is stored using **SAVE** **ALL** or **STATE ONLY**.

■ If the multiple corresponded points are found, the analyzer returns right-hand nearest peak of the reference point.

■ If the search fails, the analyzer returns 0,0.

Examples

```
OUTPUT @E5100;"NEXPK?"
ENTER @E5100;N_peak,F_npeak
PRINT N_peak,F_npeak
```

NUMLMAX?

Returns the number of positive peaks within the specified range. (Query only)

Syntax NUMLMAX?

Query *n*

Response Where,

Register	Parameter	Description
0	<i>n</i>	Number of peaks

Note ■ If the search fails, the analyzer returns 0.

Examples OUTPUT @E5100;"NUMLMAX?"
 ENTER @E5100;N
 PRINT N

NUMLMIN?

Returns the number of negative peaks within the specified range. (Query only)

Syntax NUMLMIN?

Query *n*

Response Where,

Register	Parameter	Description
0	<i>n</i>	Number of negative peaks

Note ■ If the search fails, the analyzer returns 0.

LMAX?

Returns the n th positive peak counted from the left end of the range.

Syntax *LMAX? n*

Where,

Register	Parameter	Description
0	n	Peak counted from the left end of the range.

Query *LMAX_n*

Response Where,

Register	Parameter	Description
0	<i>LMAX_n</i>	Value of n th peak

Note ■ If the search fails, the analyzer returns 3.40282346639E+38.

Examples OUTPUT @E5100;"LMAX? 5"
 ENTER @E5100;Lmax
 PRINT Lmax

LMIN?

Returns the n th negative peak counted from the left end of the range.

Syntax *LMIN? n*

Where,

Register	Parameter	Description
0	n	Negative peak counted from the left end of the range.

Query *LMIN_n*

Response Where,

Register	Parameter	Description
0	<i>LMIN_n</i>	Value of n th negative peak

Note ■ If the search fails, the analyzer returns 3.40282346639E+38.

TARR?

Searches to the right for the point having the specified parameter-value from the left end of the range, and returns its stimulus.

Syntax TARR? *target*

Where,

Register	Parameter	Description
0	<i>target</i>	Search value.

Query *f_{target}*

Response Where,

Register	Parameter	Description
0	<i>f_{target}</i>	Stimulus of the first point found.

Note ■ If the search fails, the analyzer returns 0.

Examples INPUT "Enter Target Value.",Target
 OUTPUT @E5100;"TARR? ";Target
 ENTER @E5100;F_target
 PRINT F_target

TARL?

Searches to the left for the point having the specified parameter-value from the right end of the range, and returns its stimulus.

Syntax TARL? *target*

Where,

Register	Parameter	Description
0	<i>target</i>	Search value.

Query *f_{target}*

Response Where,

Register	Parameter	Description
0	<i>f_{target}</i>	Stimulus of the first point found.

Note ■ If the search fails, the analyzer returns 0.

Ripple Analysis Commands

Ripple analysis commands analyze the ripples of the waveform and return the results.

- RPLPP?
- RPLHEI?
- RPLRHEI?
- RPLLHEI?
- RPLENV?
- RPLMEA?
- RPLMM?
- RPLVAL?
- POLE?

RPLPP?

Returns the maximum difference between the positive peak and the negative peak within the specified range. (Query only)

Syntax RPLPP?

Query MAX_{diff}

Response Where,

Register	Parameter	Description
0	MAX_{diff}	Maximum difference between positive and negative peak.




Figure D-2. RPLPP?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

ASSIGN @E5100 TO 717
OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLPP?"
ENTER @E5100;Max_diff
PRINT Max_diff;"[dB]"
END

```

RPLHEI?

Returns the maximum difference between adjacent positive and negative peaks. (Query only)

Syntax RPLHEI?

Query *value*

Response Where,

Register	Parameter	Description
0	<i>value</i>	Maximum difference between adjacent positive and negative peaks.




Figure D-3. RPLHEI?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

ASSIGN @E5100 TO 717
OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLHEI?"
ENTER @E5100;Adj_diff
PRINT Adj_diff;"[dB]"
END

```

RPLRHEI?

Returns the maximum difference between the positive peak and the right-hand adjacent negative peak. (Query only)

Syntax RPLRHEI?

Query *value*

Response Where,

Register	Parameter	Description
0	<i>value</i>	Maximum difference between the positive peak and the right-hand adjacent negative peak.




Figure D-4. RPLRHEI?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

ASSIGN @E5100 TO 717
OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLRHEI?"
ENTER @E5100;Adj_diff
PRINT Adj_diff;"[dB]"
END

```

RPLLHEI?

Returns the maximum difference between the positive peak and the left-hand adjacent negative peak. (Query only)

Syntax RPLLHEI?

Query *value*

Response Where,

Register	Parameter	Description
0	<i>value</i>	Maximum difference between the positive peak and the left-hand adjacent negative peak.




Figure D-5. RPLLHEI?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

ASSIGN @E5100 TO 717
OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOData"
OUTPUT @E5100;"RPLLHEI?"
ENTER @E5100;Adj_diff
PRINT Adj_diff;"[dB]"
END

```

RPLENV?

Returns the maximum height between the negative peak and the intersection of an imaginary slope line between the adjacent positive peaks. (Query only)

Syntax RPLENV?

Query *value*

Response Where,

Register	Parameter	Description
0	<i>value</i>	Maximum height between the negative peak and the intersection of an imaginary slope line between the adjacent positive peaks. (See Figure D-6.)




Figure D-6. RPLENV?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLENV?"
ENTER @E5100;Env_diff
PRINT Env_diff;"[dB]"
END

```

RPLMEA?

Returns the mean of the difference between the adjacent positive and negative peaks within the specified range. (Query only)

Syntax RPLMEA?

Query *value*

Response Where,

Register	Parameter	Description
0	<i>value</i>	Mean of the difference between the adjacent positive and negative peaks. (See Figure D-7)




Figure D-7. RPLMEA?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLMEA?"
ENTER @E5100;Mean_diff
PRINT Mean_diff;"[dB]"
END

```

RPLMM

Outputs the difference value between the maximum and minimum values within the range specified with the ANARANG command. (The maximum and minimum values are same as ones OTUPMINMAX? outputs.) (Query only)

Syntax RPLMM?

Query *value*

Response Where,

Register	Parameter	Description
0	<i>value</i>	difference value between the maximum and minimum values

Note ■ If the search fails, the analyzer returns 0.

Examples

```

OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLMM?"
ENTER @E5100;Max_min
PRINT Max_min;"[dB]"
END

```

RPLVAL?

Returns the maximum total of the differences between the negative peaks and the adjacent positive peaks on both sides and the stimulus of the corresponding negative peak. (Query only)

Syntax RPLVAL?

Query *Rpl_val, stimulus*

Response Where,

Register	Parameter	Description
0	<i>Rpl_val</i>	Maximum total of the differences between the negative peaks and the adjacent positive peaks on both sides. (See Figure D-8)
1	<i>stimulus</i>	Stimulus of the corresponding negative peak




Figure D-8. RPLVAL?

Note ■ If the search fails, the analyzer returns 0.

Examples

```

OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
OUTPUT @E5100;"RPLVAL?"
ENTER @E5100;Val,Stim
PRINT Val;"[dB]";Stim;"[Hz]"
END

```

POLE?

Returns the stimulus and value of the first negative peak found on each side of the maximum point that are below the specified value from the maximum peak. (Query only)

Syntax POLE? *D*

Where,

Register	Parameter	Description
0	<i>D</i>	Difference from the maximum peak.

Query $x_1, stim_1, x_2, stim_2$

Response Where,

Register	Parameter	Description
0	x_1	Left negative peak value.
1	$stim_1$	Stimulus of x_1 .
2	x_2	Right negative peak value.
3	$stim_2$	Stimulus of x_2 .




Figure D-9. POLE?

Note

- If the search fails, the analyzer returns 0.
- Give the command parameter as a negative value. For instance, to specify 50 dB down from the maximum peak as a reference level, the parameter is -50.

Examples

```
ASSIGN @E5100 TO 717
OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOATA"
OUTPUT @E5100;"POLE? -50"
ENTER @E5100;X1,S1,X2,S2
PRINT "LEFT :";X1;"[dB]";S1;"[Hz]"
PRINT "RIGHT:";X2;"[dB]";S2;"[Hz]"
END
```

Filter and Resonator Analysis Commands

The following commands are device related. They are easy to use for specific device analysis because they can output many parameters using only a single command.

- `OUTPFILT?`
- `OUTPXFIL?`
- `OUTPXF2?`
- `OUTPCFIL?`
- `OUTPCF2?`
- `OUTPRES0?`
- `OUTPRESR?`
- `OUTPRESF?`
- `OUTPCERR?`

OUTPFILT?

Analyzes the filter and returns the parameters.

Syntax OUTPFILT? *x*

Where,

Register	Parameter	Description
0	<i>x</i>	The dB value down the bandwidth filter.

Query Response *Loss, BW, fcent, Q, Δfleft, Δfright* (Total6)

Register	Parameter	Description
0	<i>Loss</i>	Insertion loss
1	<i>BW</i>	<i>x</i> dB down bandwidth
2	<i>fcent</i>	Center frequency
3	<i>Q</i>	Q (Quality factor)
4	<i>Δfleft</i>	Frequency difference between the left cutoff point and the middle of the range.
5	<i>Δfright</i>	Frequency difference between the right cutoff point and the middle of the range.




Figure D-10. OUTPFILT?

- Semantics**
- Insertion loss is the maximum value within the specified range.
 - *x* dB bandwidth is the frequency difference between both of the *x*dB down cutoff points.
 - Center frequency is the middle point of both cutoff points.
 - Q is calculated using the following equation:

$$Q = \frac{\sqrt{f_{cl} \times f_{cr}}}{BW}$$

Note

- If both of the two cutoff points are not found, the analyzer returns 0 for all values of the query response.

Examples

```

10 ASSIGN @E5100 TO 717
20 CALL Sweep(1) ! Goes to the subroutine.
30 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
40 OUTPUT @E5100;"OUTPFILT? -3"
50 ENTER @E5100;Loss,Bw,Fc,Q,Dfl,Dfr
60 PRINT "Loss: ";Loss;" [dB] BW: ";Bw;" [Hz] "
70 PRINT "fc: ";Fc;" [Hz] Q: ";Q
80 PRINT "Dfl: ";Dfl;" [Hz] Dfr: ";Dfr;" [Hz] "
90 END
100 SUB Sweep(Ch)! Sweep End Detection Subroutine
101      ! (Parameter: No. of channel)
110  ASSIGN @E5100 TO 717
120  ON INTR 7 GOTO Sweep_end
130  OUTPUT @E5100;"TRGS BUS"
140  OUTPUT @E5100;"ESNB 2; *SRE 4"
150  FOR I=1 TO Ch
160  OUTPUT @E5100;"*CLS;*OPC?"
170  ENTER @E5100;Opc
180  ENABLE INTR 7;2
190  TRIGGER @E5100
200  Waiting:GOTO Waiting
210  Sweep_end:!
220  NEXT I
230  SUBEND

```

OUTPXFIL?

Outputs filter parameters within the range specified by the ANARANG command.

Syntax OUTPXFIL? x_1, x_2, D, f_1, f_2

Where,

Register	Parameter	Description
0	x_1	The dB value down the bandwidth filter. (1) x_1 [dB]
1	x_2	The dB value down the bandwidth filter. (2) x_2 [dB]
2	D	Difference from maximum value. (Same as POLE? parameter.)
3	f_1	Stop frequency of the range for the rejection level.
4	f_2	Start frequency of the range for the spurious level.

Query

$Loss, BW, f_{cent}, Q, \Delta f_{left1}, \Delta f_{right1}, \Delta f_{left2}, \Delta f_{right2}, Pass, Reject, Spurious, Pole_{x1}, Pole_{stim1}, Pole_{x2}, Pole_{stim2}$ (15)

Response

Register	Parameter	Description
0	$Loss$	Insertion loss
1	BW	x_1 dB down bandwidth
2	f_{cent}	Center frequency
3	Q	Q
4	Δf_{left}	Frequency difference between the left cutoff point (f_{cl}) and the middle of the range.
5	Δf_{right}	Frequency difference between the right cutoff point (f_{cr}) and the middle of the range.
6	Δf_{left2}	Frequency difference between the left cutoff point (f_{cl2}) and the middle of the range.
7	Δf_{right2}	Frequency difference between the right cutoff point (f_{cr2}) and the middle of the range.
8	$Pass$	Passband ripple
9	$Reject$	Rejection level
10	$Spurious$	Spurious level
11	$Pole_{x1}$	First negative peak found to the left of the maximum point.
12	$Pole_{stim1}$	Stimulus of $Pole_{x1}$.
13	$Pole_{x2}$	First negative peaks found to the right of the maximum point.
14	$Pole_{stim2}$	Stimulus of $Pole_{x2}$.




Figure D-11. OUTPXFL?

- Semantics**
- Insertion loss, x_1 dB bandwidth, center frequency, Q , Δf_{left} , and Δf_{right} are the same as the responses of OUTPFILT?.
 - Δf_{left2} and Δf_{right2} are the frequency differences between both sides at the x_2 dB down cutoff points (f_{cl2} and f_{cr2}) and the middle of the range.
 - Passband ripple is the frequency difference of the maximum positive peak and the minimum negative peak between the x_1 dB down cutoff points (f_{cl} , f_{cr}).
 - Rejection level is the frequency difference from the insertion loss to the maximum level in the range from the left edge of analysis range to f_1 .
 - Spurious level is the frequency difference from the insertion loss to the maximum level between f_2 and the right edge of analysis range.
 - $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, $Pole_{stim2}$ are the same as the query response of POLE? with the parameter D .
- Note**
- If both of the two x_1 dB down cutoff points are not found, the analyzer returns 0 for all values of the query response.
 - If both of the two x_2 dB down cutoff points are not found, the analyzer returns 0 for Δf_{left2} and Δf_{right2} .
 - If the corresponding peak for POLE? is not found, the analyzer returns 0 for $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, and $Pole_{stim2}$.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"CENT 70MHZ; SPAN 100KHZ"
30 CALL Sweep(1) ! Goes to sub routine.
40 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
50 OUTPUT @E5100;"OUTPXFL? -3,-10,-50,69.98MHz,70.02MHz"
60 ENTER @E5100;Loss,Bw,Fc,Q,Df1,Dfr,Df12,Dfr2,Pass,Reject,
Spurious,Pole1,Fp1,Pole2,Fp2
70 PRINT "Loss: ";Loss;" [dB] BW: ";Bw;" [Hz] fc: ";Fc;" [Hz]"
80 PRINT "Q: ";Q;" Df1: ";Df1;" [Hz] Dfr: ";Dfr;" [Hz]"
90 PRINT "Df12: ";Df12;" [Hz] Dfr2: ";Dfr2;" [Hz] Pass: ";Pass;" [dB]"
100 PRINT "Reject: ";Reject;" [dB] Spurious: ";Spurious;" [dB]"

```

```
110 PRINT "Pole (left):";Pole1;"[dB] ";Fp1;"[Hz]"
120 PRINT "Pole (right):";Pole2;"[dB] ";Fp2;"[Hz]"
130 END
```

OUTPXF2?

Outputs filter parameters within the range specified by the ANARANG command. This command outputs the same parameters as OUTPXFIL and outputs up to 20 sets of frequency offsets from center of the analysis range to left and right cutoff points. (Query only)

Syntax OUTPXF2? $D, f_1, f_2, x_1, x_2, \dots, x_n$ Where,

Register	Parameter	Description
0.	D	Difference from maximum value. (Same as POLE? parameter.)
1.	f_1	Stop frequency of the range for the rejection level.
2.	f_2	Start frequency of the range for the spurious level.
3.	x_1	The dB value down the bandwidth filter. (1) x_1 [dB]
4.	x_2	The dB value down the bandwidth filter. (2) x_2 [dB]
⋮	⋮	⋮
n+2.	x_n	The dB value down the bandwidth filter. (n) x_n [dB]

Query Response $Loss, BW, f_{cent}, Q, Pass, Reject, Spurious, Pole_{x1}, Pole_{stim1}, Pole_{x2}, Pole_{stim2}, \Delta f_{left1}, \Delta f_{right1}, \Delta f_{left2}, \Delta f_{right2}, \dots, \Delta f_{left n}, \Delta f_{right n}$ (2n+11)

Register	Parameter	Description
0	$Loss$	Insertion loss
1	BW	x_1 dB down bandwidth
2	f_{cent}	Center frequency
3	Q	Q
4	$Pass$	Passband ripple
5	$Reject$	Rejection level
6	$Spurious$	Spurious level
7	$Pole_{x1}$	First negative peak found to the left of the maximum point.
8	$Pole_{stim1}$	Stimulus of $Pole_{x1}$.
9	$Pole_{x2}$	First negative peaks found to the right of the maximum point.
10	$Pole_{stim2}$	Stimulus of $Pole_{x2}$.
11	Δf_{left}	Frequency difference between the left cutoff point (f_{cl}) and the middle of the range.
12	Δf_{right}	Frequency difference between the right cutoff point (f_{cr}) and the middle of the range.
13	Δf_{left2}	Frequency difference between the left cutoff point (f_{cl2}) and the middle of the range.
14	Δf_{right2}	Frequency difference between the right cutoff point (f_{cr2}) and the middle of the range.
⋮	⋮	⋮
2n+9	$\Delta f_{left n}$	Frequency difference between the left cutoff point ($f_{cl n}$) and the middle of the range.
2n+10	$\Delta f_{right n}$	Frequency difference between the right cutoff point ($f_{cr n}$) and the middle of the range.

- Semantics**
- Insertion loss, x_1 dB bandwidth, center frequency, Q , Δf_{left} , and Δf_{right} are the same as the responses of OUTPFILT?
 - $\Delta f_{left\ n}$ and $\Delta f_{right\ n}$ are the frequency differences between both sides at the x_n dB down cutoff points ($f_{cl\ n}$ and $f_{cr\ n}$) and the middle of the range.
 - Passband ripple is the frequency difference of the maximum positive peak and the minimum negative peak between the x_1 dB down cutoff points (f_{cl} , f_{cr}).
 - Rejection level is the frequency difference from the insertion loss to the maximum level in the range from the left edge of analysis range to f_1 .
 - Spurious level is the frequency difference from the insertion loss to the maximum level between f_2 and the right edge of analysis range.
 - $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, $Pole_{stim2}$ are the same as the query response of POLE? with the parameter D .

- Note**
- If both of the two x_1 dB down cutoff points are not found, the analyzer returns 0 for all values of the query response.
 - If both of the two x_n dB down cutoff points are not found, the analyzer returns 0 for $\Delta f_{left\ n}$ and $\Delta f_{right\ n}$.
 - If the corresponding peak for POLE? is not found, the analyzer returns 0 for $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, and $Pole_{stim2}$.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"CENS 70MHz,100kHz"
30 OUTPUT @E5100;"SING?"
40 ENTER @E5100;Tmp
50 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOADATA"
60 OUTPUT @E5100;"OUTPXF2? -3,69.98MHz,70.02MHz,-3,-10,-12,-50"
70 ENTER @E5100;Loss,Bw,Fc,Q,Pass,Reject,Suprious,Pole1,Fp1,Pole2,Fp2,
   Dfl,Dfr,Dfl2,Dfr2,Dfl3,Dfr3,Dfl4,Dfr4
80 PRINT "Loss: ";Loss;"[dB] BW: ";Bw;"[Hz] fc: ";Fc;"[Hz]"
90 PRINT "Q: ";Q;" Dfl: ";Dfl;"[Hz] Dfr: ";Dfr;"[Hz]"
100 PRINT "Dfl2: ";Dfl2;"[Hz] Dfr2: ";Dfr2;"[Hz] Pass: ";Pass;"[dB]"
110 PRINT "Dfl3: ";Dfl3;"[Hz] Dfr3: ";Dfr3;"[Hz] Pass: ";Pass;"[dB]"
120 PRINT "Dfl4: ";Dfl4;"[Hz] Dfr4: ";Dfr4;"[Hz] Pass: ";Pass;"[dB]"
130 PRINT "Reject: ";Reject;"[dB] Suprious: ";Suprious;"[dB]"
140 PRINT "Pole (left): ";Pole1;"[dB] ";Fp1;"[Hz]"
150 PRINT "Pole (right): ";Pole2;"[dB] ";Fp2;"[Hz]"
160 END

```

OUTPCFIL?

Analyzes the filter at the nominal frequency, and returns the parameters.

Syntax OUTPCFIL? $f_c, x_1, x_2, D, f_1, f_2$

Where,

Register	Parameter	Description
0	f_c	Nominal frequency
1	x_1	The dB value down the bandwidth filter. (1) x_1 [dB]
2	x_2	The dB value down the bandwidth filter. (2) x_2 [dB]
3	D	Difference from maximum value. (Same as POLE? parameter.)
4	f_1	Stop frequency of the range for the rejection level.
5	f_2	Start frequency of the range for the spurious level.

Query Response $Loss, Loss_c, BW, f_{cent}, Q, \Delta f_{left1}, \Delta f_{right1}, \Delta f_{left2}, \Delta f_{right2}, Pass, Reject, Spurious, Pole_{x1}, Pole_{stim1}, Pole_{x2}, Pole_{stim2}$ (Total 16)

Register	Parameter	Description
0	$Loss$	Insertion loss
1	$Loss_c$	Const Loss
2	BW	x_1 dB down bandwidth
3	f_{cent}	Center frequency
4	Q	Q
5	Δf_{left}	Frequency difference between the left cutoff point (f_{cl}) and the middle of the range.
6	Δf_{right}	Frequency difference between the right cutoff point (f_{cr}) and the middle of the range.
7	Δf_{left2}	Frequency difference between the left cutoff point (f_{cl2}) and the middle of the range.
8	Δf_{right2}	Frequency difference between the right cutoff point (f_{cr2}) and the middle of the range.
9	$Pass$	Passband ripple
10	$Reject$	Rejection level
11	$Spurious$	Spurious level
12	$Pole_{x1}$	First negative peaks found to the left of the maximum point.
13	$Pole_{stim1}$	Stimulus of $Pole_{x1}$.
14	$Pole_{x2}$	First negative peak found to the right of the maximum point.
15	$Pole_{stim2}$	Stimulus of $Pole_{x2}$.




Figure D-12. OUTPCFIL?

- Semantics**
- Insertion loss, rejection level, spurious level, $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, and $Pole_{stim2}$ are the same as the responses of OUTPXFIL?.
 - The const loss is the value of the point that is specified by command parameter, f_c .
 - x_1 dB bandwidth is the frequency difference between two x_1 dB down cutoff points (f_{cl} , f_{cr}) from the const loss point.
 - Center frequency is the middle point of f_{cl} and f_{cr} .
 - Q is calculated using the following equation:

$$Q = \frac{\sqrt{f_{cl} \times f_{cr}}}{BW}$$

- Δf_{left} and Δf_{right} are the frequency differences between both sides at the x_1 dB down cutoff points (f_{cl} and f_{cr}) and f_c .
- Δf_{left2} and Δf_{right2} are the frequency differences between both sides at the x_2 dB down cutoff points (f_{cl2} and f_{cr2}) and f_c .
- Passband ripple is the frequency difference of maximum positive peak and minimum negative peak between x_1 dB down cutoff points (f_{cl1} , f_{cr1}).

- Note**
- If both of the two x_1 dB down cutoff points are not found, the analyzer returns 0 for all values of the query response.
 - If both of the two x_2 dB down cutoff points are not found, the analyzer returns 0 for Δf_{left2} and Δf_{right2} .
 - If the corresponding peak for POLE? is not found, the analyzer returns 0 for $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, and $Pole_{stim2}$.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"CENT 70MHZ; SPAN 100KHZ"
30 CALL Sweep(1) ! Goes to sub routine. (See OUTPXFIL?)
40 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAO DATA"
50 OUTPUT @E5100;"OUTPCFIL? 70MHz,-3,-10,-50,69.98MHz,70.02MHz"

```

```
60 ENTER @E5100;Loss,Lc,Bw,Fc,Q,Df1,Dfr,Df12,Dfr2,Pass,Reject,  
Spurious,Pole1,Fp1,Pole2,Fp2  
70 PRINT "Loss:";Loss;"[dB] Const Loss:";Lc;"[dB]"  
80 PRINT "BW:";Bw;"[Hz] Fc:";Fc;"[Hz]"  
90 PRINT "Q:";Q;" DF1:";Df1;"[Hz] DFR:";Dfr;"[Hz]"  
100 PRINT "Df12:";Df12;"[Hz] Dfr2:";Dfr2;"[Hz] Pass:";Pass;"[dB]"  
110 PRINT "Reject:";Reject;"[dB] Spurious:";Spurious;"[dB]"  
120 PRINT "Pole (left):";Pole1;"[dB] ";Fp1;"[Hz]"  
130 PRINT "Pole (right):";Pole2;"[dB] ";Fp2;"[Hz]"  
140 END
```

OUTPCF2?

Analyzes the filter at the nominal frequency, and returns the parameters. This command outputs the same parameters as OUTPCFIL and outputs up to 20 sets of frequency offsets from center frequency (f_c) to left and right cutoff points ($\Delta f_{\text{left } n}$, $\Delta f_{\text{right } n}$).

Syntax OUTPCFIL? f_c , D , f_1 , f_2 , x_1 , x_2 , ..., x_n

Where,

Register	Parameter	Description
0	f_c	Nominal frequency
1	D	Difference from maximum value. (Same as POLE? parameter.)
2	f_1	Stop frequency of the range for the rejection level.
3	f_2	Start frequency of the range for the spurious level.
4	x_1	The dB value down the bandwidth filter. (1) x_1 [dB]
5	x_2	The dB value down the bandwidth filter. (2) x_2 [dB]
⋮	⋮	⋮ level.
n+3	x_n	The dB value down the bandwidth filter. (n) x_n

Query $Loss$, $Loss_c$, BW , f_{cent} , Q , $Pass$, $Reject$, $Spurious$, $Pole_{x1}$, $Pole_{stim1}$, $Pole_{x2}$, $Pole_{stim2}$,
Response $\Delta f_{\text{left}1}$, $\Delta f_{\text{right}1}$, $\Delta f_{\text{left}2}$, $\Delta f_{\text{right}2}$, ..., $\Delta f_{\text{left } n}$, $\Delta f_{\text{right } n}$ (2n+11)

Register	Parameter	Description
0	<i>Loss</i>	Insertion loss
1	<i>Loss_c</i>	Const Loss
2	<i>BW</i>	x_1 dB down bandwidth
3	<i>f_{cent}</i>	Center frequency
4	<i>Q</i>	Q
5	<i>Pass</i>	Passband ripple
6	<i>Reject</i>	Rejection level
7	<i>Spurious</i>	Spurious level
8	<i>Pole_{x1}</i>	First negative peaks found to the left of the maximum point.
9	<i>Pole_{stim1}</i>	Stimulus of <i>Pole_{x1}</i> .
10	<i>Pole_{x2}</i>	First negative peak found to the right of the maximum point.
11	<i>Pole_{stim2}</i>	Stimulus of <i>Pole_{x2}</i> .
12	Δf_{left}	Frequency difference between the left cutoff point (f_{cl}) and the middle of the range.
13	Δf_{right}	Frequency difference between the right cutoff point (f_{cr}) and the middle of the range.
14	Δf_{left2}	Frequency difference between the left cutoff point (f_{cl2}) and the middle of the range.
15	Δf_{right2}	Frequency difference between the right cutoff point (f_{cr2}) and the middle of the range.
⋮	⋮	⋮
2n+10.	$\Delta f_{left\ n}$	Frequency difference between the left cutoff point ($f_{cl\ n}$) and the middle of the range.
2n+11.	$\Delta f_{right\ n}$	Frequency difference between the right cutoff point ($f_{cr\ n}$) and the middle of the range.

Semantics ■ Insertion loss, constant loss, x_1 dB bandwidth, Center frequency, Q, passband ripple, rejection level, spurious level, *Pole_{x1}*, *Pole_{stim1}*, *Pole_{x2}* are the same as the responses of OUTPCFILT?.

■ $\Delta f_{left\ n}$ and $\Delta f_{right\ n}$ are the frequency differences between both sides at the x_n dB down cutoff points ($f_{cl\ n}$ and $f_{cr\ n}$) and f_c .

Note ■ If both of the two x_1 dB down cutoff points are not found, the analyzer returns 0 for all values of the query response.

■ If both of the two x_n dB down cutoff points are not found, the analyzer returns 0 for $\Delta f_{left\ n}$ and $\Delta f_{right\ n}$.

■ If the corresponding peak for POLE? is not found, the analyzer returns 0 for *Pole_{x1}*, *Pole_{stim1}*, *Pole_{x2}*, and *Pole_{stim2}*.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"CENS 70MHZ,100kHz"
30 OUTPUT @E5100;"SING?"
40 ENTER @E5100;Tmp
50 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOATA"
60 OUTPUT @E5100;"OUTPCF2? 70MHZ,-50,69.98MHZ,70.02MHZ,-3,-10,-20"

```

```
70 ENTER @E5100;Loss,Lc,Bw,Fc,Q,Pass,Reject,Suprious,Pole1,Fp1,Pole2,Fp2,  
    Df1,Dfr,Df12,Dfr2,Df13,Dfr3  
80 PRINT "Loss: ";Loss;" [dB] Const Loss: ";Lc;" [dB]"  
90 PRINT "BW: ";Bw;" [Hz] Fc: ";Fc;" [Hz]"  
100 PRINT "Q: ";Q;" DF1: ";Df1;" [Hz] DFr: ";Dfr;" [Hz]"  
110 PRINT "Df12: ";Df12;" [Hz] Dfr2: ";Dfr2;" [Hz]"  
110 PRINT "Df13: ";Df13;" [Hz] Dfr3: ";Dfr3;" [Hz] Pass: ";Pass;" [dB]"  
120 PRINT "Reject: ";Reject;" [dB] Suprious: ";Suprious;" [dB]"  
130 PRINT "Pole (left): ";Pole1;" [dB] ";Fp1;" [Hz]"  
140 PRINT "Pole (right): ";Pole2;" [dB] ";Fp2;" [Hz]"  
150 END
```

OUTPRESO?

Returns resonator specific parameters. (Query only)

Syntax OUTPRESO?

Query Z_r, f_r, Z_a, f_a (Total 4)

Response Where,

Register	Parameter	Description
0	Z_r	Resonant impedance
1	f_r	Resonant frequency
2	Z_a	Anti-resonant impedance
3	f_a	Anti-resonant frequency




Figure D-13. OUTPRESO?

- Semantics** ■ OUTPRESO? executes the following actions and returns their values:
1. Searches for the 0° phase point from the left edge of the analysis range.
 2. Defines the first point found as the resonant point, and then returns its impedance and its frequency.
 3. Defines the next point found as the anti-resonant point, and then returns its impedance and its frequency.

- Note** ■ You must select the following conditions to use this command:
- Dual Channel & Coupled Channel: ON
 - Impedance Conversion: ON
 - Analysis channel: LOG MAG format
 - Non-analysis channel: Phase format
- OUTPRESO? returns the first two found 0° phase point events if there are more than three corresponding points.
- If there is only one 0° phase point in the range, OUTPRESO? defines that point as a resonant point and returns 0 for Z_a and f_a .

- If there is no 0° point, OUTPRES0? returns 0 for all parameters.
- If the impedance conversion is off, OUTPRES0? returns the magnitude (dB) at the 0° phase point.

Examples

```
10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"DUAC ON; COUC ON"
30 OUTPUT @E5100;"CHAN2; FMT PHAS"
40 OUTPUT @E5100;"CHAN1; FMT LOGM; CONV ZTRA"
50 OUTPUT @E5100;"CENT 70MHZ; SPAN 100KHZ"
60 CALL Sweep(2) ! Goes to sub routine. (See OUTPFILT?)
61             ! Parameter is 2 because of Dual Channel ON.
70 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
80 OUTPUT @E5100;"OUTPRES0?"
90 ENTER @E5100;Zr,Fr,Za,Fa
100 PRINT "Resonant: ";Zr;"[ohm] , ";Fr;"[Hz]"
110 PRINT "Anti-Resonant: ";Za;"[ohm] , ";Fa;"[Hz]"
120 END
```

OUTPRESR?

Returns the resonator specific parameters. (Query only)

Syntax **OUTPRESR?**

Query $Z_r, f_r, Z_a, f_a, Rpl_1, Rpl_2, Rpl_3$ (Total 7)

Response Where,

Register	Parameter	Description
0	Z_r	Resonant impedance
1	f_r	Resonant frequency
2	Z_a	Anti-resonant impedance
3	f_a	Anti-resonant frequency
4	Rpl_1	Maximum left height of the ripple where is on the left side of the resonant point.
5	Rpl_2	Maximum height right of the ripple that is between the resonant and anti-resonant points.
6	Rpl_3	Maximum height left of the ripple that is on the right side of the resonant point.




Figure D-14. OUTPRESR?

Semantics ■ **OUTPRESR?** executes the following actions:

1. Searches for the 0° phase point from the left edge of the analysis range.
2. Defines the first point found as the resonant point, and then returns its impedance and its frequency.
3. Defines the next point found as the anti-resonant point, and then returns its impedance and its frequency.
4. Returns the maximum height of the ripple, Rpl_1 , that is the difference between the peak and left adjacent negative peak.
5. Returns the maximum height of the ripple, Rpl_2 , that is the difference between the peak and right adjacent negative peak.

6. Returns the maximum height of the ripple, Rpl_3 , that is the difference between the peak and left adjacent negative peak.

Note

- You must select the following conditions to use this command:
 - Dual Channel & Coupled Channel: ON
 - Impedance Conversion: ON
 - Analysis channel: LOG MAG format
 - Non-analysis channel: Phase format
- OUTPRESR? returns the first two 0° phase point events found if there are more than three corresponding points points.
- If there is only one 0° phase point in the range, OUTPRESR? defines that point as a resonant point and returns 0 for Z_a , f_a , Rpl_1 , Rpl_2 , and Rpl_3 .
- If there is no 0° point, OUTPRESR? returns 0 for all parameters.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"DUAC ON; CGUC ON"
30 OUTPUT @E5100;"CHAN2; FMT PHAS"
40 OUTPUT @E5100;"CHAN1; FMT LOGM; CONV ZTRA"
50 OUTPUT @E5100;"CENT 70MHZ; SPAN 100KHZ"
60 CALL Sweep(2) ! Goes to sub routine. (See OUTPFILT?)
70 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
80 OUTPUT @E5100;"OUTPRESR?"
90 ENTER @E5100;Zr,Fr,Za,Fa,R1,R2,R3
100 PRINT "Resonant: ";Zr;"[ohm], ";Fr;"[Hz]"
110 PRINT "Anti-Resonant: ";Za;"[ohm], ";Fa;"[Hz]"
120 PRINT "Ripple L: ";R1;"[dB]"
130 PRINT "Ripple M: ";R2;"[dB]"
140 PRINT "Ripple R: ";R3;"[dB]"
150 END

```

OUTPRESF?

Returns the resonator specific parameters. (Query only)

Syntax OUTPRESF? x_1 , x_2

Where,

Register	Parameter	Description
0	x_1	Value down from the maximum peak.
1	x_2	Value above the maximum peak.

Query Response f_s , f_p , f_{s1} , f_{s2} , f_{p1} , f_{p2} (Total 6)

Where,

Register	Parameter	Description
0	f_s	Middle point frequency between f_{s1} and f_{s2} .
1	f_p	Middle point frequency between f_{p1} and f_{p2} .
2	f_{s1}	Left one of the two points x_1 dB down from the maximum peak.
3	f_{s2}	Right one of the two points x_1 dB down from the maximum peak.
4	f_{p1}	Left one of the two points x_2 dB above the minimum negative peak.
5	f_{p2}	Right one of the two points x_2 dB above the minimum negative peak.




Figure D-15. OUTPRESF?

Semantics ■ OUTPRESF? executes the following actions:

1. Searches for the maximum peak in the analysis range.
2. Searches for the x_1 dB below points on both sides, and defines the first found left and right side points as f_{s1} and f_{s2} , respectively.

3. Defines the middle point between f_{s1} and f_{s2} to f_s .
4. Searches for the x_2 dB above points on both sides, and defines the first found left and right side points as f_{p1} and f_{p2} , respectively.
5. Defines the middle point between f_{p1} and f_{p2} as f_p .

Note

- If there is no corresponding peak in the range, OUTPRESF? returns 0 for all parameters.
- If the maximum peak cannot be found, OUTPRESF? returns 0 for f_s , f_{s1} , and f_{s2} .
- If the minimum negative peak cannot be found, OUTPRESF? returns 0 for f_p , f_{p1} , and f_{p2} .
- Specify the negative value for x_1 and positive value for x_2 .

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"FMT LOGM; CENT 60.06MHz; SPAN 20kHz"
30 CALL Sweep(1) ! Goes to sub routine. (See OUTPFILT?)
40 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOData"
50 OUTPUT @E5100;"OUTPRESF? -3dB,3dB"
60 ENTER @E5100;Fs,Fp,Fs1,Fs2,Fp1,Fp2
70 PRINT "Series-Resonant: ";Fs;" [Hz] "
80 PRINT "Parallel-Resonant: ";Fp;" [Hz] "
90 END

```

OUTPCERR?

Returns the ceramic resonator specific parameters. (Query only)

Syntax OUTPCERR?

Query $Z_r, f_r, Z_a, f_a, Rpl_1, Rpl_2, Rpl_3$ (Total7)

Response Where,

Register	Parameter	Description
0	Z_r	Resonant impedance
1	f_r	Resonant frequency
2	Z_a	Anti-resonant impedance
3	f_a	Anti-resonant frequency
4	Rpl_1	Maximum height of the ripple that is on the left side of the resonant point.
5	Rpl_2	Maximum height of the ripple that is between the resonant and anti-resonant points.
6	Rpl_3	Maximum height of the ripple that is on the right side of the resonant point.




Figure D-16. OUTPCERR?

- Semantics**
- You need to select the LOG MAG format (FMT LOGM) and turn impedance conversion on (CONV ZTRA) to use this command.
 - OUTPCERR? executes the following actions:
 1. Searches for the minimum negative peak in the range and defines it as a resonant point. Then returns the resonant impedance, Z_r , and resonant frequency, f_r .
 2. Searches for the maximum peak in the range and defines it as a anti-resonant point. Then returns the anti-resonant impedance, Z_a , and anti-resonant frequency, f_p .

3. Returns the maximum height of the ripple, Rpl_1 , that is the difference between the peak and left adjacent negative peak.
4. Returns the maximum height of the ripple, Rpl_2 , that is the difference between the peak and right adjacent negative peak.
5. Returns the maximum height of the ripple, Rpl_3 , that is the difference between the peak and left adjacent negative peak.

Note

- This command can be used when the LOG MAG format (FMT LOGM) is selected. If another format is selected, OUTPCERR? returns 0 for all parameters.
- If no corresponding ripple is found, OUTPCERR? returns 0.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"FMT LOGM; CONV ZTRA; CENT 60.02MHz; SPAN 20kHz"
30 CALL Sweep(1) ! Goes to sub routine. (See OUTPFILT?)
40 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOADATA"
50 OUTPUT @E5100;"OUTPCERR?"
60 ENTER @E5100;Zr,Fr,Za,Fa,R1,R2,R3
70 PRINT "Resonant: ";Zr;" [ohm] , ";Fr;" [Hz] "
80 PRINT "Anti-Resonant: ";Za;" [ohm] , ";Fa;" [Hz] "
90 PRINT "Ripple L: ";R1;" [dB] "
100 PRINT "Ripple M: ";R2;" [dB] "
110 PRINT "Ripple R: ";R3;" [dB] "
120 END

```

Equivalent circuit analysis commands

The following commands are for the equivalent circuit analysis. They are easy to use for specific device analysis because they can output many parameters using only a single command.

- EQUCPARA?
 - EQUM
- EQUCPARA5?
- EQUCPARS?
- EQUCO?
- EQUCPARS4?

EQUCPARA?

Returns the six-device equivalent circuit parameters of the crystal resonator. (Query only)

Syntax EQUCPARA?

Query $C_0, C_1, L_1, R_1, G_0, R_0$ (Total 6)

Response Where,

Register	Parameter	Description
0	C_0	Parallel capacitance
1	C_1	Motional capacitance
2	L_1	Motional inductance
3	R_1	Motional resistance
4	G_0	Electrode conductance
5	R_0	Electrode resistance




Figure D-17. Six-Device Equivalent Circuit of Crystal Resonator

Semantics ■ EQUCPARA? executes the following actions:

1. Obtains the admittance characteristic circle diagram.
2. Obtains the maximum conductance. (G_{max})
3. Obtains frequencies f_1 and f_2 ($f_1 < f_2$) of the two points where the conductance is half the maximum conductance (G_{max}).
4. Calculates f_s by $f_s = \sqrt{f_1 \times f_2}$.
5. Obtains susceptance B_{fs} at f_s .
6. Calculates ω_s by $\omega_s = 2 \times \pi \times f_s$.
7. Assumes that the frequency at which the phase becomes 0° near the parallel resonance frequency is f_a , and obtains its conductance G_a .
8. Calculates ω_a by $\omega_a = 2 \times \pi \times f_a$.

9. Assumes that the frequency at which the phase becomes 0° * near the series resonance frequency is f_r .
10. Calculates the constants using the above values and the following equations:

$$Q_s = \frac{f_s}{f_2 - f_1} \quad C_o' = \frac{B_1 + B_2}{2\omega_s}$$

$$L_1 = \frac{Q_s}{\omega_s G_{max}} \quad R_1 = \frac{C_o'}{C_o G_{max}}$$

$$C_1 = \frac{G_{max}}{\omega_s Q_s} \quad R_o = \frac{1}{G_{max}} - R_1$$

$$C_o = \frac{B_{fs}}{\omega_s} \quad G_o = G_a - \frac{R_1 \omega_a^2 C_o^2}{1 + R_o R_1 \omega_a^2 C_o^2}$$

* "EQUCPARA?" interpolates the 0° phase points even if it does not exist in measured data.

- If the number of points between the maximum peak point (f_{Bmax}) and the minimum peak point (f_{Bmin}) of the conductance is less than 10 points, EQUCPARA? approximates an admittance circle. The circle approximation can be performed if there are 3 points for analysis. You can specify how many points are used for circle approximation using the EQU command to reduce the analysis time.
- If EQUCPARA? fails the circle approximation, 0 will be return for all parameters.
- If there are only 2 points for analysis, EQUCPARA? returns four-device equivalent circuit parameters. In this case, EQUCPARA? returns 0 for G_o and R_o .
- If there is only 1 point for analysis, EQUCPARA? returns 0 for all parameters.

EQU value

Specifies how many points are used for an approximation of a circle with the EQUCPARA? and EQUCPARS? commands. EQUCPARA? (or EQUCPARS?) thins the measured points out for the specified points, then makes a circle approximation. When the EQU parameter is set greater than the number of points, EQUCPARA? uses all the points for the circle approximation. Default value is 8.
value 2 to 801

Note

- You must select the following conditions or Polar format to use this command:
 - Dual Channel & Coupled Channel: ON
 - Impedance Conversion: ON
 - Analysis channel: LOG MAG format
 - Non-analysis channel: Phase format

If another format is selected, 0 will be returned for query response.

Examples

```
10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"FMT POLA; CONV YTRA; CENT 60.06MHz; SPAN 20kHz"
30 CALL Sweep(1) ! Goes to sub routine. (See OUTPFILT?)
40 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
50 OUTPUT @E5100;"EQUCPARA?"
60 ENTER @E5100;C0,C1,L1,R1,GO,RO
70 PRINT "C0:";C0;" C1:";C1
80 PRINT "L1:";L1;" R1:";R1
90 PRINT "GO:";GO;" RO:";RO
100 END
```

EQUCPARA5?

Executes the equivalent circuit analysis for a resonator using the same equivalent circuit of EQUCPARA?, but does not output G0. EQUCPARA5? does not display the warning even an anti-resonance frequency is not in the analysis range. (Query only)

Syntax EQUCPARA5?

Query C_0, C_1, L_1, R_1, R_0 (Total 5)

Response Where,

Register	Parameter	Description
0	C_0	Parallel capacitance
1	C_1	Motional capacitance
2	L_1	Motional inductance
3	R_1	Motional resistance
4	R_0	Electrode resistance

EQUCPARS?

Outputs the six-device equivalent circuit parameters of the crystal resonator. (Query only)

Syntax EQUCPARS?

Query $C_0, C_1, L_1, R_1, f_s, f_a, f_r, f_1^*, f_2^*, G_0, R_0$ (Total 11)

Response * $f_1 < f_2$

For information about each parameter, see "EQUCPARA?".

Note ■ You must select the following conditions or Polar format to use this command:

- Dual Channel & Coupled Channel: ON
- Impedance Conversion: ON
- Analysis channel: LOG MAG format
- Non-analysis channel: Phase format

If another format is selected, 0 will be returned for query response.

EQUCO? value

Returns the parallel capacitance (C_0) of the equivalent circuit of the resonator at the specified frequency. (Query only)

Syntax EQUCO? *value*

Where,

Register	Parameter	Description
0	<i>value</i>	Frequency for C_0

Query C_0

Response Where,

Register	Parameter	Description
0	C_0	Parallel capacitance

Semantics ■ C_0 is calculated using the following equation:

$$C_0 = \frac{B_s}{\omega_s}$$

Where,

B_s Imaginary part of the point on f_s .

$\omega_s = 2\pi f_s$

f_s Frequency that is specified by the command parameter.

- If the impedance conversion is selected, C_0 is calculated using the following equation:

$$C_0 = \frac{-1}{B_s \times \omega_s}$$

Note ■ You must select the following conditions or Polar format to use this command:

- Dual Channel & Coupled Channel: ON
- Impedance Conversion: ON
- Analysis channel: LOG MAG format
- Non-analysis channel: Phase format

If another format is selected, 0 will be returned for query response.

- If the specified frequency is out of analysis range, 0 will be returned.
- If B_s is 0 when the impedance conversion is selected, EQUCO? returns 0.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"DUAC ON; COUC ON"
30 OUTPUT @E5100;"CHAN2; FMT PHAS"
40 OUTPUT @E5100;"CHAN1; FMT LOGM; CONV ZTRA"
50 OUTPUT @E5100;"CENT 60.06MHz; SPAN 20kHz"
60 CALL Sweep(2) ! Goes to sob routine. (See OUTPFILT?)
70 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAODATA"
80 OUTPUT @E5100;"EQUCO? 60.06MHz"
90 ENTER @E5100;CO
100 PRINT "CO:";CO

```

110 END

EQUCPARS4?

Returns the 4-device equivalent circuit parameters of the crystal resonator. (Query only)

Syntax EQUCPARS4?

Query $C_0, C_1, L_1, R_1, f_s, f_a, f_r, f_1, f_2$ (Total9)

Response Where,

Register	Parameter	Description
0	C_0	Parallel capacitance
1	C_1	Motional capacitance
2	L_1	Motional inductance
3	R_1	Motional resistance
4	f_s	Motional (parallel) resonant frequency
5	f_a	Anti-resonant frequency
6	f_r	Resonant frequency
7	f_1	Frequency at the point where the half of maximum conductance.
8	f_2	Frequency at the point where the half of maximum conductance. ($f_1 < f_2$)




Figure D-18. Four-Device Equivalent Circuit of Crystal Resonator

- Semantics**
- You need to select the polar format (FMT POLA) and turn the admittance conversion on to use this command.
 - EQUCPARS4? executes the following actions:
 1. Obtains the admittance characteristic circle diagram. (See Figure D-19.)
 2. Obtains the susceptance (B_{fs}) and its frequency (f_s) at the maximum conductance (G_{max}) point.
 3. Obtains frequencies f_1 and f_2 ($f_1 < f_2$) of the two points where the conductance is half the maximum conductance (G_{max}).

4. Assumes that the frequency at which the phase becomes 0° near the parallel resonance frequency is f_a .
5. Assumes that the frequency at which the phase becomes 0° near the series resonance frequency is f_r .
6. Calculates the constants using the above values and the following equations:

$$C_0 = \frac{f_r^2}{f_a^2 - f_r^2} \times C_1$$

$$C_1 = \frac{1}{QR_1 2\pi f_s}$$

$$L_1 = \frac{QR_1}{2\pi f_s}$$

$$Q = \left| \frac{f_s}{f_2 - f_1} \right|$$

$$R_1 = \frac{1}{G_{max}}$$

If there are no f_r and f_a points on the admittance chart, C_0 is calculated using the following equation:

$$C_0 = \frac{B_{fs}}{2\pi f_s}$$

Where, B_{fs} is the susceptance at the G_{max} point.




Figure D-19. Admittance Characteristic Circle Diagram

Note

* This command is only available when Polar format and the admittance conversion is on. If these are not selected, 0 will be returned.

Examples

```

10 ASSIGN @E5100 TO 717
20 OUTPUT @E5100;"FMT POLA; CONV YTRA; CENT 60.06MHz; SPAN 20kHz"
30 CALL Sweep(1) ! Goes to sub routine. (See OUTPFILT?)
40 OUTPUT @E5100;"ANAOCH1;ANARFULL;ANAOData"
50 OUTPUT @E5100;"EQUCPARS4?"
60 ENTER @E5100;C0,C1,L1,R1
70 PRINT "C0: ";C0;" C1: ";C1
80 PRINT "L1: ";L1;" R1: ";R1
90 END

```


E

Command HP 87510A v.s. HP E5100A

COMMAND

COMMAND	HP 87510A	HP E5100A
*CLS	SUPPORT	SUPPORT
*ESE	SUPPORT	SUPPORT
*ESR?	SUPPORT	SUPPORT
*IDN?	SUPPORT	SUPPORT
*PCB	SUPPORT	SUPPORT
*RST	SUPPORT	SUPPORT
*SRE	SUPPORT	SUPPORT
*STB?	SUPPORT	SUPPORT
*TRG	SUPPORT	SUPPORT
*TST?	SUPPORT	SUPPORT
*WAI	SUPPORT	SUPPORT
ADDRCONT	SUPPORT	NO SUPPORT
ADDRPIN	SUPPORT	NO SUPPORT
ADDRPLOT	NO SUPPORT	NO SUPPORT
ADTOTRAC	NO SUPPORT	SUPPORT
ANAMODE	NO SUPPORT	SUPPORT
ANA OCH	SUPPORT	SUPPORT
ANAODATA	SUPPORT	SUPPORT
ANAOMEMO	SUPPORT	SUPPORT
ANARANG	SUPPORT	SUPPORT
ANARANGP	NO SUPPORT	SUPPORT
ANARFULL	SUPPORT	SUPPORT
AR	SUPPORT	SUPPORT
ASCE	NO SUPPORT	NO SUPPORT
ATTI	NO SUPPORT	SUPPORT
ATTIAAUTO	NO SUPPORT	SUPPORT
ATTIBAUTO	NO SUPPORT	SUPPORT
ATTICAUTO	NO SUPPORT	SUPPORT
ATTIRAUTO	NO SUPPORT	SUPPORT

E: Command HP 87510A
v.s. HP E5100A

COMMAND	HP 87510A	HP E5100A
ATTN	SUPPORT	NO SUPPORT
ATTW	SUPPORT	SUPPORT
AUTO	SUPPORT	SUPPORT
AVER	NO SUPPORT	NO SUPPORT
AVERFACT	NO SUPPORT	NO SUPPORT
AVERREST	NO SUPPORT	NO SUPPORT
BASL	NO SUPPORT	SUPPORT
BEEPDONE	SUPPORT	SUPPORT
BEEFFAIL	SUPPORT	SUPPORT
BEEPWARN	SUPPORT	SUPPORT
BINCLEL	NO SUPPORT	NO SUPPORT
BINEDONE	NO SUPPORT	NO SUPPORT
BINESB	NO SUPPORT	NO SUPPORT
BINL	NO SUPPORT	NO SUPPORT
BINO	NO SUPPORT	NO SUPPORT
BINOA	NO SUPPORT	NO SUPPORT
BINOB	NO SUPPORT	NO SUPPORT
BINP	NO SUPPORT	NO SUPPORT
BINS	NO SUPPORT	NO SUPPORT
BINSADD	NO SUPPORT	NO SUPPORT
BINSDEL	NO SUPPORT	NO SUPPORT
BINSDON	NO SUPPORT	NO SUPPORT
BINSEDI	NO SUPPORT	NO SUPPORT
BINSIZE	SUPPORT	SUPPORT
BINSLINE	NO SUPPORT	NO SUPPORT
BINU	NO SUPPORT	NO SUPPORT
C0	SUPPORT	NO SUPPORT
C1	SUPPORT	NO SUPPORT
C2	SUPPORT	NO SUPPORT
CALCASSI	SUPPORT	NO SUPPORT
CALCOPY	NO SUPPORT	SUPPORT
CALI	SUPPORT	SUPPORT
CALIRAI	SUPPORT	NO SUPPORT
CALIRESP	SUPPORT	NO SUPPORT
CALIS111	SUPPORT	NO SUPPORT
CALK	SUPPORT	SUPPORT
CALK7MM	NO SUPPORT	NO SUPPORT
CALKN50	NO SUPPORT	NO SUPPORT
CALKN75	NO SUPPORT	NO SUPPORT
CALKUSED	SUPPORT	NO SUPPORT
CALN	SUPPORT	NO SUPPORT
CALS	SUPPORT	NO SUPPORT
CENS	SUPPORT	SUPPORT
CENT	SUPPORT	SUPPORT

COMMAND	HP 87510A	HP E5100A
CHAD	SUPPORT	SUPPORT
CHAIRANG	SUPPORT	NO SUPPORT
CHAN	NO SUPPORT	SUPPORT
CHAN1	SUPPORT	SUPPORT
CHAN2	SUPPORT	SUPPORT
CIN	SUPPORT	SUPPORT
CIVAL	NO SUPPORT	SUPPORT
CLAD	SUPPORT	NO SUPPORT
CLASS11	SUPPORT	SUPPORT
CLEL	SUPPORT	SUPPORT
CLEM	SUPPORT	SUPPORT
CLEMNU3	NO SUPPORT	SUPPORT
CLES	SUPPORT	SUPPORT
CONT	SUPPORT	SUPPORT
CONV	SUPPORT	SUPPORT
CONVIDS	NO SUPPORT	NO SUPPORT
CONVMP16	SUPPORT	NO SUPPORT
CONVMP4	SUPPORT	NO SUPPORT
CONVMP8	SUPPORT	NO SUPPORT
CONVOFF	SUPPORT	NO SUPPORT
CONVYREF	SUPPORT	NO SUPPORT
CONVYTRA	SUPPORT	NO SUPPORT
CONVZREF	SUPPORT	NO SUPPORT
CONVZTRA	SUPPORT	NO SUPPORT
COPA	SUPPORT	NO SUPPORT
COPT	SUPPORT	NO SUPPORT
CORR	SUPPORT	SUPPORT
COUC	SUPPORT	SUPPORT
COUT	SUPPORT	SUPPORT
CRED	SUPPORT	SUPPORT
CURD?	SUPPORT	SUPPORT
CURMPOIN?	NO SUPPORT	SUPPORT
CWFREQ	SUPPORT	SUPPORT
DATAM3TER	SUPPORT	NO SUPPORT
DATAMNONE	SUPPORT	NO SUPPORT
DATAMTHRU	SUPPORT	NO SUPPORT
DATI	NO SUPPORT	SUPPORT
DAYMYEAR	SUPPORT	SUPPORT
DCBUS	SUPPORT	NO SUPPORT
DEFS	SUPPORT	NO SUPPORT
DELA	SUPPORT	SUPPORT
DELO	SUPPORT	SUPPORT
DELR	SUPPORT	SUPPORT
DELRFIXM	SUPPORT	SUPPORT
DESTOFF	SUPPORT	NO SUPPORT
DESTON	SUPPORT	NO SUPPORT

E: Command HP 87510A
v.s. HP E5100A

COMMAND	HP 87510A	HP E5100A
DFLT	NO SUPPORT	NO SUPPORT
DIN	SUPPORT	SUPPORT
DISA	SUPPORT	SUPPORT
DISAALLB	SUPPORT	NO SUPPORT
DISAALLI	SUPPORT	SUPPORT
DISABASS	SUPPORT	SUPPORT
DISAHIHB	SUPPORT	SUPPORT
DISBLIST	NO SUPPORT	SUPPORT
DISFDOS	SUPPORT	NO SUPPORT
DISFLIF	NO SUPPORT	NO SUPPORT
DISG	SUPPORT	SUPPORT
DISL1	SUPPORT	NO SUPPORT
DISL2	SUPPORT	NO SUPPORT
DISLLIST	SUPPORT	NO SUPPORT
DISMCTSP	SUPPORT	NO SUPPORT
DISMMD	SUPPORT	NO SUPPORT
DISMNUM	SUPPORT	NO SUPPORT
DISMSTEP	SUPPORT	NO SUPPORT
DISMSTSP	SUPPORT	NO SUPPORT
DISMUL	SUPPORT	NO SUPPORT
DISP	SUPPORT	SUPPORT
DISPDATA	SUPPORT	NO SUPPORT
DISPDATM	SUPPORT	NO SUPPORT
DISPDDM	SUPPORT	NO SUPPORT
DISPDMM	SUPPORT	NO SUPPORT
DISPMEMO	SUPPORT	NO SUPPORT
DIST	SUPPORT	NO SUPPORT
DONE	SUPPORT	SUPPORT
DOUT	SUPPORT	SUPPORT
DSKEY	SUPPORT	SUPPORT
DUAC	SUPPORT	SUPPORT
EDITBINL	NO SUPPORT	NO SUPPORT
EDITDONE	SUPPORT	SUPPORT
EDITLIML	SUPPORT	NO SUPPORT
EDITLIS1	SUPPORT	SUPPORT
EDITLIS2	SUPPORT	SUPPORT
EDITLIST	SUPPORT	SUPPORT
ELED	SUPPORT	SUPPORT
ENKEY	SUPPORT	SUPPORT
EQUO?	SUPPORT	SUPPORT
EQUCPARA5	NO SUPPORT	SUPPORT
EQUCPARA?	SUPPORT	SUPPORT
EQUCPARS4?	SUPPORT	SUPPORT
EQUCPARS?	SUPPORT	SUPPORT
EQU	SUPPORT	SUPPORT

COMMAND	HP 87510A	HP E5100A
ERRH	SUPPORT	NO SUPPORT
ESB?	SUPPORT	SUPPORT
ESNB	SUPPORT	SUPPORT
EXET	SUPPORT	SUPPORT
EXPP	SUPPORT	NO SUPPORT
EXPZP	NO SUPPORT	SUPPORT
EXTRLOCK?	SUPPORT	SUPPORT
EXTT	SUPPORT	SUPPORT
EXTTOFF	SUPPORT	NO SUPPORT
EXTTON	SUPPORT	NO SUPPORT
EXTTPOIN	SUPPORT	NO SUPPORT
FBUS	SUPPORT	NO SUPPORT
FILC	SUPPORT	SUPPORT
FIRLPNOR	SUPPORT	NO SUPPORT
FIRLPOPE	SUPPORT	NO SUPPORT
FIRR?	SUPPORT	NO SUPPORT
FMT	SUPPORT	SUPPORT
FMTMAGZDF	NO SUPPORT	SUPPORT
FNDAUTO	SUPPORT	NO SUPPORT
FNDMANU	SUPPORT	NO SUPPORT
FNDVALU	SUPPORT	NO SUPPORT
FNVNARR	SUPPORT	NO SUPPORT
FNVNORM	SUPPORT	NO SUPPORT
FNVOPEN	SUPPORT	NO SUPPORT
FNVWIDE	SUPPORT	NO SUPPORT
FORM2	SUPPORT	SUPPORT
FORM3	SUPPORT	SUPPORT
FORM4	SUPPORT	SUPPORT
FORM5	SUPPORT	SUPPORT
FREELIST	SUPPORT	NO SUPPORT
FREO	SUPPORT	NO SUPPORT
FULP	NO SUPPORT	NO SUPPORT
GRAE	NO SUPPORT	NO SUPPORT
GRAPFORM	NO SUPPORT	SUPPORT
GRODAPER	SUPPORT	SUPPORT
HOLD	SUPPORT	SUPPORT
IFBW	SUPPORT	SUPPORT
IFBWAUTO	SUPPORT	SUPPORT
IFRCH?	SUPPORT	NO SUPPORT
IFRX1	SUPPORT	NO SUPPORT
IFRX1X8	SUPPORT	NO SUPPORT
IFRX64	SUPPORT	NO SUPPORT
IFRX8X1	SUPPORT	NO SUPPORT
IMAG	SUPPORT	NO SUPPORT
INID	SUPPORT	SUPPORT

ECommand HP 87510A
 v.s. HP E5100A

COMMAND	HP 87510A	HP E5100A
INP8IO	SUPPORT	SUPPORT
INPT?	SUPPORT	SUPPORT
INPUCALC{01-03}	SUPPORT	SUPPORT
INPUCALK	SUPPORT	NO SUPPORT
INPUD	SUPPORT	NO SUPPORT
INPUDATA	NO SUPPORT	SUPPORT
INPUDATM1	SUPPORT	NO SUPPORT
INPUDATM2	SUPPORT	NO SUPPORT
INPUDATM3	SUPPORT	NO SUPPORT
INPUDATM4	SUPPORT	NO SUPPORT
INPUDATTP	SUPPORT	NO SUPPORT
INPUFORM	SUPPORT	SUPPORT
INPUIFORM	NO SUPPORT	SUPPORT
INPULOAA	SUPPORT	SUPPORT
INPUMEMO	NO SUPPORT	SUPPORT
INPUMEMTP	SUPPORT	NO SUPPORT
INPUOPEA	SUPPORT	NO SUPPORT
INPURAW	NO SUPPORT	SUPPORT
INPURFORM	NO SUPPORT	SUPPORT
INPURTMEM	NO SUPPORT	SUPPORT
INPUSHOA	SUPPORT	NO SUPPORT
INPUTMEM	SUPPORT	NO SUPPORT
INPUTRAC	SUPPORT	SUPPORT
INPUTRACB	SUPPORT	SUPPORT
IOPO?	SUPPORT	SUPPORT
KEY	SUPPORT	NO SUPPORT
KITD	SUPPORT	NO SUPPORT
LABERES{P I}	SUPPORT	NO SUPPORT
LABES11{A B C}	SUPPORT	NO SUPPORT
LABK	SUPPORT	NO SUPPORT
LABS	SUPPORT	NO SUPPORT
LEFL	NO SUPPORT	NO SUPPORT
LEFU	NO SUPPORT	NO SUPPORT
LIMCLEL	SUPPORT	NO SUPPORT
LIMD	SUPPORT	NO SUPPORT
LIMEDONE	SUPPORT	NO SUPPORT
LIMIAMPO	SUPPORT	NO SUPPORT
LIMILINE	SUPPORT	SUPPORT
LIMMAOF	SUPPORT	NO SUPPORT
LIMIOPOIN	SUPPORT	NO SUPPORT
LIMIOSEND	SUPPORT	NO SUPPORT
LIMISTIO	SUPPORT	NO SUPPORT
LIMITEST	SUPPORT	SUPPORT
LIML	SUPPORT	NO SUPPORT
LIMM	SUPPORT	NO SUPPORT

COMMAND	HP 87510A	HP E5100A
LIMS	SUPPORT	NO SUPPORT
LIMSADD	SUPPORT	NO SUPPORT
LIMSDEL	SUPPORT	NO SUPPORT
LIMSDON	SUPPORT	NO SUPPORT
LIMSEDI	SUPPORT	NO SUPPORT
LIMU	SUPPORT	NO SUPPORT
LINFREQ	SUPPORT	SUPPORT
LINM	SUPPORT	SUPPORT
LINM{F P}	SUPPORT	NO SUPPORT
LINT{DATA MEMO}	NO SUPPORT	NO SUPPORT
LISDFBASE	SUPPORT	SUPPORT
LISDOBASE	SUPPORT	SUPPORT
LISFREQ	SUPPORT	SUPPORT
LISSLIS1	SUPPORT	SUPPORT
LISSLIS2	SUPPORT	SUPPORT
LISV	SUPPORT	NO SUPPORT
LMAX?	SUPPORT	SUPPORT
LMAXS?	SUPPORT	NO SUPPORT
LMIN?	SUPPORT	SUPPORT
LMINS?	SUPPORT	NO SUPPORT
LOGFREQ	SUPPORT	NO SUPPORT
LOGM	SUPPORT	SUPPORT
LOGMD	SUPPORT	SUPPORT
LOGMP	SUPPORT	SUPPORT
LOWLIMI	NO SUPPORT	SUPPORT
MANTRIG	SUPPORT	NO SUPPORT
MARD	SUPPORT	SUPPORT
MARKBUCK	SUPPORT	SUPPORT
MARKCENT	SUPPORT	SUPPORT
MARKCONT	SUPPORT	SUPPORT
MARKCOUP	SUPPORT	SUPPORT
MARKDISC	SUPPORT	SUPPORT
MARKFAUV	SUPPORT	NO SUPPORT
MARKFSTI	SUPPORT	SUPPORT
MARKFVAL	SUPPORT	SUPPORT
MARKL	SUPPORT	SUPPORT
MARKMIDD	SUPPORT	NO SUPPORT
MARKODATA	SUPPORT	SUPPORT
MARKOFF	SUPPORT	SUPPORT
MARKOMEMO	SUPPORT	SUPPORT
MARKPEAD	SUPPORT	NO SUPPORT
MARKREF	SUPPORT	SUPPORT
MARKSPAN	SUPPORT	SUPPORT
MARKSTAR	SUPPORT	SUPPORT
MARKSTIM	SUPPORT	NO SUPPORT

COMMAND	HP 87510A	HP E5100A
MARKSTOP	SUPPORT	SUPPORT
MARKTIME	SUPPORT	SUPPORT
MARKUNCO	SUPPORT	SUPPORT
MARKZERO	SUPPORT	SUPPORT
MARK{1-8}	SUPPORT	SUPPORT
MAXD?	SUPPORT	NO SUPPORT
MAXPOIN?	NO SUPPORT	SUPPORT
MAXPORT?	NO SUPPORT	SUPPORT
MEAS	SUPPORT	SUPPORT
MEAS1PT?	NO SUPPORT	SUPPORT
MEASA	SUPPORT	SUPPORT
MEASR	SUPPORT	SUPPORT
MEASTAT	SUPPORT	SUPPORT
MENU3	NO SUPPORT	SUPPORT
MOD11	SUPPORT	NO SUPPORT
MONDYEAR	SUPPORT	SUPPORT
MOVADARY	NO SUPPORT	SUPPORT
MULC	NO SUPPORT	SUPPORT
NEGL	SUPPORT	SUPPORT
NEXP	SUPPORT	NO SUPPORT
NEXPK?	SUPPORT	SUPPORT
NUMC	NO SUPPORT	SUPPORT
NUMG	SUPPORT	NO SUPPORT
NUMLMAX?	SUPPORT	SUPPORT
NUMLMIN?	SUPPORT	SUPPORT
NUMLMINMAX?	SUPPORT	NO SUPPORT
OFSD	SUPPORT	NO SUPPORT
OFSL	SUPPORT	NO SUPPORT
OFSZ	SUPPORT	NO SUPPORT
OPEP	SUPPORT	NO SUPPORT
OPTI?	SUPPORT	NO SUPPORT
OSE	SUPPORT	SUPPORT
OSER?	SUPPORT	SUPPORT
OSNT	SUPPORT	SUPPORT
OSPT	SUPPORT	SUPPORT
OSR?	SUPPORT	SUPPORT
OUT1ENVH	SUPPORT	SUPPORT
OUT1ENVL	SUPPORT	SUPPORT
OUT1H	SUPPORT	SUPPORT
OUT1L	SUPPORT	SUPPORT
OUT2ENVH	SUPPORT	SUPPORT
OUT2ENVL	SUPPORT	SUPPORT
OUT2H	SUPPORT	SUPPORT
OUT2L	SUPPORT	SUPPORT
OUT8IO	SUPPORT	SUPPORT
OUTAIO	SUPPORT	SUPPORT
OUTBIO	SUPPORT	SUPPORT

COMMAND	HP 87510A	HP E5100A
OUTCIO	SUPPORT	SUPPORT
OUTDIO	SUPPORT	SUPPORT
OUTEIO	SUPPORT	SUPPORT
OUTFIO	SUPPORT	SUPPORT
OUTGIO	SUPPORT	SUPPORT
OUTHIO	SUPPORT	SUPPORT
OUTPCALC{01-03}?	SUPPORT	SUPPORT
OUTPCALK?	SUPPORT	NO SUPPORT
OUTPCERR?	SUPPORT	SUPPORT
OUTPCF2	NO SUPPORT	SUPPORT
OUTPCFIL?	SUPPORT	SUPPORT
OUTPDATA?	NO SUPPORT	SUPPORT
OUTPDATAT?	SUPPORT	SUPPORT
OUTPDATTP?	SUPPORT	SUPPORT
OUTPERRO?	SUPPORT	SUPPORT
OUTPFAIP?	SUPPORT	NO SUPPORT
OUTPFBUS?	SUPPORT	NO SUPPORT
OUTPFILT?	SUPPORT	SUPPORT
OUTPFORM?	SUPPORT	SUPPORT
OUTPFORMP?	SUPPORT	SUPPORT
OUTPIFORM?	SUPPORT	SUPPORT
OUTPINP8IO?	SUPPORT	SUPPORT
OUTPINPCIO?	SUPPORT	SUPPORT
OUTPINPDIO?	SUPPORT	SUPPORT
OUTPINPEIO?	SUPPORT	SUPPORT
OUTPIRFORM?	SUPPORT	SUPPORT
OUTPIRTMEM?	SUPPORT	SUPPORT
OUTPITMEM?	SUPPORT	NO SUPPORT
OUTPLIMF?	SUPPORT	NO SUPPORT
OUTPLIML?	SUPPORT	NO SUPPORT
OUTPLIMM?	SUPPORT	NO SUPPORT
OUTPMARK?	SUPPORT	SUPPORT
OUTPMAX?	SUPPORT	SUPPORT
OUTPMEAN?	SUPPORT	SUPPORT
OUTPMEMO?	SUPPORT	SUPPORT
OUTPMEMOT?	SUPPORT	SUPPORT
OUTPMEMTP?	SUPPORT	NO SUPPORT
OUTPMIN?	SUPPORT	SUPPORT
OUTPMINMAX?	SUPPORT	SUPPORT
OUTPMSTA?	SUPPORT	SUPPORT
OUTPMWID?	SUPPORT	SUPPORT
OUTPMWIL?	SUPPORT	SUPPORT
OUTPMWLF?	SUPPORT	SUPPORT
OUTPPEAK?	SUPPORT	NO SUPPORT
OUTPRAW?	NO SUPPORT	SUPPORT

E Command HP 87510A
v.s. HP E5100A

COMMAND	HP 87510A	HP E5100A
OUTPRESF?	SUPPORT	SUPPORT
OUTPRESO?	SUPPORT	SUPPORT
OUTPRESR?	SUPPORT	SUPPORT
OUTPRFORM?	SUPPORT	SUPPORT
OUTPRFORMP?	NO SUPPORT	SUPPORT
OUTPRTMEM?	SUPPORT	SUPPORT
OUTPRTMEMP?	NO SUPPORT	SUPPORT
OUTPSTIM?	SUPPORT	SUPPORT
OUTPTESS?	SUPPORT	NO SUPPORT
OUTPTITL?	SUPPORT	SUPPORT
OUTPTMEM?	SUPPORT	NO SUPPORT
OUTPTMEMP?	SUPPORT	NO SUPPORT
OUTPTRAC	SUPPORT	SUPPORT
OUTPTRACB	SUPPORT	SUPPORT
OUTPXF2	NO SUPPORT	SUPPORT
OUTPXFIL?	SUPPORT	SUPPORT
PARS	SUPPORT	SUPPORT
PEADX	SUPPORT	NO SUPPORT
PEADY	SUPPORT	NO SUPPORT
PEAK?	SUPPORT	SUPPORT
PEAKLIST?	NO SUPPORT	SUPPORT
PHAO	SUPPORT	SUPPORT
PHAS	SUPPORT	SUPPORT
PICIRC	NO SUPPORT	SUPPORT
PLOALL	NO SUPPORT	NO SUPPORT
PLOC	NO SUPPORT	NO SUPPORT
PLODONLY	NO SUPPORT	NO SUPPORT
PLOGRAT	NO SUPPORT	NO SUPPORT
PLOS	NO SUPPORT	NO SUPPORT
PLOT	NO SUPPORT	NO SUPPORT
POIN	SUPPORT	SUPPORT
POLA	SUPPORT	NO SUPPORT
POLE?	SUPPORT	SUPPORT
POLM	SUPPORT	NO SUPPORT
POLMLIN	SUPPORT	NO SUPPORT
POLMLOG	SUPPORT	NO SUPPORT
POLMRI	SUPPORT	NO SUPPORT
PORE	SUPPORT	NO SUPPORT
PORTA	SUPPORT	NO SUPPORT
PORTR	SUPPORT	NO SUPPORT
POSL	SUPPORT	SUPPORT
POWDAUTO	SUPPORT	NO SUPPORT
POWDMANU	SUPPORT	NO SUPPORT
POWDVALU	SUPPORT	NO SUPPORT
POWE	SUPPORT	SUPPORT

COMMAND	HP 87510A	HP E5100A
POWF	SUPPORT	NO SUPPORT
POWL	SUPPORT	NO SUPPORT
POWO?	SUPPORT	NO SUPPORT
POWS	SUPPORT	SUPPORT
POWU	NO SUPPORT	SUPPORT
PREP	SUPPORT	NO SUPPORT
PRES	SUPPORT	SUPPORT
PRINALL	SUPPORT	SUPPORT
PRIR	NO SUPPORT	SUPPORT
PSOFT	NO SUPPORT	NO SUPPORT
PTABORT	NO SUPPORT	SUPPORT
PTFOVHD	SUPPORT	SUPPORT
PTFR	SUPPORT	SUPPORT
PTFRBW	SUPPORT	NO SUPPORT
PTFRSR	SUPPORT	SUPPORT
PTPARA	NO SUPPORT	SUPPORT
PTPMNUM	SUPPORT	NO SUPPORT
PTRACK	SUPPORT	SUPPORT
PTREP	NO SUPPORT	SUPPORT
PTRUPD	SUPPORT	NO SUPPORT
PTSTAT	SUPPORT	SUPPORT
PTTRGLMT	SUPPORT	SUPPORT
PTTRGPHS	SUPPORT	SUPPORT
PURG	SUPPORT	SUPPORT
QUAD	NO SUPPORT	NO SUPPORT
RAID	SUPPORT	SUPPORT
RAISOL	SUPPORT	SUPPORT
RAIRESP	SUPPORT	SUPPORT
REAL	SUPPORT	NO SUPPORT
RECCOFF	SUPPORT	NO SUPPORT
RECCON	SUPPORT	NO SUPPORT
RECD	SUPPORT	SUPPORT
REFP	SUPPORT	SUPPORT
REFV	SUPPORT	SUPPORT
RESAVD	SUPPORT	SUPPORT
RESC	SUPPORT	NO SUPPORT
RESD	SUPPORT	NO SUPPORT
RESPDONE	SUPPORT	SUPPORT
REST	SUPPORT	SUPPORT
RFOPNORM	SUPPORT	NO SUPPORT
RFOPEN	SUPPORT	NO SUPPORT
RIGL	NO SUPPORT	NO SUPPORT
RIGU	NO SUPPORT	NO SUPPORT
RPLENV?	SUPPORT	SUPPORT
RPLHEI?	SUPPORT	SUPPORT

E: Command HP 87510A
v.s. HP E5100A

COMMAND	HP 87510A	HP E5100A
RPLLHEI?	SUPPORT	SUPPORT
RPLMEA?	SUPPORT	SUPPORT
RPLMM	NO SUPPORT	SUPPORT
RPLPP?	SUPPORT	SUPPORT
RPLPPS?	SUPPORT	SUPPORT
RPLRHEI?	SUPPORT	SUPPORT
RPLVAL?	SUPPORT	SUPPORT
SADD	SUPPORT	SUPPORT
SAVI	SUPPORT	SUPPORT
SAVC	SUPPORT	NO SUPPORT
SAVCA	SUPPORT	SUPPORT
SAVDA	NO SUPPORT	SUPPORT
SAVDALL	SUPPORT	SUPPORT
SAVDASC	SUPPORT	SUPPORT
SAVDDAT	SUPPORT	SUPPORT
SAVDGRA	NO SUPPORT	NO SUPPORT
SAVDGRAP	NO SUPPORT	SUPPORT
SAVDMNU3	NO SUPPORT	SUPPORT
SAVDSTA	SUPPORT	SUPPORT
SAVEUSEK	SUPPORT	NO SUPPORT
SAVFA	NO SUPPORT	SUPPORT
SAVMA	NO SUPPORT	SUPPORT
SAVRA	NO SUPPORT	SUPPORT
SAVTA	SUPPORT	SUPPORT
SAVTMA	SUPPORT	SUPPORT
SCAC	SUPPORT	NO SUPPORT
SCAFDATA	SUPPORT	SUPPORT
SCAFMEMO	SUPPORT	SUPPORT
SCAL	SUPPORT	SUPPORT
SCAPFULL	NO SUPPORT	NO SUPPORT
SCAPGL	NO SUPPORT	NO SUPPORT
SCAPGU	NO SUPPORT	NO SUPPORT
SCAU	SUPPORT	NO SUPPORT
SCAY	NO SUPPORT	SUPPORT
SDEL	SUPPORT	SUPPORT
SDON	SUPPORT	SUPPORT
SEAL	SUPPORT	SUPPORT
SEALMAX	SUPPORT	SUPPORT
SEALMIN	SUPPORT	SUPPORT
SEAM	SUPPORT	SUPPORT
SEAMAX	SUPPORT	SUPPORT
SEAMEAN	SUPPORT	SUPPORT
SEAMIN	SUPPORT	SUPPORT
SEAOFF	SUPPORT	SUPPORT
SEAPPEAK	SUPPORT	SUPPORT

COMMAND	HP 87510A	HP E5100A
SEAR	SUPPORT	SUPPORT
SEARSTOR	SUPPORT	SUPPORT
SEATARG	SUPPORT	SUPPORT
SEDI	SUPPORT	SUPPORT
SEET	SUPPORT	SUPPORT
SELD	SUPPORT	NO SUPPORT
SERM?	SUPPORT	NO SUPPORT
SETCDATE	SUPPORT	SUPPORT
SETCTIME	SUPPORT	SUPPORT
SETZ	SUPPORT	SUPPORT
SINDTMEM	SUPPORT	NO SUPPORT
SING	SUPPORT	SUPPORT
SINSPEAK	SUPPORT	SUPPORT
SMOO	SUPPORT	SUPPORT
SMOOPER	SUPPORT	SUPPORT
SOUCOFF	SUPPORT	NO SUPPORT
SOUCON	SUPPORT	NO SUPPORT
SPAN	SUPPORT	SUPPORT
SPECRESI	SUPPORT	NO SUPPORT
SPECRESP	SUPPORT	NO SUPPORT
SPECS11A	SUPPORT	NO SUPPORT
SPECS11B	SUPPORT	NO SUPPORT
SPECS11C	SUPPORT	NO SUPPORT
SPLD	SUPPORT	SUPPORT
SRCHFR	NO SUPPORT	SUPPORT
STAN	SUPPORT	NO SUPPORT
STAR	SUPPORT	SUPPORT
STAS	SUPPORT	SUPPORT
STAW	NO SUPPORT	SUPPORT
STDD	SUPPORT	NO SUPPORT
STDT	SUPPORT	NO SUPPORT
STDTARBI	SUPPORT	NO SUPPORT
STDTDELA	SUPPORT	NO SUPPORT
STDTLOAD	SUPPORT	NO SUPPORT
STDTOPEN	SUPPORT	NO SUPPORT
STDTSHOR	SUPPORT	NO SUPPORT
STED	NO SUPPORT	SUPPORT
STIDROUT{1-16}	SUPPORT	SUPPORT
STIMROUT{1-16}	SUPPORT	SUPPORT
STOD{DISK MEMO}	SUPPORT	SUPPORT
STOP	SUPPORT	SUPPORT
STPSIZE	SUPPORT	SUPPORT
STR	NO SUPPORT	SUPPORT
SWET	SUPPORT	SUPPORT
SWETAUTO	SUPPORT	SUPPORT

E Command HP 87510A
v.s. HP E5100A

COMMAND	HP 87510A	HP E5100A
SWPT	SUPPORT	SUPPORT
TARL?	SUPPORT	SUPPORT
TARLRP?	SUPPORT	NO SUPPORT
TARR?	SUPPORT	SUPPORT
TERI	SUPPORT	NO SUPPORT
TESC	SUPPORT	NO SUPPORT
TEST	SUPPORT	NO SUPPORT
THRR	SUPPORT	SUPPORT
TIMO	SUPPORT	SUPPORT
TITL	SUPPORT	SUPPORT
TOTIME	SUPPORT	SUPPORT
TRABGE	NO SUPPORT	SUPPORT
TRABLE	NO SUPPORT	SUPPORT
TRACK	SUPPORT	SUPPORT
TRAFDATA	SUPPORT	SUPPORT
TRAFMEMO	SUPPORT	SUPPORT
TRAP	SUPPORT	SUPPORT
TRAR	SUPPORT	SUPPORT
TRIGMEAS	NO SUPPORT	SUPPORT
TRIM	NO SUPPORT	SUPPORT
UPPD	SUPPORT	SUPPORT
UPPELIMI	NO SUPPORT	SUPPORT
VELOFACT	SUPPORT	NO SUPPORT
WIDSIN	SUPPORT	NO SUPPORT
WIDSOUT	SUPPORT	NO SUPPORT
WIDT	SUPPORT	SUPPORT
WIDV	SUPPORT	SUPPORT
WRIT16	SUPPORT	SUPPORT
WRIT24	SUPPORT	SUPPORT

Keyword Guide to Porting

The following sections summarize the differences in the HP-IB commands of the HP 87510A and the HP E5100A/B. This appendix is intended only as a quick reference to the keywords and their compatibility. For detailed information, see the Chapter 8.

The following table is summary of this appendix

HP 87510A	HP E5100A
Trigger	
SRQ	SING?
SING	TRIM SING
Calibration	
*OPC?	STANC?, RAIRESP?, RAIISO?, CLASS11[A B C]?
*OPC?	RESPDONE, RAID?, SAV1
Data Transfer	
FORM2, FORM3, FORM5	No support for data input. (use FORM4)
Meas/Format	
AR, MEASA, MEASR	Full support
LOGM, PHAS, DELA, LINM, LOGMP, LOGMD, REAL, IMAG	Full support
EXPP	FMT EXPP
POLAR	No support
CONVZTRA, CONVZREF, CONVYTRA, CONVYREF	ANAMODE ZREFL, ANAMODE YREFL
Analysis Function	
BINSLINE, BINS, BINOA, BINOB, EDITBINL	No support
EDITLIS1, EDITLIS2	UPELIMI, LOWELIMI
Other	
KEY	No support

Trigger Function

Single Trigger (SRQ)

Use the SING? command instead of SRQ to generate a single trigger. SING? initiates a single measurement sweep and returns 1 when the sweep is completed.

```
OUTPUT @E5100;"SING?"  
ENTER @E5100;Tmp$
```

Single Trigger Mode (SING)

Use the TRIM SING command instead of SING to select a single trigger as the trigger mode.

Calibration

Calibration Measurement (OPC)

Use the STANC?, RAIRESP?, RAIISO?, CLASS11[A|B|C]? query commands instead of the OPC command (which is used to detect that calibration measurement is completed). These query commands make a calibration measurement and return 1 when the calibration measurement is completed.

Calibration Done (OPC)

Use the RESPDONE?, RAID? and SAV1 query commands instead of the OPC command (which is used to detect that the calibration coefficient calculation is completed). These query commands calculate and save the calibration coefficient into the internal array and return 1 when the operation is completed.

Data Transfer Command

Binary Transfer Commands (FORM2, FORM3, FORM4)

Use ASCII transfer (FORM4) when data is entered from an external controller to the HP E5100A/B. The HP E5100A/B does not support binary transfer for data input.

Meas/Format

Input Port (AR, MEASA, MEASR)

The following commands are used to select an input port for the HP 87510A. The AR, MEASA, and MEASR commands are supported (see the following table)

Table F-1.

Equivalent Softkey	HP 87510A	HP E5100A
A/R	MEAS AR or AR	MEAS AR or AR
B/R	-	MEAS BR
C/R	-	MEAS CR
B/A	-	MEAS BA
C/A	-	MEAS CA
R/A	-	MEAS RA
A/B	-	MEAS AB
C/B	-	MEAS CB
R/B	-	MEAS RB
A/C	-	MEAS AC
B/C	-	MEAS BC
R/C	-	MEAS RC
R	MEAS R or MEASR	MEAS R or MEASR
A	MEAS A or MEASA	MEAS A or MEASA
B	-	MEAS B
C	-	MEAS C

Use the MEAS? query command to confirm the current port.

```
OUTPUT @E5100;MEAS?
ENTER @E5100; Port$
```

Format (LOGM, PHAS, DELA, LINM, EXPP, LOGMP, LOGMD, REAL, IMAG, POLAR)

The following commands are supported :

LOGM, PHAS, DELA, LINM, LOGMP, LOGMD, REAL, IMAG

Use the FMT EXPP command instead of EXPP for selecting Expanded Phase format. EXPP is not supported.

Polar chart format (POLAR) is not supported. (Refer to the following table.)

Table F-2.

Equivalent Softkey	HP 87510A	HP E5100A
LOG MAG & PHASE	FMT LOGMP or LOGMP	FMT LOGMP or LOGMP
LOG MAG & DELAY	FMT LOGMD or LOGMD	FMT LOGMD or LOGMD
LIN MAG & PHASE	-	FMT LINMP
LIN MAG & DELAY	-	FMT LINMD
REAL & IMAGINARY	-	FMT RIMAG
LDG MAG	FMT LOGM or LOGM	FMT LOGM or LOGM
LIN MAG	FMT LINM or LINM	FMT LINM or LINM
PHASE	FMT PHAS or PHAS	FMT PHAS or PHAS
DELAY	FMT PHAS or PHAS	FMT PHAS or PHAS
REAL	FMT REAL or REAL	FMT REAL or REAL
IMAGINARY	FMT IMAG or IMAG	FMT IMAG or IMAG
EXPANDED PHASE	FMT EXPP or EXPP	FMT EXPP (EXPP is not supported)
Z & PHASE z	-	FMT MAGZP
Y & PHASE y	-	FMT MAGYP
R-X	-	FMT IMPRX
G-B	-	FMT ADMGB
Z	-	FMT MAGZ
Y	-	FMT MAGY
PHASE z	-	FMT PHAZ
PHASE y	-	FMT PHAY
R	-	FMT IMPR
G	-	FMT ADMG
B	-	FMT ADMB
X	-	FMT IMPX
POLAR	POLA or FMT POLA	No support

Use the FMT? query command to confirm the current format.

```
OUTPUT @E5100;FMT?
ENTER @E5100; Format$
```

Impedance Conversion Function (CONVZTRA, CONVZREF, CONVYTRA, CONVYREF)

Use the impedance measurement mode ANAMODE instead of the conversion function, even though the HP E5100A/B supports the CONV commands.

Table F-3.

Impedance Conversion (HP 87510A)	Impedance Measurement Mode (HP E5100A)
CONVZTRA	ANAMODE ZTRAN
CONVZREF	ANAMODE ZREFL
CONVYTRA	ANAMODE ZTRAN;FMT MAGY
CONVYREF	ANAMODE ZREFL;FMT MAGY

Analysis Commands

BIN sort (BINSLINE, BINS, BINOA, BINO, EDITBINL)

The HP E5100A/B does not support the BIN sort function.

Limit Line Edit (EDITLIS1, EDITLIS2)

Use UPELIMI and LOWELIMI commands to define a limit line. The HP E5100A/B does not support the EDITLIS1 and EDITLIS2 commands of the HP 87510A.

Other Commands

KEY command

The HP E5100A/B does not support the KEY command.

Option 022

Overview

This option enables you to control metal deposition during a deposition process of crystal resonators.

Metal deposition is used during a manufacturing to adjust the frequency of a crystal resonator. The resonant frequency of a crystal resonator becomes lower as deposition proceeds. HP E5100A/B OPT. 022, as described in Figure G-1, monitors the phase at a specified frequency f_1 and outputs preset data to the I/O port when the measured phase goes under 0 (or any value you specify as you wish).

HP E5100A/B OPT. 022, then, moves its focus to the next frequency f_2 and performs another measurement. When the measured phase reaches the specified value again, it outputs different data to the I/O port. This output to the I/O port controls the amount of metal deposition to achieve efficient and accurate metal deposition. This option's function is called the Trap Function because it awaits the phase to match the preset phase at a certain frequency.




Figure G-1. The Trap Function

Using HP E5100A/B Option 022

You must use a program to use the trap function to control a deposition process.

A sample program is provided in the sample program disk. This chapter explains how to measure a device using the sample program.

The sample program for the trap function is prepared under the following conditions.

- A π network test fixture is used.
- Characteristic impedance is set at 12.5Ω to support standard π network test fixture.

This sample program measures a device using the trap function on the channel 1. When the program sees the resonant frequency of the resonator matching the final trap frequency, it switches the focus to the channel 2 to perform a frequency sweep and verifies the resonant frequency.

HP E5100A/B option 022 starts performing the trap function from the measured point at the right of the screen. As deposition proceeds, the resonant frequency of resonator becomes lower and the measured frequency goes from right to left. This is because the function waits for the measured phase to go under 0° (actually the phase value specified by INPUTRAC) at the measured point so that it can output certain data to the I/O port and move to the next measured point to repeat the process.

The more detailed description of the program function is discussed for the sample program with its default setup. The default setup is completed between lines 1750 and 2060 in "The Sample Program List". Please refer to this section for the setup. Figure G-2 describes the status of the channel 1 with this setup.




Figure G-2. How HP E5100A/B option 022 Works with the Default Setup

An Example of Trapping

1. The leftmost frequency on the screen is the stop frequency for the trap function.
(19.979MHz)
2. The rightmost frequency on the screen is the start frequency for the trap function.
(19.998979 MHz = $19.979 \times (1 + 1000E-6)$)
3. A phase is continuously measured at the rightmost frequency on the screen.
4. When deposition proceeds and the phase becomes 0, HP E5100A/B option 022 outputs 8 to the I/O port.
5. HP E5100A/B option 022 moves the measured point to the second point from the right (it changes the measured frequency) and measures the phase.
6. Since the output data is not specified for this I/O port, the output data to the I/O port remain unchanged when the phase becomes 0.
7. HP E5100A/B option 022 goes to the third measured point from the right to perform the trap function.
8. At the 4th measured point from the right, HP E5100A/B option 022 outputs 4 to the I/O port.
9. HP E5100A/B option 022 continues to measure at the rest of measured points until it reaches the stop frequency for the trap function.

Note

You can specify at which measured point and what data to output to the I/O port



As soon as HP E5100A/B option 022 completes the trap function on the channel 1, it performs a normal frequency sweep on the channel 2 to verify the resonant point. When HP E5100A/B option 022 completes the sweep, it calculates the resonant frequency and resonant impedance and displays them at the lower half of the screen. (Figure G-3)

There are 4 waveforms on the screen because the measurement screen at the upper half of the screen displays the waveforms of the channel 1 over those of the channel 2.




Figure G-3. The Displayed Measurement Result

Required Parameters

Following parameters are required to execute this program.

- Parameters Required for the Trap Function (Channel 1)
 - The stop frequency for the trap function
 - The start frequency for the trap function (Input in the ratio (ppm) to the stop frequency for the trap function)
 - The IF bandwidth
 - The number of measured points
 - The phase value to trap at each measured point and data to output when the value matches
- Parameters Specified for a Frequency Sweep (Channel 2)
 - The frequency sweep span
 - The IF bandwidth
 - The number of measured points
 - The wait time interval between the completion of the trap function and the start of frequency sweep.

The Sample Program List

The list of a sample program 022 is explained below. This list contains the important portion only.

```

490  ASSIGN @E5100 TO 800
500  CLEAR @E5100
510  !
520  DIM Bin(1:20)
530  DIM Dat$[80],Lw$[20],Err$[50],Lmt_flag$(1:20)[5]
540  !
550  OUTPUT @E5100;"PRES"
560  GOSUB Setting
570  GOSUB Power_setup
580  GOSUB Set_ch1
590  GOSUB Pi_cal
600  CLEAR SCREEN
610  GOSUB Set_ch1_part2
620  OUTPUT @E5100;"COUC OFF"
630  GOSUB Setup_monitor
640  GOSUB Set_ch2
650  GOSUB Measurement
660  !
670  STOP
680  !
690  !
700  !
710  ! ***** SURBRoutines *****
720  !
730  ! ***** DEFAULT VALUES *****
740  !
750  Default_set: !
760  Z0=12.5
770  F_nominal=19.979 ! (MHz)
780  Span_adj=1000 ! (ppm)
790  Span_fr=100 ! (ppm)
800  Ifbw1=1000 ! (Hz)
810  Ifbw2=1000 ! (Hz)
820  Nop1=11 ! NOP for CH1
830  Nop2=51 ! NOP for CH2
840  Wait_check=100 ! WAIT TIME for Fr check after EVAPOLATION
850  RETURN
860  !
870  !
880  Default_power: !
890  W=10 ! POWER = 10 (uW)
900  R=10 ! CI = 10 (ohm)
910  RETURN
920  !
930  !
940  Lmt_flag: !
950  !
960  ! Lmt,Flag
970  ! -----
980  DATA "0,ON","0000000000000001" ! 1
990  DATA "0,OFF"
1000 DATA "0,OFF"
1010 DATA "0,OFF"
1020 DATA "0,ON","0000000000000010" ! 2
1030 DATA "0,OFF"
1040 DATA "0,OFF"
1050 DATA "0,ON","0000000000000100" ! 4
1060 DATA "0,OFF"
1070 DATA "0,OFF"
1080 DATA "0,ON","0000000000001000" ! 8

```

```

1090 !
1100 !
1110 !
1120 ! *****
1130 !
1140 Setting: !
1150 !
1160 GOSUB Default_set
1170 ! OUTPUT @E5100;"DSE 16384;DSNT 16384;*SRE 128"
      ! SET INTERRUPT for DATA TRANSFER between NA and EXT PC
1180 OUTPUT @E5100;"*CLS"
1190 OUTPUT @E5100;"DISAHIHB"
1200 OUTPUT @E5100;"CHAN1;HOLD"
1210 OUTPUT @E5100;"TRAP OFF"
1220 OUTPUT @E5100;"CHAN2;HOLD"
1230 OUTPUT @E5100;"TRAP OFF"
1240 OUTPUT @E5100;"CHAN1"
1250 OUTPUT @E5100;"DUAC ON"
1260 OUTPUT @E5100;"COUC ON"
1270 OUTPUT @E5100;"LINFREQ"
1280 INPUT "Nominal frequency (MHz) ?",F_nominal
1290 F_nominal=F_nominal*1.E+6
1300 INPUT "Adjustment Span (ppm) ?",Span_adj
1310 Span_adj=Span_adj/1.E+6
1320 Span_adj=F_nominal*Span_adj
1330 Startf=F_nominal
1340 Stopf=F_nominal+Span_adj
1350 INPUT "Span for Fr check (ppm) ?",Span_fr
1360 Span_fr=Span_fr/1.E+6
1370 Span_fr=F_nominal*Span_fr
1380 INPUT "IFBW at CH1 (Hz) ?",Ifbw1
1390 INPUT "IFBW at CH2 (Hz) ?",Ifbw2
1400 ! INPUT "NOP for EVAPOLATION at CH1 ? ",Nop1
1410 INPUT "NOP for Fr check at CH2 ? ",Nop2
1420 INPUT "WAIT TIME after EVAPOLATION (msec) ?",Wait_check
1430 Wait_check=Wait_check/1000
1440 EXECUTE "ANAOCH2"
1450 EXECUTE "ANAO DATA"
1460 EXECUTE "ANARFULL"
1470 RETURN
1480 !
1490 !
1500 Power_setup: !
1510 GOSUB Default_power
1520 INPUT "POWER (uW)",W
1530 W=W/1.E+6
1540 INPUT "RESONANT RESISTANCE (ohm) ",R
1550 !
1560 ! Power level calculation
1570 !
1580 Pi_r3=12.5
1590 R0=50
1600 R1=50
1610 R2=50
1620 R3=83.3
1630 R4=159
1640 R5=66.2
1650 R6=14.2
1660 Pi_r2=R6*(R3*R4+R5*(R3+R4))/(R3*R4+(R3+R4)*(R5+R6))
1670 V0_a=(Pi_r2+R+Pi_r3)*SQRT(W/R)
1680 Vp=(R3*R4+(R3+R4)*(R5+R6))/(R4*R6)*V0_a
1690 V0=(R0+R1+R2)/(R1+R2)*Vp
1700 V1=V0/4
1710 Power=10*LGT(1000*V1^2/50)
1720 RETURN
1730 !

```

```

1740 !
1750 Set_ch1:!
1760 OUTPUT @E5100;"CHAN1;AR;LOGMP"
1770 OUTPUT @E5100;"HOLD"
1780 OUTPUT @E5100;"POWE ";Power
1790 OUTPUT @E5100;"STAR ";Startf-(Span_fr)*(7/8) ! SET START FREQ for CAL
1800 OUTPUT @E5100;"STOP ";Stopf
1810 OUTPUT @E5100;"POIN ";Nop1
1820 OUTPUT @E5100;"IFBW ";Ifbw1
1830 OUTPUT @E5100;"DISG OFF"
1840 OUTPUT @E5100;"SWED DOWN"
1850 RETURN
1860 !
1870 !
1880 Set_ch1_part2:!
1890 OUTPUT @E5100;"STAR ";Startf
1900 RETURN
1910 !
1920 !
1930 Set_ch2:!
1940 OUTPUT @E5100;"CHAN2;AR;LOGMP"
1950 OUTPUT @E5100;"HOLD"
1960 ! OUTPUT @E5100;"POWE ";Power
1970 ! OUTPUT @E5100;"LINFREQ"
1980 OUTPUT @E5100;"STOP ";Startf+(Span_fr)*(1/8)
! SET STOP FREQ for CH2 based on START FREQ at CH1
1990 OUTPUT @E5100;"STAR ";Startf-(Span_fr)*(7/8)
2000 OUTPUT @E5100;"POIN ";Nop2
2010 OUTPUT @E5100;"IFBW ";Ifbw2
2020 OUTPUT @E5100;"DISG OFF"
2030 OUTPUT @E5100;"SETZ ";Z0
2040 OUTPUT @E5100;"CONVZTRA"
2050 RETURN
2060 !
2070 !
2080 ! ***** MONITOR FUNCTION SETUP *****
2090 !
2100 Setup_monitor:!          MONITOR function setip
2110 !
2120 Start_pt=1              ! Set Start Point
2130 Stop_pt=Nop1           ! Set Stop Point
2140 !
2150 OUTPUT @E5100;"WRIT16" ! Set the Number of PIN of I/O Port
2160 !
2170 RESTORE Lmt_flag
2180 !
2190 FOR I=Start_pt TO Stop_pt
2200     READ Lmt_flag$(I)
2210     OUTPUT @E5100;"INPUTRAC "&VAL$(I)&","&Lmt_flag$(I)
! Set I/O output Points
2220     IF Lmt_flag$(I)[3;2]="ON" THEN
2230         READ Lw$
2240         Bin(I)=DVAL(Lw$,2)
2250         OUTPUT @E5100;"INPUTRACB "&VAL$(I)&","1,";Bin(I)
! Set I/O output info
2260     ELSE
2270         OUTPUT @E5100;"INPUTRACB "&VAL$(I)&","1,0"
2280     END IF
2290 NEXT I
2300 !
2310 OUTPUT @E5100;"TRAR ";Start_pt,Stop_pt
2320 OUTPUT @E5100;"TRAFMEMO"
2330 OUTPUT @E5100;"TRAP ON"
2340 !
2350 RETURN
2360 !

```

```

2370 !
2380 ! ***** MEASUREMENT *****
2390 !
2400 Measurement: !
2410 !
2420 LOOP
2430 DISP "CONNECT DEVICE, and PRESS CONTINUE."
2440 PAUSE
2450 DISP "MEASURING"
2460 !
2470 OUTPUT @E5100;"UPDD OFF"
2480 OUTPUT @E5100;"CHAN1"
2490 OUTPUT @E5100;"UPDD ON"
2500 EXECUTE "SING"
2510 !
2520 WAIT Wait_check
2530 !
2540 OUTPUT @E5100;"UPDD OFF"
2550 OUTPUT @E5100;"CHAN2"
2560 OUTPUT @E5100;"UPDD ON"
2570 EXECUTE "SING"
2580 !
2590 GOSUB Analysis
2600 GOSUB Printing
2610 END LOOP
2620 !
2630 RETURN
2640 !
2650 ! *****
2660 !
2670 Analysis: !
2680 EXECUTE "OUTPRESO?"
2690 Ci=READIO(8,0)
2700 Fr=READIO(8,1)
2710 RETURN
2720 !
2730 !
2740 Printing: !
2750 CLEAR SCREEN
2760 PRINT USING "25A,4D.8D,6A";"RESONANT FREQUENCY ";Fr/1.E+6;" (MHz)"
2770 PRINT USING "25A,5X,6D.D,6A";"RESONANT RESISTANCE ";Ci;" (ohm)"
2780 RETURN
2790 !
2800 !
2810 !
2820 ! ***** CALIBRATION *****
2830 !
2840 Pi_cal: !
2850 !
2860 GOSUB Declarations
2870 GOSUB Setup_pi
2880 GOSUB Ckt_const_pi
2890 GOSUB Modify_calkit
2900 GOSUB Meas_pi
2910 RETURN
2920 !
2930 Declarations: !
2940 INTEGER Yes_pi,No_pi,Ans_pi
2950 Yes_pi=1
2960 No_pi=0
2970 INTEGER Open_pi,Short_pi,Load_pi
2980 Open_pi=1
2990 Short_pi=2
3000 Load_pi=3
3010 Maxstd_pi=3
3020 INTEGER RO_pi,C0_pi,LO_pi,Fc_pi

```

```

3030     RO_pi=1
3040     CO_pi=2
3050     LO_pi=3
3060     Fc_pi=4
3070     Maxele_pi=4
3080     INTEGER Real_pi,Imag_pi
3090     Real_pi=1
3100     Imag_pi=2
3110     File_pi$="PI_DATA"
3120     INTEGER Nop_pi
3130     REAL Const_pi(1:3,1:4)! (STD TYPE, L C R Fc)
3140     !
3150     RETURN
3160     !
3170     !*****
3180     ! SET UP INSTRUMENT
3190     !*****
3200 Setup_pi: !
3210     OUTPUT @E5100;"DISAHIB"
3220     RETURN
3230     !
3240     !*****
3250     ! EQUIVALENT CIRCUIT CONSTANTS
3260     !*****
3270 Ckt_const_pi: !
3280     GOSUB Init_const_pi
3290     Modify_pi=No_pi
3300     LOOP
3310     PRINT "CONSTANTS"
3320     PRINT ""
3330     PRINT "OPEN CO",Const_pi(Open_pi,CO_pi);"pF"
3340     PRINT "SHORT RO",Const_pi(Short_pi,RO_pi);"Ohm"
3350     PRINT "SHORT LO",Const_pi(Short_pi,LO_pi);"nH"
3360     PRINT "LOAD RO",Const_pi(Load_pi,RO_pi);"Ohm"
3370     PRINT "LOAD LO",Const_pi(Load_pi,LO_pi);"nH"
3380     !
3390     Ans_pi=No_pi
3400     INPUT "MODIFY ? (Yes=1,No=0,DEFAULT=0)",Ans_pi
3410     EXIT IF Ans_piYes_pi
3420     !
3430     !
3440     ! MODIFY
3450     !
3460     Modify_pi=Yes_pi
3470     INPUT "OPEN CO (pF)",Const_pi(Open_pi,CO_pi)
3480     INPUT "SHORT RO (Ohm)",Const_pi(Short_pi,RO_pi)
3490     INPUT "SHORT LO (nH)",Const_pi(Short_pi,LO_pi)
3500     INPUT "LOAD RO (Ohm)",Const_pi(Load_pi,RO_pi)
3510     INPUT "LOAD LO (nH)",Const_pi(Load_pi,LO_pi)
3520     !
3530     END LOOP
3540     !
3550     IF Modify_pi THEN GOSUB Backup_pi
3560     !
3570     ! UNIT CONVERSION
3580     !
3590     ! pF --> F
3600     Const_pi(Open_pi,CO_pi)=Const_pi(Open_pi,CO_pi)*1.E-12
3610     ! nH --> H
3620     Const_pi(Short_pi,LO_pi)=Const_pi(Short_pi,LO_pi)*1.E-9
3630     Const_pi(Load_pi,LO_pi)=Const_pi(Load_pi,LO_pi)*1.E-9
3640     !
3650     ! NORMALIZE BY CHARACTERISTIC IMPEDANCE
3660     !
3670     REAL ZO_pi
3680     OUTPUT @E5100;"SETZ?"

```

```

3690 ENTER @E5100;Z0_pi
3700 !
3710 Const_pi(Open_pi,CO_pi)=Const_pi(Open_pi,CO_pi)*Z0_pi
3720 Const_pi(Short_pi,CO_pi)=Const_pi(Short_pi,CO_pi)/Z0_pi
3730 Const_pi(Short_pi,CO_pi)=Const_pi(Short_pi,CO_pi)/Z0_pi
3740 Const_pi(Load_pi,CO_pi)=Const_pi(Load_pi,CO_pi)/Z0_pi
3750 Const_pi(Load_pi,CO_pi)=Const_pi(Load_pi,CO_pi)/Z0_pi
3760 !
3770 RETURN
3780 !
3790 Init_const_pi: !
3800 ON ERROR GOTO Cannot_open_pi
3810 !
3820 ! FROM BACKUP DATA FILE
3830 !
3840 ASSIGN @File_pi TO File_pi$
3850 ENTER @File_pi;Const_pi(*)
3860 ASSIGN @File_pi TO *
3870 OFF ERROR
3880 GOTO Getdata_end_pi
3890 Cannot_open_pi: !
3900 OFF ERROR
3910 DISP "FILE NOT FOUND. USE CURRENT SETTING."
3920 File_notfound: !
3930 LOOP
3940 ON KEY 1 LABEL " E5100A DEFAULT" GOTO E5100a_data
3950 ON KEY 5 LABEL " 87510A DEFAULT" GOTO Hp87510a_data
3960 END LOOP
3970 !
3980 E5100a_data: !
3990 OUTPUT @E5100;"CALKOC?"
4000 ENTER @E5100;Const_pi(Open_pi,CO_pi)
4010 OUTPUT @E5100;"CALKSRS?"
4020 ENTER @E5100;Const_pi(Short_pi,RO_pi)
4030 OUTPUT @E5100;"CALKSLS?"
4040 ENTER @E5100;Const_pi(Short_pi,LO_pi)
4050 OUTPUT @E5100;"CALKLRS?"
4060 ENTER @E5100;Const_pi(Load_pi,RO_pi)
4070 OUTPUT @E5100;"CALKLLS?"
4080 ENTER @E5100;Const_pi(Load_pi,LO_pi)
4090 RETURN
4100 !
4110 Hp87510a_data: !
4120 RETURN
4130 !
4140 !*****
4150 ! RE-SAVE BACKUP DATA FILE
4160 !*****
4170 Backup_pi: !
4180 Ans_pi=Yes_pi
4190 INPUT "SAVE THOSE DATA ? (YES=1,NO=0,DEFAULT=1)",Ans_pi
4200 IF Ans_pi=No_pi THEN RETURN
4210 !
4220 ON ERROR GOSUB Createfile_pi
4230 ASSIGN @File_pi TO File_pi$
4240 OFF ERROR
4250 !
4260 ON ERROR GOTO Create_err_pi
4270 OUTPUT @File_pi;Const_pi(*)
4280 OUTPUT @File_pi;END
4290 ASSIGN @File_pi TO *
4300 OFF ERROR
4310 GOTO Backup_end_pi
4320 Create_err_pi: !
4330 OFF ERROR
4340 PRINT "CAN'T CREATE BAKC-UP DATA FILE"

```

```

4350 Backup_end_pi: !
4360 RETURN
4370 !
4380 Createfile_pi: !
4390 ! REAL IS 8 BYTES
4400 CREATE BDAT File_pi$,Maxstd_pi*Maxele_pi,8
4410 RETURN
4420 !
4430 !*****
4440 ! MODIFY CAL KIT
4450 !*****
4460 Modify_calkit: !
4470 IF Modify_pi THEN
4480 PRINT "MODIFYING..."
4490 OUTPUT @E5100;"CALKQCP ";Const_pi(Open_pi,CO_pi)
4500 OUTPUT @E5100;"CALKSRS ";Const_pi(Short_pi,RO_pi)
4510 OUTPUT @E5100;"CALKSLS ";Const_pi(Short_pi,LO_pi)
4520 OUTPUT @E5100;"CALKLRs ";Const_pi(Load_pi,RO_pi)
4530 OUTPUT @E5100;"CALKLLS ";Const_pi(Load_pi,LO_pi)
4540 END IF
4550 DISP
4560 RETURN
4570 !
4580 !*****
4590 ! MEASURE SEQUENTIAL
4600 !*****
4610 Meas_pi: !
4620 !
4630 OUTPUT @E5100;"CALIS111"
4640 Onkey_loop: !
4650 LOOP
4660 IF Open_done_pi THEN
4670 ON KEY 1 LABEL " ((OPEN))" GOTO Meas_open_pi
4680 ELSE
4690 ON KEY 1 LABEL " OPEN " GOTO Meas_open_pi
4700 END IF
4710 IF Short_done_pi THEN
4720 ON KEY 2 LABEL " ((SHORT))" GOTO Meas_short_pi
4730 ELSE
4740 ON KEY 2 LABEL " SHORT " GOTO Meas_short_pi
4750 END IF
4760 IF Load_done_pi THEN
4770 ON KEY 3 LABEL " ((LOAD))" GOTO Meas_load_pi
4780 ELSE
4790 ON KEY 3 LABEL " LOAD " GOTO Meas_load_pi
4800 END IF
4810 !
4820 ON KEY 5 LABEL " CAL BREAK" GOSUB Break_pi
4830 !
4840 All_done_pi=Open_done_pi*Short_done_pi*Load_done_pi
4850 IF All_done_pi THEN
4860 ON KEY 8 LABEL " DONE:3 TERM CAL" GOTO Meas_end_pi
4870 DISP "PRESS 'DONE' IF FINISHED WITH CAL."
4880 ELSE
4890 OFF KEY 8
4900 DISP "CONNECT STD THEN PRESS KEY TO MEASURE."
4910 END IF
4920 !
4930 END LOOP
4940 !
4950 Meas_open_pi: !
4960 OUTPUT @E5100;"CLASS11A;*OPC?"
4970 ENTER @E5100;Opc
4980 Open_done_pi=1
4990 GOTO Onkey_loop
5000 !

```

```
5010 Meas_short_pi: !
5020 OUTPUT @E5100;"CLASS11B;*OPC?"
5030 ENTER @E5100;Qpc
5040 Short_done_pi=1
5050 GOTO Onkey_loop
5060 !
5070 Meas_load_pi: !
5080 OUTPUT @E5100;"CLASS11C;*OPC?"
5090 ENTER @E5100;Qpc
5100 Load_done_pi=1
5110 GOTO Onkey_loop
5120 !
5130 Break_pi: !
5140 DISP "CAL BREAK."
5150 GOSUB Quit_pi
5160 STOP
5170 !
5180 Meas_end_pi: !
5190 OFF KEY
5200 BEEP 500,.3
5210 OUTPUT @E5100;"SAV1"
5220 DISP ""
5230 RETURN
5240 !
5250 !*****
5260 ! EXIT PROCEDURE
5270 !*****
5280 Quit_pi: !
5290 OFF KEY
5300 BEEP 500,.2
5310 RETURN
5320 !
5330 END
```

HP-IB Commands of Option 022

BINSIZE□<value>

Specify the number of continuous data outputted to the I/O port by INPUTRACB.

Specify the maximum value in the second parameter of INPUTRACB. (Option 022 only)

Parameter Range	0 to 6
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"BINSIZE 2" OUTPUT @E5100;"BINSIZE7" ENTER @E5100;A</pre>

INPUTRAC□<value1>,<value2>,<value3>

Pass the phase value to trap at a measured point and specify whether data is outputted to the I/O port or not when the measured value reaches the phase value. The data to be outputted to the I/O port is specified by INPUTRACB. (No Query,Option 022 only)

Parameter Description	<p><value1> : The measured point number <value2> : The trap phase value <value3> : I/O output ON/OFF</p>
Examples	<pre>OUTPUT @E5100;"INPUTRAC 10, 0, ON"</pre>


INPUTRACB□<value1>,<value2>,<value3>

Pass the data outputted to the I/O port when the measured value of phase reaches the limit value specified by the INPUTRAC command. Multiple data up to 6 data can be outputted continuously to the I/O port. (No Query,Option 022 only)

Parameter Description	<p><value1> : The measured point number <value2> : The order of output I/O data (1 to 6) <value3> : Data Query Response</p>
Examples	<pre>OUTPUT @E5100;"INPUTRACB 5, 1, ";DVAL("1010",2) OUTPUT @E5100;"INPUTRACB 5, 2, ";DVAL("0100",2) OUTPUT @E5100;"INPUTRACB 5, 3, ";DVAL("0000",2)</pre>

If HP E5100A/B OPT. 022 is set up by the above commands, the data described in the following figure will be outputted to the I/O port when the measured phase reaches the target value at the measured point 5.

INPUTRACB \square <value1>, <value2>, <value3>



C40S2001

Figure G-4. The Data Output Timing Chart

MEAS1PT? <value>

Outputs measurement value at the point number specified by the parameter. When the trap function is turn on (TRAP ON), MEAS1PT waits to start a measurement until the phase value is in the condition specified by INPUTRAC command, and returns the query response.

Parameter Range	1 to NOP
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"MEAS1PT? 10000000" OUTPUT @E5100;"MEAS1PT?" ENTER @E5100;A</pre>

OUTPTRAC? \square <value>

Outputs the phase value to trap at a measured point and the setting of I/O port at the point specified by the parameter. The phase value and the I/O port setting are specified by INPUTRAC command. (Query only, Option 022 only)

Parameter Description	{value} : measurement point number
Query Response	{value} {0 1} {value} : Phase value {1 0} : I/O port setting (ON/OFF)
Examples	<pre>OUTPUT @E5100;"OUTPTRAC? 10" ENTER @E5100;A,B</pre>

OUTPTRACB?□<value1>,<value2>

Outputs the data outputted to the I/O port when the measured value of phase reaches the limit value. The data outputted to the I/O port is specified by INPUTRACB command. (Query only, Option 022 only)

Parameter Description	<value1> : Measurement point number <value2> : Output port number (1 to 6)
Query Response	{value} {value} : Data
Examples	OUTPUT @E5100;"INPUTRACB? 10,1" ENTER @E5100;A

TIMO□{ON|OFF|0|1}

Sets the time limit for the trapping on/off. (Option 022 only)

Parameter Description	ON or 1 : ON OFF or 0 : OFF
Query Response	{1 0}
Examples	OUTPUT @E5100;"TIMO ON" OUTPUT @E5100;"TIMO?" ENTER @E5100;A

TOTIME□<value>

Sets the limit time for the trapping. (Option 022 only)

Parameter Range	0 to
Query Response	{value} (ms)
Examples	OUTPUT @E5100;"TOTIME 2000" OUTPUT @E5100;"TOTIME?" ENTER @E5100;A

TRABGE

Sets the condition so that the measurement is advanced to the next point when a measurement value (phase value) is greater than or equal to the threshold value specified by the INPUTRAC command. (Option 022 only)

TRABGE

Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"TRABGE" OUTPUT @E5100;"TRABGE?" ENTER @E5100;A</pre>

TRABLE

Sets the condition so that the measurement is advanced to the next point when a measurement value (phase value) is less than or equal to the threshold value specified by the INPUTRAC command. (Option 022 only)

Query Response	{1 0}
Examples	<pre>OUTPUT @E5100;"TRABLE" OUTPUT @E5100;"TRABLE?" ENTER @E5100;A</pre>

TRAFDATA

Set the trap function on the data trace. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"TRAFDATA" OUTPUT @E5100;"TRAFDATA?" ENTER @E5100;A</pre>

TRAFMEMO

Set the trap function on the sub trace. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"TRAFMEMO" OUTPUT @E5100;"TRAFMEMO?" ENTER @E5100;A</pre>

TRAP □ {OFF|ON|0|1}

Set the trap function on/off. (Option 022 only)

Parameter Description	OFF or 0 : Set the trap function off. ON or 1 : Set the trap function on.
Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"TRAP ON" OUTPUT @E5100;"TRAP?" ENTER @E5100;A</pre>

TRAR □ <value1>, <value2>

Set the start and the end points for the partial sweep for the trap function.

If the start point and the end point are not between 1 to the value specified by **Number of Point**, HP E5100A/B option 022 will perform a sweep within the possible range. Use this function to reserve wider sweep range so that you can change the sweep range by simply changing this setup. This way, you can save more measurement time than changing the whole sweep setup. (Option 022 only)

Parameter Description	<value1> : The start point for the partial sweep <value2> : The end point for the partial sweep
Query Response	{value1}{value2}
Examples	<pre>OUTPUT @E5100;"TRAR 10, 20" OUTPUT @E5100;"TRAR?" ENTER @E5100;A,B</pre>

WRIT16

Set the bit width of the data outputted to the I/O port while performing the trap function to 16 bit. The port F is used. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"WRIT16" OUTPUT @E5100;"WRIT16?" ENTER @E5100;A</pre>

WRIT24

Set the bit width of the data outputted to the I/O port while performing the trap function to 24 bit. The port H is used. (Option 022 only)

Query Response	{0 1}
Examples	<pre>OUTPUT @E5100;"WRIT24" OUTPUT @E5100;"WRIT24?" ENTER @E5100;A</pre>


H

Option 023

Overview

This option enables you to measure the drive level characteristics of the resonant frequency (F_r) and the resonant impedance (CI) of crystal resonators quickly and precisely.

To realize a quick measurement of resonant characteristics, this option adopts the Phase Tracking. The phase tracking will control the source signal to trace a specific resonant phase value (generally 0°) to calculate the frequency and the impedance at the phase. The drive level characteristic measurement, the function's most typical usage, realizes a very quick measurement because it measures only the resonant frequency and the resonant impedance as it sweeps the drive level. HP E5100A option 023 will display the result as described in Figure H-1 where X-axis representing drive level and Y-axis representing resonant frequency and resonant impedance. Instead of an absolute value, a relative value based on the resonant frequency at the minimum level or the nominal resonant frequency is used as the resonant frequency.



G40S3004

Figure H-1. Measuring Drive Level Characteristics

You can also measure the aging characteristics of the resonant frequency and the resonant impedance by setting the drive level fixed and performing the phase tracking. One application of this aging characteristic measurement may be that you can change the temperature around a resonator with time to measure characteristics of temperature. Instead of an absolute value, a

relative value based on the resonant frequency at the minimum level or the nominal resonant frequency is used as the resonant frequency.




Figure H-2. The Aging Characteristic Measurement

Measuring Drive Level Characteristics using Phase Tracking Function

You must use a program to use the phase tracking to measure drive level characteristics or aging characteristics (the temperature characteristic measurement). Sample programs are provided in the sample program disk. This chapter explains how to measure these characteristics using a sample program.

The following procedure shows how to set and measure the drive level characteristics using the phase tracking function. HP-IB command are shown in parentheses.

1. Measurement Settings for Phase Tracking

- Select the ratio measurement (MEAS AR)
- Select the impedance measurement mode (ANAMODE ZTRAN)
- Select the power sweep (SWPT POWE)
- Select the order base display (LISDOBASE)
- Select Z-ΔF format (FMT MAGZDF)

2. Setting for Using π Network Test Fixture

The following settings are required when a π Network Test Fixture is used.

- Set Z0 (SETZ)
- Select Watt as the power unit (POWU WATT)
- Select using π network test fixture (PICIRC ON)
- Set CI value (CIVAL)


In the sample program, Characteristic impedance is set at 12.5Ω to support a standard π test fixture.

3. Settings for the Drive Level Measurement

The following settings are required for measuring the drive level characteristics.

- Make a drive level table and input it as the sweep table. (INPUSTIM)

The result will be similar to Figure G-3 because both the up sweep and the down sweep measurements are performed at the same time for a drive level.



G40S3003

Figure H-3. An Example of the Measurement Result

Note



Set enough power (For example, SPAN=0 dBm, CENTER=0 dBm) and set the CW frequency as you measure DUT when you execute calibration measurement.

If the calibration is measured under the low power level, measurement results are not stable.

4. Settings Parameters for Tracking Function

The following process is required for the tracking function.

- Set the start frequency for tracking (PTFR, unit:Hz)
- Set the range from the start frequency for tracking to search the resonant point F_r (PTFRSR, unit:Hz)
- Set the range value for a phase (PTTRGLMT, unit:deg.)
- Set the measurement abort ON/OFF when the phase is over the limit (PTABORT)
- Define the phase at the resonant point. (PTTRGPHS)
- Set the number of tracking for each measurement point (PTREPN)

5. Searching for Resonant Point

HP E5100A/B searches the range specified by PTFRSR command for the resonant point, which is defined by PTTRGHS. When the resonant point is measured, HP E5100A/B automatically sets the test signal level to the same level as START and uses the calibration data at the start point. Then Get F_r , CI, and the tracking parameter at the resonant point. (SRCHFR?)

6. Tracking Measurement

To start the tracking, the following procedure is required.

- Set the phase tracking ON (PTRACK)
- Set the reference frequency (PTFR)
- Set the tracking parameter given by the SRCHFR? query response (PTPARA)
- Check the status of phase tracking (PTSTAT?)


7. The Measurement Result

This program gives you the following result.

Figure G-2 describes the result screen. The upper half of the screen shows the measurement and the lower half shows the resonant frequency at the minimum drive level and resonant frequencies and resonant impedance of typical drive levels (in [W]) are displayed.

Values along the horizontal axis on the screen (drive levels) are displayed in the unit specified by the marker. The data trace indicates resonant impedance and the sub trace indicates resonant frequency (Δf).

When HP E5100A option 023 displays a resonant frequency (Δf), it displays the value relative to the reference frequency specified by PTFR command.



C72H001

Figure H-4. A Displayed Result

Sample Program List

```

10 CLEAR SCREEN
20 ! MAIN *****
30 !
40 GOSUB Constants
50 GOSUB Setting
60 GOSUB Cal_setting
70 GOSUB Pi_cal
80 CLEAR SCREEN
90 GOSUB Set_watt
100 GOSUB Set_watt_list
110 GOSUB Set_phase_track
120 GOSUB Start_loop
130 STOP
140 !
150 ! DECLARE CONSTANTS *****
160 Constants: !
170 Z0=12.5
180 Ci=24
190 W_min=.001 !uW
200 W_max=1000 !uW
210 Ifbw=1000
220 Tgt_phas=0
230 Track_num=2
240 P_nop=100
250 Tgt_limit=8
260 !
270 Srch_rng=2000 !ppm
280 Srch_mode=1
290 Srch_wait=0
300 Fr_trg=6.E+7
310 RETURN
320 !

```

```

330 ! INSTRUMENT SETUP *****
340 Setting: !
350 ASSIGN @E5100 TO 800
360 OUTPUT @E5100;"CHAN1;MEAS AR;ANAMODE ZTRAN"
370 OUTPUT @E5100;"HOLD"
380 OUTPUT @E5100;"COUC OFF"
390 OUTPUT @E5100;"IFBW ";Ifbw
400 OUTPUT @E5100;"POIN ";P_nop*2
410 OUTPUT @E5100;"FMT MAGZDF"
420 OUTPUT @E5100;"MULC OFF;SPLD ON;DISAHIHB"
430 RETURN
440 !
450 Cal_setting: !
460 OUTPUT @E5100;"SWPT LINP"
470 OUTPUT @E5100;"CENT ";Fr_trg
480 OUTPUT @E5100;"SPAN ";Srch_rng*1.E-6*Fr_trg
490 OUTPUT @E5100;"POWE ";0
500 RETURN
510 !
520 Pi_cal: !
530 PRINT TABXY(0,11);"PERFORM PI-CAL"
540 INPUT "CONNECT OPEN, THEN PRESS ENTER",Tmp$
550 OUTPUT @E5100;"CALI ONEP;CLASS11A?"
560 ENTER @E5100;Tmp
570 INPUT "CONNECT SHORT, THEN PRESS ENTER",Tmp$
580 OUTPUT @E5100;"CLASS11B?"
590 ENTER @E5100;Tmp
600 INPUT "CONNECT LOAD, THEN PRESS ENTER",Tmp$
610 OUTPUT @E5100;"CLASS11C?"
620 ENTER @E5100;Tmp
630 OUTPUT @E5100;"SAV1?"
640 ENTER @E5100;Tmp
650 INPUT "CONNECT DEVICE, THEN PRESS ENTER",Tmp$
660 RETURN
670 !
680 Set_watt: !
690 OUTPUT @E5100;"SWPT POWE"
700 OUTPUT @E5100;"LISDOBASE;SWED UP"
710 OUTPUT @E5100;"ATTW .002"
720 OUTPUT @E5100;"SETZ ";Z0
730 OUTPUT @E5100;"POWU WATT;PICIRC ON"
740 OUTPUT @E5100;"CIVAL ";Ci
750 RETURN
760 !
770 ! SETTING POWER LIST (WATT) *****
780 Set_watt_list: !
790 ALLOCATE Pwr(1:2*P_nop)
800 P_min=W_min*1.E-6
810 P_max=W_max*1.E-6
820 OUTPUT @E5100;"STAS ";P_min,P_max
830 P_step=(P_max-P_min)/(P_nop-1)
840 K=2*P_nop
850 FOR I=1 TO P_nop
860 P=P_min+P_step*(I-1)
870 Pwr(I)=P
880 Pwr(K)=P
890 K=K-1
900 NEXT I
910 OUTPUT @E5100;"STIM";Pwr(*)
920 RETURN
930 !
940 ! PHASE TRACKING SETTING *****
950 Set_phase_track: !
960 OUTPUT @E5100;"PTFR ";Fr_trg!
970 OUTPUT @E5100;"PTFRSR ";Srch_rng*Fr_trg*1.E-6
980 OUTPUT @E5100;"PTTRGLMT ";Tgt_limit

```

```

990  OUTPUT @E5100;"PTABORT ON"
1000 OUTPUT @E5100;"PTTRGPHS ";Tgt_phase
1010 OUTPUT @E5100;"PTREPN ";Track_num
1020 OUTPUT @E5100;"PTRACK ON"! TRACKING MODE ON
1030 RETURN
1040 !
1050 ! SEARCH "ACTUAL" FR *****
1060 Start_loop: !
1070 LOOP
1080 BEEP
1090 OUTPUT @E5100;"PTFR ";Fr_trg!
1100 OUTPUT @E5100;"SRCHFR? ";Srch_mode,Srch_wait
1110 ENTER @E5100;Srched_fr,Ci,Pt_param
1120 IF Srched_fr=-1 THEN
1130 PRINT TABXY(0,10),"Fr Search FAILED "
1140 BEEP
1150 BEEP
1160 GOTO Prompt
1170 END IF
1180 !
1190 ! DLD CHARACTERISTICS MEASUREMENT *****
1200 !
1210 OUTPUT @E5100;"PTFR ";Srched_fr
1220 OUTPUT @E5100;"PTPARA ";Pt_param
1230 OUTPUT @E5100;"SING?"
1240 ENTER @E5100;Dummy
1250 !
1260 OUTPUT @E5100;"PTSTAT?"
1270 ENTER @E5100;Result
1280 IF Result=0 THEN
1290 BEEP
1300 BEEP
1310 PRINT TABXY(0,10),"Phase Track FAILED"
1320 GOTO Prompt
1330 END IF
1340 !
1350 GOSUB Analysis
1360 !
1370 GOSUB Printing
1380 !
1390 Prompt: !
1400 INPUT "CONNECT NEXT DEVICE and Press Continue.",Tmp$
1410 !
1420 END LOOP
1430 !
1440 Analysis: ! DATA ANALYSIS
1450 OUTPUT @E5100;"ANARFULL"
1460 OUTPUT @E5100;"ANAOMEMO"
1470 OUTPUT @E5100;"OUTPMINMAX?"
1480 ENTER @E5100;Min_df,Mindf_p,Max_df,Maxdf_p
1490 OUTPUT @E5100;"ANAODATA"
1500 OUTPUT @E5100;"OUTPMINMAX?"
1510 ENTER @E5100;Min_z,Min_zp,Max_z,Max_zp
1520 RETURN
1530 !
1540 Printing: ! PRINTING ROUTINE
1550 PRINT TABXY(0,1)
1560 PRINT USING "27A,5X,3D.9D,X,5A";"RESONANT FREQ at MIN power",Srched_fr*
1.E-6,"(MHz)"
1570 PRINT USING "11A,2X,6A,X,4Z.2D,X,6A,3X,6A,X,4Z.3D,X,5A";"DELTA FREQ:","
[MIN]",Min_df,"(Hz)","[MAX]",Max_df,"(Hz)"
1580 PRINT USING "X,10A,2X,5A,2X,3Z.2D,X,5A,5X,5A,2X,3Z.2D,X,5A";"IMPEDANCE:
","[MIN]",Min_z,"(ohm)","[MAX]",Max_z,"(ohm)"
1590 !
1600 PRINT TABXY(0,6)
1610 PRINT USING "5A,X,4A,4X,10A,X,5A,2X,20A,X,4A";"POWER","(uW)","RESISTANC

```

```
E","(ohm)","FREQUENCY DIFFERENCE","(Hz)"
1620 FOR I=1 TO 3
1630   Point=10^(I-1)
1640   OUTPUT @E5100;"OUTPFORMP? ";Point
1650   ENTER @E5100;Z,Df
1660   PRINT USING "X,4D.4D,11X,3D.2D,12X,4D.3D";Pwr(Point)*1.E+6,Z,Df
1670 NEXT I
1680 PRINT TABXY(0,10),"
1690 !
1700 RETURN
1710 !
1720 END
```

Measuring Aging Characteristics

You can measure the aging characteristics by setting the drive level fixed and performing the phase tracking. You can also read out the time value using the marker time mode. Basically, the measurement setting is almost same as that of the drive level characteristics measurement except for the signal level is constant.

Compensation of Sweep Time for Aging Characteristics

The actual measurement time contents not only the sweep time but also the overhead time to process signal during measurement. PTFOVHD adds the overhead time to the marker time value as to compensate the measurement time value.

The following list shows how to use PTFOVHD command.

Sub-routine for Measuring Aging Characteristics

```

770 Meas_time: !
780  OUTPUT @E5100;"PTRACK ON"
790  OUTPUT @E5100;"SWETAUTO"
800  T1=TIMEDATE
810  OUTPUT @E5100;"SING?"
820  ENTER @E5100;Dummy
830  Meas_time=TIMEDATE-T1
840  OUTPUT @E5100;"SWET?"
850  ENTER @E5100;Anal_sweep_t
860  Pt_ovhd=((Meas_time-.005)/Track_num-Anal_sweep_t)/Nop
870  OUTPUT @E5100;"PTFOVHD ";Pt_ovhd
880  OUTPUT @E5100;"MARKTIME ON"
890  RETURN

```

Notes on the Phase Tracking

When the Phase Tracking Fails

Followings are problems and solutions caused during the phase tracking

- The resonator cannot keep up with a sweep because there are not enough drive level measurement points.

Solution	Increase the number of measurement points and repeat a measurement Increase the number of tracking at each point (PTREPN)
----------	--
- The actual resonant frequency is out of the sweep span.

Solution	Check the resonant frequency and repeat a measurement.
----------	--
- The resonator cannot keep up with a sweep because it is too fast.

Solution	Set the sweep time slower on the front panel and repeat a measurement. Increase the number of tracking at each point (PTREPN)
----------	--
- A tangent response is observed during a sweep.

Solution	Set the attenuator switching time slower and repeat a measurement.
----------	--
- HP E5100A can not find the resonance point (F_r)

Solution	Change the search mode of SRCHFR? to 3 or 4.
----------	--

Increase the number of parameter of PTFRSR.

The above list shows the possible causes of problems. You may want to investigate the cause by setting the display to "Z-Phase" to check at which point tracking exceeded the range and failed.

Limitations on the Phase Tracking

The following is a list of limitations on a measurement when the phase tracking is ON (PTRACK ON).

- The time specified by the SWEEP TIME and the actual sweep time differ.

HP-IB Commands for Option 023

FMTLMAGZDF

Sets the format as Z-Δ format for tracking measurement. (Option 023 only)

Examples

```
OUTPUT @E5100;"FMT MAGZDF"
```

PTABORTL{ON|OFF}

Sets the measurement abort ON/OFF when the phase value is over the limit during the tracking. (Option 023 only)

Parameter Description	ON : Abort a measurement when the tracking is failed OFF : Continue a measurement even the tracking is failed
Query Response	{1 0}
Examples	OUTPUT @E5100;"PTABORT ON" OUTPUT @E5100;"PTABORT?" ENTER @E5100;A

PTFOVHDL<value>

Input the parameters required to display the time base at the temperature characteristic measurement (the aging measurement). Refer to "Compensation of Sweep Time for Aging Characteristics" for inputting. (Option 023 only)

Parameter Range	0 to 1 sec
Query Response	{value}

PTFR□<value>

Input the start frequency for tracking. (Option 023 only)

Parameter Range	10 kHz to 300 MHz
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"PTFR 199 kHz" OUTPUT @E5100;"PTFR?" ENTER @E5100;A</pre>

PTFRSR□<value>

Sets the range for searching for F_r . (Option 023 only)

Parameter Range	0 Hz to 100 kHz
Query Response	{value} (Hz)
Examples	<pre>OUTPUT @E5100;"PTFRSR 5000" OUTPUT @E5100;"PTFRSR?" ENTER @E5100;A</pre>

PTPARA□<value>

Sets the tracking parameter. The tracking parameter is given by the SRCHFR? command query. (Option 023 only)

Parameter Range	-1000 to 1000
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"SRCHFR?" ENTER @E5100;Fr,Ci,Param OUTPUT @E5100;"PTPARA";Param OUTPUT @E5100;"PTPARA?" ENTER @E5100;A</pre>

PTRACK□{OFF|ON}

Set the phase tracking ON/OFF. (Option 023 only)

PTRACK \square {OFF|ON}

Parameter Description	OFF : The phase tracking is OFF ON : The phase tracking is ON
Query Response	{0 1}
Examples	OUTPUT @E5100;"PTRACK ON" OUTPUT @E5100;"PTRACK?" ENTER @E5100;A

PTREPN \square <value>

Sets the number of tracking on each point. (Option 023 only)

Parameter Range	1 to 1,000,000
Query Response	{value}
Examples	OUTPUT @E5100;"PTREPN 5" OUTPUT @E5100;"PTREPN?" ENTER @E5100;A

PTSTAT?

Returns the status of the phase tracking. (Query only, Option 023 only)

Query Response	{0 1} 0 : Error encountered during phase tracking 1 : The phase tracking was successful.
Examples	OUTPUT @E5100;"PTSTAT?" ENTER @E5100;A

PTTRGLMT \square <value>

Defines the range value for a phase, which is used for the phase tracking. (Option 023 only)

Parameter Range	0° to 180°
Query Response	{value}
Examples	OUTPUT @E5100;"PTTRGLMT 8" OUTPUT @E5100;"PTTRGLMT?" ENTER @E5100;A

PTTRGPHS□<value>

Defines the phase at the resonant point. (Option 023 only)

Parameter Range	-180° to 180°
Query Response	{value}
Examples	<pre>OUTPUT @E5100;"PTTRGPHS 0" OUTPUT @E5100;"PTTRGPHS?" ENTER @E5100;A</pre>

SRCHFR?□{1|2|3|4},<value>

Search for the resonance frequency (F_r). (Query only, Option 023 only)

Parameter Description	<p>Search mode:</p> <p>1 : Rough (High speed)</p> <p>2 : Normal</p> <p>3 : Fine</p> <p>4 : Finer (Slow)</p> <p><value> : Waiting time during searching (sec)</p>
Query Response	<p>{value1}, {value2}, {value3}</p> <p>{value1} : F_r [Hz]</p> <p>{value2} : CI [Ω]</p> <p>{value3} : Tracking parameter</p>
Examples	<pre>OUTPUT @E5100;"SRCHFR? 2,0" ENTER @E5100;A,B,C</pre>

Index

A

ABORT , 1-4
 accessing calibration coefficient arrays , 3-5
 active channel , 2-3
 active controller , 1-4
 ASCII data transfer , 3-7

B

binary data transfer , 3-8

C

calibration coefficient , 3-5
 clearing register , 5-10
 *CLS , 5-10
 condition register , 5-3
 controller , 1-4

D

data array , 3-4
 data arrays , 3-1
 data header , 3-9
 data processing flow , 3-1
 data transfer methods , 3-7
 device selector , 1-3, 1-5
 display allocation , 1-6
 DOS , 1-8

E

enable register , 5-2
 event register , 5-2
 external controller , 1-7

F

formatted data array , 3-5

G

general status register model , 5-1
 getting data from the analyzer , 3-10

H

HP BASIC , 1-9
 HP-IB , 1-4, 1-9
 HP-IB address
 setting , 1-6, 1-7
 HP-IB address , 1-5
 HP-IB cable , 1-7

HP-IB command , 2-1
 HP-IB control , 2-1
 HP Instrument BASIC , 1-5

I

IEEE 32-bit floating point format , 3-8
 IEEE 488 bus , 1-4
 IEEE 64-bit floating point format , 3-8
 initializing module , 1-2
 Instrument BASIC
 preparing , 1-6
 Instrument BASIC , 1-6
 Instrument BASIC editor , 1-8
 Interrupt , 5-10

L

LIF , 1-8

M

manual changes , A-1
 manual overview , 1-1
 memory trace , 3-6
 module , 1-2
 MS-DOS format , 3-9
 MSS , 5-2

O

option UKR , 1-6
 OUTPUT , 2-1

P

PASS CONTROL , 1-4
 preparation for operation , 1-6
Program Module
 detecting sweep end using SRQ and
 interrupt , 5-10
 getting data from analyzer using binary
 transfer , 3-10
 getting data trace array , 3-5
 getting memory array data , 3-6
 getting raw data array , 3-4
 initialize for the external controller , 1-3
 initialize module for Instrument BASIC ,
 1-2
 presetting the analyzer , 2-1
 querying active channel setting , 2-2

- reading an event register , 5-9
- retrieving data from the analyzer using
ASCII transfer , 3-8
- sending data to analyzer by ASCII transfer
, 3-8
- sending multiple commands in a line , 2-1
- Setting data to the data array, 3-4
- setting the active channel , 2-3
- setting the Measurement and Format , 2-3
- setting trace array data , 3-6
- program module , 1-2
- PROGram subsystem command, 8-101

Q

- query , 2-2
- query response , 2-2

R

- raw data , 3-4
- reading an event register directly , 5-9
- related documentation , 1-9
- remote mode , 1-7
- required instrument , 1-1
- RQS , 5-2, 5-10

S

- sample program
 - loading , 1-8
 - reading on PC , 1-8
- sample program disk , 1-8
- SCPI , 1-9, 8-101
- select code , 1-3
- serial number , A-2
- serial poll , 5-2
- service request , 5-2
- SPOLL , 5-2
- SRQ , 5-2, 5-10
- SRQ generation , 5-10
- status byte register , 5-2
- status register
 - using , 5-9
- status register structure , 5-4
- status reporting system , 5-1
- *STB , 5-2
- system controller , 1-4

T

- transition filter , 5-3

U

- unformatted data , 3-4

REGIONAL SALES AND SUPPORT OFFICES

If you need technical assistance with a Hewlett-Packard test and measurement product or application, please contact the Hewlett-Packard office or distributor in your country.

Asia Pacific:

Hewlett-Packard Asia Pacific Ltd
17-21/F Shell Tower, Times Square,
1 Matheson Street, Causeway Bay,
Hong Kong
(852) 2599 7070

Australia/New Zealand:

Hewlett-Packard Australia Ltd.
31-41 Joseph Street
Blackburn, Victoria 3130
Australia
131 347 ext. 2902

Canada:

Hewlett-Packard Canada Ltd.
5150 Spectrum Way
Mississauga, Ontario
L4W 5G1
(905) 206 4725

**In Europe, Africa and Middle East
please call your local HP sales office
or representative:**

Austria/East Central Europe:
(1) 25000-0

Belgium and Luxembourg:
(02) 778 34 17

Czech Republic:
(2) 471 7327

Denmark:
45 99 10 00

Finland:
(90) 88 721

France:
(1) 69.82.65.00

Germany:
(0180) 532 62-33

Greece:
(1) 7264045

Hungary:
(1) 252 4705

Ireland:
(01) 284 4633

Israel:
(03) 5380 333

Italy:
02 - 92 122 999

Netherlands:
(020) 547 6669

Norway:
(22) 73 56 00

Poland:
(22) 37 50 65

Portugal:
(11) 301 73 30

Russia:
(095) 956 2126

South Africa:
(011) 806 1000

Spain:
900 123 123

Sweden:
(08) 444 20 00

Switzerland:
(01) 735 7111

Turkey:
(212) 224 59 25

United Kingdom:
(01344) 366 666

**For European countries
not listed,
contact:**

Hewlett-Packard
International Sales Europe
Geneva, Switzerland
Tel: +41-22-780-4111
Fax: +41-22-780-4770

Japan:

Hewlett-Packard Japan Ltd.
Measurement Assistance Center
9-1, Takakura-cho, Hachioji-shi,
Tokyo 192, Japan
Tel: (81) 426 48 0722
Fax: (81) 426 48 1073

Latin America:

Hewlett-Packard
Latin American Region Headquarters
5200 Blue Lagoon Drive
9th Floor
Miami, Florida 33126
U.S.A.
(305) 267 4245/4220

United States:

Hewlett-Packard Company
Test and Measurement Organization
5301 Stevens Creek Blvd.
Bldg. 51L-SC
Santa Clara, CA 95052-8059
1 800 452 4844

