
User's Quick Start Guide

Publication number 54855-97016
December 2004

© Copyright Agilent Technologies 2002-2004
All Rights Reserved

Infinium Oscilloscopes

In This Book

This book gives you the information you need to begin using the Infiniium Oscilloscopes. It contains four chapters:

Setting up the Oscilloscope Chapter 1 contains inspection, power requirements, air flow, and setup information.

Working in Comfort Chapter 2 contains recommendations for working comfortably and safely while operating the Infiniium Oscilloscope.

Using the Oscilloscope Chapter 3 gives an overview of the front panel and the graphical user interface, and tells you how to perform basic operations with the oscilloscope.

Using the Built-In Information System Chapter 4 describes the built-in information system contents and navigation. The built-in information system contains all of the information that is generally found in a User's Guide.

- For detailed information on how the oscilloscope makes measurements and how to use the oscilloscope, see the built-in information system in the oscilloscope.
- For information on programming the oscilloscope using a computer with a GPIB interface card or LAN interface card, see the *Oscilloscopes Programmer's Reference* for your model number oscilloscope found on the manuals CD-ROM.
- For information on testing and servicing the oscilloscope, see the *Infiniium Service Guide* for your model number oscilloscope found on the manuals CD-ROM.

Contents

1 Setting Up the Oscilloscope

- To inspect package contents 1-3
- To inspect options and accessories 1-7
- To connect power 1-11
- To connect the mouse, the keyboard, a LAN cable, a printer, and a GPIB cable 1-14
-  To connect oscilloscope probes 1-16
- To connect SMA Cables 1-19
-  To connect the digital probe 1-20
- Digital probe lead set 1-21
- To tilt the oscilloscope upward for easier viewing 1-25
- To turn on the oscilloscope 1-26
- To turn off the oscilloscope 1-27
- To verify basic oscilloscope operation 1-28
- Installing application programs on Infiniium 1-30
- Changing Windows System Settings 1-31
- To clean the oscilloscope 1-32

2 Working in Comfort

- About Repetitive Strain Injury 2-3
- Mice and Other Input Devices 2-4

3 Using the Oscilloscope

- Front Panel 3-4
- Acquisition and General Controls 3-5
- Horizontal Controls 3-5
- Trigger Controls 3-5
- Vertical Controls 3-5
- Marker and Measurement Controls 3-6
- To set the oscilloscope to a known starting condition 3-7
- To start and stop waveform acquisition 3-8
- To clear the waveform display 3-9
- To turn an analog channel on or off 3-10
- To turn digital channels on or off 3-11
- To change input impedance and input coupling 3-12
- To adjust analog channel's vertical scale and offset 3-13
- To adjust digital channel's vertical size and offset 3-14
- To adjust sweep speed and horizontal position 3-15
- To magnify a part of the waveform using delayed sweep 3-16
- To set the oscilloscope to trigger on an edge 3-17
- To use the markers 3-18
- To use the quick measurements 3-19

Contents

- To reinitialize the oscilloscope 3-20
- Full-Screen Mode 3-21
- Graphical Interface Mode 3-23
- To perform basic user interface operations 3-32
- To select a command from the menu bar 3-35
- To select a command from a context-sensitive menu 3-36
- To start and stop waveform acquisition 3-39
- To clear the waveform display 3-40
- To print the screen 3-41
- To turn a channel on or off 3-42
- To adjust the vertical offset 3-43
- To adjust vertical scaling 3-45
- To access the channel setup 3-46
- To set the horizontal reference point 3-47
- To adjust sweep speed 3-48
- To adjust horizontal position 3-49
- To access the horizontal setup 3-50
- To zoom on a section of the waveform 3-51
- To move the markers using the graphical interface 3-53
- To make a measurement on a waveform 3-54
- To access the trigger setup 3-56
- To set an edge trigger 3-57
- To set dialog box preferences 3-58
- To recover your Infinium hard disk 3-61

4 Using the Built-In Information System

- To access the information system 4-4
- Topic Types 4-7
- To navigate through the information system 4-9
- To access context-sensitive information 4-10

1

Setting Up the Oscilloscope

Setting Up the Oscilloscope

This chapter shows you how to set up your Infiniium oscilloscope, connect power and accessories, and verify general operation.

To inspect package contents

Inspect the shipping container for damage.

Keep a damaged shipping container or cushioning material until you have inspected the contents of the shipment for completeness and have checked the oscilloscope mechanically and electrically.

Verify that you received the following items in the Infiniium Oscilloscope packaging.

- Infiniium Oscilloscope
- Mouse
- (4) 1165A 10:1 10 M Ω passive probes (54831 model only)
- (2) 1165A 10:1 10 M Ω passive probes (54830 model only)
- Digital Probe Kit (5483xD models only)
- Calibration Cable (54855A and DSO80000 series models only)
- BNC Shorting Cap for calibrating oscilloscope (54850 series models only)
- Precision 3.5 mm Adaptors (DSO80000 series, 54854, and 54855 models only)
- Accessory Pouch
- Front Panel Cover
- Keyboard
- Power cord
- Probe De-skew and Performance Verification Kit (54850 and DSO80000 series models only)
- *User's Quick Start Guide*

See Figure 1-1, Figure 1-2, or Figure 1-3 for your model of Infiniium Oscilloscope. (See table 1-3 for the power cord.) If anything is missing, contact your nearest Agilent Technologies Sales Office. If the shipment was damaged, contact the carrier, then contact the nearest Agilent Technologies Sales Office.


Inspect the oscilloscope.

- If there is mechanical damage or a defect, or if the oscilloscope does not operate properly or does not pass performance tests, notify your Agilent Technologies Sales Office.
- If the shipping container is damaged, or the cushioning materials show signs of stress, notify the carrier and your Agilent Technologies Sales Office. Keep the shipping materials for the carrier's inspection. The Agilent Technologies

Setting Up the Oscilloscope To inspect package contents


Sales Office will arrange for repair or replacement at Agilent's option without waiting for claim settlement.

Figure 1-1




Setting Up the Oscilloscope
To inspect package contents

Figure 1-2



Setting Up the Oscilloscope To inspect package contents

Figure 1-3



Package Contents for the 54830 Series Infiniium Oscilloscopes

To inspect options and accessories

- ❑ Verify that you received the options and accessories you ordered and that none were damaged.

If anything is missing, contact your nearest Agilent Technologies Sales Office. If the shipment was damaged, or the cushioning materials show signs of stress, contact the carrier and your Agilent Technologies Sales Office.

Some accessories that will enhance your work with the DSO80000 and 54850 series oscilloscopes are listed in table 1-1.

Table 1-1

Accessories for the DSO80000 and 54850 Series Infiniium Oscilloscopes

| Agilent Model Number | Description |
|----------------------|---|
| 54855-67602 | 18 GHz BNC-compatible to APC 3.5 mm adaptor |
| 1250-2427 | PC Board Mini-Probe Socket (horizontal mount) |
| 1250-2428 | PC Board Mini-Probe Socket (vertical mount) |
| 1250-1454 | BNC to Miniature Probe Adapter |
| 10240B | BNC Blocking Capacitor |
| 11094B | 75Ω Feedthrough Termination |
| 10024A | 16-pin IC clip |
| 10211A | 24-pin IC clip |
| 10833A | GPIB cable, 1 m |
| 10833B | GPIB cable, 2 m |
| 10833C | GPIB cable, 4 m |
| 10833D | GPIB cable, 0.5 m |
| 1131A | 3.5 GHz InfiniiMax Active Probe |
| 1132A | 5 GHz InfiniiMax Active Probe |
| 1134A | 7 GHz InfiniiMax Active Probe |
| 1156A | 1.5 GHz Active Probe |
| 1157A | 2.5 GHz Active Probe |
| 1158A | 4 GHz Active Probe |
| 1168A | 10 GHz Active Probe |
| 1169A | 12 GHz Active Probe |

Setting Up the Oscilloscope
To inspect options and accessories

| Agilent Model Number | Description |
|----------------------|--|
| 1184A | Testmobile with keyboard and mouse tray and drawer for accessories |
| 34398A | RS-232-C printer cable |
| 34399A | RS-232-C Adapter kit |
| 54006A | 6 GHz probe, 10:1 (500 Ω) or 20:1 (1 k Ω), .25 pf |
| C2950A | Parallel printer cable, 2 m |
| C2951A | Parallel printer cable, 3 m |
| E2609B | Rackmount Kit |
| E2621A | 75 Ω terminator |
| E2622A | 100/110/120 Ω differential terminator |
| E2646A | SQiDD Fixture for USB option |
| E2654A | EZ-Probe [®] Positioner |
| E2655A | Additional probe deskew and performance verification kit |
| E2680A | 1 MB Memory upgrade |
| E2681A | EZJIT Jitter Analysis Software |
| E2682A | VoiceControl Retrofit Kit |
| E2683A | USB Test Option |
| E2685A | Telecommunications Mask Test Template Kit |
| E2688A | Serial Data Analysis Software Option |
| E5850A | Time-correlation fixture, integrates Infiniium oscilloscope and 16900 logic analyzer |

Some accessories that will enhance your work with the 54830 series oscilloscopes are listed in table 1-2.

Table 1-2

Accessories for the 54830 Series Infiniium Oscilloscopes

| Agilent Model Number | Description |
|----------------------|----------------------------------|
| 10076A | 4 kV, 250 MHz High Voltage Probe |
| 10467-68701 | 4 IC Clips, 0.5 mm |
| N2771A | 10 kV, 50 MHz High Voltage Probe |
| 10833A | GPIB cable, 1 m |
| 10833B | GPIB cable, 2 m |
| 10833C | GPIB cable, 4 m |

Setting Up the Oscilloscope
To inspect options and accessories


| Agilent Model Number | Description |
|----------------------|---|
| 10833D | GPIB cable, 0.5 m |
| 1184A | Testmobile with keyboard and mouse tray, accessories draw |
| 1250-2427 | PC Board Mini-Probe Socket (horizontal mount) |
| 1250-2428 | PC Board Mini-Probe Socket (vertical mount) |
| 34398A | RS-232-C printer cable |
| 34399A | RS-232-C Adapter kit |
| 1130A | 1.5 GHz InfiniiMax Active Probe |
| 1131A | 3.5 GHz InfiniiMax Active Probe |
| 1146A | AC/DC Current Probe 100 kHz |
| 1147A | AC/DC Current Probe 50 MHz |
| 1153A | 200 MHz Differential Probe |
| 1154A | 500 MHz Differential Probe |
| 1155A | 750 MHz 2-Channel, Low-Mass Active Probe |
| 1156A | 1.5 GHz Active Probe |
| 1159A | 1 GHz Differential Probe |
| 1162A | 25 MHz, 1:1, 1 M Ω Passive Probe |
| 1163A | 1.5 GHz, 10:1, 50 Ω Passive Probe |
| 1165A | 500 MHz, 10:1, 10 M Ω Passive Probe |
| 1171A | 500 MHz, 10:1, 10 M Ω Low-mass Passive Probe |
| 1173A | 500 MHz, 20:1, 10 M Ω Low-mass Passive Probe |
| 1250-1454 | BNC to Miniature Probe Adapter |
| E2609B | Rackmount Kit |
| E2614A | Wedge Adapter, 0.5 mm spacing, 8 signals |
| E2615A | Wedge Adapter, 0.65 mm spacing, 3 signals |
| E2615B | 2 Wedge Adapters, 0.65 mm spacing, 3 signals |
| E2616A | Wedge Adapter, 0.65 mm spacing, 8 signals |
| E2621A | 75 Ω terminator |
| E2622A | 100/110/120 Ω differential terminator |
| E2685A | Communication Mask Test Kit Option |
| E2682A | VoiceControl Upgrade Kit |
| E2643A | Wedge Adapter, 0.5 mm spacing, 16 signals |

**Setting Up the Oscilloscope
To inspect options and accessories**

| Agilent Model Number | Description |
|-----------------------------|--|
| E2644A | Wedge Adapter, 0.65 mm spacing, 16 signals |
| E2683A | USB Test Option (for 54831 and 54832 models only) |
| E2646A | SQiDD Fixture for USB option |
| E2654A | EZ-Probe® Positioner |
| E2813A | Wedge Adapter, 0.5 mm spacing, 3 signals |
| E2813B | 2 Wedge Adapters, 0.5 mm spacing, 3 signals |
| E2681A | EZJIT Jitter Analysis Software |
| E2698A | Ethernet Masks |
| E2699A | My Infiniium Integration Package |
| E5850A | Time Correlation Fixture (Integrates the Infiniium oscilloscope and the 16900 logic analyzer) |
| N2845A | Upgrades standard memory to 4 Mpts/Ch. |
| N2846A | Upgrades standard memory to 8 Mpts/Ch. |
| N2847A | Upgrades memory from 4 Mpts. to 8 Mpts/Ch. |
| N5384A | Serial Data Analysis Software Option |

**Setting Up the Oscilloscope
To connect power****To connect power**

- 1 Position the oscilloscope where it will have sufficient clearance for airflow around the top, back, and sides.
- 2 Position the oscilloscope so that it is not difficult to unplug the power cord.


Figure 1-4**Positioning the Infiniium Oscilloscope with Sufficient Clearance**

Setting Up the Oscilloscope To connect power

- 3 Connect the power cord to the rear of the oscilloscope, then to a suitable ac voltage source (100 to 240 VAC \pm 10%, 47 to 440 Hz all models except for the DSO80000 series and 100 to 240 VAC \pm 10%, 50 or 60 Hz for the DSO80000 series)

Maximum power dissipation: 548000 series is 550 W, 54850 series is 475 W, and 54830 series is 440 W.

Figure 1-5



Infiniium Oscilloscope Power Cord Connection


The oscilloscope power supply automatically adjusts for line input voltages in the range 100 to 240 VAC. Therefore, you do not need to adjust an input line voltage setting. The line cord provided is matched by Agilent Technologies to the country of origin of the order.

- 4 Ensure that you have the correct line cord. See table 1-3.

Setting Up the Oscilloscope
To connect power

Table 1-3

Power Cords

| | Plug Type | Cable Part No. | Plug Description | Length (in/cm) | Color | Country |
|------|---|----------------|--------------------------|----------------|------------|---|
| 250V |  | 8120-1351 | Straight *BS1363A | 90/228 | Gray | United Kingdom, Cyprus, Nigeria, Zimbabwe, Singapore |
| | | 8120-1703 | 90° | 90/228 | Mint Gray | |
| 250V |  | 8120-1369 | Straight *NZSS198/ASC | 79/200 | Gray | Australia, New Zealand |
| | | 8120-0696 | 90° | 87/221 | Mint Gray | |
| 250V |  | 8120-1689 | Straight *CEE7-Y11 | 79/200 | Mint Gray | East and West Europe, Saudi Arabia, So. Africa, India (unpolarized in many nations) |
| | | 8120-1692 | 90° | 79/200 | Mint Gray | |
| | | 8120-2857 | Straight (Shielded) | 79/200 | Coco Brown | |
| 125V |  | 8120-1378 | Straight *NEMA5-15P | 90/228 | Jade Gray | United States, Canada, Mexico, Philippines, Taiwan |
| | | 8120-1521 | 90° | 90/228 | Jade Gray | |
| | | 8120-1992 | Straight (Medical) UL544 | 96/244 | Black | |
| 250V |  | 8120-2104 | Straight *SEV1011 | 79/200 | Mint Gray | Switzerland |
| | | 8120-2296 | 1959-24507 Type 12 90° | 79/200 | Mint Gray | |
| 220V |  | 8120-2956 | Straight *DHCK107 | 79/200 | Mint Gray | Denmark |
| | | 8120-2957 | 90° | 79/200 | Mint Gray | |
| 250V |  | 8120-4211 | Straight SABS164 | 79/200 | Jade Gray | Republic of South Africa India |
| | | 8120-4600 | 90° | 79/200 | Jade Gray | |
| 100V |  | 8120-4753 | Straight MITI | 90/230 | Dark Gray | Japan |
| | | 8120-4754 | 90° | 90/230 | Dark Gray | |

* Part number shown for plug is the industry identifier for the plug only. Number shown for cable is the Agilent part number for the complete cable including the plug.

Setting Up the Oscilloscope

To connect the mouse, the keyboard, a LAN cable, a printer, and a GPIB cable

To connect the mouse, the keyboard, a LAN cable, a printer, and a GPIB cable

See Figure 1-6 for the location of the connectors.

Mouse Plug the mouse into the matching connector on the back panel of the oscilloscope. The connectors are labeled 1 and 5.

While you can operate many oscilloscope functions using only the front-panel keys and knobs, you will need the mouse to access advanced oscilloscope functions through the graphical interface, or to find out more about the oscilloscope through the built-in information system.

Keyboard Plug the keyboard cable into the connector labeled 2 on the back panel of the oscilloscope.

A keyboard cannot be plugged into the oscilloscope after the Windows operating system has started booting.

LAN Cable Connect your LAN cable to the RJ-45 connector labeled 3 on the back panel of the oscilloscope.

After you have connected to the LAN card, you must set up the network. Before you try to setup your network, you should exit the oscilloscope application. If you do not know how to setup a network in Windows XP, see your network administrator or use the Windows XP on-line help.


Printer Cable If you have a parallel (Centronics) printer, you will need a parallel printer cable, such as an C2950A (2 m) or C2951A (3 m) cable. Connect cable into the connector labeled 4.

If you have a serial printer, you will need a 9-pin to 25-pin serial printer cable, such as an 34398A cable, plus the 34399A adapter kit. Some printers may require other cable configurations, but the oscilloscope has a 9-pin serial connector. Connect cable into the connector labeled 7.

GPIB Cable Attach the GPIB connector to the GPIB interface card connector labeled 6 on the rear of the oscilloscope.

Setting Up the Oscilloscope
To connect the mouse, the keyboard, a LAN cable, a printer, and a GPIB cable

Figure 1-6



54830e52


Back Panel

Setting Up the Oscilloscope
To connect oscilloscope probes

To connect oscilloscope probes

- 1 Attach the probe connector to the desired oscilloscope channel or trigger input. Push it straight on until it latches into place.

Figure 1-7




Attaching the Probe Connector

Setting Up the Oscilloscope
To connect oscilloscope probes

- 2 Connect the probe to the circuit of interest using the browser or other probing accessories.

Figure 1-8




Probing the Circuit

- 3 To disconnect the probe, push the small latch on top of the probe connector to the left, then pull the connector body away from the front panel of the oscilloscope without twisting it.

Setting Up the Oscilloscope To connect oscilloscope probes

Figure 1-9



Disconnecting the Oscilloscope Probe

CAUTION

Do not attempt to twist the snap-on probes on or off the oscilloscope's BNC connector. Twisting the probe connector body will damage it.

CAUTION



For the 54830 series oscilloscopes do not exceed the maximum input voltage rating. The maximum input voltage for 50 Ω inputs is 5 Vrms, CAT I. Maximum voltage at 1 M Ω impedance is ± 100 V (dc + ac) [ac < 10 kHz], CAT I.

CAUTION



For the 54850 series oscilloscopes do not exceed the maximum input voltage rating. The maximum input voltage for the 50 Ω inputs is 5 Vrms, CAT I.

CAUTION




For the DSO80000 series oscilloscopes do not exceed the maximum input voltage rating. The maximum input voltage for the 50 Ω inputs is ± 5 Vpeak, CAT I.

**Setting Up the Oscilloscope
To connect SMA Cables****To connect SMA Cables**

You can connect an SMA cable to the DSO80000 and the 54850 series Infiniium oscilloscopes by using precision 3.5 mm to BNC compatible adaptors.

- 1** Attach the two precision 3.5 mm to BNC compatible adaptors to the ends of an SMA cable.
- 2** Push the precision 3.5 mm to BNC compatible adaptors onto the oscilloscope BNC connectors.
- 3** Tighten the thumbscrews until they are snug.

Figure 1-10**Connecting SMA to BNC Adaptors**

Setting Up the Oscilloscope
To connect the digital probe


To connect the digital probe

The MSO series Infiniium oscilloscopes are the only oscilloscopes that have 16 digital channels.

The digital clip lead marked clk (clock) is unused. All the other digital clip leads are used for the digital channels.

- 1 Push the small connector end of the digital cable with the tab key facing left into the digital connector.
- 2 Tighten both thumb screws.

Figure 1-11



Connecting the Digital Cable to the Oscilloscope


Digital probe lead set

The MSO Infiniium oscilloscopes are the only oscilloscopes that have 16 digital channels.

The digital clip lead marked clk (clock) is unused. All the other digital clip leads are used for the digital channels.

The probe lead set has 16 digital channels with a ground lead for each channel.

Figure 1-12



Digital Probe Lead Set

If a 0.63 mm square pin or a 0.66 diameter round pin is installed on the circuit under test, the signal and ground leads can be directly connect to these pins. Otherwise, the IC clips can be used to connect to the circuit.

CAUTION 


Do not exceed the maximum input voltage rating of ± 40 V peak, CAT I.

Probe tip isolation network and equivalent load

The probe tips of the probe lead set contain an isolation network which serves to minimize the loading effect of the digital channels on the circuit under test. The isolation network schematic is shown in Figure 1-13.

Setting Up the Oscilloscope
Digital probe lead set


Figure 1-13



Probe Tip Isolation Network

The loading effect of the probe tip on the circuit under test is represented by the circuit shown in the equivalent load schematic in Figure 1-14.

Figure 1-14




Equivalent Load including oscilloscope

Direct connection through 40-pin connector

The probe cable can also be directly plugged into various 40-pin connectors on the circuit board under test. This requires each signal pin of the 40-pin connector to have an isolation network (Figure 1-13) on the circuit board. The pinout of the 40-pin connector is shown in Figure 1-15.

Figure 1-15



40-pin Connector Pinout

Note: +5 V is supplied by the oscilloscope to provide power for the demo board. DO NOT connect these pins to the circuit board under test.

CAUTION

Do not exceed the maximum input voltage rating of ±40 V peak, CAT I. The isolation network must be used on all digital channels for this to be valid.

Setting Up the Oscilloscope
Digital probe lead set


For more information on digital probing solutions, search for the document titled "Probing Solutions for Logic Analysis Systems" (Agilent part number 5968-4632E) on the Agilent Technologies web site at www.agilent.com.

Setting Up the Oscilloscope
To tilt the oscilloscope upward for easier viewing

To tilt the oscilloscope upward for easier viewing

- 1** Lift up the front of the oscilloscope, grasp the wire bail near the center, and pull it down and forward until it latches into place.

Figure 1-16



Latching the Oscilloscope Front Feet

Setting Up the Oscilloscope


To turn on the oscilloscope

To turn on the oscilloscope

The first time that you turn on the oscilloscope, you may need to have a keyboard and mouse connected. The keyboard and mouse are needed to enter the Product Key from the Microsoft Certificate of Authenticity for Windows XP if prompted to do so. This label is located on the rear panel of the Infiniium.

- 1 Depress the power switch in the lower left-hand corner of the oscilloscope front panel.

Figure 1-17



Turning on the Oscilloscope

After a short initialization period, the oscilloscope display appears. The oscilloscope is ready to use.

- 2 Hook up all cables and accessories before applying power. You can connect and disconnect probes while the oscilloscope is turned on.

Setting Up the Oscilloscope
To turn off the oscilloscope

To turn off the oscilloscope

- 1** Momentarily depress the power switch at the lower left-hand corner of the oscilloscope front panel. The oscilloscope will go through a normal Windows shutdown process.


Setting Up the Oscilloscope
To verify basic oscilloscope operation

To verify basic oscilloscope operation

- 1 Connect an oscilloscope probe to channel 1.
- 2 Attach the probe to the calibration output on the front panel of the oscilloscope.

Use a probe grabber tip so you do not need to hold the probe. The calibration output is marked with a square wave symbol.

Figure 1-18




54830e04

Verifying Basic Oscilloscope Operation for DS080000 Series and for 54830 Series Oscilloscopes.

Setting Up the Oscilloscope To verify basic oscilloscope operation

Figure 1-19



Verifying Basic Oscilloscope Operation for 54850 Series Oscilloscopes.

3 Press the Default Setup key on the front panel.

The display will pause momentarily while the oscilloscope is configured to its default settings.

4 Press the Autoscale key on the front panel.

The display will pause momentarily while the oscilloscope adjusts the sweep speed and vertical scale. You should then see a square wave with peak-to-peak amplitude of approximately 5 divisions and a period of almost 7 divisions. If you do not see the waveform, ensure your power source is adequate, the oscilloscope is properly powered-on, and the probe is connected securely to the front-panel channel input BNC and to the probe calibration output.

5 Move the mouse around the mouse surface and verify that the on screen mouse pointer follows moves with the mouse movement.

Setting Up the Oscilloscope
Installing application programs on Infiniium

Installing application programs on Infiniium

Infiniium is an open Windows system. This allows you to install your own application software. Agilent has verified that the following applications are compatible with the Infiniium oscilloscope application.

- Microsoft Office 2000
- MathWorks MATLAB
- Mathsoft MathCad 2001i
- McAfee VirusScan
- Symantec Norton AntiVirus

Before installing any software, you should exit the oscilloscope application.

If you install an application other than those which Agilent has tested, it is possible that it could break the oscilloscope application. This would require you to reinstall the oscilloscope application using the recovery disks.

Changing Windows System Settings

Before changing any Windows System settings outside of the oscilloscope application you should **Exit** the oscilloscope application.

There are several Windows System settings that can be changed to suit your own personal preferences. However, there are some system settings that you should avoid changing because it will interfere with the proper operation of the oscilloscope.

- Do not change the Power Options.
- Do not change the System Properties Hardware Tab settings.
- Do not change the System Properties Advanced Tab settings.
- Do not change the Regional and Language Options Advanced Tab settings.
- Do not remove Fonts.
- Display Settings
 - Do not change or turn off the default screen saver. The screen saver turns off the display's backlights extending their life.
 - Do not change the screen resolution from 640 by 480 pixels or the color quality from High (24 bit).
 - Do not use the Windows XP Theme.
 - Do not change "Windows and buttons" from the "Windows Classic Style".
 - Do not change the Font size to Extra Large.
 - Do not use a Menu font size greater than 12 points.
 - Do not modify "1. Digital Flat Panel (640x480) on Chips and Technologies (Asilant) 65550".
 - Do not set "Intel (r) 82815 Graphics Controller" to "Use this device as the primary monitor".
- Do not use the Administrative Tools to enable or disable Internet Information Services (Web Server). Use the Infiniium Web Control dialog box to enable or disable the Web Server.
- Do not delete or modify the InfiniiumAdmin user account.

Setting Up the Oscilloscope
To clean the oscilloscope

To clean the oscilloscope

- Clean the oscilloscope with a soft cloth dampened with a mild soap and water solution.

CAUTION

Do not use too much liquid in cleaning the oscilloscope. Water can enter the Infiniium front panel, damaging sensitive electronic components.

Working in Comfort

Introduction

To optimize your comfort and productivity, it is important that you set up your work area correctly and use your Infiniium oscilloscope properly. With that in mind, we have developed some set-up and use recommendations for you to follow based on established ergonomic principles.

Improper and prolonged use of keyboards and input devices are among those tasks that have been associated with repetitive strain injury (RSI) to soft tissues in the hands and arms. If you experience discomfort or pain while using the oscilloscope, discontinue use immediately and consult your physician as soon as possible. For more information on RSI you may wish to consult the *About Repetitive Strain Injury* section.

Please study the recommendations offered here in this chapter. Included there are references to relevant parts of international standards, regulations and guidelines, such as ISO 9241 and the European Community Display Screen Equipment directive. You may also wish to consult your employer's human resources department or other relevant departments for guidance specific to your company.

About Repetitive Strain Injury

Because your comfort and safety are our primary concern, we strongly recommend that you use the Infinium oscilloscope in accordance with established ergonomic principles and recommendations. Scientific literature suggests that there may be a relationship between injury to soft tissues—especially in the hands and arms—and prolonged improper use of keyboards or other equipment requiring repeated motions of the hands and forearms. This literature also suggests that there are many other risk factors that may increase the chance of such injury, commonly called Repetitive Strain Injury.

What is RSI?

Repetitive Strain Injury (RSI—also known as cumulative trauma disorder or repetitive motion injury) is a type of injury where soft tissues in the body, such as muscles, nerves, or tendons, become irritated or inflamed. RSI has been a reported problem for those who perform repetitive tasks such as assembly line work, meatpacking, sewing, playing musical instruments, and computer work. RSI also has been observed in those who frequently engage in activities such as carpentry, knitting, housework, gardening, tennis, windsurfing and lifting children.

What causes RSI?

The specific causes of RSI have not been established. Nevertheless, the incidence of RSI has been associated with a variety of risk factors, including:

- Too many uninterrupted repetitions of an activity or motion.
- Performing an activity in an awkward or unnatural posture.
- Maintaining static posture for prolonged periods.
- Failing to take frequent short breaks.
- Other environmental and psychosocial factors.

In addition, there have been reports associating the occurrence of RSI with the use of keyboards, mice, and other input devices. Also, certain medical conditions, such as rheumatoid arthritis, obesity and diabetes, may predispose some people to this type of injury.

What if I experience discomfort?

If you are experiencing any discomfort, seek professional medical advice immediately. Typically, the earlier a problem is diagnosed and treated, the easier it is to resolve.

Working in Comfort
Mice and Other Input Devices

Mice and Other Input Devices

Various aspects of using mice and other input devices may increase your risk of discomfort or injury. Observing the following recommendations may reduce that risk.

- Try to keep your hand, wrist, and forearm in a neutral position while using your mouse or other input device.
- If you use your thumb to rotate the ball on a trackball or spaceball, keep it in a relaxed, natural shape, and maintain a neutral posture in your hand, wrist, and forearm.
- Hold the mouse gently by draping your fingers over it. Keep your hand relaxed and fingers loose. Do not grip the mouse tightly.
- It takes very little pressure or force from your fingers to activate the buttons or scroll wheel on your mouse, scrolling mouse, trackball, or other input device. Using too much force can place unnecessary stress on the tendons and muscles in your hands, wrists, and forearms.
- If you are using a scrolling mouse, be sure to keep your fingers and hand in a relaxed, neutral position when activating the scroll wheel. Also, this type of mouse features software that can minimize the number of mouse movements or button clicks.
- When using a mouse, trackball, or other input device, position it as close to the keyboard as possible, and keep it at the same level as you do not have to stretch while using it.
- Be sure to keep your mouse and trackball clean. Regular removal of accumulated dust and dirt helps ensure proper tracking and reduces unnecessary hand and wrist motions.

Using the Oscilloscope

Using the Oscilloscope

The Infiniium Oscilloscope is designed to be easy to use.

- The familiar front-panel oscilloscope interface with knobs and keys is optimized for the most common kinds of troubleshooting tasks and basic measurements. See “Using the Front Panel” on page 3-3.
- The graphical interface with menus, windows, dialogs, and toolbars provides easy logical access to dozens of configuration and analysis tools, making it easy for you to set up and make the most complex measurements. The interface also allows you to use the Infiniium oscilloscope’s built-in information system, which gives detailed information on using the oscilloscope to make measurements. See “Using the Graphical Interface” on page 3-21.

Using the Front Panel

The Infiniium Oscilloscope front panel has been designed to give you direct access to the functions needed to perform the most common measurements needed in troubleshooting, using a traditional oscilloscope interface. Knobs and keys are included to enable direct setting of vertical and horizontal parameters. In addition, the front panel has a set of LED (Light-Emitting Diode) indicators; by using these and the display, you can assess the configuration of the oscilloscope at a glance—there is no need to enter a series of keystrokes to navigate through complex menus.


The Infiniium Oscilloscope uses color consistently throughout the front panel and user interface. For example, the color of the knob for channel 1 is the same color as the waveform for channel 1. All the configuration items and values related to channel 1 are displayed in the same color.

Using the Oscilloscope
Front Panel

Front Panel

Figure 3-1 shows the Infiniium oscilloscope front panel of the 54855 but all Infiniium oscilloscopes have the same set of controls.

Figure 3-1



Infiniium Oscilloscope Front Panel

Using the front panel, you can configure the Infiniium oscilloscope for most troubleshooting tasks. The control categories are:

- Acquisition and general controls
- Horizontal controls
- Trigger controls
- Vertical controls
- Marker and measurement controls

Acquisition and General Controls

Using the acquisition and general controls, you control whether the oscilloscope is running or stopped. Other keys allow you to reset the oscilloscope to its default setup, automatically configure the oscilloscope for the current input signals (Autoscale), or erase the waveforms from the display.

Horizontal Controls

Using the horizontal controls, you configure the oscilloscope's sweep speed (seconds per division) and horizontal position of the waveform. You can also view a magnified section of the waveform using the delayed sweep window, which uses software to expand part of the acquisition memory.

Trigger Controls

Using the trigger controls, you set the conditions on which the oscilloscope will trigger and acquire an input signal. You can set up a variety of trigger conditions. Edge and glitch triggers can be selected from the front panel, and the parameters for edge triggering can be set up here as well. Some glitch trigger parameters (such as glitch width) and all advanced trigger configurations are set up using the graphical interface.

Trigger configuration settings you make using the graphical interface are reflected in the front-panel status indicators, and will remain set unless you change them (either using the front panel or the graphical interface) or press the Default Setup key. See "Using the Graphical Interface" on page 3-21 for information on accessing the graphical interface.

Vertical Controls

Using the vertical controls, you set the vertical scaling (volts per division) and vertical offset. You can also turn the display on or off for a particular channel.

Using the Oscilloscope Marker and Measurement Controls

Marker and Measurement Controls

Using the marker and measurement controls, you control two sets of markers within the oscilloscope graticule. You use markers to make more accurate measurements of waveform events than you could make visually. Both time and voltage differences between the markers are updated continuously on the screen. By default, the markers track the source waveform. Voltage measurements from the markers are the value of the waveform at the time set with the marker arrow keys.

The QuickMeas key initiates five preset measurements on the waveforms. Both quick measurements and markers will function on any input waveform; simply continue to press and release one of the keys (either QuickMeas, Marker A, or Marker B) to cycle through all the waveforms on the screen, then to the off state. You choose which five measurements will be performed by using the measurement configuration commands in the graphical interface.

Using the Oscilloscope
To set the oscilloscope to a known starting condition

To set the oscilloscope to a known starting condition

- **Press the Default Setup key.**

You can set up the oscilloscope for many different kinds of complex measurements. To easily reset the oscilloscope to a known measurement configuration, use the Default Setup key.

If you use the Default Setup key with the graphical interface enabled, you can select Undo Default Setup from the Control menu to return the oscilloscope to its original configuration.

Save the Current Oscilloscope Configuration


Before using Default Setup, you may want to save the current oscilloscope configuration for later use. See the built-in information system (described in chapter 4) for instructions on saving and recalling setups, and for information on the exact configuration that is set when you press Default Setup.

Using the Oscilloscope
To start and stop waveform acquisition

To start and stop waveform acquisition

- **To start waveform acquisition, press the Run key.**
The oscilloscope begins acquiring data. When it receives a trigger signal, it finishes acquiring data, updates the display, then starts another acquisition cycle if it is in triggered or auto trigger mode. If it is in single sweep mode, it stops after updating the display.
- **To stop waveform acquisition, press the Stop key.**
The oscilloscope stops acquiring data. Whatever data was last acquired remains on the screen.

Figure 3-2



54600e33

Run and Stop Keys


Using the Oscilloscope
To clear the waveform display

To clear the waveform display

- Press the Clear Display key.

The oscilloscope clears the waveform display. If the oscilloscope is in Run mode and is receiving triggers, it will update the display as it collects new waveform data. Clearing the waveform display also resets averaging, infinite persistence, and color grade persistence, histogram, and mask testing database.

Figure 3-3



Clear Display Key

Using the Oscilloscope
To turn an analog channel on or off


To turn an analog channel on or off

- To turn an analog channel on, press the channel number key until it is illuminated. To turn it off, press the channel number key again.

If you are not using a particular channel, you can turn it off. This simplifies the waveform display and also increases the display update rate. While a channel is turned off, data acquisition continues for that analog channel. Thus, you can still use the analog channel as a source for functions.

Using an Analog Channel as External Trigger
 Any analog channel can be used as a trigger source. If you need an external trigger but do not need all analog channels, you can use an analog channel as an external trigger without displaying it by turning the analog channel display off.

Figure 3-4



54895e03

Using the Oscilloscope
To turn digital channels on or off


To turn digital channels on or off

Digital channels are only available on the MSO series Infiniium oscilloscopes.

- To turn digital channels on, press the digital channel number key until it is illuminated. To turn them off, press the digital channel number key again.

The only digital channels that are displayed are the ones that are turned on in the Digital Setup dialog box.

Figure 3-5




Using the Oscilloscope
To change input impedance and input coupling

To change input impedance and input coupling

Input impedance and input coupling are only available on the 54830 series Infiniium oscilloscopes.

- To change the input impedance, press the Input key until the LED for the desired impedance is illuminated.
 Choices are 50 Ω and 1 M Ω .
- To change the input coupling, press the Coupling key until the LED for the desired coupling is illuminated.
 Choices are AC and DC. If you change the input coupling to AC when 50 Ω impedance is selected, the input impedance changes to 1 M Ω . If you change the input impedance to 50 Ω , the input coupling changes to DC.

Figure 3-6



Input Impedance and Coupling

Using the Oscilloscope
To adjust analog channel's vertical scale and offset

To adjust analog channel's vertical scale and offset


- To make the waveform bigger, turn the vertical scale knob clockwise. To make it smaller, turn the knob counter-clockwise.

The vertical scale knob is the larger of the two knobs for a channel. It is marked with a set of sine wave symbols. Decreasing the vertical scale makes the waveform bigger. There are fewer volts displayed per division. Increasing the vertical scale makes the waveform smaller. There are more volts displayed per division.

- To move the waveform toward the top of the display, turn the vertical offset knob clockwise. To move it toward the bottom of the display, turn the knob counter-clockwise.

The vertical offset knob is the smaller of the two knobs for a channel. It is marked with a set of arrows.

Figure 3-7



Vertical Scale and Offset Controls

54855e03


Using the Oscilloscope
To adjust digital channel's vertical size and offset

To adjust digital channel's vertical size and offset

Digital channels are only available on the MSO Infiniium oscilloscopes.

- To select the digital channel which is affected by the vertical offset knob, turn the select knob clockwise to move toward the top-most digital channel. Turn the knob counter-clockwise to move toward the bottom most digital channel.
- To make the waveform bigger or smaller, press the size button.
- To move the waveform toward the top of the display, turn the vertical offset knob clockwise. To move it toward the bottom of the display, turn the knob counter-clockwise.

Figure 3-8



54830e36

Digital Channel's Vertical Size and Offset Controls

Using the Oscilloscope
To adjust sweep speed and horizontal position


To adjust sweep speed and horizontal position

- To stretch the waveform horizontally, turn the sweep speed knob clockwise. To shrink it horizontally, turn the knob counter-clockwise. The sweep speed knob is the larger of the two horizontal control knobs. It is marked with a set of sine wave symbols. Stretching the waveform means there are fewer seconds displayed per division. Shrinking the waveform means there are more seconds displayed per division.
- To move the waveform to the right, turn the horizontal position knob clockwise. To move the waveform to the left, turn the horizontal position knob counter-clockwise.

Moving the waveform to the right shows more of the pretrigger data (data acquired before the trigger event). Moving the waveform to the left shows more of the post-trigger data (data acquired after the trigger event).

The horizontal position knob is the smaller of the two horizontal control knobs. It is marked with a set of arrows. There is a detent programmed into the software so there is a momentary pause at zero while you are turning the knob. Continuing to turn the knob will move the horizontal position through zero.

Figure 3-9



Horizontal Sweep Speed and Position Controls

Using the Oscilloscope


To magnify a part of the waveform using delayed sweep**To magnify a part of the waveform using delayed sweep**

- To turn on the delayed sweep, press **Delayed**. To turn it off, press **Delayed** again.

The waveform display area splits into two regions. The top one is the main sweep. The bottom is the delayed sweep, which represents a software expansion of the acquired waveform data. A section of the waveform in the main sweep window is highlighted to indicate the part shown in the delayed sweep window.

The horizontal sweep speed and horizontal position controls now change how the waveform is shown in the delayed sweep window. The sweep speed will change the amount of magnification, while the position will change the part of the waveform in the main sweep window that is shown in the delayed sweep window.

Figure 3-10


**Magnifying Part of the Waveform with Delayed Sweep**

Using the Oscilloscope
To set the oscilloscope to trigger on an edge

To set the oscilloscope to trigger on an edge

- 1 Press and release the Mode key until the Edge LED indicator is illuminated.
- 2 Press and release the Source key until the desired source LED is illuminated.
You can choose any of the channels or the Aux Trig In as the source for an edge trigger.
- 3 Press the Slope key until the desired slope LED is illuminated.
You can have an edge trigger on a rising or falling edge.
- 4 Press the Sweep key until the Trig'd LED is illuminated.
The oscilloscope will wait for the edge before initiating a sweep.
- 5 Turn the Level knob to adjust the voltage level at which the oscilloscope will trigger.

Figure 3-11



Trigger Controls and Indicators

Using the Oscilloscope

To use the markers

To use the markers

Markers make it easier to make precise measurements because the marker measurement readouts show exact voltage and time positions for the markers. The measurements are based on actual waveform data from the acquisition system, not on approximations based on the display position, so you can be sure that the values are highly accurate.

- To turn on Marker A, press the Marker A key.

Marker A has a solid line pattern on the waveform display. It is associated with the first available source on the display. Press the key again to move to the next available source. When there are no more sources, the marker turns off.


- To turn on Marker B, press the Marker B key.

Marker B has a dashed line pattern on the waveform display. It is associated with the first available source on the display. Press the key again to move to the next available source. When there are no more sources, the marker turns off.

- To move a marker on the waveform, press and hold the left arrow or right arrow key next to the desired Marker key. Release the key when the marker is at the desired waveform event.

The marker snaps to and follows the shape of the waveform on the screen. The voltage value shown for a marker is the value of the waveform at the specified horizontal time, which is set with the marker arrow keys. This is the default mode. You can change the marker mode using the graphical interface. See the built-in information system for details.

Figure 3-12




Using the Oscilloscope To use the quick measurements

To use the quick measurements

- To turn on the quick measurement display, press the QuickMeas+ key. The five preset measurements defined in the Quick Measurement configuration are enabled and results are displayed on the screen for the first waveform source. The default measurements are: V_{p-p} , Period, Frequency, Rise time, and Fall time.
- To measure parameters for another waveform, press the QuickMeas+ key until that waveform is the one shown in the measurement readout. Continuing to press the QuickMeas+ key cycles through each of the waveforms available.
- To turn off the quick measurement display, press and release the QuickMeas+ key until the measurements are turned off. The measurement results disappear from the screen.

See the built-in information system (described in Chapter 4) for information on how to configure the Quick Measurement capability, using the Customize Measurement feature of the graphical interface.

Figure 3-13



54800e16

Quick Measurement Key

Using the Oscilloscope To reinitialize the oscilloscope


To reinitialize the oscilloscope

When you need to restore the oscilloscope to a known configuration, use the Default Setup key. If you press the Default Setup key and the oscilloscope does not seem to be functioning properly, try cycling power. If the oscilloscope still does not seem to function properly, use the following key-down powerup procedure.

- 1 Turn off the power to the oscilloscope.
- 2 Turn on the power to the oscilloscope.
- 3 Hold down any one of the arrow keys next to the Marker A and Marker B keys.
- 4 When the oscilloscope display appears, release the key you held down in step 3.

A key-down powerup completely reinitializes the oscilloscope, including the configuration RAM. It does not affect saved waveforms or setups, which are stored on the hard disk drive.

Figure 3-14



54800e16

Key-Down Powerup


Using the Graphical Interface

With the graphical interface for the Infiniium Oscilloscope, you can access all the configuration and measurement features of the oscilloscope through an easy-to-use system of menus, tool bars, dialog boxes, icons, and buttons.

Full-Screen Mode

Full-screen mode maximizes the waveform viewing area and removes the graphical interface menus and toolbars so you can concentrate on your measurement. In full-screen mode, the display looks like the following two figures.


Figure 3-15



Infiniium Oscilloscope Top of Display in Full-Screen Mode

Using the Oscilloscope
Full-Screen Mode

Figure 3-16




Infiniium Oscilloscope Bottom of Display in Full-Screen Mode

Using the Oscilloscope
Graphical Interface Mode

Graphical Interface Mode

Click the graphical interface enable button to switch to the graphical interface. When the graphical interface is enabled, the display looks like the following two figures.


Figure 3-17



Infiniium Oscilloscope Top of Display in Graphical Interface Mode

Using the Oscilloscope
Graphical Interface Mode

Figure 3-18



Infiniium Oscilloscope Bottom of Display in Graphical Interface Mode

To make it easy to see which controls affect each waveform, the oscilloscope uses color consistently throughout the graphical interface. These colors match the ones used on the front-panel knobs. For example, the color of the waveform for channel 1 matches the color of the knobs for that channel. If channel 1 is the trigger signal, all of the trigger configuration items, including the trigger level reference icon (at the right side of the waveform display area), will match that color. The buttons associated with that channel, vertical scaling and offset settings, ground reference indicator, and measurements done on that channel also have the same color.

You can still use the front panel when the graphical interface is enabled. All changes made to the front-panel settings are reflected in the graphical interface, and changes made using the graphical interface are reflected in the front panel where applicable. Use whichever interface is easiest for you in a particular measurement situation. For example, it might be easiest to set a coarse vertical scale using the knobs, then fine-tune the setting using the graphical interface.

Using the Oscilloscope Graphical Interface Mode

The graphical interface is arranged so the most common functions that affect the waveform display are located around the edge of the waveform display area. These include the measurement toolbar, horizontal and trigger toolbar, and vertical toolbar.

Measurement Toolbar

The measurement toolbar contains icons representing the most commonly used automatic measurements built into the oscilloscope.

Drag and Drop Measurements By dragging one of the measurement icons to a waveform in the waveform display area, you can make that measurement on the waveform. As you drag a measurement icon around the screen, the icon outline changes color to match the color of the closest waveform. This makes it easy to see which waveform will be measured when you drop the icon. For those measurements that are done on waveform features, the measurement is made at the feature closest to the location where you dropped the icon. For example, you might want to measure the rise time of the fifth rising edge; dropping the rise time measurement icon at that edge will cause the measurement to be made on that edge.

You can also make a measurement by simply clicking the icon on the measurement toolbar, then selecting the source to be measured in the dialog box that appears. When you start a measurement this way, any measurements on waveform-specific features will measure the feature closest to the horizontal reference indicator.

Each waveform can have multiple simultaneous measurements and the measurements can all be of the same type, if desired. For example, you can have 3 pulse width measurements on different parts of the same waveform.

Geometric Measurement Indicators For each measurement currently running, a geometric indicator at the measurement location on the waveform corresponds to an identical indicator in the measurement results readout. This makes it easy for you to verify that the readout shows results for the correct waveform and the correct feature on that waveform. See figure 3-38 for an example.

Tool Tips To find out what a particular measurement tool does, hold the mouse pointer over it for a moment. A small popup will appear that describes the measurement.

Using the Oscilloscope Graphical Interface Mode

Other Measurement Features There are more measurements available than will fit on a single toolbar. Click the More (1 of 2) or More (2 of 2) icons to see other measurements. Clicking Clear All will remove all selected measurements from the waveform display area.


You can turn off the measurement toolbar to remove it from the screen and enlarge the waveform viewing area. Use the Customize Display Layout command on the Measure menu.

Using the Oscilloscope Graphical Interface Mode

Tab Display Area

The tab display area located beneath the waveform viewing area appears when a measurement is on, mask testing is enabled, a histogram is enabled, markers are on, or color graded persistence is on.

Figure 3-19



Tab Display Area

The display area shows information and statistics for the particular tab that is selected. The type of markers that are shown in the waveform viewing area depend on the tab that you have selected. The selected tab has an orange border to reflect the type of markers being displayed. For example, when the Histogram tab is selected, the markers are histogram markers and are used to define the histogram window.

Waveform Display Area

The waveform display area shows the waveforms, and optionally, the results of your measurements. Several display options, including a grid, are available and can be configured using the graphical interface.

Waveform Manipulation When the graphical interface is enabled, two features are available that can simplify your work with waveforms:

- **Direct Manipulation**—you can use the mouse to click and drag waveforms to new vertical positions, which changes the vertical offset, or to new horizontal positions, which changes the horizontal position or delay value.
- **Zoom**—you can click and drag a rectangular area on the display, then click inside it to zoom on that section of the waveforms. The oscilloscope does this in one of two ways. If acquisition is stopped, the magnification is done by the oscilloscope software. If acquisition is running, the oscilloscope automatically adjusts the vertical scaling and offset and the horizontal sweep speed and position to present the zoomed section of the waveforms.

Using the Oscilloscope Graphical Interface Mode

See “To zoom on a section of the waveform” on page 3-51.

Avoid Overdriving Vertical Input Amplifiers

When zooming on a waveform with the oscilloscope running, be careful to keep the signal within the screen vertically to avoid overdriving the vertical input amplifiers. Overdriving causes waveform distortion and erroneous measurement results.

Ground Reference Indicators A small symbol is shown at the right side of the waveform display area for each waveform that is on, including channels, waveform memories, and functions. This symbol represents the ground reference point for each channel; it moves when you change the vertical offset. You can also drag this symbol up and down using the mouse; doing so automatically changes the vertical offset for that waveform.

Menu Control and Menus

The graphical interface control button, in the upper right-hand corner of the display, enables the graphical interface of the oscilloscope. When the graphical interface is enabled, the display looks like figure 3-17 and figure 3-18, including a menu bar, measurement toolbar (if enabled), and graphical controls for vertical, horizontal, trigger, and acquisition. You can switch to full-screen mode to maximize the waveform viewing area and eliminate the menu bar, measurement toolbar, and other graphical controls.

You can use the menu bar for most oscilloscope configuration functions. *Context-sensitive* menus, which pop up to provide a selection of commands within particular regions of the user interface, are available in the following regions:

- Memory bar
- Waveform display area
- Measurement toolbar
- Horizontal and acquisition controls

You display a context-sensitive menu by clicking the right mouse button with the pointer in one of these regions. For more information on context-sensitive menus, see “To select a command from a context-sensitive menu” on page 3-36.

Using the Oscilloscope Graphical Interface Mode

Vertical Settings and Controls The top of the waveform display area includes the vertical settings and controls. In full-screen mode, only channels that are turned on are shown, with the corresponding vertical scaling settings in volts per division. When the graphical interface is enabled, all channels are shown. Each has a checkbox allowing you to turn that channel on or off, and a set of controls allowing you to change the vertical scaling. Clicking directly on the vertical scaling value displays a pop-up numeric keypad allowing you to set a precise vertical scale.

Horizontal and Trigger Toolbar

At the bottom of the waveform display area is the horizontal and trigger toolbar. This includes the run/stop controls, the horizontal controls, and the trigger controls.

Using the Oscilloscope Graphical Interface Mode

Run/Stop Controls See figure 3-25. At the left side of the bar are four icons:

- The leftmost is a blue-green octagon. Clicking on this starts an acquisition. (Same as pressing the Run key on the front panel.)
- The middle is a red octagon. Clicking on this stops acquisition. (Same as pressing the Stop key on the front panel.)
- The next control is a small windshield wiper. Clicking on this clears acquired waveform data from the display. (Same as pressing the Clear Display key on the front panel.)
- The rightmost is a printer. Clicking prints the screen to the default printer.

Horizontal settings and controls The middle of the bar contains the horizontal settings and controls. Leftmost is a button, labeled with an “H.” Clicking on this will display the horizontal setup dialog box.

Next is the current sweep speed. Clicking on this displays a pop-up numeric keypad so you can set a particular sweep speed. Or, you can click on the two icons to the right of the sweep speed setting to cycle through the preset speeds. The leftmost icon shrinks the waveform, which decreases the sweep speed and increases the time per division. The rightmost icon stretches the waveform, which increases the sweep speed and decreases the time per division.

Next is the horizontal position (delay) setting. Clicking on this displays a pop-up numeric keypad that allows you to set a particular position. Or, you can use the three icons to the right. The left arrow moves the waveform to the left, the center “0” resets the delay to zero, and the right arrow moves the waveform to the right.

Across the toolbar are three vertical arrows. These are the left, center, and right horizontal reference indicators. Clicking on one of these arrows moves the horizontal position to the respective horizontal reference position on the display—left, center, or right. Assuming the horizontal position is at zero:

- Left means the information on the display is all post-trigger.
- Center means the information to the left of center is pretrigger; to the right is post-trigger.
- Right means the information on the display is pretrigger.

The horizontal position value represents the time relative to the trigger at the respective horizontal reference. When you change the horizontal sweep speed, the waveforms expand and contract about this reference position.

Using the Oscilloscope Graphical Interface Mode

Trigger settings and controls The right side of the bar contains the trigger settings and controls. These will vary depending on the current trigger configuration, which can be set using the front panel and the graphical interface. Advanced trigger configuration items are available only through the graphical interface. You can click on the button labeled with a “T” to bring up the trigger setup dialog box.

When the scope is set for edge trigger on a particular channel, the trigger level setting is shown. You can click on it to display a pop-up numeric keypad that allows you to set a particular trigger level. You can also click on the up and down arrows to the right of the setting to increase or decrease the trigger level, respectively. Or, you can click on the trigger reference indicator at the right side of the display and drag it up or down to change the trigger level.

Using the Oscilloscope


To perform basic user interface operations

To perform basic user interface operations

- To move the mouse pointer on the screen, move the mouse on the mouse surface.
- To *click* on an item in the graphical interface, point at that item with the mouse pointer, then press and release the left mouse button.
- To *right-click* on an item in the graphical interface, point at that item with the mouse pointer, then press and release the right mouse button.
You use the right-click operation to access context-sensitive menus. See “To select a command from a context-sensitive menu” on page 3-36.
- To use a *radio button*, click to select the desired item.
Radio buttons appear in many different dialog boxes in the oscilloscope graphical interface. See the Persistence radio buttons in figure 3-20. You can choose only one option at a time.
- To use a *check box*, click the mouse button with the pointer in the box.
A check mark in the box indicates that item is selected. See the Connect Dots check box in figure 3-20. To clear the selection, click the mouse button with the pointer in the box.
- To use a *drop-down list box*, click the arrow at the right-hand side of the box. Then click on the desired choice to highlight it.
See the Language selection list box in figure 3-21.
- To use a *spin box*, click the up arrow to increase the value displayed in the box, and the down arrow to decrease it.
See the Intensity spin box in figure 3-20.
- To move a *dialog box*, press and hold the left mouse button with the pointer in the *title bar*, drag the box to a new position on the screen, then release the mouse button.
- To close a *dialog box*, click the “X” symbol in the upper right-hand corner of the box, or click the Close button in the box.

Using the Oscilloscope
To perform basic user interface operations


Figure 3-20




Dialog Box Interface Elements

Using the Oscilloscope
To perform basic user interface operations

Figure 3-21



Click the arrow in a
drop-down list box...



...to see the options you can
choose

Dialog Box with a Drop-Down List Box

Using the Oscilloscope
To select a command from the menu bar

To select a command from the menu bar


- 1 Click on a menu bar item.
- 2 Move the pointer to the desired menu item.
- 3 Click the mouse button.

The desired command is executed, or a dialog box is presented for you to configure the oscilloscope.

If you continue to hold the mouse button after step 1, release the button in step 3 to execute the command.

Some menus have submenus. These are indicated by an arrow at the right side of the command. When you move the pointer to one of these menu commands, the submenu automatically appears. You can then move the pointer to the desired command on that submenu and click the mouse button to execute the command.

Figure 3-22



Click the Measure menu,
then Customize, and then
Measurement
Thresholds to customize
the measurement setup

Selecting a Command from the Menu Bar

Using the Oscilloscope

To select a command from a context-sensitive menu

To select a command from a context-sensitive menu

- 1** Move the mouse pointer to a particular area of the display in which you want to change the oscilloscope configuration.


Context-sensitive menus provide quick access to commands and configuration items that relate only to the context of the particular graphical interface item in which they are found. They are available in the following display areas: memory bar, measurement toolbar, waveform display area, and horizontal settings and controls.

- 2** Click the right mouse button.
- 3** Move the pointer to the desired menu item.
- 4** Click the mouse button.

If you continue to hold the mouse button after step 2, release the button in step 4 to execute the command.

Using the Oscilloscope
To select a command from a context-sensitive menu

Figure 3-23



Right-click the mouse in the waveform display area, then select Autoscale from the menu to force the oscilloscope to autoscale the current waveform(s)


Selecting a Command from a Context-Sensitive Menu

Using the Oscilloscope
 To select a command from a context-sensitive menu

Figure 3-24

Right-click in this region to see the Measurement Setup context-sensitive menu

Right-click in this region to see the Acquisition Setup context-sensitive menu



Right-click in the waveform display area to see the waveform context-sensitive menu

Right-click in this region to see the horizontal, trigger, and run controls context-sensitive menu

Location of Context-Sensitive Menus

Using the Oscilloscope
To start and stop waveform acquisition


To start and stop waveform acquisition

- To start waveform acquisition, click the start button at the bottom of the waveform display area.
- To stop waveform acquisition, click the stop button at the bottom of the waveform display area.

See figure 3-25.

You can still use the Run and Stop keys on the front panel while the graphical interface is enabled. However, if you are performing many graphical interface operations, it is easier to start and stop the oscilloscope using the buttons in the interface, so you do not have to remove your hand from the mouse.

Figure 3-25



Click this button
to start waveform
acquisition

Click this button
to stop waveform
acquisition

Start and Stop Icons

Using the Oscilloscope
To clear the waveform display


To clear the waveform display

- Click the clear display button at the bottom of the waveform display. See figure 3-26.

You can still use the Clear Display key on the front panel while the graphical interface is enabled.

Clearing the display removes all displayed waveform data in preparation for another acquisition. It also resets all automatic measurements and measurement statistics.

Figure 3-26



Click this button to
clear the waveform
display


Clearing the Display

Using the Oscilloscope To print the screen

To print the screen

- Click the print button at the bottom of the waveform display. See figure 3-27. Infiniium prints the screen to the default printer according to the configuration that you have selected in the Windows Control Panel.

Figure 3-27



Click this button to
print the screen

Printing the Screen

Using the Oscilloscope
To turn a channel on or off

To turn a channel on or off

- To turn a channel on, click the check box next to the channel number so that a check mark appears in the box. To turn a channel off, click the check box again to clear it.

See figure 3-28.


If you are not using a channel, you can turn it off. This simplifies the waveform display and also increases the display update rate. When you turn off a channel, the current vertical scaling factor and vertical scale buttons for that channel disappear. They re-appear when you turn the channel on.

You can also turn a channel on or off by using the Channel Setup dialog box, or by using the front panel key for that channel.

Figure 3-28

Turn a channel on by clicking in the On check box until a check mark appears

Turn a channel off by clicking in the On check box until the box is cleared



Turning a Channel On or Off

To adjust the vertical offset

- Click and hold the left mouse button on the waveform you want to move, then drag the mouse up or down to move the waveform to the desired offset. Release the mouse button when finished.
- or
- Click and hold the left mouse button on the ground reference indicator for the waveform you want to move, then drag the mouse to move the waveform to the desired offset. Release the mouse button when finished.

or

- Click the channel button in the vertical toolbar, then set the vertical offset using the spin box control in the Channel Setup dialog box.


If you leave the mouse pointer on the waveform (without clicking the mouse), the pointer changes into crossed arrows as a visual indicator that you can drag the waveform when you click and hold the mouse button. If you move the mouse up or down, the interface constrains the movement vertically and adjusts vertical offset only. If you initially move the mouse left or right, the interface constrains the movement horizontally and changes the horizontal position setting only.

You can also adjust the vertical offset through the Channel Setup dialog box. See “To access the channel setup” on page 3-46.

Using the Oscilloscope
To adjust the vertical offset

Figure 3-29

Click to access the Channel Setup dialog, then set vertical offset using the spin box or numeric keypad...



...or click on the ground reference indicator and drag it up or down to change the vertical offset...

...or click on the waveform and drag it up or down to do the same thing.

Adjusting Vertical Offset

Using the Oscilloscope
To adjust vertical scaling

To adjust vertical scaling

- To make the waveform bigger, click on the larger waveform button next to the channel number near the top of the display. To make the waveform smaller, click on the smaller waveform button.


See figure 3-30. The number of volts per division decreases.

The number of volts per division increases. The current setting in volts per division is shown next to the waveform buttons for the channel. Notice that the setting is color-coded to match the waveform color for easy recognition.

You can also adjust the vertical scaling by clicking on the current setting, which is shown next to the waveform buttons for the channel. This displays a pop-up numeric keypad that allows you to set an exact value for the scaling. Or, you can adjust the vertical scaling through the Channel Setup dialog box. See “To access the channel setup” on page 3-46.

Figure 3-30

Click to display a pop-up numeric keypad that lets you specify a particular vertical scaling value



Click to make the waveform bigger (fewer volts per division)

Click to make the waveform smaller (more volts per division)

Adjusting Vertical Scaling


Using the Oscilloscope
To access the channel setup

To access the channel setup

- Click the channel number button at the top of the waveform display.
 or
- Select the desired channel from the Setup menu.

With the channel setup dialog, you can set the vertical scaling, offset, input coupling, and input impedance. It also lets you specify the characteristics of the probe, using the Probes button. You can set attenuation ratio, attenuation units, and measurement units for the probe, or perform a probe calibration. For Agilent Technologies probes that are compatible with the AutoProbe Interface, the oscilloscope will automatically set these characteristics (except for skew) after identifying the probe when it is connected to the channel input.

Figure 3-31




Accessing the Channel Setup

Using the Oscilloscope
To set the horizontal reference point

To set the horizontal reference point

- Click on one of the arrows at the bottom of the waveform display. The selected horizontal reference is highlighted. The vertical arrows at the bottom of the display correspond to the left, center, and right horizontal reference points. This is the position of the trigger if the horizontal position is set to zero delay. A non-zero horizontal position value is the time relative to the trigger event at the selected horizontal reference point.

Figure 3-32



Click to set the horizontal reference to the left side of the display

Click to set the horizontal reference to the center of the display

Click to set the horizontal reference to the right side of the display

Horizontal Reference Points

Using the Oscilloscope To adjust sweep speed


To adjust sweep speed

- To stretch the waveform horizontally, click the larger waveform button next to the horizontal sweep speed setting at the bottom of the waveform display. To shrink the waveform horizontally, click the smaller waveform button.

See figure 3-33. The oscilloscope decreases or increases the number of seconds per division based on your selection.

You can also adjust the sweep speed by clicking on the current setting. This displays a pop-up numeric keypad that allows you to set an exact value for the sweep speed. Or, you can adjust the sweep speed through the Horizontal Setup dialog box. See “To access the horizontal setup” on page 3-50.

Figure 3-33



Click here to display the horizontal setup dialog box

Click here to set the sweep speed directly through a pop-up numeric keypad

Click to shrink the waveform (more seconds per division)

Click to stretch the waveform (fewer seconds per division)

Adjusting Sweep Speed

Using the Oscilloscope To adjust horizontal position

To adjust horizontal position


The *horizontal position* is the time relative to the trigger at the highlighted horizontal reference point.

- To zero the horizontal position, click the 0 button next to the position value at the bottom of the waveform display. This will position the trigger event at the highlighted horizontal reference point.
- To move the waveform to the right or left, click the right or left arrow next to the position value at the bottom of the waveform display. Or, click and hold the mouse button on the desired waveform and drag the waveform to the right or left. Release the mouse button when the waveform is in the desired position.

When you click and drag a waveform, this changes the horizontal position for all channels and functions on the display, but not that of waveform memories.

You can also adjust the horizontal position through the horizontal setup dialog.

Figure 3-34



Adjusting Horizontal Position


Using the Oscilloscope
To access the horizontal setup

To access the horizontal setup

- Click the horizontal setup button at the bottom of the waveform display.
or
- Select Horizontal from the Setup menu.
See figure 3-35.

Using the Horizontal Setup menu, you can set sweep speed, position, and the horizontal reference. You can also set up the delayed time base window, which uses software expansion to present a magnified view of an area of the waveform. See the built-in information system (described in chapter 4) for details.

Figure 3-35



Accessing the Horizontal Setup

Using the Oscilloscope
To zoom on a section of the waveform

To zoom on a section of the waveform

- 1 Click and hold the left mouse button in a blank space within the waveform display area, then drag the mouse to draw a rectangle around the areas of interest on the waveform(s). Release the mouse button.
- 2 Click the mouse once anywhere inside the rectangle drawn in step 1.

How the zoom takes place depends on whether acquisition is running or stopped. If running, the oscilloscope adjusts vertical scale, vertical offset, sweep speed, and horizontal position as necessary to best rescale the area you defined in step 1. If stopped, the oscilloscope does a software magnification of the waveform.


In either case, a small dialog box appears in the corner of the screen, with a single "Undo" button; by clicking this button, you will undo the zoom and return the oscilloscope to its previous settings.

Avoid Overdriving the Vertical Input Amplifiers

When the oscilloscope is running, the input waveform can overdrive the vertical input amplifiers if you expand the vertical excessively. The result is distortion of the waveform. Ensure that you select a region that will keep all of the waveform on the screen vertically.

Using the Oscilloscope
To zoom on a section of the waveform

Figure 3-36



Click and hold the mouse button, drag to create a selection rectangle, then click inside the rectangle to zoom the display on the waveforms inside the rectangle

Zooming on a Section of the Waveform

Using the Oscilloscope
To move the markers using the graphical interface

To move the markers using the graphical interface


1 Turn on the markers.

You can use the Marker A and Marker B keys on the front panel, or select Markers from the Measure menu.

2 Click and hold the left mouse button while the mouse pointer is on one of the markers, then drag the marker to the position you want on the waveform. Release the mouse button.

Clicking and dragging a marker makes it easy to quickly move the marker to the desired waveform event. You can use the front-panel arrow keys for fine adjustment. Or, you can select the Markers command from the Measure menu and set the marker position precisely. See the built-in information system for details.

Figure 3-37



To move a marker, click the mouse with the pointer on the marker, drag the marker to the new position, then release the mouse button

Moving the Markers

Using the Oscilloscope

To make a measurement on a waveform

To make a measurement on a waveform

- Click and hold the mouse button on a measurement icon, then drag the icon to the waveform event you want to measure. Release the mouse button.

or

- Click on a measurement icon, then specify which source you want to measure in the dialog box that is displayed.

For measurements on waveform features, such as those that involve waveform edges, if you click the measurement icon and specify a source, the measurement defaults to using the feature closest to the horizontal reference point. When you make the measurement using drag-and-drop, the measurement uses the waveform feature closest to the point where you drop the icon.

The Measurement toolbar is normally displayed at the left side of the screen when the graphical interface is enabled. You can turn off the toolbar, thereby increasing the waveform viewing area, by selecting Utilities, then User Preferences, then Show Measurement Toolbar.

Only the most commonly used measurements are available on the toolbar. Others are available by selecting them from the Measure menu. A few specialized measurements are available only through the GPIB programming command set to the oscilloscope. See the *Programmer's Reference*.

When you drag and drop a measurement icon on a waveform, notice that the icon outline changes color to match the color of each waveform it touches. This enables you to easily see which waveform will be measured when you release the mouse button. Also, the measurement results are also displayed in color to make it easy to see which waveform source is being measured.


For edge-sensitive measurements, when you drop the measurement icon on a waveform, a geometric icon appears in the waveform marker color. This icon shows exactly where the measurement is being made. The icon is also displayed next to the measurement results readout to associate the readout with the waveform and feature being measured. See figure 3-38. This feature helps you distinguish measurement results from each other when you make multiple measurements on the same waveform, but at different waveform features. For example, a series of rise time measurements on several different edges of the same waveform all have unique geometric icons to show each measurement's location.

The markers default to tracking the last measurement started. The marker position readout defaults to off to reduce clutter, but can be turned on in the Measure menu if desired.

Using the Oscilloscope
To make a measurement on a waveform

Figure 3-38

Channel numbers associate the measurement results with the channel waveform being measured.



In some cases, geometric icons are also displayed to associate the measurement results with the location on the waveform being measured.

Waveform Measurements

Using the Oscilloscope
To access the trigger setup

To access the trigger setup

- Click the trigger setup button at the bottom of the waveform display.
- or
- Select Trigger from the Setup menu.



The trigger setup dialog allows you to select any of the different modes of triggering supported by the oscilloscope—Edge, Glitch, or Advanced, with choices under Advanced of Pattern, State, Delay by Time, Delay by Events, or Violation. You can set the parameters and conditions for each of those trigger modes.

Figure 3-39



Click this button...

...or select this menu item...



...to access the trigger setup dialog.

Accessing the Trigger Setup


Using the Oscilloscope To set an edge trigger

To set an edge trigger

- 1 Click on the trigger setup button at the bottom of the waveform display. This brings up the trigger configuration dialog. See “To access the trigger setup.”
- 2 Click on Edge.
- 3 Click on the source you want to use for the trigger signal.
- 4 Click to select either the rising or falling edge of the trigger source.
- 5 Set the trigger level using the spin box.

You can also click on the trigger level setting to display a pop-up numeric keypad, where you can enter the trigger level directly.

Figure 3-40



Setting an Edge Trigger

Using the Oscilloscope
To set dialog box preferences

To set dialog box preferences

- To change the background for the dialog boxes used by the oscilloscope, select Utilities, then User Preferences. Then select Opaque, Translucent, or Transparent in the User Preferences dialog box.

The oscilloscope displays a *dialog box* when it needs information from you to set up a particular aspect of an operation or measurement. You can choose opaque, translucent, or transparent backgrounds for the dialog boxes. Opaque and translucent settings are easier for most users to read and interact with because the dialog box is clearly defined against the oscilloscope display. Translucent and transparent settings make it easier to see the waveform while you are changing the oscilloscope configuration; this allows you to see the results of your changes immediately without closing the dialog box.

Windows and Dialog Boxes are Opaque

Windows and dialog boxes for file operations, error messages, and the built-in information system are always opaque.

- To allow multiple dialog boxes on the screen at a time, select Utilities, then User Preferences. Then select Allow Multiple Active Dialogs in the User Preferences dialog box.



Normally, the oscilloscope displays only one dialog box at a time, and you must close that dialog box before opening another. With Allow Multiple Active Dialogs, you can change several different aspects of the oscilloscope configuration simultaneously. This option is usually easier to use if you select Translucent or Transparent dialogs, because it will be easier to see the effects of your changes.

Remove Clutter from the Waveform Display

Having too many dialog boxes on the screen can make it difficult to see the waveform. Close dialog boxes you are not using to remove clutter from the waveform display.

Using the Oscilloscope
To set dialog box preferences


Figure 3-41



Select one of these commands to change how dialog boxes look on the screen




Select this command to allow many dialog boxes to be present on the screen at one time



Setting Dialog Box Options

Using the Oscilloscope
To set dialog box preferences

Figure 3-42



Multiple Active Dialogs (Opaque)

Using the Oscilloscope
To recover your Infinium hard disk

To recover your Infinium hard disk


The Infinium hard disk recovery system is contained in a hidden partition on the Infinium hard disk. If you have to use the recovery system, the Infinium hard disk is returned to the condition it was in when it left the factory.

- 1** Turn off the Infinium.
- 2** Connect the keyboard to the keyboard connector on the rear panel of the oscilloscope.
- 3** Connect the mouse to the mouse connector on the rear panel of the oscilloscope.
- 4** Turn on the Infinium.

Using the Oscilloscope
To recover your Infinium hard disk

- 5 When the Splash screen (Figure 3-43) appears, hold down one of the control (Ctrl) keys.

Figure 3-43



Infinium Splash Screen

- 6 Once the Recovery Screen (Figure 3-44) appears you can release the control key.

Using the Oscilloscope
To recover your Infiniium hard disk

Figure 3-44



Recovery Screen

- 7 Select Recover System from the Boot Menu and press the Enter key.
- 8 Follow the on-screen instructions.
- 9 Once the recovery process is finished and the oscilloscope is running, check in the About Infiniium dialog box under installed options to see if all of the options you ordered are installed. If the options are not installed, please install them.

4

Using the Built-In Information System

Using the Built-In Information System

Most of the documentation for using the Infiniium Oscilloscope is included in its built-in information system. The built-in information system contains the information generally found in a User's Guide.

The information system is available only when the graphical interface is enabled, and is accessible from the menu bar and from dialog boxes. This chapter explains how to use the system to find the information you need to use the oscilloscope effectively.

The following topics are covered in the information system:

- *Probing* lists the various probes available for the Infiniium Oscilloscope, discusses probing considerations, and explains how to use the probes.
- *Acquiring and Viewing Waveforms* explains how acquisition works, discusses the acquisition and display functions and controls, and tells how to use the run/stop controls.
- *Waveform Display Style and Grid* explains how to alter the display appearance, tells how to improve the display of certain kinds of waveforms, and explains the X-versus-Y display function.
- *Triggering* explains basic triggering concepts, describes the trigger modes available in the Infiniium Oscilloscope, and tells how to use the trigger modes.
- *Saving Waveforms, Saving Setups, and Printing* explains how to simplify your work with the oscilloscope by saving critical waveforms, measurement results, and the oscilloscope configuration used to obtain them. You can recall the waveforms for viewing later, or recall the setup to repeat the measurement.
- *Jitter Measurements* explains jitter measurements and how to use them. You must order the EZJIT option before jitter measurements are available.
- *Mask Testing* describes mask testing and how to use it.
- *Measurements* describes the built-in automatic measurement functions of the oscilloscope, explains how they work, and tells how to use them. Describes the markers and how to use them.

Using the Built-In Information System

- *Math, FFTs, and Histograms* discusses oscilloscope functions that perform mathematical operations on one or more waveforms, and discusses the built-in Fast-Fourier Transform functions. Describes histograms and how to use them.
- *Keyboard Control, VoiceControl, and Web Control* explains how to use the keyboard to perform basic oscilloscope operation. Discusses the operation and use of VoiceControl which is an option that you can order with your Infiniium oscilloscope. Explains Web Control and how to use it to operate the oscilloscope.
- *Optional Analysis Packages* contains information on optional analysis software that you may have ordered.
- *Locating a Control* is a visual guide to the location of every oscilloscope control in the graphical user interface.
- *When Something Goes Wrong* explains what to do when you see an error message or have a problem using the oscilloscope.

Some topics in the information system are important enough to have their own icons.

- A *Setup Guide* explains the steps you take to perform various measurement and configuration tasks using the oscilloscope. Twenty-four different procedures are covered, making it easy for you to make a measurement without having to learn a complex setup process.
- A *Technical Support* page explains how to get help from Agilent Technologies on using or servicing the oscilloscope.
- A *Glossary* defines all the major terms used in the information system and other documentation.
- A *Specifications* page lists all the performance specifications for the oscilloscope.
- An *Accessories List* describes the accessories supplied with the oscilloscope and additional accessories available from Agilent Technologies.

Using the Built-In Information System
To access the information system


To access the information system

- Select Contents from the Help menu.
The Contents page for the oscilloscope information system is displayed.

To Use the Information System

You must select the graphical interface before you can access the information system. Select the graphical interface by clicking the mouse icon in the upper right corner of the scope screen.

Figure 4-1




Using the Built-In Information System
To access the information system

Access Methods

The information system provides three methods for finding the information you need. The *Contents* topic is shown when you access the information system from the Help menu. It looks like the following figure.

Figure 4-2



Prints this topic on the local printer

Closes the information system window

Hides the view of the Contents, Index, and Search tabs.

Click on this topic for recommended setup procedures of common tasks.

Contents Topic Page

Using the Built-In Information System To access the information system

Viewing the Information System Window

The Information System window will always stay on top of the interface display, so you can refer to it while working with the oscilloscope. You can move the window around the screen or resize it to make it easier to use.

The *Index* tab is the second method of finding information. You can type in a keyword and search the index for that keyword or you scroll through the list to find a topic.


The *Search* tab is the final method of finding information. You can type in a keyword and if that word exists in the help system, a list of topics containing the keyword is shown.

Hyperlinks throughout the information system allow you to reach the information you need. Most hyperlinks will be in blue underlined text, but some hyperlinks also have buttons (many of the reference topics) or tabs (in setup guide topics). You can find out whether a particular item has a hyperlink by pointing to it with the mouse; if it is a hyperlink, the pointer will change to a hand icon. Clicking once will follow the hyperlink to display a new topic with the relevant information.

Topic Types

There are two general categories of topics in the information system. First are *reference topics*, which explain what the oscilloscope does, how it does it, or considerations you should observe in setting up or making a measurement. The following figure shows an example.

Figure 4-3




Reference Topic (Typical)

Using the Built-In Information System

Topic Types

The second are *setup guide topics*, which show you the steps to follow to perform a task with the oscilloscope. These are organized as an overview with a set of tabs that show the specific steps for each part in the process. The following figure shows an example.

Figure 4-4



Setup Guide Topic

Some setup guide topics have multiple levels because the steps depend on the type of measurement you want to make.

To navigate through the information system

Once you are in the help system windows, you may need to navigate through them to find the information you need.

For most topics:

- To display information for another topic or category described in a hyperlink, click on that hyperlink or on the button icon next to it.
- To return to the previously visited topic, click on the Back button near the top of the Help window.

For Setup Guide topics:

- To display information for another topic or category described in a hyperlink, click on that hyperlink or on the button icon next to it.
- To jump to a particular step, click on the hyperlink button next to the description of that step, or click the tab corresponding to that step in the Overview window.

The Overview window is the one you first see when you select a particular Setup Guide topic. It gives an overview of the steps in the task.

- To move to the next step, click on the tab for that step in the step description window.
- To return to the previously visited topic, click the Back button near the top of the Help window.

Using the Built-In Information System
To access context-sensitive information

To access context-sensitive information


- To see overview information for items in a dialog box, click the Help button in that dialog box.
- To get information on a particular user interface element, click on the question mark button at the upper right-hand corner of a dialog box, then point to the element.

The mouse pointer becomes a pointer plus question mark whenever you click the question mark button. After you click on the desired user interface element, a pop-up window appears describing the function or use of the element. To clear the pop-up window, click the mouse button.

First Access the Graphical Interface

You must select the graphical interface before you can access the information system.

Figure 4-5



Getting Help on the Trigger Setup Dialog Interface Elements

Index

- A**
- access methods 4-5
 - accessing
 - index 4-5, 4-6
 - information system 4-2, 4-4
 - trigger setup 3-56
 - accessories 1-7, 4-3
 - and options 1-7
 - supplied 1-3, 1-4, 1-5, 1-6
 - accurate measurements using markers 3-18
 - acquisition
 - and general controls 3-5
 - clearing display before 3-40
 - information system 4-2
 - running and stopping 3-5
 - starting and stopping 3-8, 3-30, 3-39
 - active dialog boxes
 - allowing multiple 3-58
 - Adaptors
 - connecting SMA to BNC 1-19
 - adjusting
 - digital channel's vertical offset 3-14
 - digital channel's vertical scaling 3-14
 - horizontal position 3-27, 3-49
 - sweep speed 3-15, 3-48
 - vertical offset 3-13, 3-27, 3-28, 3-43, 3-44
 - vertical scaling 3-13, 3-29, 3-45
 - advanced trigger 3-5
 - configuration 3-31
 - mode 3-56
 - airflow requirements 1-12
 - appearance
 - of display 4-2
 - Application software 1-30
 - applying power 1-26
 - arrangement of graphical interface 3-25
 - Arrow keys 3-18
 - associating measurement results with waveforms 3-54, 3-55
 - Autoprobe interface 3-46
 - Autoscale 1-28, 3-5
- B**
- back panel connections 1-14, 1-16
 - backlight 1-26
 - behavior of information system windows 4-6
- C**
- blanking
 - screen 1-26
 - BNC connectors 1-18
 - built-in information system 4-2
 - buttons 3-3
 - as hyperlink 4-6
 - channel setup 3-23, 3-24, 3-46
 - clear display 3-23, 3-24
 - external trigger setup 3-23, 3-24
 - full-screen
 - mode 3-21, 3-22
 - graphical interface mode 3-23, 3-24
 - help 4-10
 - horizontal setup 3-23, 3-24, 3-50
 - question mark 4-10
 - run 3-23, 3-24
 - stop 3-23, 3-24
 - trigger setup 3-23, 3-24, 3-56, 3-57
 - Cable Adaptors
 - connecting SMA to BNC 1-19
 - cables 1-7
 - power 1-13
 - printer 1-14
 - calibration
 - and repair service 1-7
 - MIL-STD 1-7
 - output 1-28
 - categories of information 4-7, 4-8
 - cautions
 - cleaning 1-32
 - disconnecting probes 1-18
 - center horizontal reference 3-30, 3-47
 - Centronics printers, see parallel printers
 - Channel keys 3-10, 3-11
 - channel setup
 - accessing 3-46
 - button 3-23, 3-24, 3-46
 - channels
 - displayed 3-29
 - turning on or off 3-10, 3-29, 3-42
 - turning on or off digital channels 3-11
 - characteristics 4-3
 - check box 3-32, 3-33
 - checking the oscilloscope 1-3
 - choosing
 - commands 3-35, 3-36
 - list items 3-32
 - cleaning the instrument 1-1
 - cleaning the oscilloscope 1-32
 - clear display 3-5
 - button 3-23, 3-24
 - Clear Display key 3-9
 - clearance requirements 1-12
 - clearing measurements 3-26
 - clearing the display 3-9, 3-30, 3-40
 - click 3-32
 - closing
 - dialog boxes 3-32, 3-33
 - color
 - use of 3-3, 3-24
 - use of in interface 3-54
 - command menus 3-23, 3-24
 - commands
 - selecting 3-35, 3-36
 - common procedures 4-3
 - conditions
 - trigger 3-56
 - configuration
 - Autoscale 3-5
 - channel 3-46
 - default 1-28, 3-5, 3-7
 - dialog boxes 3-58
 - edge trigger 3-17
 - horizontal 3-30, 3-50
 - through menus 3-28
 - trigger 3-5, 3-31, 3-56, 3-57
 - configuration instructions 4-3
 - configuring
 - quick measurements 3-19
 - connecting
 - GPIB 1-14
 - LAN card 1-14
 - pointing device 1-16
 - power 1-12
 - printers 1-14
 - probes 1-18
 - constraints on waveform movement 3-43
 - contents
 - information system 4-4
 - of oscilloscope package 1-3
 - contents topic 4-5
-

Index

- contents, of oscilloscope package 1-4, 1-5, 1-6
- context-sensitive
 - information 4-10
 - menus 3-28, 3-36, 3-38
- controlling the oscilloscope 1-14
- controls
 - acquisition and general 3-5
 - graphical interface 4-3
 - horizontal 3-5
 - marker and measurement 3-6
 - trigger 3-5
 - vertical 3-5
- coupling
 - trigger input 3-17
 - vertical input 3-5
- Coupling key 3-17
- customizing display layout 3-26, 3-54
- cycling
 - markers through channels 3-18
 - measurements through channels 3-19
 - through waveforms 3-6
- D**
- default setup 1-28, 3-5
- Default Setup key 3-7
- definitions of terms 4-3
- Delayed key 3-16
- delayed sweep 3-16, 3-50
- delayed time base 3-5
- dialog boxes 3-32, 3-33
 - information on 4-10
 - preferences for 3-58, 3-59
- direct manipulation of waveforms 3-43, 3-49
- Disk
 - Hard Disk Recovery 3-61
- display
 - clearing 3-5, 3-9, 3-30, 3-40
 - information system 4-2
- display area
 - waveforms 3-22, 3-27
- display format 4-2
- display layout
 - customizing 3-26, 3-54
- display of channels 3-29
- display update rate
 - increasing 3-10, 3-11, 3-42
- drag-and-drop measurements 3-25, 3-54
- drag-and-drop waveforms 3-27, 3-43, 3-49
- drop-down list box 3-32, 3-34
- E**
- edge trigger 3-5
 - mode 3-56, 3-57
 - setting 3-17
- erasing the display 3-9, 3-40
- error messages 4-3
- external trigger setup button 3-23, 3-24
- F**
- feet
 - using 1-25
- FFT functions
 - information system 4-3
- finding information 4-2, 4-5, 4-9
- front panel 3-4, 3-6
 - connections 1-18
 - interface 3-2, 3-3
 - using with graphical interface 3-24
- full-screen
 - mode 3-21, 3-22
 - mode button 3-21, 3-22
- functions
 - information system 4-3
- G**
- general controls 3-5
- geometric measurement indicators 3-25
- glitch trigger 3-5
 - mode 3-56
- glossary 4-3
- GPIB
 - cables 1-7
 - connecting 1-14
- graphical interface 3-2, 3-21
 - arrangement of 3-25
 - controls 4-3
 - mode 3-23, 3-24
 - pointing device for 1-16
 - using with front panel 3-24
- graticule 3-6
- grid 3-27
- ground reference indicator 3-21, 3-22, 3-28, 3-43
- guide to setup 4-3
- H**
- Hard Disk
 - Recovery 3-61
- help
 - button 4-10
- help, see also information system
- horizontal
 - adjusting position 3-15
 - adjusting sweep speed 3-15
 - controls 3-5
 - delayed sweep 3-16
 - horizontal and trigger toolbar 3-29
 - horizontal position 3-5, 3-21, 3-22, 3-50
 - adjusting 3-15, 3-27, 3-30, 3-49
 - in delayed sweep 3-16
 - setting 3-23, 3-24
 - with respect to horizontal reference 3-47
 - horizontal reference 3-50
 - indicators 3-21, 3-22, 3-30
 - setting 3-47
 - horizontal setup 3-30
 - accessing 3-50
 - button 3-23, 3-24, 3-50
 - horizontal sweep speed 3-21, 3-22, 3-23, 3-24, 3-30
 - adjusting 3-48
- hyperlinks 4-6
 - index 4-6
 - using to find information 4-9
- I**
- icons
 - dragging to waveform 3-25
 - learning functions of 3-25
 - measurement 3-23, 3-24, 3-25, 3-54, 3-55
- impedance
 - input 3-5
- increasing display update rate 3-10, 3-11, 3-42
- index
 - accessing 4-5, 4-6
- indicators
 - ground reference 3-21, 3-22, 3-28,

Index

- 3-43
- horizontal reference 3-21, 3-22, 3-30
- measurement 3-25
- run/stop mode 3-21, 3-22
- trigger level 3-21, 3-22
- trigger level reference 3-31
- information
 - methods for accessing 4-5
 - on dialog boxes 4-10
- information system
 - accessing 4-4
 - context-sensitive 4-10
 - index 4-6
 - navigating 4-9
 - overview 4-2
 - topic types 4-7
 - windows 4-6
- input
 - coupling 3-5, 3-46
 - impedance 3-5, 3-46
- inspecting the oscilloscope 1-3
- Installing
 - Application software 1-30
 - instructions for setup 4-3
 - instrument controller 1-14
 - instrument, see oscilloscope
 - interface elements
 - information on 4-10
- K**
- keys 3-3
 - Arrow 3-18
 - Channel 3-10
 - Clear Display 3-9
 - Coupling 3-17
 - Default Setup 3-7
 - Delayed 3-16
 - Digital Channel 3-11
 - marker 3-6
 - Marker A 3-18
 - Marker B 3-18
 - Mode 3-17
 - QuickMeas 3-19
 - Run 3-8
 - Slope 3-17
 - Source 3-17
 - Stop 3-8
 - Sweep 3-17
- kinds of information 4-7, 4-8
- knobs 3-3
 - color of 3-24
 - digital channel's vertical offset 3-14
 - digital channel's vertical scaling 3-14
 - horizontal position 3-15
 - sweep speed 3-15
 - vertical offset 3-13
 - vertical scaling 3-13
- L**
- LAN card 1-14
- learning functions of icons 3-25
- LED indicators 3-3
- left horizontal reference 3-30, 3-47
- level
 - trigger 3-17, 3-21, 3-22, 3-31, 3-57
- line voltage 1-13
- locating controls 4-3
- location of trigger 3-30
- M**
- magnifying
 - a section of the waveform 3-16
 - waveforms 3-5, 3-51, 3-52
- main sweep 3-16
- making
 - measurements 3-54
 - oscilloscope easier to view 1-25
 - quick measurements 3-19
 - waveform bigger or smaller 3-13, 3-14, 3-45
- manipulating waveforms 3-27, 3-43
- manuals 1-3, 1-4, 1-5, 1-6, 1-7
- Marker A key 3-18
- marker and measurement controls 3-6
- marker arrow keys 3-6
- Marker B key 3-18
- markers 3-6
 - and measurements 3-54
 - moving 3-53
 - positioning 3-18
 - turning on or off 3-18
- math functions
 - information system 4-3
- maximizing waveform viewing area 3-21, 3-22
- measurement icons 3-23, 3-24, 3-25, 3-55
- measurement indicators
 - geometric 3-25
- measurement results
 - associating with waveform 3-54, 3-55
- measurement toolbar 3-23, 3-24, 3-25, 3-54, 3-55
- removing 3-26
- measurements 3-6, 3-26, 3-54, 3-55
- and markers 3-54
- drag and drop 3-25
- information system 4-2
- Quick 3-19
- quick 3-6
- reset 3-40
- using markers 3-18, 3-53
- memory
 - bar 3-21, 3-22
- menu bar 3-23, 3-24, 3-28, 3-35
- menus 3-35
 - context-sensitive 3-28, 3-36, 3-38
- methods for accessing information 4-5
- MIL-STD calibration 1-7
- Mode key 3-17
- modes
 - edge trigger 3-57
 - full-screen 3-21, 3-22
 - graphical interface 3-23, 3-24
 - trigger 3-56
- mouse
 - click 3-32
 - connecting 1-16
 - moving 3-32
 - right-click 3-32, 3-36
 - verifying 1-28
- moving
 - dialog boxes 3-32, 3-33
 - markers 3-18, 3-53
 - mouse 3-32
 - waveform horizontally 3-15, 3-49
 - waveform vertically 3-13, 3-14
- multiple active dialog boxes 3-58
- N**
- navigating the information system 4-9
- notes
 - dialog boxes 3-58
 - information system access 4-4, 4-10

Index

- information system windows 4-6
 - overdriving vertical amplifiers 3-51
 - screen saver 1-26
 - zooming 3-28
- O**
- obtaining information within a particular context 4-10
 - offset
 - adjusting digital channel's vertical 3-14
 - adjusting vertical 3-13, 3-43, 3-44
 - vertical 3-5
 - opaque dialog boxes 3-58, 3-59
 - opening the information system 4-4
 - operation
 - verifying 1-28
 - options 1-7, 4-3
 - and accessories 1-7
 - display 3-27
 - power cable 1-13
 - oscilloscope
 - cleaning 1-32
 - inspecting 1-3
 - resetting 3-7
 - using 3-2
 - oscilloscope controls 4-3
 - oscilloscope feet 1-25
 - oscilloscope front panel 3-3, 3-4, 3-6
 - oscilloscope graticule 3-6
 - oscilloscope operation
 - verifying 1-28
 - oscilloscope probes, see probes
 - overdriving vertical amplifiers 3-28, 3-51
- P**
- package contents 1-3, 1-4, 1-5, 1-6
 - parallel printers 1-14
 - parameters
 - trigger 3-56
 - performance specifications 4-3
 - plugs
 - power 1-13
 - pointer
 - moving 3-32
 - pointing device 1-7
 - connecting 1-16
 - verifying operation 1-28
 - position
 - adjusting digital channel's vertical 3-14
 - adjusting horizontal 3-15, 3-30, 3-49
 - adjusting vertical 3-13
 - horizontal 3-21, 3-22, 3-23, 3-24, 3-50
 - positioning markers 3-18
 - power
 - applying 1-26
 - connecting 1-12
 - requirements 1-12
 - turning off 1-27
 - power cables 1-13
 - preferences for dialog boxes 3-58
 - preset measurements 3-19
 - printer
 - cables 1-7, 1-14
 - printers 1-7
 - connecting 1-14
 - probes 1-7
 - connecting 1-18
 - information system 4-2
 - setting characteristics of 3-46
 - probing a circuit 1-18
 - problems
 - information system 4-3
- Q**
- question mark button 4-10
 - quick measurements 3-6, 3-19
 - QuickMeas key 3-19
- R**
- radio button 3-32, 3-33
 - recalling waveforms and setups
 - information system 4-2
 - Recovery Disks 3-61
 - reducing screen clutter 3-58
 - reference
 - horizontal 3-47
 - reference topics 4-7
 - removing measurement toolbar 3-26
 - repair and calibration services 1-7
 - Repetitive Strain Injury 2-2
 - Description 2-3
 - Using the Mouse 2-4
 - reset
 - measurements and statistics 3-40
 - oscilloscope 3-7
 - right horizontal reference 3-30, 3-47
 - right-click 3-32, 3-36
 - RJ-45 connector 1-14
 - LAN card 1-14
 - RSI
 - Description 2-3
 - Introduction 2-2
 - Using the Mouse 2-4
 - run 3-39
 - button 3-23, 3-24
 - Run key 3-8
 - run/stop
 - controls 3-30
 - mode indicators 3-21, 3-22
 - running acquisition 3-5
- S**
- sampling rate 3-21, 3-22
 - saving waveforms and setups
 - information system 4-2
 - scaling
 - adjusting vertical 3-13, 3-45
 - vertical 3-5, 3-21, 3-22, 3-23, 3-24
 - screen clutter
 - simplifying 3-58
 - screen saver 1-26
 - seconds per division
 - adjusting 3-15
 - in delayed sweep 3-16
 - setting 3-48
 - seconds per division setting 3-21, 3-22, 3-23, 3-24
 - selecting
 - commands 3-35, 3-36
 - waveform features for measurement 3-25
 - waveforms for measurement 3-6
 - serial printers 1-14
 - service
 - repair and calibration 1-7
 - set up trigger 3-5
 - setting
 - dialog box preferences 3-58
 - edge trigger 3-17
 - horizontal reference 3-47
 - sweep speed 3-48

Index

- trigger position 3-47
- setup
 - channel 3-46
 - default 1-28, 3-5, 3-7
 - dialog boxes 3-58
 - horizontal 3-30, 3-50
 - trigger 3-31, 3-56, 3-57
- Setup Guide 4-3
 - steps 4-9
 - topics 4-8
- sizing
 - adjusting digital channel's vertical 3-14
- slope
 - trigger 3-17, 3-57
- Slope key 3-17
- SMA to BNC Adaptors
 - connecting 1-19
- software magnification of waveforms 3-51
- solving problems
 - information system 4-3
- source
 - trigger 3-17, 3-57
- Source key 3-17
- specifications 4-3
- spin box 3-32, 3-33
- standard accessories 1-3, 1-4, 1-5, 1-6
- starting and stopping acquisition 3-8, 3-39
- statistics
 - reset 3-40
- status indicators
 - trigger 3-5
- steps
 - in Setup Guide 4-9
- Stop
 - key 3-8
- stop 3-39
 - button 3-23, 3-24
- stopping acquisition 3-5
- stretching and shrinking the waveform 3-15, 3-30, 3-48
- submenus 3-35, 3-36
- support information 4-3
- sweep
 - key 3-17
 - triggered 3-17, 3-57
- sweep speed 3-5, 3-30, 3-50
 - adjusting 3-15, 3-48
 - horizontal 3-21, 3-22, 3-23, 3-24
 - in delayed sweep 3-16
- T**
- Tab display area 3-27
- tabs
 - as hyperlink 4-6
- tasks
 - topics regarding 4-8
- technical support information 4-3
- terms
 - defined 4-3
- tilting the oscilloscope 1-25
- time base
 - adjusting 3-48
 - see also horizontal
- time measurements 3-6
- title bar 3-32, 3-33
- tool tips 3-25
- toolbars 3-28
 - horizontal and trigger 3-21, 3-22, 3-29
 - measurement 3-23, 3-24, 3-25, 3-54, 3-55
 - removing measurement 3-26
 - vertical 3-21, 3-22, 3-29
- topics
 - contents 4-5
 - navigating 4-9
 - reference 4-7
 - Setup Guide 4-8
 - types of 4-7, 4-8
- touchpad 1-7
 - connecting 1-16
- trackball 1-7
 - connecting 1-16
- tracking waveforms
 - using markers 3-18
- translucent dialog boxes 3-58, 3-59
- transparent dialog boxes 3-58, 3-59
- trigger
 - controls 3-5
 - information system 4-2
 - input coupling 3-17
 - level 3-17, 3-21, 3-22, 3-31, 3-57
 - level reference indicator 3-21, 3-22, 3-31
 - location 3-30
 - parameters, where set up 3-5
 - position 3-47
 - setting edge 3-17
 - setup 3-31, 3-56, 3-57
 - setup button 3-23, 3-24, 3-56, 3-57
 - slope 3-17, 3-57
 - source 3-17, 3-57
 - status indicators 3-5
- triggered sweep 3-17, 3-57
- turning channels on or off 3-10, 3-29, 3-42
- turning digital channels on or off 3-11
- turning markers on and off 3-18
- types of topics 4-7, 4-8
- U**
- undo
 - default setup 3-7
- use of color in interface 3-3, 3-54
- user interface
 - front panel 3-3, 3-4, 3-6
 - use of color 3-3, 3-24
- user interface elements 4-10
- using
 - feet 1-25
 - hyperlinks 4-9
 - markers 3-18
 - oscilloscope 3-2
- V**
- verifying basic operation 1-28
- vertical amplifiers
 - overdriving 3-51
- vertical controls 3-5
- vertical inputs
 - overdriving 3-28
- vertical offset 3-5, 3-46
 - adjusting 3-13, 3-27, 3-28, 3-43, 3-44
 - adjusting digital channels 3-14
- vertical scaling 3-5, 3-21, 3-22, 3-23, 3-24, 3-46
 - adjusting 3-13, 3-29, 3-45
 - adjusting digital channels 3-14
- vertical settings 3-29
- vertical toolbar 3-29
- voltage measurements 3-6

Index

-
- volts per division
 - adjusting 3-13, 3-45
 - volts per division setting 3-21, 3-22, 3-23, 3-24

 - W**
 - waveform
 - magnifying 3-5, 3-16
 - stretching and shrinking 3-15
 - waveform display area 3-22, 3-27
 - waveform features
 - measurements on 3-54
 - moving markers to 3-53
 - selecting for measurement 3-25
 - waveforms
 - acquiring 3-8, 3-39
 - adjusting horizontal position 3-49
 - associating with measurement results 3-54, 3-55
 - drag and drop 3-43, 3-49
 - drag-and-drop measurements on 3-54
 - dragging measurement icons to 3-25
 - information system 4-2
 - making bigger or smaller 3-45
 - making quick measurements 3-19
 - manipulating 3-27
 - matching to measurement results 3-25
 - maximizing viewing area 3-21, 3-22
 - measuring using markers 3-18
 - selecting for measurement 3-6
 - stretching and shrinking 3-30, 3-48
 - tracking with markers 3-6
 - turning off 3-10, 3-11
 - zooming on 3-28, 3-51, 3-52
 - window
 - delayed sweep 3-16
 - Window system settings 1-31
 - windows
 - information system 4-6
 - working with dialog boxes 3-58

 - X**
 - X-versus-Y function 4-2

 - Z**
 - zooming on waveforms 3-28, 3-51, 3-52

Safety Notices

This apparatus has been designed and tested in accordance with IEC Publication 1010, Safety Requirements for Measuring Apparatus, and has been supplied in a safe condition. This is a Safety Class I instrument (provided with terminal for protective earthing). Before applying power, verify that the correct safety precautions are taken (see the following warnings). In addition, note the external markings on the instrument that are described under "Safety Symbols."

Warnings

- Before turning on the instrument, you must connect the protective earth terminal of the instrument to the protective conductor of the (mains) power cord. The mains plug shall only be inserted in a socket outlet provided with a protective earth contact. You must not negate the protective action by using an extension cord (power cable) without a protective conductor (grounding). Grounding one conductor of a two-conductor outlet is not sufficient protection.
- Only fuses with the required rated current, voltage, and specified type (normal blow, time delay, etc.) should be used. Do not use repaired fuses or short-circuited fuseholders. To do so could cause a shock or fire hazard.
- If you energize this instrument by an auto transformer (for voltage reduction or mains isolation), the common terminal must be connected to the earth terminal of the power source.

- Whenever it is likely that the ground protection is impaired, you must make the instrument inoperative and secure it against any unintended operation.

- Service instructions are for trained service personnel. To avoid dangerous electric shock, do not perform any service unless qualified to do so. Do not attempt internal service or adjustment unless another person, capable of rendering first aid and resuscitation, is present.

- Do not install substitute parts or perform any unauthorized modification to the instrument.

- Capacitors inside the instrument may retain a charge even if the instrument is disconnected from its source of supply.

- Do not operate the instrument in the presence of flammable gasses or fumes. Operation of any electrical instrument in such an environment constitutes a definite safety hazard.

- Do not use the instrument in a manner not specified by the manufacturer.

To clean the instrument

If the instrument requires cleaning: (1) Remove power from the instrument. (2) Clean the external surfaces of the instrument with a soft cloth dampened with a mixture of mild detergent and water. (3) Make sure that the instrument is completely dry before reconnecting it to a power source.

Safety Symbols



Instruction manual symbol: the product is marked with this symbol when it is necessary for you to refer to the instruction manual in order to protect against damage to the product.



Hazardous voltage symbol.



Earth terminal symbol: Used to indicate a circuit common connected to grounded chassis.

Notices

© Agilent Technologies, Inc.
2002-2004

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

54855-97016, December 2004

Print History

54855-97000, December 2002
54855-97004, May 2003
54855-97005, July 2003
54855-97009, October 2003
54855-97011, April 2004
54855-97014, August 2004
54855-97016, December 2004

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Document Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

WARNING

A WARNING notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood and met.

CAUTION

A CAUTION notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood and met.

Trademark Acknowledgements

Windows and MS Windows are U.S. registered trademarks of Microsoft Corporation.

MATLAB is a U.S. registered trademark of MathWorks.

Mathcad is a U.S. registered trademark of Mathsoft.